




Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG083120\
Data File : BG046469.D
Acq On : 31 Aug 2020 14:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 18:02:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 31 16:05:33 2020
Response via : Initial Calibration



TIC: BG046469.D

(77) Benzidine




19.533min (-0.003) 0.00ng

response 502040

Ion	Exp%	Act%
184.00	100	100
185.00	14.80	15.61
183.00	13.40	12.52
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG083120\
 Data File : BG046469.D
 Acq On : 31 Aug 2020 14:28
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 31 18:02:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 31 16:05:33 2020
 Response via : Initial Calibration



TIC: BG046469.D

(70) Pentachlorophenol (C)

16.960min (-0.003) 40027.15ng

response 80254

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	62.47
264.00	62.70	64.44
0.00	0.00	0.00