

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG090116\
 Data File : BG023937.D
 Acq On : 1 Sep 2016 17:55
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 02 01:14:17 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 31 20:05:47 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.470	0.344	26.8#	52	0.00
3	Pyridine	1.412	1.266	10.3	62	-0.01
4	n-Nitrosodimethylamine	0.744	0.924	-24.2	96	0.00
5 S	2-Fluorophenol	1.139	1.087	4.6	68	0.00
6	Aniline	2.328	2.301	1.2	74	0.00
7 S	Phenol-d6	1.754	1.731	1.3	72	0.00
8	2-Chlorophenol	1.320	1.326	-0.5	74	0.00
9	Benzaldehyde	1.246	1.264	-1.4	74	0.00
10 C	Phenol	1.845	1.737	5.9	68	0.00
11	bis(2-Chloroethyl)ether	1.451	1.246	14.1	68	0.00
12	1,3-Dichlorobenzene	1.545	1.492	3.4	74	0.00
13 C	1,4-Dichlorobenzene	1.636	1.517	7.3	74	0.00
14	1,2-Dichlorobenzene	1.536	1.427	7.1	72	0.00
15	Benzyl Alcohol	1.546	1.639	-6.0	76	0.00
16	2,2'-oxybis(1-Chloropropane	1.945	1.892	2.7	77	0.00
17	2-Methylphenol	1.243	1.235	0.6	73	0.00
18	Hexachloroethane	0.630	0.650	-3.2	79	0.00
19 P	n-Nitroso-di-n-propylamine	1.549	1.549	0.0	76	0.00
20	3+4-Methylphenols	1.732	1.784	-3.0	73	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	78	-0.01
22	Acetophenone	0.518	0.498	3.9	75	0.00
23 S	Nitrobenzene-d5	0.448	0.444	0.9	79	0.00
24	Nitrobenzene	0.482	0.472	2.1	79	0.00
25	Isophorone	0.869	0.837	3.7	76	0.00
26 C	2-Nitrophenol	0.183	0.184	-0.5	79	0.00
27	2,4-Dimethylphenol	0.311	0.315	-1.3	74	0.00
28	bis(2-Chloroethoxy)methane	0.439	0.399	9.1	73	0.00
29 C	2,4-Dichlorophenol	0.330	0.340	-3.0	78	0.00
30	1,2,4-Trichlorobenzene	0.362	0.349	3.6	78	0.00
31	Naphthalene	1.031	0.999	3.1	79	0.00
32	Benzoic acid	0.253	0.250	1.2	76	0.00
33	4-Chloroaniline	0.430	0.442	-2.8	78	0.00
34 C	Hexachlorobutadiene	0.272	0.275	-1.1	78	0.00
35	Caprolactam	0.116	0.121	-4.3	78	-0.02
36 C	4-Chloro-3-methylphenol	0.441	0.455	-3.2	79	0.00
37	2-Methylnaphthalene	0.784	0.816	-4.1	80	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
39	1,2,4,5-Tetrachlorobenzene	0.664	0.658	0.9	80	0.00
40 P	Hexachlorocyclopentadiene	0.335	0.324	3.3	80	0.00
41 S	2,4,6-Tribromophenol	0.360	0.353	1.9	76	0.00
42 C	2,4,6-Trichlorophenol	0.443	0.439	0.9	79	0.00
43	2,4,5-Trichlorophenol	0.448	0.433	3.3	75	0.00
44 S	2-Fluorobiphenyl	1.395	1.367	2.0	81	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG090116\
 Data File : BG023937.D
 Acq On : 1 Sep 2016 17:55
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 02 01:14:17 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG083116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 31 20:05:47 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.609	1.556	3.3	79	0.00
46	2-Chloronaphthalene	1.181	1.131	4.2	80	0.00
47	2-Nitroaniline	0.465	0.500	-7.5	89	0.00
48	Acenaphthylene	1.969	1.939	1.5	80	0.00
49	Dimethylphthalate	1.715	1.640	4.4	79	0.00
50	2,6-Dinitrotoluene	0.362	0.353	2.5	79	0.00
51 C	Acenaphthene	1.217	1.199	1.5	82	0.00
52	3-Nitroaniline	0.339	0.359	-5.9	82	0.00
53 P	2,4-Dinitrophenol	0.226	0.238	-5.3	81	0.00
54	Dibenzofuran	1.900	1.854	2.4	80	0.00
55 P	4-Nitrophenol	0.269	0.245	8.9	77	0.00
56	2,4-Dinitrotoluene	0.532	0.535	-0.6	81	0.00
57	Fluorene	1.647	1.651	-0.2	83	0.00
58	2,3,4,6-Tetrachlorophenol	0.455	0.474	-4.2	84	0.00
59	Diethylphthalate	1.831	1.802	1.6	80	0.00
60	4-Chlorophenyl-phenylether	0.887	0.919	-3.6	83	0.00
61	4-Nitroaniline	0.342	0.354	-3.5	83	0.00
62	Azobenzene	1.773	1.838	-3.7	86	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
64	4,6-Dinitro-2-methylphenol	0.139	0.143	-2.9	84	0.00
65 c	n-Nitrosodiphenylamine	0.568	0.570	-0.4	83	0.00
66	4-Bromophenyl-phenylether	0.243	0.237	2.5	79	0.00
67	Hexachlorobenzene	0.285	0.271	4.9	78	0.00
68	Atrazine	0.244	0.253	-3.7	83	0.00
69 C	Pentachlorophenol	0.152	0.149	2.0	77	0.00
70	Phenanthrene	1.040	1.046	-0.6	85	0.00
71	Anthracene	1.061	1.069	-0.8	85	0.00
72	Carbazole	0.963	0.950	1.3	83	0.00
73	Di-n-butylphthalate	1.149	1.144	0.4	84	0.00
74 C	Fluoranthene	1.253	1.263	-0.8	83	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
76	Benzidine	0.619	0.684	-10.5	89	0.00
77	Pyrene	1.230	1.220	0.8	82	0.00
78 S	Terphenyl-d14	0.878	0.912	-3.9	86	0.00
79	Butylbenzylphthalate	0.474	0.473	0.2	88	0.00
80	Benzo(a)anthracene	1.120	1.155	-3.1	87	0.00
81	3,3'-Dichlorobenzidine	0.448	0.475	-6.0	86	0.00
82	Chrysene	1.066	1.105	-3.7	88	0.00
83	Bis(2-ethylhexyl)phthalate	0.649	0.645	0.6	89	0.00
84 c	Di-n-octyl phthalate	1.065	1.087	-2.1	90	0.00
85	Indeno(1,2,3-cd)pyrene	1.180	1.234	-4.6	91	0.00
86 I	Perylene-d12	1.000	1.000	0.0	87	0.00
87	Benzo(b)fluoranthene	1.136	1.194	-5.1	93	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG090116\
Data File : BG023937.D
Acq On : 1 Sep 2016 17:55
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 02 01:14:17 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG083116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 31 20:05:47 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.127	1.152	-2.2	89	0.00
89 C	Benzo(a)pyrene	1.079	1.120	-3.8	92	-0.02
90	Dibenzo(a,h)anthracene	1.061	1.065	-0.4	89	-0.02
91	Benzo(g,h,i)perylene	1.020	1.060	-3.9	92	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0