

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090122\
 Data File : BG054817.D
 Acq On : 1 Sep 2022 11:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 14:33:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	36.842	7.9	86	0.00
3	Pyridine	40.000	37.220	7.0	84	0.00
4	n-Nitrosodimethylamine	40.000	36.159	9.6	82	0.00
5 S	2-Fluorophenol	80.000	77.471	3.2	87	0.00
6	Aniline	40.000	38.362	4.1	87	0.00
7 S	Phenol-d6	80.000	76.893	3.9	86	0.00
8	2-Chlorophenol	40.000	38.981	2.5	89	0.00
9	Benzaldehyde	40.000	35.463	11.3	87	0.00
10 C	Phenol	40.000	38.607	3.5	87	0.00
11	bis(2-Chloroethyl)ether	40.000	36.545	8.6	83	0.00
12	1,3-Dichlorobenzene	40.000	38.368	4.1	89	0.00
13 C	1,4-Dichlorobenzene	40.000	38.555	3.6	88	0.00
14	1,2-Dichlorobenzene	40.000	38.693	3.3	89	0.00
15	Benzyl Alcohol	40.000	38.665	3.3	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.538	13.7	79	0.00
17	2-Methylphenol	40.000	38.920	2.7	87	0.00
18	Hexachloroethane	40.000	37.084	7.3	84	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.475	8.8	82	0.00
20	3+4-Methylphenols	40.000	39.301	1.7	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	86	0.00
22	Acetophenone	40.000	38.488	3.8	86	0.00
23 S	Nitrobenzene-d5	80.000	77.082	3.6	86	0.00
24	Nitrobenzene	40.000	38.047	4.9	85	0.00
25	Isophorone	40.000	37.311	6.7	82	0.00
26 C	2-Nitrophenol	40.000	40.900	-2.2	89	0.00
27	2,4-Dimethylphenol	40.000	38.912	2.7	86	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.646	8.4	81	0.00
29 C	2,4-Dichlorophenol	40.000	40.556	-1.4	88	0.00
30	1,2,4-Trichlorobenzene	40.000	39.746	0.6	88	0.00
31	Naphthalene	40.000	38.757	3.1	86	0.00
32	Benzoic acid	40.000	25.470	36.3#	52	0.03
33	4-Chloroaniline	40.000	38.343	4.1	84	0.00
34 C	Hexachlorobutadiene	40.000	41.289	-3.2	93	-0.01
35	Caprolactam	40.000	39.720	0.7	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.697	0.8	87	0.00
37	2-Methylnaphthalene	40.000	38.902	2.7	86	0.00
38	1-Methylnaphthalene	40.000	38.744	3.1	86	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.261	-0.7	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.956	0.1	88	-0.01
42 S	2,4,6-Tribromophenol	80.000	85.564	-7.0	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.861	0.3	87	0.00
44	2,4,5-Trichlorophenol	40.000	39.289	1.8	86	0.00
45 S	2-Fluorobiphenyl	80.000	77.855	2.7	88	0.00
46	1,1'-Biphenyl	40.000	38.153	4.6	86	0.00
47	2-Chloronaphthalene	40.000	38.159	4.6	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090122\
 Data File : BG054817.D
 Acq On : 1 Sep 2022 11:17
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 01 14:33:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 25 17:50:55 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.157	2.1	86	0.00
49	Acenaphthylene	40.000	38.799	3.0	87	0.00
50	Dimethylphthalate	40.000	38.537	3.7	87	0.00
51	2,6-Dinitrotoluene	40.000	39.708	0.7	86	0.00
52 C	Acenaphthene	40.000	39.036	2.4	87	0.00
53	3-Nitroaniline	40.000	40.198	-0.5	88	0.00
54 P	2,4-Dinitrophenol	40.000	38.966	2.6	89	0.00
55	Dibenzofuran	40.000	39.159	2.1	88	0.00
56 P	4-Nitrophenol	40.000	40.729	-1.8	85	0.01
57	2,4-Dinitrotoluene	40.000	41.104	-2.8	88	0.00
58	Fluorene	40.000	38.852	2.9	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.117	-0.3	89	0.00
60	Diethylphthalate	40.000	38.502	3.7	86	0.00
61	4-Chlorophenyl-phenylether	40.000	39.991	0.0	90	0.00
62	4-Nitroaniline	40.000	41.243	-3.1	89	0.00
63	Azobenzene	40.000	36.511	8.7	82	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.569	1.1	85	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.022	4.9	88	0.00
67	4-Bromophenyl-phenylether	40.000	40.466	-1.2	93	0.00
68	Hexachlorobenzene	40.000	40.354	-0.9	93	0.00
69	Atrazine	40.000	38.954	2.6	88	0.00
70 C	Pentachlorophenol	40.000	40.167	-0.4	88	0.00
71	Phenanthrene	40.000	38.399	4.0	89	0.00
72	Anthracene	40.000	38.818	3.0	89	0.00
73	Carbazole	40.000	39.232	1.9	90	0.00
74	Di-n-butylphthalate	40.000	38.267	4.3	87	0.00
75 C	Fluoranthene	40.000	39.920	0.2	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	40.496	-1.2	89	0.00
78	Pyrene	40.000	38.822	2.9	92	0.00
79 S	Terphenyl-d14	80.000	77.225	3.5	93	0.00
80	Butylbenzylphthalate	40.000	37.782	5.5	91	0.00
81	Benzo(a)anthracene	40.000	38.886	2.8	94	0.00
82	3,3'-Dichlorobenzidine	40.000	39.438	1.4	93	0.00
83	Chrysene	40.000	38.656	3.4	94	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.478	8.8	88	-0.01
85 c	Di-n-octyl phthalate	40.000	36.223	9.4	87	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.491	3.8	93	0.00
88	Benzo(b)fluoranthene	40.000	37.299	6.8	89	0.00
89	Benzo(k)fluoranthene	40.000	38.843	2.9	95	0.00
90 C	Benzo(a)pyrene	40.000	38.034	4.9	92	0.00
91	Dibenzo(a,h)anthracene	40.000	38.879	2.8	94	-0.01
92	Benzo(g,h,i)perylene	40.000	38.407	4.0	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090122\
Data File : BG054817.D
Acq On : 1 Sep 2022 11:17
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 01 14:33:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 25 17:50:55 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0