

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
 Data File : BG062618.D
 Acq On : 3 Sep 2024 9:49
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 04:25:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	41.475	-3.7	111	0.00
3	Pyridine	40.000	42.893	-7.2	111	0.00
4	n-Nitrosodimethylamine	40.000	44.350	-10.9	112	0.00
5 S	2-Fluorophenol	80.000	82.754	-3.4	106	0.00
6	Aniline	40.000	41.275	-3.2	103	0.00
7 S	Phenol-d6	80.000	83.764	-4.7	106	0.00
8	2-Chlorophenol	40.000	40.784	-2.0	105	0.00
9	Benzaldehyde	40.000	41.542	-3.9	111	0.00
10 C	Phenol	40.000	41.833	-4.6	105	0.00
11	bis(2-Chloroethyl)ether	40.000	41.134	-2.8	106	0.00
12	1,3-Dichlorobenzene	40.000	41.755	-4.4	109	0.00
13 C	1,4-Dichlorobenzene	40.000	40.815	-2.0	106	0.00
14	1,2-Dichlorobenzene	40.000	41.985	-5.0	108	0.00
15	Benzyl Alcohol	40.000	41.936	-4.8	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.856	0.4	108	0.00
17	2-Methylphenol	40.000	43.229	-8.1	110	0.00
18	Hexachloroethane	40.000	39.863	0.3	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.544	1.1	99	0.00
20	3+4-Methylphenols	40.000	40.656	-1.6	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	39.314	1.7	101	0.00
23 S	Nitrobenzene-d5	80.000	84.306	-5.4	107	0.00
24	Nitrobenzene	40.000	42.156	-5.4	105	0.00
25	Isophorone	40.000	39.570	1.1	100	0.00
26 C	2-Nitrophenol	40.000	38.292	4.3	95	0.00
27	2,4-Dimethylphenol	40.000	38.892	2.8	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.466	1.3	98	0.00
29 C	2,4-Dichlorophenol	40.000	41.203	-3.0	103	0.00
30	1,2,4-Trichlorobenzene	40.000	40.600	-1.5	106	0.00
31	Naphthalene	40.000	40.927	-2.3	105	0.00
32	Benzoic acid	40.000	36.125	9.7	90	0.00
33	4-Chloroaniline	40.000	42.191	-5.5	107	0.00
34 C	Hexachlorobutadiene	40.000	41.813	-4.5	111	0.00
35	Caprolactam	40.000	39.329	1.7	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.261	-3.2	103	0.00
37	2-Methylnaphthalene	40.000	40.986	-2.5	104	0.00
38	1-Methylnaphthalene	40.000	40.926	-2.3	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.852	-2.1	105	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.717	0.7	98	0.00
42 S	2,4,6-Tribromophenol	80.000	83.152	-3.9	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.763	-4.4	104	0.00
44	2,4,5-Trichlorophenol	40.000	41.633	-4.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	82.540	-3.2	105	0.00
46	1,1'-Biphenyl	40.000	40.912	-2.3	104	0.00
47	2-Chloronaphthalene	40.000	40.763	-1.9	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
 Data File : BG062618.D
 Acq On : 3 Sep 2024 9:49
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 04:25:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.588	-4.0	103	0.00
49	Acenaphthylene	40.000	40.633	-1.6	101	0.00
50	Dimethylphthalate	40.000	39.208	2.0	99	0.00
51	2,6-Dinitrotoluene	40.000	42.984	-7.5	103	0.00
52 C	Acenaphthene	40.000	40.658	-1.6	104	0.00
53	3-Nitroaniline	40.000	42.978	-7.4	103	0.00
54 P	2,4-Dinitrophenol	40.000	41.009	-2.5	102	0.00
55	Dibenzofuran	40.000	41.095	-2.7	105	0.00
56 P	4-Nitrophenol	40.000	42.862	-7.2	103	0.00
57	2,4-Dinitrotoluene	40.000	43.142	-7.9	106	0.00
58	Fluorene	40.000	41.692	-4.2	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.634	-4.1	108	0.00
60	Diethylphthalate	40.000	40.778	-1.9	105	0.00
61	4-Chlorophenyl-phenylether	40.000	41.817	-4.5	108	0.00
62	4-Nitroaniline	40.000	43.715	-9.3	108	0.00
63	Azobenzene	40.000	40.704	-1.8	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.364	-3.4	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.540	-3.8	106	0.00
67	4-Bromophenyl-phenylether	40.000	41.714	-4.3	107	0.00
68	Hexachlorobenzene	40.000	41.399	-3.5	106	0.00
69	Atrazine	40.000	34.736	13.2	104	0.00
70 C	Pentachlorophenol	40.000	40.422	-1.1	100	0.00
71	Phenanthrene	40.000	40.586	-1.5	104	0.00
72	Anthracene	40.000	40.884	-2.2	103	0.00
73	Carbazole	40.000	40.300	-0.7	104	0.00
74	Di-n-butylphthalate	40.000	39.648	0.9	100	0.00
75 C	Fluoranthene	40.000	40.257	-0.6	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	54.299	-35.7#	119	0.00
78	Pyrene	40.000	40.089	-0.2	105	0.00
79 S	Terphenyl-d14	80.000	77.551	3.1	105	0.00
80	Butylbenzylphthalate	40.000	37.517	6.2	97	0.00
81	Benzo(a)anthracene	40.000	40.283	-0.7	106	0.00
82	3,3'-Dichlorobenzidine	40.000	39.903	0.2	104	0.00
83	Chrysene	40.000	40.508	-1.3	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.217	4.5	99	0.00
85 c	Di-n-octyl phthalate	40.000	37.213	7.0	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.678	-1.7	106	0.00
88	Benzo(b)fluoranthene	40.000	40.781	-2.0	106	0.00
89	Benzo(k)fluoranthene	40.000	41.363	-3.4	106	0.00
90 C	Benzo(a)pyrene	40.000	40.402	-1.0	105	0.00
91	Dibenzo(a,h)anthracene	40.000	41.165	-2.9	108	0.01
92	Benzo(g,h,i)perylene	40.000	40.979	-2.4	106	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
Data File : BG062618.D
Acq On : 3 Sep 2024 9:49
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 04 04:25:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 02:36:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0