

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
 Data File : BG062628.D
 Acq On : 3 Sep 2024 16:30
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 04:31:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.482	0.457	5.2	110	0.00
3	Pyridine	1.389	1.364	1.8	110	0.00
4	n-Nitrosodimethylamine	0.675	0.665	1.5	108	0.00
5 S	2-Fluorophenol	1.157	1.160	-0.3	111	0.00
6	Aniline	1.562	1.499	4.0	104	0.00
7 S	Phenol-d6	1.672	1.669	0.2	109	0.00
8	2-Chlorophenol	1.242	1.220	1.8	110	0.00
9	Benzaldehyde	0.972	1.024	-5.3	122	0.00
10 C	Phenol	1.692	1.670	1.3	107	0.00
11	bis(2-Chloroethyl)ether	1.296	1.270	2.0	109	0.00
12	1,3-Dichlorobenzene	1.363	1.324	2.9	110	0.00
13 C	1,4-Dichlorobenzene	1.403	1.365	2.7	109	0.00
14	1,2-Dichlorobenzene	1.379	1.349	2.2	109	0.00
15	Benzyl Alcohol	1.237	1.245	-0.6	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.469	2.258	8.5	107	0.00
17	2-Methylphenol	1.180	1.149	2.6	107	0.00
18	Hexachloroethane	0.526	0.506	3.8	108	0.00
19 P	n-Nitroso-di-n-propylamine	1.317	1.233	6.4	101	0.00
20	3+4-Methylphenols	1.740	1.650	5.2	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.540	0.526	2.6	103	0.00
23 S	Nitrobenzene-d5	0.371	0.388	-4.6	109	0.00
24	Nitrobenzene	0.395	0.403	-2.0	104	0.00
25	Isophorone	0.815	0.789	3.2	101	0.00
26 C	2-Nitrophenol	0.149	0.161	-8.1	111	0.00
27	2,4-Dimethylphenol	0.219	0.221	-0.9	104	0.00
28	bis(2-Chloroethoxy)methane	0.436	0.445	-2.1	104	0.00
29 C	2,4-Dichlorophenol	0.306	0.323	-5.6	109	0.00
30	1,2,4-Trichlorobenzene	0.362	0.368	-1.7	109	0.00
31	Naphthalene	0.991	0.997	-0.6	106	0.00
32	Benzoic acid	0.235	0.236	-0.4	103	0.00
33	4-Chloroaniline	0.390	0.393	-0.8	106	0.00
34 C	Hexachlorobutadiene	0.253	0.261	-3.2	113	0.00
35	Caprolactam	0.118	0.115	2.5	101	0.00
36 C	4-Chloro-3-methylphenol	0.369	0.365	1.1	102	0.00
37	2-Methylnaphthalene	0.769	0.767	0.3	104	0.00
38	1-Methylnaphthalene	0.772	0.770	0.3	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.601	0.638	-6.2	110	0.00
41 P	Hexachlorocyclopentadiene	0.172	0.192	-11.6	111	0.00
42 S	2,4,6-Tribromophenol	0.281	0.294	-4.6	106	0.00
43 C	2,4,6-Trichlorophenol	0.381	0.404	-6.0	107	0.00
44	2,4,5-Trichlorophenol	0.432	0.454	-5.1	105	0.00
45 S	2-Fluorobiphenyl	1.235	1.275	-3.2	106	0.00
46	1,1'-Biphenyl	1.334	1.359	-1.9	104	0.00
47	2-Chloronaphthalene	1.081	1.106	-2.3	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
 Data File : BG062628.D
 Acq On : 3 Sep 2024 16:30
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 04:31:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.284	0.310	-9.2	108	0.00
49	Acenaphthylene	1.601	1.656	-3.4	104	0.00
50	Dimethylphthalate	1.462	1.462	0.0	102	0.00
51	2,6-Dinitrotoluene	0.277	0.307	-10.8	107	0.00
52 C	Acenaphthene	1.001	1.022	-2.1	105	0.00
53	3-Nitroaniline	0.268	0.288	-7.5	104	0.00
54 P	2,4-Dinitrophenol	0.155	0.166	-7.1	108	0.00
55	Dibenzofuran	1.717	1.743	-1.5	104	0.00
56 P	4-Nitrophenol	0.227	0.243	-7.0	103	0.00
57	2,4-Dinitrotoluene	0.377	0.422	-11.9	110	0.00
58	Fluorene	1.347	1.356	-0.7	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.409	0.421	-2.9	107	0.00
60	Diethylphthalate	1.449	1.437	0.8	103	0.00
61	4-Chlorophenyl-phenylether	0.773	0.787	-1.8	106	0.00
62	4-Nitroaniline	0.295	0.310	-5.1	104	0.00
63	Azobenzene	1.361	1.340	1.5	101	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	0.089	0.096	-7.9	110	0.00
66 c	n-Nitrosodiphenylamine	0.481	0.484	-0.6	104	0.00
67	4-Bromophenyl-phenylether	0.206	0.214	-3.9	107	0.00
68	Hexachlorobenzene	0.244	0.252	-3.3	107	0.00
69	Atrazine	0.163	0.135	17.2	100	0.00
70 C	Pentachlorophenol	0.157	0.161	-2.5	103	0.00
71	Phenanthrene	0.934	0.938	-0.4	104	0.00
72	Anthracene	0.918	0.924	-0.7	103	0.00
73	Carbazole	0.844	0.836	0.9	103	0.00
74	Di-n-butylphthalate	0.952	0.946	0.6	102	0.00
75 C	Fluoranthene	1.181	1.172	0.8	104	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	0.360	0.461	-28.1#	113	0.00
78	Pyrene	1.271	1.261	0.8	105	0.00
79 S	Terphenyl-d14	1.009	0.973	3.6	106	0.00
80	Butylbenzylphthalate	0.373	0.373	0.0	104	0.00
81	Benzo(a)anthracene	1.261	1.268	-0.6	107	0.00
82	3,3'-Dichlorobenzidine	0.411	0.416	-1.2	107	0.00
83	Chrysene	1.208	1.225	-1.4	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.567	0.565	0.4	104	0.00
85 c	Di-n-octyl phthalate	0.993	0.980	1.3	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	1.282	1.296	-1.1	109	0.01
88	Benzo(b)fluoranthene	1.105	1.114	-0.8	109	0.00
89	Benzo(k)fluoranthene	1.081	1.097	-1.5	108	0.00
90 C	Benzo(a)pyrene	0.976	0.981	-0.5	108	0.00
91	Dibenzo(a,h)anthracene	1.056	1.074	-1.7	110	0.00
92	Benzo(g,h,i)perylene	1.065	1.081	-1.5	109	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090324\
Data File : BG062628.D
Acq On : 3 Sep 2024 16:30
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 04 04:31:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 02:36:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0