

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090424\
 Data File : BG062636.D
 Acq On : 4 Sep 2024 9:44
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 11:25:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	40.617	-1.5	105	0.00
3	Pyridine	40.000	42.539	-6.3	106	0.00
4	n-Nitrosodimethylamine	40.000	41.689	-4.2	102	0.00
5 S	2-Fluorophenol	80.000	79.225	1.0	98	0.00
6	Aniline	40.000	39.036	2.4	94	0.00
7 S	Phenol-d6	80.000	80.173	-0.2	97	0.00
8	2-Chlorophenol	40.000	39.669	0.8	98	0.00
9	Benzaldehyde	40.000	37.542	6.1	96	0.00
10 C	Phenol	40.000	40.374	-0.9	98	0.00
11	bis(2-Chloroethyl)ether	40.000	38.987	2.5	96	0.00
12	1,3-Dichlorobenzene	40.000	39.953	0.1	101	0.00
13 C	1,4-Dichlorobenzene	40.000	40.185	-0.5	100	0.00
14	1,2-Dichlorobenzene	40.000	40.978	-2.4	101	0.00
15	Benzyl Alcohol	40.000	40.656	-1.6	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.254	6.9	97	0.00
17	2-Methylphenol	40.000	39.912	0.2	98	0.00
18	Hexachloroethane	40.000	38.293	4.3	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.452	3.9	92	0.00
20	3+4-Methylphenols	40.000	39.927	0.2	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	40.363	-0.9	94	0.00
23 S	Nitrobenzene-d5	80.000	85.691	-7.1	99	0.00
24	Nitrobenzene	40.000	42.741	-6.9	97	0.00
25	Isophorone	40.000	36.249	9.4	84	0.00
26 C	2-Nitrophenol	40.000	40.001	-0.0	91	0.00
27	2,4-Dimethylphenol	40.000	38.938	2.7	89	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.996	2.5	88	0.00
29 C	2,4-Dichlorophenol	40.000	42.250	-5.6	97	0.00
30	1,2,4-Trichlorobenzene	40.000	41.072	-2.7	98	0.00
31	Naphthalene	40.000	41.497	-3.7	97	0.00
32	Benzoic acid	40.000	39.954	0.1	91	0.00
33	4-Chloroaniline	40.000	42.411	-6.0	98	0.00
34 C	Hexachlorobutadiene	40.000	42.616	-6.5	103	0.00
35	Caprolactam	40.000	41.926	-4.8	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.893	-2.2	93	0.00
37	2-Methylnaphthalene	40.000	41.213	-3.0	95	0.00
38	1-Methylnaphthalene	40.000	40.912	-2.3	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.702	0.7	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.093	-5.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	85.553	-6.9	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.325	-0.8	94	0.00
44	2,4,5-Trichlorophenol	40.000	40.191	-0.5	93	0.00
45 S	2-Fluorobiphenyl	80.000	77.123	3.6	92	0.00
46	1,1'-Biphenyl	40.000	38.891	2.8	93	0.00
47	2-Chloronaphthalene	40.000	39.169	2.1	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090424\
 Data File : BG062636.D
 Acq On : 4 Sep 2024 9:44
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 04 11:25:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.046	-7.6	100	0.00
49	Acenaphthylene	40.000	41.036	-2.6	96	0.00
50	Dimethylphthalate	40.000	39.435	1.4	94	0.00
51	2,6-Dinitrotoluene	40.000	44.634	-11.6	100	0.00
52 C	Acenaphthene	40.000	40.749	-1.9	97	0.00
53	3-Nitroaniline	40.000	45.307	-13.3	102	0.00
54 P	2,4-Dinitrophenol	40.000	43.850	-9.6	102	0.00
55	Dibenzofuran	40.000	40.646	-1.6	97	0.00
56 P	4-Nitrophenol	40.000	45.673	-14.2	102	0.00
57	2,4-Dinitrotoluene	40.000	46.864	-17.2	108	0.00
58	Fluorene	40.000	40.942	-2.4	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.283	-5.7	102	0.00
60	Diethylphthalate	40.000	40.653	-1.6	98	0.00
61	4-Chlorophenyl-phenylether	40.000	40.864	-2.2	99	0.00
62	4-Nitroaniline	40.000	45.297	-13.2	104	0.00
63	Azobenzene	40.000	40.351	-0.9	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.944	-12.4	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.851	-2.1	100	0.00
67	4-Bromophenyl-phenylether	40.000	41.324	-3.3	101	0.00
68	Hexachlorobenzene	40.000	41.247	-3.1	101	0.00
69	Atrazine	40.000	34.831	12.9	100	0.00
70 C	Pentachlorophenol	40.000	41.881	-4.7	100	0.00
71	Phenanthrene	40.000	40.910	-2.3	100	0.00
72	Anthracene	40.000	41.389	-3.5	100	0.00
73	Carbazole	40.000	41.439	-3.6	102	0.00
74	Di-n-butylphthalate	40.000	40.854	-2.1	99	0.00
75 C	Fluoranthene	40.000	41.550	-3.9	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	108	0.00
77	Benzidine	40.000	52.865	-32.2#	117	0.00
78	Pyrene	40.000	39.507	1.2	105	0.00
79 S	Terphenyl-d14	80.000	76.817	4.0	106	0.00
80	Butylbenzylphthalate	40.000	38.630	3.4	101	0.00
81	Benzo(a)anthracene	40.000	39.486	1.3	106	0.00
82	3,3'-Dichlorobenzidine	40.000	39.998	0.0	106	0.00
83	Chrysene	40.000	40.185	-0.5	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.860	2.9	102	0.00
85 c	Di-n-octyl phthalate	40.000	37.918	5.2	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.806	-2.0	109	0.00
88	Benzo(b)fluoranthene	40.000	39.751	0.6	106	0.00
89	Benzo(k)fluoranthene	40.000	41.064	-2.7	108	0.00
90 C	Benzo(a)pyrene	40.000	40.071	-0.2	107	0.00
91	Dibenzo(a,h)anthracene	40.000	41.082	-2.7	110	0.00
92	Benzo(g,h,i)perylene	40.000	41.062	-2.7	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090424\
Data File : BG062636.D
Acq On : 4 Sep 2024 9:44
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 04 11:25:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 02:36:01 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0