

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090524\
 Data File : BG062661.D
 Acq On : 5 Sep 2024 11:10
 Operator : MA/JU
 Sample : P3390-03
 Misc : MDL-SOIL-4 ng
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Sep 05 12:56:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.923	152	53576	20.000	ng	0.00
21) Naphthalene-d8	10.714	136	226159	20.000	ng	0.00
39) Acenaphthene-d10	14.545	164	186253	20.000	ng	0.00
64) Phenanthrene-d10	17.288	188	514802	20.000	ng	0.00
76) Chrysene-d12	21.530	240	580074	20.000	ng	0.00
86) Perylene-d12	24.603	264	654115	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.502	112	353394	114.017	ng	0.00
7) Phenol-d6	7.089	99	502864	112.305	ng	0.00
23) Nitrobenzene-d5	9.074	82	342508	81.563	ng	0.00
42) 2,4,6-Tribromophenol	16.031	330	341122	130.141	ng	0.00
45) 2-Fluorobiphenyl	13.170	172	910935	79.221	ng	0.00
79) Terphenyl-d14	19.909	244	2036152	69.601	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.411	88	5657	4.381	ng	98
3) Pyridine	3.810	79	14096	3.789	ng	# 88
4) n-Nitrosodimethylamine	3.716	42	7312	4.043	ng	# 82
6) Aniline	7.247	93	17828	4.262	ng	92
8) 2-Chlorophenol	7.488	128	12986	3.902	ng	82
9) Benzaldehyde	7.065	77	13436	5.159	ng	92
10) Phenol	7.112	94	16783	3.702	ng	84
11) bis(2-Chloroethyl)ether	7.341	93	13651	3.933	ng	92
12) 1,3-Dichlorobenzene	7.811	146	13589	3.721	ng	89
13) 1,4-Dichlorobenzene	7.958	146	14868	3.956	ng	96
14) 1,2-Dichlorobenzene	8.275	146	13608	3.684	ng	94
15) Benzyl Alcohol	8.158	79	11211	3.384	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.440	45	23879	3.610	ng	99
17) 2-Methylphenol	8.364	107	9907	3.135	ng	91
18) Hexachloroethane	9.004	117	4987	3.539	ng	# 75
19) n-Nitroso-di-n-propyla...	8.716	70	10738	3.043	ng	# 95
20) 3+4-Methylphenols	8.687	107	14036	3.011	ng	97
22) Acetophenone	8.740	105	23015	3.772	ng	# 97
24) Nitrobenzene	9.116	77	17512	3.922	ng	97
25) Isophorone	9.639	82	31177	3.382	ng	96
26) 2-Nitrophenol	9.827	139	5750	3.408	ng	91
27) 2,4-Dimethylphenol	9.885	122	7770	3.143	ng	98
28) bis(2-Chloroethoxy)met...	10.120	93	17071	3.459	ng	99
29) 2,4-Dichlorophenol	10.361	162	12147	3.515	ng	91
30) 1,2,4-Trichlorobenzene	10.579	180	16774	4.094	ng	99
31) Naphthalene	10.761	128	44075	3.932	ng	97
33) 4-Chloroaniline	10.872	127	16812	3.811	ng	99
34) Hexachlorobutadiene	11.055	225	10900	3.809	ng	94
35) Caprolactam	11.601	113	4596	3.449	ng	# 88
36) 4-Chloro-3-methylphenol	11.989	107	12872	3.088	ng	98
37) 2-Methylnaphthalene	12.371	142	31458	3.618	ng	95
38) 1-Methylnaphthalene	12.594	142	30581	3.503	ng	# 97
40) 1,2,4,5-Tetrachloroben...	12.741	216	21741	3.881	ng	98
41) Hexachlorocyclopentadiene	12.723	237	5189	3.247	ng	96
43) 2,4,6-Trichlorophenol	13.052	196	13872	3.907	ng	98
44) 2,4,5-Trichlorophenol	13.052	196	13872	3.450	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090524\
 Data File : BG062661.D
 Acq On : 5 Sep 2024 11:10
 Operator : MA/JU
 Sample : P3390-03
 Misc : MDL-SOIL-4 ng
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :


Quant Time: Sep 05 12:56:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 31 02:36:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	13.375	154	50457	4.061	ng	93
47) 2-Chloronaphthalene	13.422	162	37703	3.744	ng	94
48) 2-Nitroaniline	13.622	65	9121	3.444	ng	96
49) Acenaphthylene	14.268	152	59811	4.011	ng	98
50) Dimethylphthalate	13.992	163	51584	3.788	ng	# 99
51) 2,6-Dinitrotoluene	14.116	165	8660	3.362	ng	92
52) Acenaphthene	14.609	154	35132	3.768	ng	99
53) 3-Nitroaniline	14.445	138	8268	3.317	ng	96
54) 2,4-Dinitrophenol	14.662	184	1701	1.177	ng	92
55) Dibenzofuran	14.944	168	63050	3.944	ng	99
56) 4-Nitrophenol	14.756	139	6357	3.009	ng	95
57) 2,4-Dinitrotoluene	14.897	165	11919	3.391	ng	# 93
58) Fluorene	15.590	166	48014	3.827	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.167	232	13960	3.665	ng	# 98
60) Diethylphthalate	15.361	149	49686	3.681	ng	98
61) 4-Chlorophenyl-phenyle...	15.585	204	26927	3.738	ng	# 93
62) 4-Nitroaniline	15.602	138	9494	3.454	ng	94
63) Azobenzene	15.878	77	46277	3.651	ng	97
65) 4,6-Dinitro-2-methylph...	15.667	198	5455	2.374	ng	84
66) n-Nitrosodiphenylamine	15.796	169	46012	3.718	ng	97
67) 4-Bromophenyl-phenylether	16.478	248	20125	3.788	ng	92
68) Hexachlorobenzene	16.595	284	24232	3.851	ng	91
69) Atrazine	16.742	200	19847	4.721	ng	98
70) Pentachlorophenol	16.948	266	7567	1.878	ng	91
71) Phenanthrene	17.330	178	95509	3.975	ng	98
72) Anthracene	17.424	178	91974	3.893	ng	98
73) Carbazole	17.688	167	79457	3.658	ng	98
74) Di-n-butylphthalate	18.252	149	85251	3.481	ng	98
75) Fluoranthene	19.339	202	121235	3.989	ng	99
77) Benzidine	19.521	184	42411	4.056	ng	95
78) Pyrene	19.703	202	131356	3.562	ng	99
80) Butylbenzylphthalate	20.596	149	33235	3.069	ng	97
81) Benzo(a)anthracene	21.513	228	140677	3.845	ng	98
82) 3,3'-Dichlorobenzidine	21.431	252	44079	3.698	ng	# 91
83) Chrysene	21.577	228	138951	3.965	ng	97
84) Bis(2-ethylhexyl)phtha...	21.436	149	51223	3.113	ng	# 94
85) Di-n-octyl phthalate	22.594	149	83194	2.888	ng	98
87) Indeno(1,2,3-cd)pyrene	28.052	276	160105	3.817	ng	# 87
88) Benzo(b)fluoranthene	23.628	252	127179	3.520	ng	99
89) Benzo(k)fluoranthene	23.693	252	138172	3.910	ng	98
90) Benzo(a)pyrene	24.456	252	118721	3.720	ng	99
91) Dibenzo(a,h)anthracene	28.117	278	128918	3.731	ng	# 95
92) Benzo(g,h,i)perylene	29.133	276	121969	3.501	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 05 12:56:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG083024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 31 02:36:01 2024
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 7.923 min Scan# 70

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:152 Resp: 53576

Ion Ratio Lower Upper

152 100

150 151.4 119.9 179.9

115 61.8 46.6 69.8



#2

1,4-Dioxane

Concen: 4.381 ng

RT: 3.411 min Scan# 21

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion: 88 Resp: 5657


Ion Ratio Lower Upper

88 100

58 76.7 62.3 93.5


43 36.9 27.5 41.3








#3
Pyridine
Concen: 3.789 ng
RT: 3.810 min Scan# 87
Delta R.T. 0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Instrument :
BNA_G
ClientSampleId :





Tgt Ion: 79 Resp: 14096
Ion Ratio Lower Upper
79 100
52 82.6 60.9 91.3
51 52.3 31.5 47.3#




#4
n-Nitrosodimethylamine
Concen: 4.043 ng
RT: 3.716 min Scan# 73
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10



Tgt Ion: 42 Resp: 7312
Ion Ratio Lower Upper
42 100
74 108.1 98.0 147.0
44 39.3 7.4 11.2#









#5
2-Fluorophenol
Concen: 114.017 ng
RT: 5.502 min Scan# 377
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:112 Resp: 353394
Ion Ratio Lower Upper
112 100
64 65.1 50.2 75.2
63 33.5 26.6 39.8



#6
Aniline
Concen: 4.262 ng
RT: 7.247 min Scan# 674
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Tgt Ion: 93 Resp: 17828
Ion Ratio Lower Upper
93 100
66 43.3 29.1 43.7
65 21.5 16.6 25.0








#7
Phenol-d6
Concen: 112.305 ng
RT: 7.089 min Scan# 647
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Instrument :
BNA_G
ClientSampleId :





Tgt Ion: 99 Resp: 502864
Ion Ratio Lower Upper
99 100
42 20.6 16.9 25.3
71 33.7 27.9 41.9




#8
2-Chlorophenol
Concen: 3.902 ng
RT: 7.488 min Scan# 715
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10



Tgt Ion: 128 Resp: 12986
Ion Ratio Lower Upper
128 100
130 43.0 10.0 50.0
64 59.1 28.8 68.8









#9
Benzaldehyde
Concen: 5.159 ng
RT: 7.065 min Scan# 642
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 13436
Ion Ratio Lower Upper
77 100
105 99.7 70.2 110.2
106 80.5 65.4 105.4



#10
Phenol
Concen: 3.702 ng
RT: 7.112 min Scan# 651
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion: 94 Resp: 16783
Ion Ratio Lower Upper
94 100
65 39.8 8.6 48.6
66 45.9 18.2 58.2





#11

bis(2-Chloroethyl)ether

Concen: 3.933 ng

RT: 7.341 min Scan# 690

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Instrument :

BNA_G

ClientSampleId :



Tgt Ion:	93	Resp:	13651
Ion	Ratio	Lower	Upper
93	100		
63	76.3	61.9	101.9
95	39.5	12.9	52.9



#12

1,3-Dichlorobenzene

Concen: 3.721 ng




RT: 7.811 min Scan# 770


Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Tgt Ion:	146	Resp:	13589
Ion	Ratio	Lower	Upper
146	100		
148	78.8	53.9	80.9
75	35.6	27.0	40.4









#13
1,4-Dichlorobenzene
Concen: 3.956 ng
RT: 7.958 min Scan# 795
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:146 Resp: 14868
Ion Ratio Lower Upper
146 100
148 70.6 53.0 79.6
111 44.6 34.9 52.3



#14
1,2-Dichlorobenzene
Concen: 3.684 ng
RT: 8.275 min Scan# 849
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:146 Resp: 13608
Ion Ratio Lower Upper
146 100
148 72.6 52.5 78.7
111 44.0 35.5 53.3









#15
Benzyl Alcohol
Concen: 3.384 ng
RT: 8.158 min Scan# 811
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 79 Resp: 11211
Ion Ratio Lower Upper
79 100
108 71.5 57.9 86.9
77 62.1 53.4 80.2



#16
2,2'-oxybis(1-Chloropropane)
Concen: 3.610 ng
RT: 8.440 min Scan# 877
Delta R.T. -0.015 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Tgt Ion: 45 Resp: 23879
Ion Ratio Lower Upper
45 100
77 11.9 0.0 31.7
79 7.6 0.0 28.3








#17
2-Methylphenol
Concen: 3.135 ng
RT: 8.364 min Scan# 80
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:107 Resp: 9907
Ion Ratio Lower Upper
107 100
108 99.2 91.5 137.3
77 49.4 40.0 60.0
79 45.1 38.4 57.6



#18
Hexachloroethane
Concen: 3.539 ng
RT: 9.004 min Scan# 973
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:117 Resp: 4987
Ion Ratio Lower Upper
117 100
119 96.3 73.4 110.0
201 162.8 93.9 140.9#









#19
n-Nitroso-di-n-propylamine
Concen: 3.043 ng
RT: 8.716 min Scan# 91
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 70 Resp: 10738
Ion Ratio Lower Upper
70 100
42 68.8 52.3 78.5
101 14.3 6.6 9.8#
130 19.5 14.6 22.0



#20
3+4-Methylphenols
Concen: 3.011 ng
RT: 8.687 min Scan# 919
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Tgt Ion:107 Resp: 14036
Ion Ratio Lower Upper
107 100
108 87.6 72.7 112.7
77 35.9 16.5 56.5
79 30.8 10.7 50.7








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.714 min Scan# 1264
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:	136	Resp:	226159
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	9.4	7.8	11.6
68	5.9	5.1	7.7



#22
Acetophenone
Concen: 3.772 ng
RT: 8.740 min Scan# 928
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:	105	Resp:	23015
Ion Ratio	Lower	Upper	
105	100		
71	1.7	3.5	5.3#
51	33.6	25.7	38.5
120	21.2	15.7	23.5





#23

Nitrobenzene-d5

Concen: 81.563 ng

RT: 9.074 min Scan# 90

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion: 82 Resp: 342508

Ion Ratio Lower Upper

82 100

128 38.0 32.5 48.7

54 53.1 42.5 63.7



#24

Nitrobenzene

Concen: 3.922 ng

RT: 9.116 min Scan# 992

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion: 77 Resp: 17512


Ion Ratio Lower Upper

77 100

123 42.3 33.1 49.7

65 11.3 11.0 16.4





#25

Isophorone

Concen: 3.382 ng

RT: 9.639 min Scan# 1081

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion: 82 Resp: 31177

Ion Ratio Lower Upper

82 100

95 6.7 5.4 8.0

138 17.9 12.6 18.8



#26

2-Nitrophenol

Concen: 3.408 ng

RT: 9.827 min Scan# 1113

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:139 Resp: 5750


Ion Ratio Lower Upper

139 100

109 27.0 26.6 39.8

65 52.7 46.5 69.7





#27

2,4-Dimethylphenol

Concen: 3.143 ng

RT: 9.885 min Scan# 1123

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:122 Resp: 7770

Ion Ratio Lower Upper

122 100

107 116.3 95.8 143.6

121 60.9 48.6 72.8



#28

bis(2-Chloroethoxy)methane

Concen: 3.459 ng

RT: 10.120 min Scan# 1163

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion: 93 Resp: 17071

Ion Ratio Lower Upper

93 100

95 32.4 25.4 38.0

123 10.7 7.9 11.9





#29

2,4-Dichlorophenol

Concen: 3.515 ng

RT: 10.361 min Scan# 1204

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:162 Resp: 12147

Ion Ratio Lower Upper

162 100

164 61.6 46.8 86.8

98 28.4 16.4 56.4



#30

1,2,4-Trichlorobenzene

Concen: 4.094 ng

RT: 10.579 min Scan# 1241

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:180 Resp: 16774


Ion Ratio Lower Upper

180 100

182 94.3 75.1 112.7

145 24.5 20.8 31.2





#31

Naphthalene

Concen: 3.932 ng

RT: 10.761 min Scan# 1272

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:128 Resp: 44075

Ion Ratio Lower Upper

128 100

129 12.4 8.8 13.2

127 13.0 11.3 16.9



#33

4-Chloroaniline

Concen: 3.811 ng

RT: 10.872 min Scan# 1291

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Tgt Ion:127 Resp: 16812




Ion Ratio Lower Upper


127 100

129 34.5 27.3 40.9

65 35.1 29.0 43.4

92 21.0 17.0 25.4





#34

Hexachlorobutadiene

Concen: 3.809 ng

RT: 11.055 min Scan# 1415

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:225 Resp: 10900

Ion Ratio Lower Upper

225 100

223 59.8 50.1 75.1

227 67.7 48.9 73.3



#35

Caprolactam

Concen: 3.449 ng

RT: 11.601 min Scan# 1415

Delta R.T. -0.038 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:113 Resp: 4596


Ion Ratio Lower Upper

113 100

55 191.6 174.1 214.1


56 117.2 129.8 169.8#








#36
4-Chloro-3-methylphenol
Concen: 3.088 ng
RT: 11.989 min Scan# 1482
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:107 Resp: 12872
Ion Ratio Lower Upper
107 100
144 24.5 19.8 29.8
142 82.8 64.7 97.1



#37
2-Methylnaphthalene
Concen: 3.618 ng
RT: 12.371 min Scan# 1546
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:142 Resp: 31458
Ion Ratio Lower Upper
142 100
141 80.3 68.9 103.3
115 35.1 27.5 41.3





#38

1-Methylnaphthalene

Concen: 3.503 ng

RT: 12.594 min Scan# 1

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:142 Resp: 30581

Ion Ratio Lower Upper

142 100

141 88.6 73.0 109.4

116 4.8 3.2 4.8#



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.545 min Scan# 1916

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:164 Resp: 186253


Ion Ratio Lower Upper

164 100

162 100.7 83.1 124.7

160 44.3 36.2 54.4









#40
1,2,4,5-Tetrachlorobenzene
Concen: 3.881 ng
RT: 12.741 min Scan# 1
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:216 Resp: 21741
Ion Ratio Lower Upper
216 100
214 79.4 61.6 92.4
179 19.6 15.9 23.9
108 13.6 11.4 17.0



#41
Hexachlorocyclopentadiene
Concen: 3.247 ng
RT: 12.723 min Scan# 1606
Delta R.T. 0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Tgt Ion:237 Resp: 5189
Ion Ratio Lower Upper
237 100
235 56.9 40.4 80.4
272 13.7 0.0 32.7








#42
2,4,6-Tribromophenol
Concen: 130.141 ng
RT: 16.031 min Scan# 2169
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Instrument :
BNA_G
ClientSampleId :





Tgt Ion:330 Resp: 341122
Ion Ratio Lower Upper
330 100
332 96.1 77.2 115.8
141 25.4 22.2 33.4




#43
2,4,6-Trichlorophenol
Concen: 3.907 ng
RT: 13.052 min Scan# 1662
Delta R.T. 0.073 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Tgt Ion:196 Resp: 13872
Ion Ratio Lower Upper
196 100
198 92.8 72.8 109.2
200 27.5 23.5 35.3








#44
2,4,5-Trichlorophenol
Concen: 3.450 ng
RT: 13.052 min Scan# 1662
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:196 Resp: 13872
Ion Ratio Lower Upper
196 100
198 92.8 75.5 113.3
97 39.8 34.1 51.1
132 26.0 17.4 26.0



#45
2-Fluorobiphenyl
Concen: 79.221 ng
RT: 13.170 min Scan# 1682
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:172 Resp: 910935
Ion Ratio Lower Upper
172 100
171 37.3 29.7 44.5
170 25.2 20.0 30.0





#46

1,1'-Biphenyl

Concen: 4.061 ng

RT: 13.375 min Scan# 1

Delta R.T. -0.009 min

Lab File: BG062661.D





Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :

Tgt Ion:154	Resp:	50457
Ion Ratio	Lower	Upper
154	100	
153	35.4	20.8 60.8
76	13.0	0.0 33.4



#47

2-Chloronaphthalene

Concen: 3.744 ng




RT: 13.422 min Scan# 1725


Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Tgt Ion:162	Resp:	37703
Ion Ratio	Lower	Upper
162	100	
127	30.5	28.7 43.1
164	32.4	26.9 40.3





#48

2-Nitroaniline

Concen: 3.444 ng

RT: 13.622 min Scan# 1759

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Instrument :

BNA_G

ClientSampleId :



Tgt Ion:	65	Resp:	9121
Ion Ratio	Lower	Upper	
65	100		
92	71.7	56.6	84.8
138	93.1	78.7	118.1



#49

Acenaphthylene

Concen: 4.011 ng




RT: 14.268 min Scan# 1869


Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Tgt Ion:	152	Resp:	59811
Ion Ratio	Lower	Upper	
152	100		
151	18.9	16.6	24.8
153	12.7	10.2	15.4





#50

Dimethylphthalate

Concen: 3.788 ng

RT: 13.992 min Scan# 1822

Delta R.T. -0.009 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:163 Resp: 51584

Ion Ratio Lower Upper

163 100

194 4.4 2.9 4.3#

164 10.5 8.3 12.5



#51

2,6-Dinitrotoluene

Concen: 3.362 ng

RT: 14.116 min Scan# 1843

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:165 Resp: 8660


Ion Ratio Lower Upper

165 100

63 49.6 46.6 69.8

89 55.4 47.2 70.8





#52

Acenaphthene

Concen: 3.768 ng

RT: 14.609 min Scan# 1927

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:154 Resp: 35132

Ion Ratio Lower Upper

154 100

153 118.0 95.0 142.4

152 57.1 45.0 67.6



#53

3-Nitroaniline

Concen: 3.317 ng

RT: 14.445 min Scan# 1899

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:138 Resp: 8268


Ion Ratio Lower Upper

138 100

108 10.7 8.6 13.0

92 123.9 95.7 143.5









#54
2,4-Dinitrophenol
Concen: 1.177 ng
RT: 14.662 min Scan# 1936
Delta R.T. 0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:184 Resp: 1701
Ion Ratio Lower Upper
184 100
63 50.2 49.0 73.4
154 62.1 50.5 75.7



#55
Dibenzofuran
Concen: 3.944 ng
RT: 14.944 min Scan# 1984
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:168 Resp: 63050
Ion Ratio Lower Upper
168 100
139 35.6 28.6 42.8
169 12.1 11.0 16.6









#56
4-Nitrophenol
Concen: 3.009 ng
RT: 14.756 min Scan# 1952
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:139 Resp: 6357
Ion Ratio Lower Upper
139 100
109 77.1 61.2 101.2
65 116.1 90.2 130.2



#57
2,4-Dinitrotoluene
Concen: 3.391 ng
RT: 14.897 min Scan# 1976
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10


Tgt Ion:165 Resp: 11919
Ion Ratio Lower Upper
165 100
63 48.3 33.8 50.6
89 68.1 57.9 86.9
182 0.0 1.8 2.8#








#58
Fluorene
Concen: 3.827 ng
RT: 15.590 min Scan# 2094
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:166 Resp: 48014
Ion Ratio Lower Upper
166 100
165 100.4 82.0 123.0
167 15.2 11.3 16.9



#59
2,3,4,6-Tetrachlorophenol
Concen: 3.665 ng
RT: 15.167 min Scan# 2022
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:232 Resp: 13960
Ion Ratio Lower Upper
232 100
131 34.1 28.3 42.5
130 0.7 1.7 2.5#
166 26.4 22.0 33.0









#60
Diethylphthalate
Concen: 3.681 ng
RT: 15.361 min Scan# 2056
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:149 Resp: 49686
Ion Ratio Lower Upper
149 100
177 23.1 18.0 27.0
150 10.7 9.8 14.8



#61
4-Chlorophenyl-phenylether
Concen: 3.738 ng
RT: 15.585 min Scan# 2093
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:204 Resp: 26927
Ion Ratio Lower Upper
204 100
206 39.0 25.9 38.9#
141 54.2 41.0 61.6









#62
4-Nitroaniline
Concen: 3.454 ng
RT: 15.602 min Scan# 2096
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:138 Resp: 9494
Ion Ratio Lower Upper
138 100
92 55.7 38.4 78.4
108 82.1 69.0 109.0



#63
Azobenzene
Concen: 3.651 ng
RT: 15.878 min Scan# 2143
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion: 77 Resp: 46277
Ion Ratio Lower Upper
77 100
182 26.7 5.0 45.0
105 22.8 0.5 40.5
51 31.9 11.4 51.4





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.288 min Scan# 2383

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:188 Resp: 514802

Ion Ratio Lower Upper

188 100

94 7.8 6.3 9.5

80 8.4 7.0 10.4



#65

4,6-Dinitro-2-methylphenol

Concen: 2.374 ng

RT: 15.667 min Scan# 2107

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:198 Resp: 5455


Ion Ratio Lower Upper

198 100

51 50.0 25.1 65.1

105 54.4 18.3 58.3





#66

n-Nitrosodiphenylamine

Concen: 3.718 ng

RT: 15.796 min Scan# 2130

Delta R.T. -0.009 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:169 Resp: 46012

Ion Ratio Lower Upper

169 100

168 67.5 55.4 83.2

167 38.8 29.0 43.6



#67

4-Bromophenyl-phenylether

Concen: 3.788 ng

RT: 16.478 min Scan# 2245

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:248 Resp: 20125


Ion Ratio Lower Upper

248 100

250 94.6 77.9 116.9

141 51.0 50.2 75.2





#68

Hexachlorobenzene

Concen: 3.851 ng

RT: 16.595 min Scan# 2265

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:284 Resp: 24232

Ion Ratio Lower Upper

284 100

142 24.1 23.5 35.3

249 26.9 25.4 38.0



#69

Atrazine

Concen: 4.721 ng

RT: 16.742 min Scan# 2290

Delta R.T. -0.009 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:200 Resp: 19847


Ion Ratio Lower Upper

200 100

173 25.2 5.2 45.2

215 45.4 27.3 67.3





#70

Pentachlorophenol

Concen: 1.878 ng

RT: 16.948 min Scan# 2324

Delta R.T. 0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:266 Resp: 7567

Ion Ratio Lower Upper

266 100

268 64.2 52.6 79.0

264 51.6 51.6 77.4



#71

Phenanthrene

Concen: 3.975 ng

RT: 17.330 min Scan# 2390

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:178 Resp: 95509


Ion Ratio Lower Upper

178 100

176 18.4 15.4 23.0

179 16.5 12.2 18.4





#72

Anthracene

Concen: 3.893 ng

RT: 17.424 min Scan# 2406

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:178 Resp: 91974

Ion Ratio Lower Upper

178 100

176 18.3 15.2 22.8

179 14.3 12.0 18.0



#73

Carbazole

Concen: 3.658 ng

RT: 17.688 min Scan# 2451

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:167 Resp: 79457


Ion Ratio Lower Upper

167 100

166 22.3 18.2 27.2

139 13.8 10.1 15.1





#74

Di-n-butylphthalate

Concen: 3.481 ng

RT: 18.252 min Scan# 2547

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:149 Resp: 85251

Ion Ratio Lower Upper

149 100

150 9.9 7.3 10.9

104 5.5 4.2 6.2



#75

Fluoranthene

Concen: 3.989 ng

RT: 19.339 min Scan# 2732

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:202 Resp: 121235


Ion Ratio Lower Upper

202 100

101 9.3 0.0 29.6

203 16.9 0.0 37.4









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.530 min Scan# 3105
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:240 Resp: 580074
Ion Ratio Lower Upper
240 100
120 7.9 6.6 9.8
236 26.5 20.2 30.4



#77
Benzidine
Concen: 4.056 ng
RT: 19.521 min Scan# 2763
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:184 Resp: 42411
Ion Ratio Lower Upper
184 100
185 15.9 11.0 16.6
183 14.1 9.7 14.5





#78

Pyrene

Concen: 3.562 ng

RT: 19.703 min Scan# 21

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:202 Resp: 131356

Ion Ratio Lower Upper

202 100

200 22.0 17.4 26.2

203 16.9 14.1 21.1



#79

Terphenyl-d14

Concen: 69.601 ng

RT: 19.909 min Scan# 2829

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:244 Resp: 2036152


Ion Ratio Lower Upper

244 100

212 8.0 6.6 10.0

122 7.8 7.2 10.8









#80
Butylbenzylphthalate
Concen: 3.069 ng
RT: 20.596 min Scan# 2946
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
149	100		
91	70.3	54.6	81.8
206	20.9	18.4	27.6



#81
Benzo(a)anthracene
Concen: 3.845 ng
RT: 21.513 min Scan# 3102
Delta R.T. -0.003 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.7	22.3	33.5
229	21.7	16.2	24.2





#82

3,3'-Dichlorobenzidine

Concen: 3.698 ng

RT: 21.431 min Scan# 3088

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:252 Resp: 44079

Ion Ratio Lower Upper

252 100

254 57.5 52.2 78.2

126 6.4 6.7 10.1#



#83

Chrysene

Concen: 3.965 ng

RT: 21.577 min Scan# 3113

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:228 Resp: 138951


Ion Ratio Lower Upper

228 100

226 29.5 24.9 37.3

229 18.5 16.0 24.0





#84

Bis(2-ethylhexyl)phthalate

Concen: 3.113 ng

RT: 21.436 min Scan# 3088

Delta R.T. 0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:149 Resp: 51223

Ion Ratio Lower Upper

149 100

167 28.9 20.5 30.7

279 4.8 3.1 4.7#



#85

Di-n-octyl phthalate

Concen: 2.888 ng

RT: 22.594 min Scan# 3286

Delta R.T. -0.003 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:149 Resp: 83194


Ion Ratio Lower Upper

149 100

167 1.7 1.4 2.0

43 12.0 8.8 13.2





#86

Perylene-d12

Concen: 20.000 ng

RT: 24.603 min Scan# 3

Delta R.T. -0.003 min


Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:264 Resp: 654115

Ion Ratio Lower Upper

264 100

260 22.9 19.0 28.4

265 21.9 17.3 25.9



#87

Indeno(1,2,3-cd)pyrene

Concen: 3.817 ng

RT: 28.052 min Scan# 4215

Delta R.T. -0.021 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:276 Resp: 160105


Ion Ratio Lower Upper

276 100

138 5.8 14.9 22.3#

277 24.4 20.3 30.5









#88
Benzo(b)fluoranthene
Concen: 3.520 ng
RT: 23.628 min Scan# 3463
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:252 Resp: 127179
Ion Ratio Lower Upper
252 100
253 22.8 18.2 27.2
125 6.9 6.2 9.2



#89
Benzo(k)fluoranthene
Concen: 3.910 ng
RT: 23.693 min Scan# 3473
Delta R.T. -0.009 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Tgt Ion:252 Resp: 138172
Ion Ratio Lower Upper
252 100
253 21.3 17.8 26.8
125 7.5 6.1 9.1





#90

Benzo(a)pyrene

Concen: 3.720 ng

RT: 24.456 min Scan# 3

Delta R.T. -0.009 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:252 Resp: 118721

Ion Ratio Lower Upper

252 100

253 20.7 17.1 25.7

125 8.6 6.6 10.0



#91

Dibenzo(a,h)anthracene

Concen: 3.731 ng

RT: 28.117 min Scan# 4226

Delta R.T. -0.015 min

Lab File: BG062661.D

Acq: 5 Sep 2024 11:10




Tgt Ion:278 Resp: 128918


Ion Ratio Lower Upper

278 100

139 7.8 9.2 13.8#


279 22.8 19.7 29.5





#92
Benzo(g,h,i)perylene
Concen: 3.501 ng
RT: 29.133 min Scan# 44
Delta R.T. -0.021 min
Lab File: BG062661.D
Acq: 5 Sep 2024 11:10

Instrument :
BNA_G
ClientSampleId :



Tgt Ion:276 Resp: 121969
Ion Ratio Lower Upper
276 100
277 24.5 18.8 28.2
138 12.2 12.7 19.1#

