


Data Path : Z:\HPCHEM1\BNA_G\Data\BG090617\
 Data File : BG028579.D
 Acq On : 6 Sep 2017 12:06
 Operator : SJ/JU
 Sample : DFTPP76
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP76

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG080217.M
 Title : SVOA CALIBRATION
 Last Update : Thu Sep 07 02:23:39 2017



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	31261	PASS
68	69	0.00	2	1.2	372	PASS
69	198	0.00	100	36.2	29877	PASS
70	69	0.00	2	0.7	208	PASS
127	198	10	80	40.6	33525	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	82568	PASS
199	198	5	9	7.0	5792	PASS
275	198	10	60	27.9	23052	PASS
365	198	1	100	3.2	2682	PASS
441	442	0.01	100	15.6	10532	PASS
442	198	50	100	81.8	67544	PASS
443	442	15	24	20.7	14008	PASS

Scan 2462 (18.056 min): BG028579.D
DFTPP76

m/z	Abundance
44.00	839.0
49.10	311.0
72.60	221.0
73.20	161.0
84.00	249.0
207.10	242.0
271.10	179.0
318.50	184.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

m/z	Abundance
38.10	210.0
39.10	1973.0
40.20	180.0
44.00	757.0
49.10	259.0
50.10	5906.0
51.10	28512.0
52.00	1272.0
55.00	217.0
56.00	537.0
57.10	2133.0
61.10	550.0
62.00	238.0
63.00	1324.0
65.10	449.0
68.10	577.0
69.00	25024.0
70.10	175.0
73.00	229.0
74.00	2558.0
75.10	4418.0
76.10	1234.0
77.10	27152.0
78.10	1906.0
79.10	1279.0
80.10	1572.0
81.00	1880.0
82.10	448.0
83.00	536.0
84.00	183.0
85.00	436.0
86.00	477.0
87.00	269.0
91.10	570.0
91.90	661.0
93.00	3285.0
93.90	501.0
95.10	160.0
98.00	1831.0
99.10	2196.0
100.00	272.0
101.00	1387.0
101.90	176.0
103.10	301.0
104.10	842.0
105.00	891.0
106.00	207.0
107.00	7267.0
108.00	1331.0
109.20	690.0
110.00	14786.0
111.00	2122.0
116.10	671.0
117.00	4826.0
118.10	266.0
122.10	677.0
123.00	999.0
124.10	337.0
125.00	427.0
127.00	24920.0
128.00	2115.0
129.00	9587.0
130.00	1157.0
132.90	164.0
134.00	271.0
135.00	983.0
136.00	389.0
137.00	710.0
141.00	1730.0
142.00	881.0
143.10	393.0
146.10	266.0
147.10	755.0
148.00	1312.0
148.90	256.0
150.00	160.0
151.40	238.0
153.10	301.0
153.90	263.0
155.00	743.0
156.00	1208.0
157.10	396.0
157.90	329.0
158.80	164.0
160.00	341.0
161.00	543.0
161.90	191.0
165.10	473.0
166.00	449.0
167.00	2586.0
168.00	1145.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

169.10	387.0
171.80	322.0
173.00	455.0
174.10	562.0
175.00	1099.0
176.00	419.0
177.00	739.0
179.00	2754.0
180.00	1484.0
181.00	534.0
182.20	234.0
184.30	298.0
185.10	1117.0
186.00	8762.0
187.00	2255.0
189.00	738.0
189.70	189.0
191.00	629.0
192.00	538.0
193.00	882.0
195.90	2248.0
198.00	56984.0
199.00	4199.0
199.90	394.0
201.10	213.0
201.40	257.0
203.10	736.0
204.00	2195.0
205.00	4365.0
206.00	15871.0
207.10	1919.0
207.90	746.0
208.90	181.0
210.00	197.0
210.80	560.0
211.80	156.0
215.00	338.0
215.90	416.0
217.00	4535.0
218.00	907.0
220.90	2367.0
222.00	617.0
223.00	949.0
224.10	8160.0
225.00	2079.0
226.20	309.0
227.00	4378.0
227.90	467.0
229.00	841.0
231.00	739.0
234.10	239.0
237.00	455.0
240.00	161.0
240.90	282.0
242.00	444.0
243.00	788.0
244.00	6866.0
245.00	731.0
246.00	880.0
247.10	307.0
248.90	393.0
252.20	229.0
252.80	162.0
253.30	302.0
255.00	35544.0
256.00	5925.0
257.20	321.0
258.00	1857.0
259.00	273.0
264.90	687.0
265.90	184.0
273.00	984.0
274.00	1716.0
275.00	14724.0
276.00	2619.0
276.90	1701.0
277.80	227.0
283.10	309.0
285.10	215.0
293.00	269.0
294.00	163.0
296.00	3794.0
297.00	631.0
297.90	160.0
303.00	514.0
314.00	413.0
314.90	422.0
316.10	183.0
320.90	176.0
321.90	203.0
323.00	1091.0
323.90	270.0
327.00	182.0
328.10	166.0
333.00	269.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

334.00	881.0
335.10	427.0
341.00	209.0
346.00	537.0
352.10	574.0
352.90	257.0
365.00	1896.0
366.00	192.0
372.00	605.0
372.90	253.0
390.00	151.0
401.90	269.0
402.90	258.0
420.70	163.0
422.00	296.0
423.00	1788.0
441.00	6279.0
442.00	34288.0
443.00	7450.0
444.00	744.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

m/z	Abundance
38.10	314.0
39.10	2125.0
40.30	155.0
41.20	170.0
44.00	995.0
49.00	484.0
50.10	10390.0
51.10	36560.0
52.10	1512.0
53.10	176.0
54.90	273.0
56.10	1326.0
57.00	2408.0
58.00	169.0
61.10	454.0
62.10	454.0
63.10	1461.0
64.00	211.0
65.00	548.0
69.00	34040.0
70.10	200.0
73.00	433.0
74.10	3918.0
75.10	6077.0
76.10	2660.0
77.10	40856.0
78.10	2569.0
79.10	3027.0
80.00	1714.0
81.00	3300.0
82.00	700.0
83.10	725.0
83.90	150.0
85.00	563.0
86.00	757.0
87.00	518.0
87.90	151.0
91.00	970.0
92.10	562.0
93.00	4205.0
93.90	267.0
94.90	175.0
95.90	226.0
98.10	3140.0
99.00	2676.0
99.90	159.0
101.00	1687.0
103.10	384.0
104.00	1472.0
105.00	1199.0
106.10	387.0
107.00	12460.0
108.00	1688.0
110.00	21512.0
111.00	3348.0
112.00	462.0
113.10	209.0
116.00	591.0
117.00	8462.0
118.00	576.0
122.00	672.0
123.10	1367.0
123.90	467.0
125.00	721.0
127.00	40872.0
128.10	3105.0
129.00	16074.0
130.10	1483.0
131.10	165.0
132.00	380.0
134.00	793.0
135.00	934.0
136.00	655.0
137.10	607.0
137.80	169.0
140.00	281.0
141.00	1893.0
142.00	472.0
143.10	416.0
144.00	279.0
146.00	426.0
147.00	1090.0
148.00	2380.0
149.10	543.0
151.00	530.0
152.00	451.0
153.10	759.0
154.10	413.0
155.00	1574.0
156.00	1890.0
157.00	257.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

157.90	318.0
159.00	283.0
160.00	536.0
161.00	1632.0
161.70	316.0
165.00	1046.0
166.00	569.0
167.10	3820.0
168.10	1582.0
169.10	457.0
170.10	206.0
170.90	157.0
172.00	615.0
172.90	574.0
174.00	1218.0
175.00	1544.0
176.00	289.0
177.00	1066.0
178.00	440.0
178.90	3660.0
180.10	2874.0
181.00	1458.0
182.10	205.0
183.90	311.0
185.10	1502.0
186.10	12576.0
187.00	4616.0
188.00	295.0
188.90	766.0
189.80	179.0
191.00	484.0
192.00	872.0
193.00	1421.0
194.10	178.0
196.00	2293.0
196.60	958.0
198.00	98072.0
199.00	6294.0
199.90	609.0
200.90	337.0
201.50	297.0
203.00	1133.0
204.00	3667.0
205.00	6377.0
206.00	28920.0
207.00	4158.0
207.90	795.0
208.90	302.0
210.50	615.0
211.10	1180.0
215.00	375.0
216.10	769.0
217.00	7870.0
218.10	798.0
221.00	3590.0
221.90	1123.0
222.90	1414.0
224.00	14671.0
225.10	3803.0
226.00	515.0
227.00	6763.0
228.00	1064.0
229.00	1550.0
229.90	235.0
231.10	709.0
233.90	462.0
235.00	703.0
236.00	400.0
237.00	642.0
239.00	317.0
240.00	278.0
240.90	531.0
242.10	589.0
243.00	1009.0
244.00	10795.0
245.00	1280.0
246.00	2452.0
247.10	500.0
247.80	150.0
248.90	446.0
249.80	260.0
253.00	387.0
253.40	405.0
255.00	64488.0
256.00	8980.0
257.00	855.0
258.00	3702.0
259.00	645.0
260.00	273.0
265.00	1648.0
266.10	471.0
268.90	166.0
269.90	185.0
271.00	209.0
272.00	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

273.00	1730.0
274.00	4899.0
275.00	25368.0
276.00	3901.0
277.00	2424.0
277.90	178.0
283.00	308.0
284.00	234.0
285.00	626.0
290.80	252.0
291.90	152.0
293.00	699.0
294.10	229.0
294.90	225.0
296.00	7991.0
297.00	1133.0
298.00	173.0
301.00	191.0
301.80	153.0
303.00	956.0
307.80	222.0
314.10	396.0
315.00	929.0
316.00	763.0
321.00	257.0
323.00	1913.0
324.00	360.0
327.00	581.0
329.00	307.0
333.10	387.0
334.10	1640.0
335.10	494.0
341.10	284.0
345.90	155.0
352.00	866.0
353.00	400.0
353.20	408.0
354.00	1175.0
365.00	3063.0
366.00	313.0
371.00	183.0
372.00	1599.0
373.00	233.0
382.90	494.0
401.90	407.0
402.90	884.0
404.00	280.0
421.00	601.0
422.00	530.0
423.00	3943.0
424.00	679.0
441.00	10885.0
442.00	75744.0
443.00	16528.0
444.00	1413.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

m/z	Abundance
38.00	265.0
39.10	1926.0
40.20	193.0
40.80	164.0
44.10	910.0
49.00	318.0
50.10	8131.0
51.10	28712.0
52.10	1915.0
55.10	284.0
56.00	770.0
57.00	2416.0
58.10	279.0
60.90	422.0
62.10	493.0
63.10	1183.0
64.20	192.0
65.10	524.0
68.10	541.0
69.00	30568.0
70.00	249.0
73.10	187.0
74.10	2608.0
75.00	5686.0
76.10	1601.0
77.10	33976.0
78.10	2281.0
79.10	2470.0
80.00	1548.0
81.00	2102.0
82.10	688.0
83.00	667.0
84.90	339.0
86.00	663.0
87.10	347.0
91.10	673.0
92.10	710.0
93.00	3797.0
96.00	168.0
98.00	2955.0
99.00	2905.0
100.00	180.0
101.10	1630.0
103.00	682.0
104.10	1033.0
105.00	829.0
106.00	442.0
107.00	10267.0
108.00	1240.0
108.90	537.0
110.00	19512.0
111.00	2689.0
112.00	509.0
112.80	181.0
116.10	787.0
117.00	8579.0
118.00	528.0
118.90	232.0
122.10	1064.0
123.00	1121.0
123.90	744.0
125.00	444.0
127.00	34784.0
128.00	2896.0
129.00	14901.0
130.00	1346.0
131.10	309.0
132.00	150.0
134.10	439.0
135.10	1013.0
136.00	565.0
137.00	694.0
138.00	178.0
140.00	279.0
141.00	1878.0
141.90	622.0
143.00	337.0
145.00	158.0
146.10	389.0
146.90	907.0
148.10	2485.0
149.00	651.0
150.00	204.0
150.90	190.0
151.90	150.0
152.90	401.0
154.10	486.0
155.00	1163.0
156.00	2154.0
158.00	546.0
158.90	372.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

159.90	512.0
161.10	891.0
162.00	343.0
163.80	192.0
165.00	1096.0
166.00	541.0
167.10	3376.0
168.00	1661.0
169.10	280.0
171.00	324.0
172.10	464.0
173.00	597.0
174.00	910.0
175.00	1924.0
176.10	484.0
176.90	907.0
178.00	651.0
179.00	3546.0
180.10	2761.0
181.00	849.0
182.10	354.0
183.90	322.0
185.10	1313.0
186.10	12217.0
187.00	3404.0
187.90	389.0
189.10	763.0
190.90	554.0
191.90	1222.0
193.00	1507.0
193.90	413.0
195.00	328.0
196.00	2483.0
198.00	92648.0
199.00	6884.0
200.00	748.0
201.50	708.0
203.00	859.0
204.00	3668.0
205.00	6780.0
206.00	25744.0
207.00	3808.0
207.90	717.0
209.00	364.0
210.30	522.0
211.00	1079.0
215.00	154.0
216.10	1064.0
217.00	6494.0
217.90	990.0
221.00	3099.0
221.70	850.0
223.00	1596.0
224.00	13945.0
225.00	3255.0
226.00	461.0
227.00	6545.0
227.90	1173.0
229.00	1391.0
231.00	885.0
232.10	218.0
234.00	182.0
235.00	474.0
235.90	379.0
237.00	767.0
239.00	371.0
240.90	257.0
242.00	516.0
243.00	683.0
244.00	10989.0
245.00	1535.0
246.00	2668.0
246.80	453.0
248.90	377.0
251.10	214.0
252.00	465.0
253.10	345.0
255.00	63960.0
256.00	8863.0
257.00	697.0
258.00	3618.0
258.90	511.0
261.00	174.0
262.90	188.0
264.10	163.0
265.00	1466.0
265.90	321.0
269.90	174.0
271.20	152.0
271.90	172.0
273.00	2298.0
274.00	5523.0
275.00	29064.0
276.00	4176.0
277.00	2628.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76

278.00	542.0
283.10	215.0
284.00	242.0
285.00	627.0
290.90	153.0
293.00	763.0
294.00	292.0
296.00	7921.0
297.00	1147.0
297.90	154.0
302.90	1097.0
304.00	431.0
308.20	219.0
310.10	241.0
314.00	584.0
315.00	1118.0
316.10	591.0
319.00	156.0
320.90	151.0
322.00	257.0
323.00	2409.0
324.10	675.0
326.90	498.0
327.90	397.0
332.00	257.0
333.10	387.0
334.00	2866.0
335.00	489.0
340.90	372.0
341.90	193.0
346.00	762.0
352.10	595.0
353.00	510.0
354.00	806.0
355.00	194.0
365.00	3087.0
366.00	611.0
371.00	270.0
372.00	1072.0
383.10	331.0
390.10	218.0
392.00	190.0
401.10	204.0
402.00	640.0
403.10	694.0
404.00	566.0
421.00	658.0
421.90	738.0
423.00	5065.0
424.10	1298.0
441.00	14432.0
442.00	92600.0
443.00	18048.0
444.00	1793.0
474.10	173.0

Instrument :
BNA_G
ClientSampleId :
DFTPP76