

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG090717\
 Data File : BG028599.D
 Acq On : 7 Sep 2017 18:34
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG090717

Quant Time: Sep 07 19:22:30 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 18:39:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	39.314	1.7	104	0.00
3	Pyridine	40.000	38.205	4.5	101	0.00
4	n-Nitrosodimethylamine	40.000	37.462	6.3	99	0.00
5 S	2-Fluorophenol	80.000	76.855	3.9	101	0.00
6	Aniline	40.000	38.901	2.7	99	0.00
7 S	Phenol-d6	80.000	78.263	2.2	100	0.00
8	2-Chlorophenol	40.000	38.453	3.9	101	0.00
9	Benzaldehyde	40.000	38.318	4.2	101	0.00
10 C	Phenol	40.000	39.147	2.1	101	0.00
11	bis(2-Chloroethyl)ether	40.000	37.952	5.1	100	0.00
12	1,3-Dichlorobenzene	40.000	37.715	5.7	101	0.00
13 C	1,4-Dichlorobenzene	40.000	39.046	2.4	101	0.00
14	1,2-Dichlorobenzene	40.000	38.256	4.4	100	0.00
15	Benzyl Alcohol	40.000	38.148	4.6	97	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	37.401	6.5	98	0.00
17	2-Methylphenol	40.000	38.381	4.0	101	0.00
18	Hexachloroethane	40.000	36.552	8.6	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.722	3.2	100	0.00
20	3+4-Methylphenols	40.000	39.383	1.5	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.935	0.2	100	0.00
23 S	Nitrobenzene-d5	80.000	79.411	0.7	99	0.00
24	Nitrobenzene	40.000	40.300	-0.7	100	0.00
25	Isophorone	40.000	39.750	0.6	98	0.00
26 C	2-Nitrophenol	40.000	40.183	-0.5	103	0.00
27	2,4-Dimethylphenol	40.000	39.349	1.6	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.647	0.9	101	0.00
29 C	2,4-Dichlorophenol	40.000	40.466	-1.2	101	0.00
30	1,2,4-Trichlorobenzene	40.000	39.651	0.9	99	0.00
31	Naphthalene	40.000	39.350	1.6	99	0.00
32	Benzoic acid	40.000	42.458	-6.1	103	0.00
33	4-Chloroaniline	40.000	39.859	0.4	98	0.00
34 C	Hexachlorobutadiene	40.000	39.779	0.6	99	0.00
35	Caprolactam	40.000	40.240	-0.6	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.167	-0.4	98	0.00
37	2-Methylnaphthalene	40.000	39.868	0.3	99	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.335	-0.8	100	0.00
40 P	Hexachlorocyclopentadiene	40.000	38.616	3.5	100	0.00
41 S	2,4,6-Tribromophenol	80.000	81.880	-2.3	101	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.210	-0.5	98	0.00
43	2,4,5-Trichlorophenol	40.000	40.343	-0.9	99	0.00
44 S	2-Fluorobiphenyl	80.000	79.708	0.4	99	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG090717\
 Data File : BG028599.D
 Acq On : 7 Sep 2017 18:34
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG090717

Quant Time: Sep 07 19:22:30 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 18:39:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	39.681	0.8	100	0.00
46	2-Chloronaphthalene	40.000	40.521	-1.3	101	0.00
47	2-Nitroaniline	40.000	39.377	1.6	99	0.00
48	Acenaphthylene	40.000	39.658	0.9	100	0.00
49	Dimethylphthalate	40.000	39.330	1.7	99	0.00
50	2,6-Dinitrotoluene	40.000	40.455	-1.1	98	0.00
51 C	Acenaphthene	40.000	40.449	-1.1	100	0.00
52	3-Nitroaniline	40.000	39.851	0.4	98	0.00
53 P	2,4-Dinitrophenol	40.000	39.862	0.3	98	0.00
54	Dibenzofuran	40.000	40.444	-1.1	101	0.00
55 P	4-Nitrophenol	40.000	40.703	-1.8	103	0.00
56	2,4-Dinitrotoluene	40.000	40.810	-2.0	104	0.00
57	Fluorene	40.000	40.246	-0.6	102	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.141	-0.4	101	0.00
59	Diethylphthalate	40.000	40.020	-0.1	102	0.00
60	4-Chlorophenyl-phenylether	40.000	40.184	-0.5	101	0.00
61	4-Nitroaniline	40.000	40.460	-1.2	100	0.00
62	Azobenzene	40.000	39.148	2.1	99	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
64	4,6-Dinitro-2-methylphenol	40.000	41.183	-3.0	105	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.113	-0.3	99	0.00
66	4-Bromophenyl-phenylether	40.000	40.473	-1.2	100	0.00
67	Hexachlorobenzene	40.000	40.476	-1.2	101	0.00
68	Atrazine	40.000	39.399	1.5	101	0.00
69 C	Pentachlorophenol	40.000	40.434	-1.1	103	0.00
70	Phenanthrene	40.000	40.661	-1.7	102	0.00
71	Anthracene	40.000	40.404	-1.0	102	0.00
72	Carbazole	40.000	39.911	0.2	100	0.00
73	Di-n-butylphthalate	40.000	39.238	1.9	99	0.00
74 C	Fluoranthene	40.000	40.179	-0.4	101	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
76	Benzidine	40.000	38.651	3.4	95	0.00
77	Pyrene	40.000	40.521	-1.3	100	0.00
78 S	Terphenyl-d14	80.000	81.469	-1.8	101	0.00
79	Butylbenzylphthalate	40.000	40.219	-0.5	102	0.00
80	Benzo(a)anthracene	40.000	40.072	-0.2	101	0.00
81	3,3'-Dichlorobenzidine	40.000	40.287	-0.7	100	0.00
82	Chrysene	40.000	39.589	1.0	99	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	39.134	2.2	99	0.00
84 c	Di-n-octyl phthalate	40.000	39.621	0.9	100	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	40.287	-0.7	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Benzo(b)fluoranthene	40.000	39.634	0.9	101	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG090717\
Data File : BG028599.D
Acq On : 7 Sep 2017 18:34
Operator : SJ/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG090717

Quant Time: Sep 07 19:22:30 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 07 18:39:50 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.327	-0.8	99	0.00
89 C	Benzo(a)pyrene	40.000	40.083	-0.2	101	0.00
90	Dibenzo(a,h)anthracene	40.000	40.342	-0.9	100	-0.01
91	Benzo(a,h,i)perylene	40.000	40.152	-0.4	101	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0