

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG090717\
 Data File : BG028597.D
 Acq On : 7 Sep 2017 17:15
 Operator : SJ/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Quant Time: Sep 07 18:02:21 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 16:57:10 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.34	152	37457	20.00	ng	0.00
21) Naphthalene-d8	11.16	136	159862	20.00	ng	0.00
38) Acenaphthene-d10	14.95	164	108460	20.00	ng	0.00
63) Phenanthrene-d10	17.70	188	270567	20.00	ng	0.00
75) Chrysene-d12	22.04	240	320559	20.00	ng	0.00
86) Perylene-d12	25.54	264	301783	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.91	112	278134	132.82	ng	0.00
7) Phenol-d6	7.50	99	415597	137.75	ng	0.00
23) Nitrobenzene-d5	9.51	82	426983	149.24	ng	0.00
41) 2,4,6-Tribromophenol	16.44	330	203060	126.33	ng	0.00
44) 2-Fluorobiphenyl	13.57	172	999978	121.59	ng	0.00
78) Terphenyl-d14	20.29	244	1718083	121.18	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.86	88	55945	71.753	ng	93
3) Pyridine	4.25	79	174865	78.190	ng	93
4) n-Nitrosodimethylamine	4.16	42	82937	66.011	ng	# 80
6) Aniline	7.67	93	264753	75.377	ng	98
8) 2-Chlorophenol	7.91	128	156392	65.288	ng	94
9) Benzaldehyde	7.48	77	106199	64.296	ng	97
10) Phenol	7.52	94	225537	71.302	ng	91
11) bis(2-Chloroethyl)ether	7.75	93	164261	70.829	ng	94
12) 1,3-Dichlorobenzene	8.23	146	181038	64.538	ng	95
13) 1,4-Dichlorobenzene	8.38	146	182093	62.748	ng	94
14) 1,2-Dichlorobenzene	8.70	146	176644	62.697	ng	99
15) Benzyl Alcohol	8.58	79	164690	89.333	ng	93
16) 2,2'-oxybis(1-Chloropropan	8.85	45	326159	69.726	ng	99
17) 2-Methylphenol	8.78	107	141173	67.845	ng	95
18) Hexachloroethane	9.42	117	70563	74.740	ng	96
19) n-Nitroso-di-n-propylamine	9.13	70	157748	74.889	ng	94
20) 3+4-Methylphenols	9.11	107	199094	68.123	ng	95
22) Acetophenone	9.16	105	275305	71.421	ng	# 92
24) Nitrobenzene	9.55	77	226511	75.273	ng	97
25) Isophorone	10.07	82	387185	70.846	ng	99
26) 2-Nitrophenol	10.27	139	94494	65.472	ng	96
27) 2,4-Dimethylphenol	10.31	122	168211	66.398	ng	95
28) bis(2-Chloroethoxy)methane	10.54	93	232006	72.649	ng	99
29) 2,4-Dichlorophenol	10.80	162	168407	61.399	ng	97
30) 1,2,4-Trichlorobenzene	11.01	180	203203	66.081	ng	98
31) Naphthalene	11.21	128	530004	66.497	ng	99
32) Benzoic acid	10.48	122	111441	90.414	ng	92
33) 4-Chloroaniline	11.31	127	228689	70.207	ng	92
34) Hexachlorobutadiene	11.48	225	148207	69.047	ng	94
35) Caprolactam	12.11	113	61658	73.479	ng	96
36) 4-Chloro-3-methylphenol	12.42	107	212415	71.784	ng	98
37) 2-Methylnaphthalene	12.80	142	392803	64.562	ng	96
39) 1,2,4,5-Tetrachlorobenzene	13.15	216	255454	59.569	ng	# 96
40) Hexachlorocyclopentadiene	13.12	237	104069	51.999	ng	96

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG090717\
 Data File : BG028597.D
 Acq On : 7 Sep 2017 17:15
 Operator : SJ/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Quant Time: Sep 07 18:02:21 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 16:57:10 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	13.39	196	168278	64.718	nq	90
43) 2,4,5-Trichlorophenol	13.47	196	167759	61.271	nq #	90
45) 1,1'-Biphenyl	13.78	154	548574	58.861	nq	98
46) 2-Chloronaphthalene	13.84	162	427541	63.001	nq	99
47) 2-Nitroaniline	14.04	65	170877	84.275	nq	94
48) Acenaphthylene	14.68	152	640819	59.340	nq	98
49) Dimethylphthalate	14.39	163	533076	58.860	nq	98
50) 2,6-Dinitrotoluene	14.52	165	121697	62.899	nq	84
51) Acenaphthene	15.02	154	393173	58.030	nq	92
52) 3-Nitroaniline	14.86	138	121390	66.199	nq #	86
53) 2,4-Dinitrophenol	15.07	184	70857	64.547	nq	95
54) Dibenzofuran	15.35	168	610320	58.942	nq	95
55) 4-Nitrophenol	15.17	139	85980	62.757	nq	91
56) 2,4-Dinitrotoluene	15.32	165	172360	64.435	nq #	94
57) Fluorene	16.00	166	561776	60.720	nq	99
58) 2,3,4,6-Tetrachlorophenol	15.57	232	148135	60.911	nq	95
59) Diethylphthalate	15.74	149	551962	62.866	nq	98
60) 4-Chlorophenyl-phenylether	15.97	204	313883	61.529	nq	92
61) 4-Nitroaniline	16.03	138	129630	66.126	nq #	85
62) Azobenzene	16.27	77	493994	71.771	nq	92
64) 4,6-Dinitro-2-methylphenol	16.09	198	116277	64.471	nq	95
65) n-Nitrosodiphenylamine	16.20	169	477064	60.401	nq	99
66) 4-Bromophenyl-phenylether	16.87	248	211647	62.707	nq	95
67) Hexachlorobenzene	17.01	284	231347	62.901	nq #	90
68) Atrazine	17.14	200	161926	53.451	nq	99
69) Pentachlorophenol	17.35	266	111866	54.325	nq	96
70) Phenanthrene	17.75	178	856642	60.528	nq	96
71) Anthracene	17.84	178	868519	61.622	nq	98
72) Carbazole	18.11	167	773554	60.973	nq #	95
73) Di-n-butylphthalate	18.63	149	937423	67.248	nq #	94
74) Fluoranthene	19.75	202	1084117	63.792	nq	93
76) Benzidine	19.91	184	352819	47.153	nq	97
77) Pyrene	20.10	202	1113215	64.974	nq	95
79) Butylbenzylphthalate	20.97	149	431950	70.181	nq	94
80) Benzo(a)anthracene	22.02	228	1110525	63.022	nq	95
81) 3,3'-Dichlorobenzidine	21.91	252	441167	63.670	nq	95
82) Chrysene	22.09	228	1059312	62.834	nq	95
83) Bis(2-ethylhexyl)phthalate	21.87	149	612331	68.340	nq	99
84) Di-n-octyl phthalate	23.17	149	1055241	69.377	nq #	90
85) Indeno(1,2,3-cd)pyrene	29.59	276	1275301	65.421	nq	99
87) Benzo(b)fluoranthene	24.43	252	1133864	64.298	nq	99
88) Benzo(k)fluoranthene	24.50	252	1110580	63.001	nq	95
89) Benzo(a)pyrene	25.39	252	1074427	64.597	nq	99
90) Dibenzo(a,h)anthracene	29.66	278	1130826	67.336	nq	99
91) Benzo(g,h,i)perylene	30.88	276	1089376	67.306	nq #	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA G\DATA\BG090717\
Data File : BG028597.D
Acq On : 7 Sep 2017 17:15
Operator : SJ/JU
Sample : SSTDICC060
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Quant Time: Sep 07 18:02:21 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 07 16:57:10 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.34 min Scan# 809
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Instrument :
BNA_G
ClientSampleId :
SSTDICC060




Tot Ion:152 Resp: 37457
Ion Ratio Lower Upper
152 100
150 151.2 114.0 171.0
115 59.4 48.1 72.1




#2

1,4-Dioxane
Concen: 71.753 ng
RT: 3.86 min Scan# 47
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion: 88 Resp: 55945
Ion Ratio Lower Upper
88 100
58 103.5 79.4 119.2
43 34.1 33.8 50.6





#3

Pvridine

Concen: 78.190 ng

RT: 4.25 min Scan# 113

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tot Ion: 79 Resp: 174865

Ion Ratio Lower Upper

79 100

52 105.1 77.8 116.6

51 49.3 43.2 64.8



#4

n-Nitrosodimethylamine

Concen: 66.011 ng

RT: 4.16 min Scan# 98

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15




Tgt Ion: 42 Resp: 82937


Ion Ratio Lower Upper

42 100

74 117.1 76.7 115.1#

44 7.5 6.1 9.1





#5

2-Fluorophenol

Concen: 132.818 ng

RT: 5.91 min Scan# 396

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:112 Resp: 278134

Ion Ratio Lower Upper

112 100

64 74.5 61.8 92.6




63 41.8 37.2 55.8



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 75.377 ng

RT: 7.67 min Scan# 694

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion: 93 Resp: 264753

Ion Ratio Lower Upper

93 100

66 44.5 35.4 53.2



65 24.2 21.2 31.8






Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 137.750 ng

RT: 7.50 min Scan# 665

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tot Ion: 99 Resp: 415597

Ion Ratio Lower Upper

99 100

42 20.3 20.5 30.7#

71 40.3 37.6 56.4



#8

2-Chlorophenol

Concen: 65.288 ng

RT: 7.91 min Scan# 735

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:128 Resp: 156392


Ion Ratio Lower Upper

128 100

130 31.8 11.4 51.4

64 59.9 46.6 86.6





#9

Benzaldehyde

Concen: 64.296 ng

RT: 7.48 min Scan# 663

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion: 77 Resp: 106199

Ion Ratio Lower Upper

77 100

105 85.9 60.9 100.9




106 75.5 55.1 95.1



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 71.302 ng

RT: 7.52 min Scan# 670

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion: 94 Resp: 225537

Ion Ratio Lower Upper

94 100

65 35.0 20.6 60.6



66 48.9 35.5 75.5




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
 bis(2-Chloroethyl)ether
 Concen: 70.829 ng
 RT: 7.75 min Scan# 709
 Delta R.T. 0.01 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Tot Ion: 93 Resp: 164261
 Ion Ratio Lower Upper
 93 100
 63 92.7 78.5 118.5
 95 32.3 9.8 49.8




Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 63.00 (62.70 to 63.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC






#12
 1,3-Dichlorobenzene
 Concen: 64.538 ng
 RT: 8.23 min Scan# 790
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion: 146 Resp: 181038
 Ion Ratio Lower Upper
 146 100
 148 65.2 49.2 73.8
 75 38.7 33.4 50.2



Abundance Ion 146.00 (145.70 to 146.70): BGC
 Ion 148.00 (147.70 to 148.70): BGC
 Ion 75.00 (74.70 to 75.70): BGC





#13

1,4-Dichlorobenzene

Concen: 62.748 ng

RT: 8.38 min Scan# 815

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:146 Resp: 182093

Ion Ratio Lower Upper

146 100

148 71.2 52.2 78.2




111 48.5 37.0 55.6



Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 62.697 ng

RT: 8.70 min Scan# 870

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:146 Resp: 176644

Ion Ratio Lower Upper

146 100

148 69.7 54.6 81.8



111 49.9 39.7 59.5




Abundance Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 89.333 ng

RT: 8.58 min Scan# 849

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tot Ion: 79 Resp: 164690

Ion Ratio Lower Upper

79 100

108 67.8 45.5 68.3

77 67.2 54.3 81.5



#16

2,2'-oxybis(1-Chloropropane)

Concen: 69.726 ng

RT: 8.85 min Scan# 896

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15




Tgt Ion: 45 Resp: 326159


Ion Ratio Lower Upper

45 100

77 11.0 0.0 30.6

79 8.0 0.0 28.5





#17

2-Methylphenol

Concen: 67.845 ng

RT: 8.78 min Scan# 883

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060

Tot Ion:107 Resp: 141173


Ion Ratio Lower Upper

107 100

108 107.7 85.6 128.4

77 57.0 51.4 77.2

79 55.3 49.2 73.8






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BG

Ion 79.00 (78.70 to 79.70): BG



#18

Hexachloroethane

Concen: 74.740 ng

RT: 9.42 min Scan# 993

Delta R.T. -0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:117 Resp: 70563

Ion Ratio Lower Upper

117 100

119 94.8 69.8 104.6



201 118.5 95.4 143.0




Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 74.889 ng

RT: 9.13 min Scan# 944

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060

Tot Ion: 70 Resp: 157748


Ion Ratio Lower Upper

70 100

42 63.3 56.1 84.1

101 10.2 7.5 11.3

130 19.0 14.6 21.8






Abundance Ion 70.00 (69.70 to 70.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 101.00 (100.70 to 101.70): BGC

Ion 130.00 (129.70 to 130.70): BGC



#20

3+4-Methylphenols

Concen: 68.123 ng

RT: 9.11 min Scan# 939

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:107 Resp: 199094


Ion Ratio Lower Upper

107 100

108 87.5 70.8 110.8

77 42.2 25.8 65.8

79 35.1 20.1 60.1






Abundance Ion 107.00 (106.70 to 107.70): BGC

Ion 108.00 (107.70 to 108.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.16 min Scan# 1289
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Tot Ion:136	Resp:	159862
Ion Ratio	Lower	Upper
136	100	
137	10.9	8.9 13.3
54	12.7	9.3 13.9
68	5.9	5.1 7.7




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#22
Acetophenone
Concen: 71.421 ng
RT: 9.16 min Scan# 949
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:105	Resp:	275305
Ion Ratio	Lower	Upper
105	100	
71	0.3	0.9 1.3#
51	34.5	34.0 51.0
120	21.7	17.3 25.9



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BGC
Ion 51.00 (50.70 to 51.70): BGC
Ion 120.00 (119.70 to 120.70): E





#23

Nitrobenzene-d5

Concen: 149.239 ng

RT: 9.51 min Scan# 1008

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion: 82 Resp: 426983

Ion Ratio Lower Upper




82 100

128 38.7 26.4 39.6

54 58.8 49.4 74.2



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 128.00 (127.70 to 128.70): BGC
Ion 54.00 (53.70 to 54.70): BGC



#24

Nitrobenzene

Concen: 75.273 ng

RT: 9.55 min Scan# 1015

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion: 77 Resp: 226511

Ion Ratio Lower Upper



77 100


123 35.0 26.1 39.1

65 14.9 12.4 18.6



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 123.00 (122.70 to 123.70): BGC
Ion 65.00 (64.70 to 65.70): BGC





#25

Isophorone

Concen: 70.846 ng

RT: 10.07 min Scan# 1104

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tot Ion: 82 Resp: 387185

Ion Ratio Lower Upper

82 100

95 8.6 7.5 11.3

138 15.7 12.3 18.5



#26

2-Nitrophenol

Concen: 65.472 ng

RT: 10.27 min Scan# 1137

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15




Tgt Ion: 139 Resp: 94494


Ion Ratio Lower Upper

139 100

109 43.8 38.6 58.0

65 69.3 57.1 85.7









#27
2,4-Dimethylphenol
Concen: 66.398 ng
RT: 10.31 min Scan# 1144
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tot Ion:122 Resp: 168211
Ion Ratio Lower Upper
122 100
107 136.0 104.2 156.4
121 60.7 46.0 69.0



#28
bis(2-Chloroethoxy)methane
Concen: 72.649 ng
RT: 10.54 min Scan# 1183
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion: 93 Resp: 232006
Ion Ratio Lower Upper
93 100
95 32.7 25.7 38.5
123 9.9 8.3 12.5





#29

2,4-Dichlorophenol

Concen: 61.399 ng

RT: 10.80 min Scan# 1228

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:162 Resp: 168407

Ion Ratio Lower Upper

162 100

164 64.2 42.4 82.4




98 37.5 20.3 60.3



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): B



#30

1,2,4-Trichlorobenzene

Concen: 66.081 ng

RT: 11.01 min Scan# 1264

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:180 Resp: 203203

Ion Ratio Lower Upper

180 100

182 94.0 77.0 115.4



145 27.8 23.4 35.0




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 66.497 ng

RT: 11.21 min Scan# 1297

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:128 Resp: 530004

Ion Ratio Lower Upper

128 100

129 11.6 9.3 13.9




127 13.7 11.4 17.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 90.414 ng

RT: 10.48 min Scan# 1173

Delta R.T. 0.02 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:122 Resp: 111441

Ion Ratio Lower Upper

122 100

105 147.1 120.1 160.1



77 113.4 104.6 144.6




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG









#33
4-Chloroaniline
Concen: 70.207 ng
RT: 11.31 min Scan# 1315
Delta R.T. -0.01 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC060


Tot Ion:	127	Resp:	228689
Ion	Ratio	Lower	Upper
127	100		
129	34.0	23.6	35.4
65	40.8	37.7	56.5
92	23.8	17.6	26.4



#34
Hexachlorobutadiene
Concen: 69.047 ng
RT: 11.48 min Scan# 1343
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:	225	Resp:	148207
Ion	Ratio	Lower	Upper
225	100		
223	60.9	50.7	76.1
227	67.8	49.0	73.6





#35

Caprolactam

Concen: 73.479 ng

RT: 12.11 min Scan# 1450

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:113 Resp: 61658

Ion Ratio Lower Upper

113 100

55 217.0 198.6 238.6




56 172.3 140.4 180.4



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#36

4-Chloro-3-methylphenol

Concen: 71.784 ng

RT: 12.42 min Scan# 1503

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:107 Resp: 212415

Ion Ratio Lower Upper

107 100

144 22.7 17.4 26.0



142 69.3 54.1 81.1




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
 2-Methylnaphthalene
 Concen: 64.562 ng
 RT: 12.80 min Scan# 1567
 Delta R.T. 0.01 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Tot Ion	Ratio	Lower	Upper
142	100		
141	89.7	70.4	105.6
115	39.8	35.5	53.3



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.95 min Scan# 1934
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
164	100		
162	98.1	85.1	127.7
160	44.9	34.7	52.1





#39

1,2,4,5-Tetrachlorobenzene

Concen: 59.569 ng

RT: 13.15 min Scan# 1628

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060

Tot Ion:216 Resp: 255454


Ion Ratio Lower Upper

216 100

214 78.0 64.4 96.6

179 19.8 17.4 26.0

108 15.2 15.6 23.4#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 51.999 ng

RT: 13.12 min Scan# 1623

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:237 Resp: 104069

Ion Ratio Lower Upper

237 100

235 62.7 39.4 79.4



272 10.6 0.0 32.0




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 126.332 ng
 RT: 16.44 min Scan# 2188
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Tot Ion	Ratio	Lower	Upper
330	100		
332	97.4	77.3	115.9
141	33.0	32.1	48.1




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#42
 2,4,6-Trichlorophenol
 Concen: 64.718 ng
 RT: 13.39 min Scan# 1669
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
196	100		
198	90.0	81.4	122.2
200	31.6	27.0	40.6



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#43
2,4,5-Trichlorophenol
Concen: 61.271 ng
RT: 13.47 min Scan# 1682
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Tot Ion	Ratio	Lower	Upper
196	100		
198	92.3	79.9	119.9
97	45.3	45.4	68.0
132	23.6	20.7	31.1






Abundance

Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E


Ion 97.00 (96.70 to 97.70): BG

Ion 132.00 (131.70 to 132.70): E



#44
2-Fluorobiphenyl
Concen: 121.591 ng
RT: 13.57 min Scan# 1699
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.7	32.2	48.2
170	22.9	21.8	32.6






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 58.861 ng

RT: 13.78 min Scan# 1735

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:154 Resp: 548574

Ion Ratio Lower Upper




154 100

153 44.4 23.0 63.0

76 12.5 0.0 33.5



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): B



#46

2-Chloronaphthalene

Concen: 63.001 ng

RT: 13.84 min Scan# 1744

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:162 Resp: 427541

Ion Ratio Lower Upper



162 100


127 39.4 32.2 48.4

164 34.8 27.3 40.9



Abundance Ion 162.00 (161.70 to 162.70): E
Ion 127.00 (126.70 to 127.70): E
Ion 164.00 (163.70 to 164.70): E









#47
2-Nitroaniline
Concen: 84.275 ng
RT: 14.04 min Scan# 1779
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tot Ion: 65 Resp: 170877
Ion Ratio Lower Upper
65 100
92 61.7 49.0 73.4
138 81.8 58.6 87.8



#48
Acenaphthylene
Concen: 59.340 ng
RT: 14.68 min Scan# 1887
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion: 152 Resp: 640819
Ion Ratio Lower Upper
152 100
151 21.6 18.2 27.2
153 13.3 11.3 16.9









#49
Dimethylphthalate
Concen: 58.860 ng
RT: 14.39 min Scan# 1839
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC060


Tot Ion:163 Resp: 533076
Ion Ratio Lower Upper
163 100
194 4.2 3.8 5.6
164 10.7 9.3 13.9



#50
2,6-Dinitrotoluene
Concen: 62.899 ng
RT: 14.52 min Scan# 1861
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:165 Resp: 121697
Ion Ratio Lower Upper
165 100
63 65.5 63.9 95.9
89 59.5 58.6 88.0





#51

Acenaphthene

Concen: 58.030 ng

RT: 15.02 min Scan# 1945

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:154 Resp: 393173

Ion Ratio Lower Upper




154 100

153 120.0 87.8 131.6

152 56.2 42.2 63.4



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 66.199 ng

RT: 14.86 min Scan# 1919

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:138 Resp: 121390

Ion Ratio Lower Upper



138 100


108 9.8 10.9 16.3#

92 126.1 115.3 172.9



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BGC









#53
2,4-Dinitrophenol
Concen: 64.547 ng
RT: 15.07 min Scan# 1955
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC060


Tot Ion:184 Resp: 70857
Ion Ratio Lower Upper
184 100
63 70.7 52.7 79.1
154 67.7 57.3 85.9



#54
Dibenzofuran
Concen: 58.942 ng
RT: 15.35 min Scan# 2001
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:168 Resp: 610320
Ion Ratio Lower Upper
168 100
139 40.3 35.3 52.9
169 13.2 11.5 17.3









#55
4-Nitrophenol
Concen: 62.757 ng
RT: 15.17 min Scan# 1971
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC060


Tot Ion:139 Resp: 85980
Ion Ratio Lower Upper
139 100
109 109.6 101.2 141.2
65 144.4 135.3 175.3



#56
2,4-Dinitrotoluene
Concen: 64.435 ng
RT: 15.32 min Scan# 1996
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:165 Resp: 172360
Ion Ratio Lower Upper
165 100
63 49.9 44.0 66.0
89 79.6 67.3 100.9
182 3.2 3.8 5.6#





#57

Fluorene

Concen: 60.720 ng

RT: 16.00 min Scan# 2112

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:166 Resp: 561776

Ion Ratio Lower Upper




166 100

165 97.4 78.6 117.8

167 12.7 11.6 17.4



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 60.911 ng

RT: 15.57 min Scan# 2040

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:232 Resp: 148135


Ion Ratio Lower Upper

232 100



131 39.2 34.9 52.3


130 2.4 2.3 3.5

166 28.8 24.2 36.4



Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E









#59
Diethylphthalate
Concen: 62.866 ng
RT: 15.74 min Scan# 2069
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampled :
SSTDIC060


Tot Ion:149 Resp: 551962
Ion Ratio Lower Upper
149 100
177 24.0 20.1 30.1
150 13.2 10.2 15.2



#60
4-Chlorophenyl-phenylether
Concen: 61.529 ng
RT: 15.97 min Scan# 2108
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion:204 Resp: 313883
Ion Ratio Lower Upper
204 100
206 33.7 30.1 45.1
141 56.6 50.6 76.0









#61
4-Nitroaniline
Concen: 66.126 ng
RT: 16.03 min Scan# 2117
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tot Ion:138 Resp: 129630
Ion Ratio Lower Upper
138 100
92 60.8 44.2 84.2
108 102.1 104.8 144.8#



#62
Azobenzene
Concen: 71.771 ng
RT: 16.27 min Scan# 2159
Delta R.T. 0.01 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion: 77 Resp: 493994
Ion Ratio Lower Upper
77 100
182 27.6 4.7 44.7
105 18.0 0.0 36.3
51 30.3 16.4 56.4





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.70 min Scan# 2402

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060



Tot Ion:188 Resp: 270567

Ion Ratio Lower Upper

188 100

94 6.2 5.8 8.8



80 7.8 7.5 11.3



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#64

4,6-Dinitro-2-methylphenol

Concen: 64.471 ng

RT: 16.09 min Scan# 2127

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15



Tgt Ion:198 Resp: 116277

Ion Ratio Lower Upper

198 100

51 45.1 22.6 62.6


105 37.9 14.4 54.4




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BGC

Ion 105.00 (104.70 to 105.70): E









#65
n-Nitrosodiphenylamine
Concen: 60.401 ng
RT: 16.20 min Scan# 2146
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tot Ion	Ratio	Lower	Upper
169	100		
168	69.1	55.7	83.5
167	37.5	29.8	44.6



#66
4-Bromophenyl-phenylether
Concen: 62.707 ng
RT: 16.87 min Scan# 2261
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
248	100		
250	100.3	75.0	112.6
141	66.7	55.2	82.8








#67
Hexachlorobenzene
Concen: 62.901 ng
RT: 17.01 min Scan# 2284
Delta R.T. 0.00 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC060


Tot Ion	Ratio	Lower	Upper
284	100		
142	28.2	29.9	44.9#
249	34.0	29.2	43.8



#68
Atrazine
Concen: 53.451 ng
RT: 17.14 min Scan# 2306
Delta R.T. 0.01 min
Lab File: BG028597.D
Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
200	100		
173	24.1	3.0	43.0
215	49.2	28.4	68.4









#69
 Pentachlorophenol
 Concen: 54.325 ng
 RT: 17.35 min Scan# 2343
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Tot Ion:266 Resp: 111866
 Ion Ratio Lower Upper
 266 100
 268 65.0 48.6 72.8
 264 65.0 50.1 75.1




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 60.528 ng
 RT: 17.75 min Scan# 2410
 Delta R.T. 0.01 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion:178 Resp: 856642
 Ion Ratio Lower Upper
 178 100
 176 21.2 17.7 26.5
 179 15.4 15.0 22.6



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 61.622 ng

RT: 17.84 min Scan# 2425

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :


BNA_G

ClientSampleId :

SSTDIC060

Tot Ion:178 Resp: 868519


Ion	Ratio	Lower	Upper
178	100		
176	19.4	16.4	24.6
179	16.2	13.8	20.8





Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



Time-->



#72

Carbazole

Concen: 60.973 ng

RT: 18.11 min Scan# 2471


Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:167 Resp: 773554


Ion	Ratio	Lower	Upper
167	100		
166	23.7	20.2	30.2
139	12.6	12.8	19.2




Abundance Ion 167.00 (166.70 to 167.70): E


Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



Time-->





#73

Di-n-butylphthalate

Concen: 67.248 ng

RT: 18.63 min Scan# 2560

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:149 Resp: 937423

Ion Ratio Lower Upper

149 100

150 9.6 9.1 13.7




104 5.7 6.7 10.1#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 63.792 ng

RT: 19.75 min Scan# 2750

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:202 Resp: 1084117

Ion Ratio Lower Upper

202 100

101 7.2 0.0 30.1



203 18.2 1.3 41.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.000 ng

RT: 22.04 min Scan# 3140

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:240 Resp: 320559

Ion Ratio Lower Upper

240 100

120 6.4 5.7 8.5




236 25.0 22.2 33.4



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 47.153 ng

RT: 19.91 min Scan# 2778

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:184 Resp: 352819

Ion Ratio Lower Upper

184 100

185 15.8 13.0 19.6



183 12.1 11.0 16.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pvrene

Concen: 64.974 ng

RT: 20.10 min Scan# 2811

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :


BNA_G

ClientSampleId :

SSTDICC060

Tot Ion:202 Resp: 1113215

Ion	Ratio	Lower	Upper
202	100		
200	21.6	19.6	29.4
203	18.0	15.8	23.8






Abundance

Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 121.184 ng

RT: 20.29 min Scan# 2842


Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:244 Resp: 1718083

Ion	Ratio	Lower	Upper
244	100		
212	7.8	10.0	15.0#
122	8.1	8.6	12.8#






Abundance

Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
 Butylbenzylphthalate
 Concen: 70.181 ng
 RT: 20.97 min Scan# 2958
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Tot Ion	Ratio	Lower	Upper
149	100		
91	72.6	63.5	95.3
206	27.1	21.0	31.6



#80
 Benzo(a)anthracene
 Concen: 63.022 ng
 RT: 22.02 min Scan# 3137
 Delta R.T. 0.01 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.0	24.9	37.3
229	20.5	18.2	27.2









#81
 3,3'-Dichlorobenzidine
 Concen: 63.670 ng
 RT: 21.91 min Scan# 3119
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Tot Ion:252 Resp: 441167
 Ion Ratio Lower Upper
 252 100
 254 61.8 53.1 79.7
 126 7.6 7.0 10.4



#82
 Chrysene
 Concen: 62.834 ng
 RT: 22.09 min Scan# 3149
 Delta R.T. 0.00 min
 Lab File: BG028597.D
 Acq: 7 Sep 2017 17:15

Tgt Ion:228 Resp: 1059312
 Ion Ratio Lower Upper
 228 100
 226 30.3 27.0 40.4
 229 20.1 17.8 26.6





#83

Bis(2-ethylhexyl)phthalate

Concen: 68.340 ng

RT: 21.87 min Scan# 3111

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:149 Resp: 612331

Ion Ratio Lower Upper

149 100

167 28.9 23.7 35.5




279 5.9 4.6 6.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 69.377 ng

RT: 23.17 min Scan# 3332

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:149 Resp: 1055241

Ion Ratio Lower Upper

149 100

167 1.8 1.4 2.2



43 10.1 11.8 17.6#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC





#85

Indeno(1,2,3-cd)pyrene

Concen: 65.421 ng

RT: 29.59 min Scan# 4425

Delta R.T. -0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDICC060


Tot Ion:276 Resp: 1275301

Ion Ratio Lower Upper

276 100

138 5.4 4.7 7.1




277 26.1 20.6 31.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.54 min Scan# 3737

Delta R.T. 0.00 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15


Tgt Ion:264 Resp: 301783

Ion Ratio Lower Upper

264 100

260 26.1 21.8 32.8



265 21.8 18.2 27.4




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 64.298 ng

RT: 24.43 min Scan# 3547

Delta R.T. 0.00 min





Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Tot Ion:252 Resp: 1133864

Ion	Ratio	Lower	Upper
252	100		
253	22.9	18.7	28.1
125	6.0	5.3	7.9



#88

Benzo(k)fluoranthene

Concen: 63.001 ng

RT: 24.50 min Scan# 3560




Delta R.T. 0.01 min


Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:252 Resp: 1110580

Ion	Ratio	Lower	Upper
252	100		
253	22.5	20.2	30.2
125	5.7	5.1	7.7





#89

Benzo(a)pyrene

Concen: 64.597 ng

RT: 25.39 min Scan# 3710

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :





BNA_G

ClientSampleId :

SSTDICC060

Tot Ion:252 Resp: 1074427

Ion	Ratio	Lower	Upper
252	100		
253	23.4	19.2	28.8
125	6.8	5.4	8.0



#90

Dibenzo(a,h)anthracene

Concen: 67.336 ng

RT: 29.66 min Scan# 4437




Delta R.T. 0.01 min


Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Tgt Ion:278 Resp: 1130826

Ion	Ratio	Lower	Upper
278	100		
139	8.8	7.7	11.5
279	23.5	19.1	28.7





#91

Benzo(a,h,i)perylene

Concen: 67.306 ng

RT: 30.88 min Scan# 4645

Delta R.T. 0.01 min

Lab File: BG028597.D

Acq: 7 Sep 2017 17:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC060


Tot Ion:276 Resp: 1089376

Ion Ratio Lower Upper

276 100

277 24.6 18.6 27.8

138 12.3 8.1 12.1#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

