

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090721\
 Data File : BG050166.D
 Acq On : 7 Sep 2021 11:46
 Operator : CG/JU
 Sample : SST01030
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST010430

Quant Time: Sep 07 13:24:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 07 13:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.949	152	44233	20.000	ng/ul	0.00
20) Naphthalene-d8	10.769	136	206851	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.611	164	117411	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.366	188	249514	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.637	240	250728	20.000	ng/ul	0.00
88) Perylene-d12	24.856	264	250385	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.303	96	3918	4.568	ng/uL	0.00
4) Pyridine-d5	3.737	84	31463	12.177	ng/ul	0.00
7) Phenol-d5	7.121	99	44651	11.888	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.274	67	24044	11.522	ng/ul	0.00
11) 2-Chlorophenol-d4	7.485	132	35259	12.546	ng/ul	0.00
15) 4-Methylphenol-d8	8.672	113	31092	10.511	ng/ul	0.00
21) Nitrobenzene-d5	9.124	128	17797	11.098	ng/ul	0.00
24) 2-Nitrophenol-d4	9.847	143	17674	9.256	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.399	165	33522	9.842	ng/ul	0.00
31) 4-Chloroaniline-d4	10.910	131	43213	9.322	ng/ul	0.00
46) Dimethylphthalate-d6	14.012	166	91910	10.229	ng/ul	0.00
49) Acenaphthylene-d8	14.305	160	108529	10.301	ng/ul	0.00
54) 4-Nitrophenol-d4	14.846	143	8334	6.411	ng/ul	0.00
60) Fluorene-d10	15.604	176	75074	9.854	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.750	200	13036	8.121	ng/ul	0.00
73) Anthracene-d10	17.466	188	117110	10.452	ng/ul	0.00
81) Pyrene-d10	19.757	212	137268	10.131	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.632	264	137029	10.316	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.344	88	4237	4.052	ng/uL#	80
5) Pyridine	3.755	79	34274	12.274	ng/ul#	88
6) Benzaldehyde	7.086	77	22796	10.546	ng/ul	95
8) Phenol	7.144	94	41496	10.938	ng/ul	96
10) Bis(2-Chloroethyl)ether	7.362	93	33355	12.736	ng/ul	96
12) 2-Chlorophenol	7.514	128	31824	11.459	ng/ul#	79
13) 2-Methylphenol	8.407	108	31506	10.764	ng/ul	95
14) 2,2'-oxybis(1-Chloropr...	8.472	45	29279	10.661	ng/ul#	94
16) Acetophenone	8.777	105	51582	10.692	ng/ul	93
17) N-Nitroso-di-n-propyla...	8.754	70	26249	10.836	ng/ul	91
18) 4-Methylphenol	8.742	108	31192	9.906	ng/ul	89
19) Hexachloroethane	9.030	117	13818	11.369	ng/ul	90
22) Nitrobenzene	9.165	77	42418	9.762	ng/ul#	94
23) Isophorone	9.682	82	72365	9.347	ng/ul	96
25) 2-Nitrophenol	9.882	139	17370	9.356	ng/ul#	79
26) 2,4-Dimethylphenol	9.946	107	38067	9.226	ng/ul	94
27) Bis(2-Chloroethoxy)met...	10.170	93	42813	10.709	ng/ul	99
29) 2,4-Dichlorophenol	10.428	162	30890	9.881	ng/ul	96
30) Naphthalene	10.822	128	99331	9.607	ng/ul	95
32) 4-Chloroaniline	10.933	127	38504	8.624	ng/ul#	85
33) Hexachlorobutadiene	11.098	225	24203	9.571	ng/ul	92
34) Caprolactam	11.697	113	4863	4.441	ng/ul	92
35) 4-Chloro-3-methylphenol	12.073	107	34035	9.110	ng/ul#	87
36) 2-Methylnaphthalene	12.437	142	73271	9.534	ng/ul	95


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090721\
 Data File : BG050166.D
 Acq On : 7 Sep 2021 11:46
 Operator : CG/JU
 Sample : SST01030
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST010430

Quant Time: Sep 07 13:24:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 07 13:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.649	142	70613	9.102	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.807	216	39660	11.500	ng/ul#	95
40) Hexachlorocyclopentadiene	12.778	237	9520	6.168	ng/ul#	94
41) 2,4,6-Trichlorophenol	13.054	196	22989	10.461	ng/ul#	92
42) 2,4,5-Trichlorophenol	13.136	196	25239	10.958	ng/ul	94
43) 1,1'-Biphenyl	13.442	154	89484	10.909	ng/ul	98
44) 2-Chloronaphthalene	13.489	162	69871	10.605	ng/ul	98
45) 2-Nitroaniline	13.694	65	20629	9.222	ng/ul	94
47) Dimethylphthalate	14.059	163	88158	9.914	ng/ul#	98
48) 2,6-Dinitrotoluene	14.194	165	18117	9.868	ng/ul#	86
50) Acenaphthylene	14.335	152	104482	10.850	ng/ul	95
51) 3-Nitroaniline	14.528	138	16012	9.038	ng/ul#	87
52) Acenaphthene	14.675	153	72516	10.381	ng/ul	97
53) 2,4-Dinitrophenol	14.769	184	2862	3.052	ng/ul#	63
55) 4-Nitrophenol	14.857	109	9203	5.583	ng/ul	89
56) Dibenzofuran	15.010	168	102240	10.569	ng/ul	94
57) 2,4-Dinitrotoluene	14.993	165	26519	10.042	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.251	232	17542	9.191	ng/ul#	96
59) Diethylphthalate	15.421	149	89885	9.881	ng/ul	96
61) Fluorene	15.662	166	83728	10.259	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.651	204	44534	10.313	ng/ul	95
63) 4-Nitroaniline	15.686	138	13956	7.894	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.762	198	11265	7.742	ng/ul#	96
67) N-Nitrosodiphenylamine	15.868	169	71443	10.876	ng/ul	98
68) 4-Bromophenyl-phenylether	16.549	248	30532	11.062	ng/ul	97
69) Hexachlorobenzene	16.673	284	36402	11.421	ng/ul	91
70) Atrazine	16.814	200	31890	10.927	ng/ul	98
71) Pentachlorophenol	17.031	266	7420	5.498	ng/ul	89
72) Phenanthrene	17.407	178	135379	10.883	ng/ul	99
74) Anthracene	17.501	178	139942	11.102	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.412	216	42008	12.549	ng/ul	92
76) Pentachlorobenzene	14.940	250	42121	11.875	ng/ul	98
77) Carbazole	17.777	167	118385	10.554	ng/ul	97
78) Di-n-butylphthalate	18.317	149	145385	10.135	ng/ul	99
80) Fluoranthene	19.422	202	169791	10.155	ng/ul#	96
82) Pyrene	19.786	202	167605	10.116	ng/ul	99
83) Butylbenzylphthalate	20.661	149	60717	9.287	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.531	252	59057	9.936	ng/ul#	98
85) Benzo(a)anthracene	21.619	228	160475	10.272	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.507	149	88675	9.876	ng/ul	99
87) Chrysene	21.684	228	149395	10.116	ng/ul	96
89) Di-n-octyl phthalate	22.712	149	148911	10.025	ng/ul	100
90) Benzo(b)fluoranthene	23.834	252	164795	10.289	ng/ul#	98
91) Benzo(k)fluoranthene	23.898	252	159807	10.497	ng/ul	99
93) Benzo(a)pyrene	24.703	252	145391	10.453	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.521	276	64524	3.517	ng/ul#	95
95) Dibenzo(a,h)anthracene	28.580	278	155269	10.109	ng/ul	98
96) Benzo(g,h,i)perylene	29.696	276	152755	9.903	ng/ul	95





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 4.052 ng/uL
RT: 3.344 min Scan# 52
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion: 88 Resp: 4237
Ion Ratio Lower Upper
88 100
43 22.2 21.6 32.4
58 84.2 52.0 78.0#



#5
Pyridine
Concen: 12.274 ng/uL
RT: 3.755 min Scan# 122
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46


Tgt Ion: 79 Resp: 34274
Ion Ratio Lower Upper
79 100
52 67.9 61.1 91.7
51 33.4 34.8 52.2#








#6
Benzaldehyde
Concen: 10.546 ng/ul
RT: 7.086 min Scan# 689
Delta R.T. 0.007 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Instrument :
BNA_G
ClientSampleId :
SSTD010430







Tgt Ion: 77 Resp: 22796
Ion Ratio Lower Upper
77 100
105 98.6 72.3 108.5
106 92.6 72.3 108.5



#8
Phenol
Concen: 10.938 ng/ul
RT: 7.144 min Scan# 699
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion: 94 Resp: 41496
Ion Ratio Lower Upper
94 100
65 33.6 27.7 41.5
66 44.4 39.0 58.4









#10

Bis(2-Chloroethyl)ether
Concen: 12.736 ng/ul
RT: 7.362 min Scan# 736
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Instrument :
BNA_G
ClientSampleId :
SSTD010430




Tgt Ion: 93	Resp: 33355
Ion Ratio	Lower Upper
93 100	
63 79.1	59.9 89.9
95 35.7	27.6 41.4




#12

2-Chlorophenol
Concen: 11.459 ng/ul
RT: 7.514 min Scan# 762
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion: 128	Resp: 31824
Ion Ratio	Lower Upper
128 100	
64 49.7	28.0 42.0#
130 38.9	24.4 36.6#









#13
2-Methylphenol
Concen: 10.764 ng/ul
RT: 8.407 min Scan# 914
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:108 Resp: 31506
Ion Ratio Lower Upper
108 100
107 93.5 71.0 106.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 10.661 ng/ul
RT: 8.472 min Scan# 925
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion: 45 Resp: 29279
Ion Ratio Lower Upper
45 100
77 18.4 16.9 25.3
79 16.6 11.0 16.4#









#16
Acetophenone
Concen: 10.692 ng/ul
RT: 8.777 min Scan# 977
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:105 Resp: 51582
Ion Ratio Lower Upper
105 100
77 79.2 67.7 101.5
51 25.7 25.3 37.9



#17
N-Nitroso-di-n-propylamine
Concen: 10.836 ng/ul
RT: 8.754 min Scan# 973
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion: 70 Resp: 26249
Ion Ratio Lower Upper
70 100
42 47.6 44.2 66.4
101 8.8 7.8 11.8
130 21.9 15.3 22.9









#18
4-Methylphenol
Concen: 9.906 ng/ul
RT: 8.742 min Scan# 971
Delta R.T. 0.007 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:108 Resp: 31192
Ion Ratio Lower Upper
108 100
107 108.3 78.3 117.5



#19
Hexachloroethane
Concen: 11.369 ng/ul
RT: 9.030 min Scan# 1020
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:117 Resp: 13818
Ion Ratio Lower Upper
117 100
201 100.3 89.3 133.9
199 69.9 50.5 75.7









#22
Nitrobenzene
Concen: 9.762 ng/ul
RT: 9.165 min Scan# 1043
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion: 77 Resp: 42418
Ion Ratio Lower Upper
77 100
123 35.1 30.4 45.6
65 10.8 11.3 16.9#



#23
Isophorone
Concen: 9.347 ng/ul
RT: 9.682 min Scan# 1131
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion: 82 Resp: 72365
Ion Ratio Lower Upper
82 100
95 8.6 7.8 11.8
138 15.8 11.4 17.0









#25
2-Nitrophenol
Concen: 9.356 ng/ul
RT: 9.882 min Scan# 1165
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:139 Resp: 17370
Ion Ratio Lower Upper
139 100
65 51.7 53.1 79.7#
109 37.9 44.1 66.1#



#26
2,4-Dimethylphenol
Concen: 9.226 ng/ul
RT: 9.946 min Scan# 1176
Delta R.T. 0.007 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:107 Resp: 38067
Ion Ratio Lower Upper
107 100
121 44.6 32.7 49.1
122 70.4 52.6 79.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.709 ng/ul
 RT: 10.170 min Scan# 1214
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010430


Tgt Ion: 93 Resp: 42813
 Ion Ratio Lower Upper
 93 100
 95 31.6 24.8 37.2
 123 9.9 8.2 12.2



#29
 2,4-Dichlorophenol
 Concen: 9.881 ng/ul
 RT: 10.428 min Scan# 1258
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46

Tgt Ion:162 Resp: 30890
 Ion Ratio Lower Upper
 162 100
 164 65.3 51.0 76.6
 98 38.6 27.8 41.8









#30
Naphthalene
Concen: 9.607 ng/ul
RT: 10.822 min Scan# 1325
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:128 Resp: 99331
Ion Ratio Lower Upper
128 100
129 9.7 9.3 13.9
127 14.0 9.8 14.6



#32
4-Chloroaniline
Concen: 8.624 ng/ul
RT: 10.933 min Scan# 1344
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:127 Resp: 38504
Ion Ratio Lower Upper
127 100
129 28.6 30.3 45.5#









#33
Hexachlorobutadiene
Concen: 9.571 ng/ul
RT: 11.098 min Scan# 1372
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:225 Resp: 24203
Ion Ratio Lower Upper
225 100
223 57.9 51.3 76.9
227 62.1 55.1 82.7



#34
Caprolactam
Concen: 4.441 ng/ul
RT: 11.697 min Scan# 1474
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:113 Resp: 4863
Ion Ratio Lower Upper
113 100
55 145.7 114.0 171.0
56 118.2 107.4 161.0









#35
4-Chloro-3-methylphenol
Concen: 9.110 ng/ul
RT: 12.073 min Scan# 1538
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:107 Resp: 34035
Ion Ratio Lower Upper
107 100
144 29.5 18.9 28.3#
142 63.2 59.2 88.8



#36
2-Methylnaphthalene
Concen: 9.534 ng/ul
RT: 12.437 min Scan# 1600
Delta R.T. 0.007 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:142 Resp: 73271
Ion Ratio Lower Upper
142 100
141 83.7 70.5 105.7









#37
1-Methylnaphthalene
Concen: 9.102 ng/ul
RT: 12.649 min Scan# 1636
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:142 Resp: 70613
Ion Ratio Lower Upper
142 100
141 95.3 75.8 113.8
116 4.6 3.7 5.5



#39
1,2,4,5-Tetrachlorobenzene
Concen: 11.500 ng/ul
RT: 12.807 min Scan# 1663
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:216 Resp: 39660
Ion Ratio Lower Upper
216 100
214 78.2 64.2 96.2
179 22.0 12.8 19.2#
108 20.1 13.1 19.7#









#40
Hexachlorocyclopentadiene
Concen: 6.168 ng/ul
RT: 12.778 min Scan# 1658
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:237 Resp: 9520
Ion Ratio Lower Upper
237 100
235 62.4 53.5 80.3
272 11.1 11.5 17.3#



#41
2,4,6-Trichlorophenol
Concen: 10.461 ng/ul
RT: 13.054 min Scan# 1705
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:196 Resp: 22989
Ion Ratio Lower Upper
196 100
198 97.9 83.7 125.5
200 25.7 25.9 38.9#









#42
2,4,5-Trichlorophenol
Concen: 10.958 ng/ul
RT: 13.136 min Scan# 1719
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:196 Resp: 25239
Ion Ratio Lower Upper
196 100
198 93.7 70.3 105.5
200 37.3 26.6 40.0



#43
1,1'-Biphenyl
Concen: 10.909 ng/ul
RT: 13.442 min Scan# 1771
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:154 Resp: 89484
Ion Ratio Lower Upper
154 100
153 39.7 32.5 48.7
76 10.7 7.8 11.8









#44
2-Chloronaphthalene
Concen: 10.605 ng/ul
RT: 13.489 min Scan# 1779
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:	162	Resp:	69871
Ion Ratio	Lower	Upper	
162	100		
164	32.3	25.2	37.8
127	38.6	31.9	47.9



#45
2-Nitroaniline
Concen: 9.222 ng/ul
RT: 13.694 min Scan# 1814
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:	65	Resp:	20629
Ion Ratio	Lower	Upper	
65	100		
92	70.4	55.6	83.4
138	89.3	78.7	118.1









#47
Dimethylphthalate
Concen: 9.914 ng/ul
RT: 14.059 min Scan# 1876
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:163 Resp: 88158
Ion Ratio Lower Upper
163 100
194 3.2 3.4 5.0#
164 10.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 9.868 ng/ul
RT: 14.194 min Scan# 1899
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:165 Resp: 18117
Ion Ratio Lower Upper
165 100
89 66.4 62.5 93.7
121 19.4 22.2 33.2#









#50
Acenaphthylene
Concen: 10.850 ng/ul
RT: 14.335 min Scan# 1923
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:152 Resp: 104482
Ion Ratio Lower Upper
152 100
151 21.6 15.2 22.8
153 14.7 10.7 16.1



#51
3-Nitroaniline
Concen: 9.038 ng/ul
RT: 14.528 min Scan# 1956
Delta R.T. 0.007 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:138 Resp: 16012
Ion Ratio Lower Upper
138 100
108 8.7 9.0 13.4#
92 135.6 96.4 144.6









#52
Acenaphthene
Concen: 10.381 ng/ul
RT: 14.675 min Scan# 1981
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:153 Resp: 72516
Ion Ratio Lower Upper
153 100
152 50.7 37.4 56.2
154 88.2 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 3.052 ng/ul
RT: 14.769 min Scan# 1997
Delta R.T. 0.018 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:184 Resp: 2862
Ion Ratio Lower Upper
184 100
63 55.5 75.8 113.8#
154 52.1 64.6 96.8#









#55
4-Nitrophenol
Concen: 5.583 ng/ul
RT: 14.857 min Scan# 2012
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:109 Resp: 9203
Ion Ratio Lower Upper
109 100
139 83.4 61.4 92.0
65 109.5 76.5 114.7



#56
Dibenzofuran
Concen: 10.569 ng/ul
RT: 15.010 min Scan# 2038
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:168 Resp: 102240
Ion Ratio Lower Upper
168 100
139 38.7 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 10.042 ng/ul
RT: 14.993 min Scan# 2035
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:165 Resp: 26519
Ion Ratio Lower Upper
165 100
63 45.1 42.3 63.5
182 3.0 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.191 ng/ul
RT: 15.251 min Scan# 2079
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:232 Resp: 17542
Ion Ratio Lower Upper
232 100
131 48.6 40.0 60.0
130 3.4 1.8 2.8#
166 27.6 24.8 37.2









#59
Diethylphthalate
Concen: 9.881 ng/ul
RT: 15.421 min Scan# 2108
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:149 Resp: 89885
Ion Ratio Lower Upper
149 100
177 22.0 15.8 23.6
150 11.6 9.7 14.5



#61
Fluorene
Concen: 10.259 ng/ul
RT: 15.662 min Scan# 2149
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:166 Resp: 83728
Ion Ratio Lower Upper
166 100
165 103.4 83.3 124.9
167 12.7 11.8 17.6









#62
4-Chlorophenyl-phenylether
Concen: 10.313 ng/ul
RT: 15.651 min Scan# 2147
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:204 Resp: 44534
Ion Ratio Lower Upper
204 100
206 36.5 25.4 38.2
141 67.3 51.8 77.6



#63
4-Nitroaniline
Concen: 7.894 ng/ul
RT: 15.686 min Scan# 2153
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:138 Resp: 13956
Ion Ratio Lower Upper
138 100
92 54.3 50.0 75.0
108 117.7 101.8 152.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 7.742 ng/ul
 RT: 15.762 min Scan# 2166
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46




Instrument :
 BNA_G
 ClientSampleId :
 SSTD010430


Tgt Ion:198 Resp: 11265
 Ion Ratio Lower Upper
 198 100
 182 6.3 3.0 4.6#
 77 31.4 23.4 35.2



#67
 N-Nitrosodiphenylamine
 Concen: 10.876 ng/ul
 RT: 15.868 min Scan# 2184
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46

Tgt Ion:169 Resp: 71443
 Ion Ratio Lower Upper
 169 100
 168 69.2 55.0 82.6
 167 39.8 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 11.062 ng/ul
RT: 16.549 min Scan# 2300
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:248 Resp: 30532
Ion Ratio Lower Upper
248 100
250 103.7 80.2 120.2
141 76.8 60.6 90.8



#69
Hexachlorobenzene
Concen: 11.421 ng/ul
RT: 16.673 min Scan# 2321
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:284 Resp: 36402
Ion Ratio Lower Upper
284 100
142 31.2 28.7 43.1
249 26.5 25.5 38.3





#70

Atrazine

Concen: 10.927 ng/ul

RT: 16.814 min Scan# 2345

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46

Tgt Ion:200 Resp: 31890

Ion Ratio Lower Upper

200 100

173 25.3 19.3 28.9





215 45.1 34.9 52.3

Instrument :

BNA_G

ClientSampleId :

SSTD010430



#71

Pentachlorophenol

Concen: 5.498 ng/ul

RT: 17.031 min Scan# 2382

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46




Tgt Ion:266 Resp: 7420


Ion Ratio Lower Upper

266 100

264 62.5 54.1 81.1

268 52.7 52.6 78.8









#72
Phenanthrene
Concen: 10.883 ng/ul
RT: 17.407 min Scan# 2446
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:178 Resp: 135379
Ion Ratio Lower Upper
178 100
179 14.9 11.7 17.5
176 20.9 16.2 24.2



#74
Anthracene
Concen: 11.102 ng/ul
RT: 17.501 min Scan# 2462
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:178 Resp: 139942
Ion Ratio Lower Upper
178 100
179 15.7 11.4 17.0
176 18.7 12.8 19.2









#75

1,2,3,4-Tetrachlorobenzene
Concen: 12.549 ng/uL
RT: 13.412 min Scan# 1766
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Instrument :
BNA_G
ClientSampleId :
SSTD010430




Tgt Ion:216 Resp: 42008
Ion Ratio Lower Upper
216 100
214 79.5 70.9 106.3
218 47.1 40.9 61.3




#76

Pentachlorobenzene
Concen: 11.875 ng/uL
RT: 14.940 min Scan# 2026
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:250 Resp: 42121
Ion Ratio Lower Upper
250 100
248 62.6 51.5 77.3
252 60.4 47.9 71.9









#77
 Carbazole
 Concen: 10.554 ng/ul
 RT: 17.777 min Scan# 2509
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46




Instrument :
 BNA_G
 ClientSampleId :
 SST010430


Tgt Ion:167 Resp: 118385
 Ion Ratio Lower Upper
 167 100
 166 24.2 17.8 26.8
 139 14.2 11.1 16.7



#78
 Di-n-butylphthalate
 Concen: 10.135 ng/ul
 RT: 18.317 min Scan# 2601
 Delta R.T. 0.001 min
 Lab File: BG050166.D
 Acq: 7 Sep 2021 11:46

Tgt Ion:149 Resp: 145385
 Ion Ratio Lower Upper
 149 100
 150 9.6 8.1 12.1
 104 6.4 5.3 7.9









#80
Fluoranthene
Concen: 10.155 ng/ul
RT: 19.422 min Scan# 2789
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:202 Resp: 169791
Ion Ratio Lower Upper
202 100
101 7.0 4.3 6.5#
100 5.6 3.5 5.3#



#82
Pyrene
Concen: 10.116 ng/ul
RT: 19.786 min Scan# 2851
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:202 Resp: 167605
Ion Ratio Lower Upper
202 100
101 6.7 5.4 8.0
100 6.9 5.0 7.6





#83

Butylbenzylphthalate

Concen: 9.287 ng/ul

RT: 20.661 min Scan# 3000

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46





Tgt Ion:149 Resp: 60717

Ion Ratio Lower Upper

149 100

91 73.9 60.6 90.8

206 21.5 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 9.936 ng/ul

RT: 21.531 min Scan# 3148

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46




Tgt Ion:252 Resp: 59057


Ion Ratio Lower Upper

252 100

254 66.1 54.2 81.2

126 8.8 5.4 8.0#









#85
Benzo(a)anthracene
Concen: 10.272 ng/ul
RT: 21.619 min Scan# 3163
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:228 Resp: 160475
Ion Ratio Lower Upper
228 100
229 20.1 16.2 24.4
226 25.3 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 9.876 ng/ul
RT: 21.507 min Scan# 3144
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:149 Resp: 88675
Ion Ratio Lower Upper
149 100
167 25.8 20.4 30.6
279 4.0 3.7 5.5





#87

Chrysene

Concen: 10.116 ng/ul

RT: 21.684 min Scan# 3174

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46

Instrument :

BNA_G

ClientSampleId :

SSTD010430





Tgt Ion:228 Resp: 149395

Ion Ratio Lower Upper

228 100

226 29.7 23.0 34.4

229 18.8 17.8 26.6



#89

Di-n-octyl phthalate

Concen: 10.025 ng/ul




RT: 22.712 min Scan# 3349


Delta R.T. 0.007 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46

Tgt Ion:149 Resp: 148911





#90

Benzo(b)fluoranthene

Concen: 10.289 ng/ul

RT: 23.834 min Scan# 3540

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46

Tgt Ion:252 Resp: 164795

Ion Ratio Lower Upper

252 100

253 21.6 18.1 27.1





125 5.7 3.4 5.0#

Instrument :

BNA_G

ClientSampleId :

SSTD010430



#91

Benzo(k)fluoranthene

Concen: 10.497 ng/ul

RT: 23.898 min Scan# 3551

Delta R.T. 0.001 min

Lab File: BG050166.D

Acq: 7 Sep 2021 11:46




Tgt Ion:252 Resp: 159807


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 5.9 4.1 6.1









#93
Benzo(a)pyrene
Concen: 10.453 ng/ul
RT: 24.703 min Scan# 3688
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46




Instrument :
BNA_G
ClientSampleId :
SSTD010430


Tgt Ion:252 Resp: 145391
Ion Ratio Lower Upper
252 100
253 21.9 17.0 25.4
125 4.6 4.4 6.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 3.517 ng/ul
RT: 28.521 min Scan# 4338
Delta R.T. -0.023 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:276 Resp: 64524
Ion Ratio Lower Upper
276 100
138 12.0 5.5 8.3#
277 22.2 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 10.109 ng/ul
RT: 28.580 min Scan# 4348
Delta R.T. -0.005 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Instrument :
BNA_G
ClientSampleId :
SSTD010430

Tgt Ion:278 Resp: 155269
Ion Ratio Lower Upper
278 100
139 7.0 5.2 7.8
279 22.8 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 9.903 ng/ul
RT: 29.696 min Scan# 4538
Delta R.T. 0.001 min
Lab File: BG050166.D
Acq: 7 Sep 2021 11:46

Tgt Ion:276 Resp: 152755
Ion Ratio Lower Upper
276 100
138 10.2 8.6 13.0
277 25.0 17.4 26.0

