

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090721\
 Data File : BG050167.D
 Acq On : 7 Sep 2021 12:27
 Operator : CG/JU
 Sample : SST02031
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020

Quant Time: Sep 07 13:24:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 07 13:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.948	152	31371	20.000	ng/u1	0.00
20) Naphthalene-d8	10.768	136	168245	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.610	164	81865	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.365	188	176353	20.000	ng/u1	0.00
79) Chrysene-d12	21.642	240	173878	20.000	ng/u1	0.00
88) Perylene-d12	24.855	264	175935	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.308	96	6885	11.317	ng/uL	0.00
4) Pyridine-d5	3.736	84	68248	37.244	ng/u1	0.00
7) Phenol-d5	7.120	99	79801	29.956	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.267	67	40321	27.244	ng/u1	0.00
11) 2-Chlorophenol-d4	7.484	132	51161	25.667	ng/u1	0.00
15) 4-Methylphenol-d8	8.677	113	62906	29.984	ng/u1	0.00
21) Nitrobenzene-d5	9.123	128	27434	21.032	ng/u1	0.00
24) 2-Nitrophenol-d4	9.852	143	38528	24.806	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	65523	23.653	ng/u1	0.00
31) 4-Chloroaniline-d4	10.909	131	88598	23.497	ng/u1	0.00
46) Dimethylphthalate-d6	14.011	166	167658	26.760	ng/u1	0.00
49) Acenaphthylene-d8	14.305	160	188743	25.693	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	17313	19.100	ng/u1	0.00
60) Fluorene-d10	15.609	176	134474	25.315	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.750	200	24465	21.563	ng/u1	0.00
73) Anthracene-d10	17.465	188	208034	26.270	ng/u1	0.00
81) Pyrene-d10	19.756	212	233148	24.813	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.632	264	238852	25.590	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	8460	11.407	ng/uL	88
5) Pyridine	3.754	79	69714	35.200	ng/u1#	83
6) Benzaldehyde	7.079	77	39065	25.483	ng/u1	94
8) Phenol	7.149	94	79411	29.513	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.367	93	52300	28.157	ng/u1	98
12) 2-Chlorophenol	7.514	128	48828	24.790	ng/u1#	90
13) 2-Methylphenol	8.407	108	47208	22.741	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.477	45	45789	23.508	ng/u1	96
16) Acetophenone	8.783	105	91956	26.876	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.759	70	50710	29.517	ng/u1	92
18) 4-Methylphenol	8.736	108	65244	29.215	ng/u1	90
19) Hexachloroethane	9.035	117	24858	28.838	ng/u1	92
22) Nitrobenzene	9.164	77	79330	22.445	ng/u1	93
23) Isophorone	9.687	82	160648	25.511	ng/u1#	95
25) 2-Nitrophenol	9.881	139	34719	22.992	ng/u1	94
26) 2,4-Dimethylphenol	9.940	107	72170	21.505	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.169	93	86540	26.613	ng/u1	96
29) 2,4-Dichlorophenol	10.427	162	59832	23.532	ng/u1	95
30) Naphthalene	10.821	128	196059	23.314	ng/u1	97
32) 4-Chloroaniline	10.938	127	84276	23.208	ng/u1	93
33) Hexachlorobutadiene	11.097	225	54092	26.298	ng/u1	94
34) Caprolactam	11.702	113	9950	11.172	ng/u1	90
35) 4-Chloro-3-methylphenol	12.078	107	76452	25.159	ng/u1	98
36) 2-Methylnaphthalene	12.431	142	141042	22.563	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090721\
 Data File : BG050167.D
 Acq On : 7 Sep 2021 12:27
 Operator : CG/JU
 Sample : SST02031
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST020


Quant Time: Sep 07 13:24:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 07 13:21:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.648	142	153220	24.283	ng/ul	92
39) 1,2,4,5-Tetrachloroben...	12.807	216	79763	33.172	ng/ul#	95
40) Hexachlorocyclopentadiene	12.777	237	21642	20.111	ng/ul#	90
41) 2,4,6-Trichlorophenol	13.053	196	45871	29.936	ng/ul	88
42) 2,4,5-Trichlorophenol	13.136	196	48744	30.352	ng/ul	92
43) 1,1'-Biphenyl	13.441	154	171189	29.931	ng/ul#	94
44) 2-Chloronaphthalene	13.488	162	131304	28.582	ng/ul	99
45) 2-Nitroaniline	13.699	65	52453	33.629	ng/ul	96
47) Dimethylphthalate	14.058	163	157252	25.361	ng/ul	99
48) 2,6-Dinitrotoluene	14.193	165	33259	25.981	ng/ul	86
50) Acenaphthylene	14.334	152	174336	25.966	ng/ul	98
51) 3-Nitroaniline	14.522	138	24151	19.551	ng/ul	94
52) Acenaphthene	14.675	153	126168	25.905	ng/ul	98
53) 2,4-Dinitrophenol	14.751	184	13004	19.889	ng/ul#	80
55) 4-Nitrophenol	14.857	109	18822	16.377	ng/ul	95
56) Dibenzofuran	15.009	168	173115	25.667	ng/ul	100
57) 2,4-Dinitrotoluene	14.992	165	45910	24.932	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.250	232	30694	23.064	ng/ul#	89
59) Diethylphthalate	15.421	149	152178	23.991	ng/ul	98
61) Fluorene	15.662	166	146986	25.831	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.650	204	75277	25.003	ng/ul	99
63) 4-Nitroaniline	15.691	138	19332	15.683	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.761	198	21017	20.437	ng/ul#	93
67) N-Nitrosodiphenylamine	15.867	169	120760	26.009	ng/ul	95
68) 4-Bromophenyl-phenylether	16.549	248	54903	28.143	ng/ul	92
69) Hexachlorobenzene	16.678	284	58649	26.035	ng/ul#	93
70) Atrazine	16.813	200	50916	24.685	ng/ul	96
71) Pentachlorophenol	17.030	266	14325	15.017	ng/ul	88
72) Phenanthrene	17.412	178	230907	26.264	ng/ul	98
74) Anthracene	17.500	178	239724	26.907	ng/ul	95
75) 1,2,3,4-Tetrachloroben...	13.412	216	76856	32.483	ng/ul	91
76) Pentachlorobenzene	14.939	250	72743	29.015	ng/ul	97
77) Carbazole	17.776	167	205063	25.866	ng/ul	97
78) Di-n-butylphthalate	18.317	149	252478	24.901	ng/ul	99
80) Fluoranthene	19.421	202	292392	25.217	ng/ul#	96
82) Pyrene	19.785	202	283805	24.701	ng/ul	99
83) Butylbenzylphthalate	20.661	149	110462	24.362	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.530	252	97086	23.555	ng/ul	97
85) Benzo(a)anthracene	21.618	228	273547	25.248	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.507	149	152846	24.547	ng/ul	100
87) Chrysene	21.683	228	261487	25.532	ng/ul	98
89) Di-n-octyl phthalate	22.705	149	263216	25.220	ng/ul	100
90) Benzo(b)fluoranthene	23.833	252	288140	25.604	ng/ul	95
91) Benzo(k)fluoranthene	23.897	252	273524	25.569	ng/ul	99
93) Benzo(a)pyrene	24.702	252	251875	25.771	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.544	276	322160	24.988	ng/ul#	96
95) Dibenzo(a,h)anthracene	28.585	278	264876	24.542	ng/ul#	92
96) Benzo(g,h,i)perylene	29.678	276	118796	10.960	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 07 13:24:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 07 13:21:30 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.407 ng/uL
RT: 3.343 min Scan# 52
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:	88	Resp:	8460
Ion Ratio	Lower	Upper	
88	100		
43	26.1	21.6	32.4
58	52.1	52.0	78.0



#5
Pyridine
Concen: 35.200 ng/uL
RT: 3.754 min Scan# 122
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:	79	Resp:	69714
Ion Ratio	Lower	Upper	
79	100		
52	61.7	61.1	91.7
51	32.2	34.8	52.2#









#6
Benzaldehyde
Concen: 25.483 ng/ul
RT: 7.079 min Scan# 688
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion: 77	Resp: 39065
Ion Ratio	Lower Upper
77 100	
105 100.5	72.3 108.5
106 89.1	72.3 108.5



#8
Phenol
Concen: 29.513 ng/ul
RT: 7.149 min Scan# 700
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion: 94	Resp: 79411
Ion Ratio	Lower Upper
94 100	
65 34.1	27.7 41.5
66 44.7	39.0 58.4









#10
Bis(2-Chloroethyl)ether
Concen: 28.157 ng/ul
RT: 7.367 min Scan# 737
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:	93	Resp:	52300
Ion	Ratio	Lower	Upper
93	100		
63	75.5	59.9	89.9
95	31.2	27.6	41.4



#12
2-Chlorophenol
Concen: 24.790 ng/ul
RT: 7.514 min Scan# 762
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:	128	Resp:	48828
Ion	Ratio	Lower	Upper
128	100		
64	45.4	28.0	42.0#
130	30.6	24.4	36.6









#13
2-Methylphenol
Concen: 22.741 ng/ul
RT: 8.407 min Scan# 914
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:108 Resp: 47208
Ion Ratio Lower Upper
108 100
107 93.1 71.0 106.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.508 ng/ul
RT: 8.477 min Scan# 926
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion: 45 Resp: 45789
Ion Ratio Lower Upper
45 100
77 23.5 16.9 25.3
79 15.1 11.0 16.4









#16
Acetophenone
Concen: 26.876 ng/ul
RT: 8.783 min Scan# 978
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:105 Resp: 91956
Ion Ratio Lower Upper
105 100
77 85.9 67.7 101.5
51 25.9 25.3 37.9



#17
N-Nitroso-di-n-propylamine
Concen: 29.517 ng/ul
RT: 8.759 min Scan# 974
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion: 70 Resp: 50710
Ion Ratio Lower Upper
70 100
42 47.3 44.2 66.4
101 11.0 7.8 11.8
130 17.9 15.3 22.9









#18
4-Methylphenol
Concen: 29.215 ng/ul
RT: 8.736 min Scan# 970
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:108 Resp: 65244
Ion Ratio Lower Upper
108 100
107 108.1 78.3 117.5



#19
Hexachloroethane
Concen: 28.838 ng/ul
RT: 9.035 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:117 Resp: 24858
Ion Ratio Lower Upper
117 100
201 119.0 89.3 133.9
199 70.6 50.5 75.7









#22
Nitrobenzene
Concen: 22.445 ng/ul
RT: 9.164 min Scan# 1043
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion: 77 Resp: 79330
Ion Ratio Lower Upper
77 100
123 42.6 30.4 45.6
65 12.5 11.3 16.9



#23
Isophorone
Concen: 25.511 ng/ul
RT: 9.687 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion: 82 Resp: 160648
Ion Ratio Lower Upper
82 100
95 7.7 7.8 11.8#
138 15.8 11.4 17.0









#25
2-Nitrophenol
Concen: 22.992 ng/ul
RT: 9.881 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:139 Resp: 34719
Ion Ratio Lower Upper
139 100
65 65.3 53.1 79.7
109 47.4 44.1 66.1



#26
2,4-Dimethylphenol
Concen: 21.505 ng/ul
RT: 9.940 min Scan# 1175
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:107 Resp: 72170
Ion Ratio Lower Upper
107 100
121 42.9 32.7 49.1
122 70.4 52.6 79.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 26.613 ng/ul
 RT: 10.169 min Scan# 1214
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020


Tgt Ion: 93 Resp: 86540
 Ion Ratio Lower Upper
 93 100
 95 33.9 24.8 37.2
 123 9.6 8.2 12.2



#29
 2,4-Dichlorophenol
 Concen: 23.532 ng/ul
 RT: 10.427 min Scan# 1258
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27

Tgt Ion:162 Resp: 59832
 Ion Ratio Lower Upper
 162 100
 164 67.7 51.0 76.6
 98 38.1 27.8 41.8









#30
Naphthalene
Concen: 23.314 ng/ul
RT: 10.821 min Scan# 1325
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:128 Resp: 196059
Ion Ratio Lower Upper
128 100
129 11.3 9.3 13.9
127 13.8 9.8 14.6



#32
4-Chloroaniline
Concen: 23.208 ng/ul
RT: 10.938 min Scan# 1345
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:127 Resp: 84276
Ion Ratio Lower Upper
127 100
129 33.5 30.3 45.5









#33
Hexachlorobutadiene
Concen: 26.298 ng/ul
RT: 11.097 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampled :
SSTD020


Tgt Ion:225 Resp: 54092
Ion Ratio Lower Upper
225 100
223 70.2 51.3 76.9
227 71.9 55.1 82.7



#34
Caprolactam
Concen: 11.172 ng/ul
RT: 11.702 min Scan# 1475
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:113 Resp: 9950
Ion Ratio Lower Upper
113 100
55 156.7 114.0 171.0
56 143.5 107.4 161.0









#35
4-Chloro-3-methylphenol
Concen: 25.159 ng/ul
RT: 12.078 min Scan# 1539
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:107 Resp: 76452
Ion Ratio Lower Upper
107 100
144 22.1 18.9 28.3
142 75.4 59.2 88.8



#36
2-Methylnaphthalene
Concen: 22.563 ng/ul
RT: 12.431 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:142 Resp: 141042
Ion Ratio Lower Upper
142 100
141 88.6 70.5 105.7









#37
1-Methylnaphthalene
Concen: 24.283 ng/ul
RT: 12.648 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:142 Resp: 153220
Ion Ratio Lower Upper
142 100
141 86.9 75.8 113.8
116 4.2 3.7 5.5



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.172 ng/ul
RT: 12.807 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:216 Resp: 79763
Ion Ratio Lower Upper
216 100
214 76.9 64.2 96.2
179 20.8 12.8 19.2#
108 17.7 13.1 19.7









#40

Hexachlorocyclopentadiene
Concen: 20.111 ng/ul
RT: 12.777 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Instrument :
BNA_G
ClientSampleId :
SSTD020




Tgt Ion:237 Resp: 21642
Ion Ratio Lower Upper
237 100
235 59.0 53.5 80.3
272 10.1 11.5 17.3#




#41

2,4,6-Trichlorophenol
Concen: 29.936 ng/ul
RT: 13.053 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:196 Resp: 45871
Ion Ratio Lower Upper
196 100
198 90.0 83.7 125.5
200 34.2 25.9 38.9









#42
2,4,5-Trichlorophenol
Concen: 30.352 ng/ul
RT: 13.136 min Scan# 1719
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:196 Resp: 48744
Ion Ratio Lower Upper
196 100
198 97.3 70.3 105.5
200 31.5 26.6 40.0



#43
1,1'-Biphenyl
Concen: 29.931 ng/ul
RT: 13.441 min Scan# 1771
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:154 Resp: 171189
Ion Ratio Lower Upper
154 100
153 44.4 32.5 48.7
76 12.8 7.8 11.8#









#44
2-Chloronaphthalene
Concen: 28.582 ng/ul
RT: 13.488 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:162 Resp: 131304
Ion Ratio Lower Upper
162 100
164 31.1 25.2 37.8
127 40.8 31.9 47.9



#45
2-Nitroaniline
Concen: 33.629 ng/ul
RT: 13.699 min Scan# 1815
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion: 65 Resp: 52453
Ion Ratio Lower Upper
65 100
92 71.5 55.6 83.4
138 93.6 78.7 118.1









#47
Dimethylphthalate
Concen: 25.361 ng/ul
RT: 14.058 min Scan# 1876
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:163 Resp: 157252
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 9.6 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 25.981 ng/ul
RT: 14.193 min Scan# 1899
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:165 Resp: 33259
Ion Ratio Lower Upper
165 100
89 63.8 62.5 93.7
121 24.0 22.2 33.2









#50
Acenaphthylene
Concen: 25.966 ng/ul
RT: 14.334 min Scan# 1923
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:152 Resp: 174336
Ion Ratio Lower Upper
152 100
151 19.9 15.2 22.8
153 14.4 10.7 16.1



#51
3-Nitroaniline
Concen: 19.551 ng/ul
RT: 14.522 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:138 Resp: 24151
Ion Ratio Lower Upper
138 100
108 13.2 9.0 13.4
92 127.0 96.4 144.6









#52
Acenaphthene
Concen: 25.905 ng/ul
RT: 14.675 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:153 Resp: 126168
Ion Ratio Lower Upper
153 100
152 48.1 37.4 56.2
154 87.8 71.5 107.3



#53
2,4-Dinitrophenol
Concen: 19.889 ng/ul
RT: 14.751 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:184 Resp: 13004
Ion Ratio Lower Upper
184 100
63 68.4 75.8 113.8#
154 70.2 64.6 96.8









#55
4-Nitrophenol
Concen: 16.377 ng/ul
RT: 14.857 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:109 Resp: 18822
Ion Ratio Lower Upper
109 100
139 81.5 61.4 92.0
65 100.6 76.5 114.7



#56
Dibenzofuran
Concen: 25.667 ng/ul
RT: 15.009 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:168 Resp: 173115
Ion Ratio Lower Upper
168 100
139 42.3 33.8 50.8









#57
2,4-Dinitrotoluene
Concen: 24.932 ng/ul
RT: 14.992 min Scan# 2035
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:165 Resp: 45910
Ion Ratio Lower Upper
165 100
63 48.5 42.3 63.5
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.064 ng/ul
RT: 15.250 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:232 Resp: 30694
Ion Ratio Lower Upper
232 100
131 39.5 40.0 60.0#
130 2.5 1.8 2.8
166 28.6 24.8 37.2









#59
Diethylphthalate
Concen: 23.991 ng/ul
RT: 15.421 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:149 Resp: 152178
Ion Ratio Lower Upper
149 100
177 21.5 15.8 23.6
150 12.1 9.7 14.5



#61
Fluorene
Concen: 25.831 ng/ul
RT: 15.662 min Scan# 2149
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:166 Resp: 146986
Ion Ratio Lower Upper
166 100
165 96.9 83.3 124.9
167 12.6 11.8 17.6









#62
4-Chlorophenyl-phenylether
Concen: 25.003 ng/ul
RT: 15.650 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:204 Resp: 75277
Ion Ratio Lower Upper
204 100
206 30.8 25.4 38.2
141 65.0 51.8 77.6



#63
4-Nitroaniline
Concen: 15.683 ng/ul
RT: 15.691 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:138 Resp: 19332
Ion Ratio Lower Upper
138 100
92 61.5 50.0 75.0
108 122.1 101.8 152.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.437 ng/ul
 RT: 15.761 min Scan# 2166
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020


Tgt Ion:198 Resp: 21017
 Ion Ratio Lower Upper
 198 100
 182 4.6 3.0 4.6#
 77 33.3 23.4 35.2



#67
 N-Nitrosodiphenylamine
 Concen: 26.009 ng/ul
 RT: 15.867 min Scan# 2184
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27

Tgt Ion:169 Resp: 120760
 Ion Ratio Lower Upper
 169 100
 168 71.8 55.0 82.6
 167 41.4 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 28.143 ng/ul
RT: 16.549 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:248 Resp: 54903
Ion Ratio Lower Upper
248 100
250 92.6 80.2 120.2
141 68.7 60.6 90.8



#69
Hexachlorobenzene
Concen: 26.035 ng/ul
RT: 16.678 min Scan# 2322
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:284 Resp: 58649
Ion Ratio Lower Upper
284 100
142 28.5 28.7 43.1#
249 32.1 25.5 38.3





#70

Atrazine

Concen: 24.685 ng/ul

RT: 16.813 min Scan# 2345

Delta R.T. 0.000 min

Lab File: BG050167.D

Acq: 7 Sep 2021 12:27

Tgt Ion:200 Resp: 50916

Ion Ratio Lower Upper

200 100

173 21.7 19.3 28.9





215 40.9 34.9 52.3

Instrument :

BNA_G

ClientSampleId :

SSTD020



#71

Pentachlorophenol

Concen: 15.017 ng/ul

RT: 17.030 min Scan# 2382

Delta R.T. 0.000 min

Lab File: BG050167.D

Acq: 7 Sep 2021 12:27




Tgt Ion:266 Resp: 14325


Ion Ratio Lower Upper

266 100

264 55.1 54.1 81.1

268 58.8 52.6 78.8









#72
Phenanthrene
Concen: 26.264 ng/u1
RT: 17.412 min Scan# 2447
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:178 Resp: 230907
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 19.5 16.2 24.2



#74
Anthracene
Concen: 26.907 ng/u1
RT: 17.500 min Scan# 2462
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:178 Resp: 239724
Ion Ratio Lower Upper
178 100
179 14.8 11.4 17.0
176 19.1 12.8 19.2









#75

1,2,3,4-Tetrachlorobenzene
Concen: 32.483 ng/uL
RT: 13.412 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Instrument :
BNA_G
ClientSampleId :
SSTD020




Tgt Ion:216 Resp: 76856
Ion Ratio Lower Upper
216 100
214 78.8 70.9 106.3
218 46.2 40.9 61.3




#76

Pentachlorobenzene
Concen: 29.015 ng/uL
RT: 14.939 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:250 Resp: 72743
Ion Ratio Lower Upper
250 100
248 65.6 51.5 77.3
252 62.7 47.9 71.9









#77
 Carbazole
 Concen: 25.866 ng/ul
 RT: 17.776 min Scan# 2509
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27




Instrument :
 BNA_G
 ClientSampleId :
 SST020


Tgt Ion:167 Resp: 205063
 Ion Ratio Lower Upper
 167 100
 166 23.9 17.8 26.8
 139 12.6 11.1 16.7



#78
 Di-n-butylphthalate
 Concen: 24.901 ng/ul
 RT: 18.317 min Scan# 2601
 Delta R.T. 0.000 min
 Lab File: BG050167.D
 Acq: 7 Sep 2021 12:27

Tgt Ion:149 Resp: 252478
 Ion Ratio Lower Upper
 149 100
 150 9.5 8.1 12.1
 104 6.6 5.3 7.9









#80
Fluoranthene
Concen: 25.217 ng/ul
RT: 19.421 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:202 Resp: 292392
Ion Ratio Lower Upper
202 100
101 6.6 4.3 6.5#
100 5.6 3.5 5.3#



#82
Pyrene
Concen: 24.701 ng/ul
RT: 19.785 min Scan# 2851
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:202 Resp: 283805
Ion Ratio Lower Upper
202 100
101 6.8 5.4 8.0
100 6.9 5.0 7.6





#83

Butylbenzylphthalate

Concen: 24.362 ng/ul

RT: 20.661 min Scan# 3000

Delta R.T. 0.000 min

Lab File: BG050167.D

Acq: 7 Sep 2021 12:27





Tgt Ion:149 Resp: 110462

Ion Ratio Lower Upper

149 100

91 74.5 60.6 90.8

206 20.5 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 23.555 ng/ul

RT: 21.530 min Scan# 3148

Delta R.T. 0.000 min

Lab File: BG050167.D

Acq: 7 Sep 2021 12:27




Tgt Ion:252 Resp: 97086


Ion Ratio Lower Upper

252 100

254 65.5 54.2 81.2

126 7.8 5.4 8.0









#85
Benzo(a)anthracene
Concen: 25.248 ng/ul
RT: 21.618 min Scan# 3163
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:228 Resp: 273547
Ion Ratio Lower Upper
228 100
229 20.9 16.2 24.4
226 26.9 21.0 31.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 24.547 ng/ul
RT: 21.507 min Scan# 3144
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:149 Resp: 152846
Ion Ratio Lower Upper
149 100
167 25.6 20.4 30.6
279 4.1 3.7 5.5









#87
Chrysene
Concen: 25.532 ng/ul
RT: 21.683 min Scan# 3174
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:228 Resp: 261487
Ion Ratio Lower Upper
228 100
226 28.7 23.0 34.4
229 20.1 17.8 26.6



#89
Di-n-octyl phthalate
Concen: 25.220 ng/ul
RT: 22.705 min Scan# 3348
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:149 Resp: 263216









#90
Benzo(b)fluoranthene
Concen: 25.604 ng/ul
RT: 23.833 min Scan# 3540
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:252 Resp: 288140
Ion Ratio Lower Upper
252 100
253 19.7 18.1 27.1
125 4.7 3.4 5.0



#91
Benzo(k)fluoranthene
Concen: 25.569 ng/ul
RT: 23.897 min Scan# 3551
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:252 Resp: 273524
Ion Ratio Lower Upper
252 100
253 21.3 17.7 26.5
125 5.4 4.1 6.1









#93
Benzo(a)pyrene
Concen: 25.771 ng/ul
RT: 24.702 min Scan# 3688
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27




Instrument :
BNA_G
ClientSampleId :
SSTD020


Tgt Ion:252 Resp: 251875
Ion Ratio Lower Upper
252 100
253 20.7 17.0 25.4
125 5.6 4.4 6.6



#94
Indeno(1,2,3-cd)pyrene
Concen: 24.988 ng/ul
RT: 28.544 min Scan# 4342
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:276 Resp: 322160
Ion Ratio Lower Upper
276 100
138 12.0 5.5 8.3#
277 22.6 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 24.542 ng/ul
RT: 28.585 min Scan# 4349
Delta R.T. 0.000 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Instrument :
BNA_G
ClientSampleId :
SSTD020

Tgt Ion:278 Resp: 264876
Ion Ratio Lower Upper
278 100
139 8.1 5.2 7.8#
279 27.8 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 10.960 ng/ul
RT: 29.678 min Scan# 4535
Delta R.T. -0.018 min
Lab File: BG050167.D
Acq: 7 Sep 2021 12:27

Tgt Ion:276 Resp: 118796
Ion Ratio Lower Upper
276 100
138 10.1 8.6 13.0
277 23.7 17.4 26.0

