


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090721\
 Data File : BG050163.D
 Acq On : 7 Sep 2021 9:42
 Operator : CG/JU
 Sample : DFTPP419
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP419

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG090721.M
 Title : SVOA CALIBRATION
 Last Update : Tue Sep 07 18:32:38 2021



AutoFind: Scans 2510, 2511, 2512; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.5	19469	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.4	22213	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.1	27531	PASS
197	198	0.00	2	0.4	226	PASS
198	198	100	100	100.0	54915	PASS
199	198	5	9	6.8	3732	PASS
275	198	10	60	31.4	17223	PASS
365	198	1	100	5.2	2837	PASS
441	198	0.01	100	14.8	8128	PASS
442	198	50	100	97.6	53576	PASS
443	442	15	24	19.0	10195	PASS

m/z	Abundance
40.05	319.0
44.15	434.0
265.45	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

m/z	Abundance
39.15	1080.0
40.15	267.0
44.15	420.0
50.15	4538.0
51.15	17096.0
52.25	918.0
56.15	559.0
57.05	1087.0
58.15	161.0
61.15	307.0
61.85	202.0
62.25	161.0
63.15	929.0
65.05	490.0
69.15	19784.0
73.35	167.0
74.15	1778.0
75.15	3160.0
76.25	709.0
77.25	19784.0
78.25	1464.0
79.15	1100.0
80.15	1258.0
81.15	1318.0
82.25	538.0
83.05	241.0
85.15	355.0
86.05	337.0
87.05	171.0
91.15	300.0
92.05	406.0
93.15	2622.0
94.15	318.0
98.15	2155.0
99.15	1638.0
101.05	1010.0
103.05	309.0
104.15	335.0
105.15	498.0
105.95	278.0
107.15	7768.0
108.15	919.0
110.15	11132.0
111.05	2214.0
112.05	222.0
116.15	391.0
117.15	5435.0
118.15	408.0
122.05	530.0
123.15	860.0
124.15	225.0
125.25	325.0
127.15	21000.0
128.05	1679.0
129.15	8281.0
130.15	593.0
130.95	312.0
132.25	212.0
134.15	215.0
135.25	731.0
136.25	305.0
137.25	437.0
138.05	177.0
140.15	158.0
141.05	795.0
142.15	183.0
143.15	364.0
147.15	851.0
148.15	1226.0
149.15	440.0
151.05	225.0
153.15	355.0
154.25	389.0
155.15	614.0
156.15	1299.0
157.45	207.0
158.15	372.0
159.35	259.0
160.15	481.0
161.25	420.0
165.05	667.0
166.05	341.0
167.15	2417.0
168.15	1123.0
169.15	214.0
171.95	190.0
172.25	210.0
173.05	258.0
174.25	817.0
175.15	1094.0
177.15	431.0
178.15	231.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

179.15	1849.0
180.15	1151.0
181.15	485.0
183.85	197.0
185.15	890.0
186.15	5590.0
187.25	2052.0
188.05	288.0
189.15	559.0
191.25	282.0
192.15	905.0
193.05	846.0
196.15	1225.0
196.95	679.0
198.15	38600.0
199.15	2443.0
200.25	272.0
203.15	385.0
204.15	1991.0
205.25	3492.0
206.25	12402.0
207.15	1579.0
208.35	436.0
209.25	206.0
211.05	568.0
215.25	299.0
216.15	359.0
217.15	4171.0
218.15	507.0
221.15	1185.0
222.25	485.0
223.25	1142.0
224.25	6451.0
225.15	1783.0
226.15	227.0
227.15	3985.0
228.05	433.0
229.15	811.0
229.95	153.0
231.25	287.0
234.05	168.0
234.35	196.0
235.15	208.0
236.15	251.0
237.25	247.0
241.25	246.0
242.25	495.0
243.15	629.0
244.15	5212.0
245.15	639.0
246.15	1266.0
247.15	327.0
249.05	245.0
255.15	29688.0
256.15	4268.0
257.05	452.0
258.15	2078.0
259.15	243.0
265.15	986.0
272.35	326.0
273.15	708.0
274.15	2085.0
275.25	12397.0
276.15	1764.0
277.15	954.0
278.25	197.0
281.05	227.0
283.25	249.0
285.05	382.0
296.15	4036.0
297.25	718.0
303.15	540.0
304.25	202.0
314.25	203.0
315.15	389.0
316.35	170.0
321.95	190.0
323.25	1428.0
324.35	175.0
327.15	288.0
333.15	180.0
334.25	626.0
341.35	239.0
346.15	267.0
352.35	270.0
353.15	179.0
354.15	243.0
354.95	162.0
365.25	1706.0
366.15	491.0
372.05	554.0
383.15	221.0
402.05	298.0
403.25	285.0
404.05	223.0
421.15	443.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

423.15	2078.0
424.25	488.0
441.15	5698.0
442.25	31272.0
443.15	5784.0
444.35	544.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

m/z	Abundance
38.05	338.0
39.25	1643.0
44.05	261.0
49.15	267.0
50.15	6476.0
51.15	22856.0
52.15	1082.0
56.15	528.0
57.15	1784.0
57.95	157.0
60.95	417.0
62.25	312.0
63.15	1110.0
65.05	545.0
69.15	24120.0
73.25	339.0
74.15	2438.0
75.15	4618.0
76.15	1474.0
77.15	27848.0
78.15	1921.0
79.15	2366.0
80.15	1488.0
81.15	2372.0
82.15	494.0
83.15	728.0
85.15	375.0
86.15	505.0
87.15	184.0
89.35	163.0
91.15	410.0
92.15	384.0
93.15	3676.0
94.05	272.0
98.15	2705.0
99.15	1984.0
99.95	228.0
101.15	920.0
102.95	231.0
104.05	789.0
105.05	827.0
106.25	393.0
107.15	9800.0
108.15	1539.0
109.15	470.0
110.15	15252.0
111.15	2793.0
113.25	262.0
116.05	491.0
117.15	7577.0
118.15	643.0
120.05	200.0
122.15	778.0
123.15	1169.0
124.05	287.0
125.05	701.0
127.15	32728.0
128.15	2320.0
129.15	13833.0
130.15	939.0
132.25	152.0
134.05	569.0
135.15	1316.0
136.15	460.0
137.05	668.0
140.15	170.0
141.15	2255.0
142.25	468.0
143.05	395.0
146.15	372.0
147.25	718.0
148.15	1824.0
149.15	404.0
151.05	307.0
151.85	264.0
152.25	159.0
153.15	696.0
154.15	394.0
155.15	1016.0
156.15	1593.0
157.45	329.0
158.15	246.0
159.15	281.0
160.05	503.0
161.15	848.0
162.15	328.0
163.25	195.0
163.95	208.0
165.15	777.0
166.15	562.0
167.15	3304.0
168.15	1032.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

169.05	225.0
172.15	353.0
173.15	499.0
174.15	866.0
175.25	1734.0
176.25	258.0
177.15	645.0
178.05	229.0
179.15	3167.0
180.15	1716.0
181.15	787.0
182.15	317.0
184.15	191.0
185.15	1871.0
186.25	10680.0
187.25	2658.0
187.95	199.0
189.25	849.0
191.15	349.0
192.15	883.0
193.25	808.0
194.15	335.0
196.15	1692.0
198.15	66184.0
199.15	4233.0
200.05	527.0
201.35	225.0
203.25	726.0
204.25	3051.0
205.15	4920.0
206.25	20200.0
207.25	2853.0
208.15	594.0
209.25	233.0
211.05	608.0
215.05	192.0
216.15	558.0
217.15	6273.0
218.15	826.0
219.25	250.0
221.15	2317.0
222.15	573.0
223.15	1384.0
224.25	11412.0
225.15	2885.0
226.25	211.0
227.15	4488.0
228.15	660.0
229.15	973.0
230.15	163.0
231.15	581.0
233.25	208.0
234.25	489.0
234.95	418.0
236.15	220.0
237.15	311.0
239.25	324.0
240.35	181.0
241.15	343.0
242.25	711.0
243.25	724.0
244.25	7882.0
245.25	1059.0
246.15	2524.0
247.15	264.0
249.15	240.0
253.05	475.0
254.05	676.0
255.15	49088.0
256.25	6095.0
257.25	528.0
258.25	3510.0
259.15	521.0
260.15	170.0
265.15	1677.0
266.15	260.0
272.05	226.0
273.15	1707.0
274.15	3350.0
275.25	19744.0
276.25	2641.0
277.15	2133.0
277.95	420.0
283.25	250.0
285.15	300.0
293.15	293.0
294.15	263.0
295.15	185.0
296.15	6580.0
297.15	700.0
302.35	156.0
303.15	902.0
304.15	208.0
305.15	210.0
312.15	206.0
314.15	388.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

315.05	903.0
316.15	399.0
321.05	303.0
323.25	1898.0
324.25	324.0
327.05	413.0
327.95	248.0
332.15	208.0
333.05	425.0
334.25	1682.0
335.25	227.0
341.15	390.0
346.15	326.0
347.15	212.0
352.15	683.0
352.95	561.0
353.45	254.0
354.25	826.0
365.15	3268.0
366.25	431.0
370.15	170.0
371.15	241.0
372.15	615.0
372.95	241.0
383.35	277.0
390.35	276.0
391.45	178.0
392.15	214.0
402.25	434.0
403.35	346.0
404.25	290.0
421.25	459.0
422.25	650.0
423.25	3375.0
424.25	747.0
441.25	9266.0
442.25	58672.0
443.25	11264.0
444.25	1216.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

m/z	Abundance
38.25	263.0
39.15	1529.0
40.05	251.0
44.15	366.0
50.15	4813.0
51.15	18456.0
52.15	1228.0
56.15	736.0
57.15	1513.0
61.05	214.0
62.15	161.0
63.15	986.0
64.25	231.0
65.25	606.0
69.15	22736.0
73.15	186.0
74.25	2897.0
75.15	4193.0
76.15	1051.0
77.15	24080.0
78.15	1833.0
79.15	1863.0
80.15	1411.0
81.15	2230.0
82.15	633.0
83.05	176.0
85.15	399.0
86.25	435.0
87.25	252.0
89.05	154.0
91.05	613.0
92.05	621.0
93.15	3734.0
94.15	240.0
96.65	156.0
97.35	192.0
98.15	2638.0
99.15	1857.0
101.15	1194.0
103.15	264.0
104.15	654.0
105.15	738.0
106.25	535.0
107.15	9633.0
108.15	1401.0
109.15	324.0
110.15	15388.0
111.15	2671.0
112.05	374.0
116.15	517.0
117.15	6294.0
118.05	366.0
119.95	248.0
122.15	782.0
123.25	837.0
124.15	293.0
125.25	257.0
127.15	28864.0
128.15	1918.0
129.15	12550.0
130.15	942.0
131.15	169.0
134.15	469.0
135.15	782.0
136.15	698.0
137.25	315.0
141.15	1896.0
142.15	497.0
143.05	260.0
144.15	172.0
146.45	229.0
147.25	1032.0
148.15	1813.0
149.25	280.0
151.15	213.0
152.15	189.0
153.15	477.0
154.15	336.0
155.15	1097.0
156.25	1737.0
157.45	196.0
158.25	308.0
159.15	358.0
160.15	539.0
161.15	646.0
162.15	396.0
164.05	156.0
165.15	510.0
166.25	496.0
167.15	3007.0
168.05	1467.0
169.15	377.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

171.15	176.0
172.15	328.0
173.15	623.0
174.15	455.0
175.25	1030.0
176.35	219.0
177.15	655.0
178.25	197.0
179.15	2438.0
180.25	1813.0
181.25	729.0
184.15	176.0
185.25	1537.0
186.25	10218.0
187.25	3245.0
188.15	297.0
189.25	950.0
191.15	377.0
192.15	1041.0
193.15	693.0
194.15	248.0
196.05	1664.0
198.15	59960.0
199.15	4521.0
200.25	475.0
203.25	498.0
204.25	2790.0
205.25	4492.0
206.25	20784.0
207.15	3021.0
208.15	894.0
210.15	183.0
211.25	810.0
215.15	267.0
216.05	450.0
217.15	5566.0
218.15	569.0
221.25	1937.0
222.15	752.0
223.15	1263.0
224.25	10622.0
225.25	2709.0
226.15	378.0
227.15	5551.0
228.25	769.0
229.15	1512.0
231.25	337.0
234.15	383.0
235.15	254.0
236.15	442.0
237.15	373.0
240.25	221.0
241.05	219.0
242.25	672.0
244.15	7486.0
245.25	1381.0
246.15	1701.0
247.25	375.0
249.05	216.0
253.25	260.0
255.15	47208.0
256.15	7589.0
257.15	344.0
258.15	3136.0
259.15	409.0
260.15	164.0
265.25	1420.0
271.05	246.0
273.15	1688.0
274.25	3780.0
275.25	19528.0
276.15	2393.0
277.25	2025.0
278.15	335.0
283.25	404.0
284.05	211.0
285.35	178.0
293.05	396.0
293.85	199.0
296.15	6811.0
297.05	883.0
301.25	172.0
302.15	200.0
303.25	757.0
304.15	197.0
308.25	170.0
314.15	496.0
315.15	844.0
316.15	300.0
321.15	235.0
323.25	2018.0
324.25	445.0
327.25	546.0
328.35	282.0
332.25	291.0
333.35	278.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419

334.15	1370.0
335.15	435.0
341.25	325.0
346.25	363.0
347.05	200.0
351.45	167.0
352.15	610.0
353.15	518.0
354.15	921.0
365.25	3537.0
366.25	630.0
371.25	426.0
372.15	1377.0
373.45	373.0
377.35	219.0
383.25	588.0
384.25	206.0
390.35	309.0
402.25	553.0
403.25	396.0
410.35	169.0
415.15	221.0
421.15	528.0
422.15	888.0
423.25	4333.0
424.25	1038.0
441.25	9421.0
442.25	70784.0
443.25	13536.0
444.25	1406.0
445.15	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP419