


Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG090916\  
 Data File : BG023970.D  
 Acq On : 9 Sep 2016 9:35  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG083116.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Sep 09 16:24:09 2016



AutoFind: Scans 2570, 2571, 2572; Background Corrected with Scan 2564

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 40.8      | 17594   | PASS             |
| 68          | 69           | 0.00         | 2            | 2.0       | 388     | PASS             |
| 69          | 198          | 0.00         | 100          | 45.1      | 19454   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 88      | PASS             |
| 127         | 198          | 10           | 80           | 43.2      | 18646   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 43152   | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 2894    | PASS             |
| 275         | 198          | 10           | 60           | 26.8      | 11554   | PASS             |
| 365         | 198          | 1            | 100          | 6.8       | 2913    | PASS             |
| 441         | 442          | 0.01         | 24           | 14.7      | 4933    | PASS             |
| 442         | 198          | 50           | 100          | 77.8      | 33584   | PASS             |
| 443         | 442          | 15           | 24           | 22.7      | 7608    | PASS             |