

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090924\
 Data File : BG062691.D
 Acq On : 9 Sep 2024 16:24
 Operator : JU/RC
 Sample : SST04055
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040441

Quant Time: Sep 10 00:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 00:41:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.907	152	66837	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	298104	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	238526	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.278	188	610028	20.000	ng/u1	0.00
79) Chrysene-d12	21.526	240	633735	20.000	ng/u1	0.00
88) Perylene-d12	24.593	264	682818	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.418	96	445	0.300	ng/uL	0.05
4) Pyridine-d5	3.770	84	192647	40.932	ng/u1	0.00
7) Phenol-d5	7.072	99	232886	39.736	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.237	67	139582	39.724	ng/u1	0.00
11) 2-Chlorophenol-d4	7.443	132	173596	40.664	ng/u1	0.00
15) 4-Methylphenol-d8	8.612	113	192974	39.511	ng/u1	0.00
21) Nitrobenzene-d5	9.064	128	88307	41.447	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	100670	43.204	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	206043	40.604	ng/u1	0.00
31) 4-Chloroaniline-d4	10.833	131	281342	40.335	ng/u1	0.00
46) Dimethylphthalate-d6	13.941	166	731671	39.834	ng/u1	0.00
49) Acenaphthylene-d8	14.229	160	829135	40.644	ng/u1	0.00
54) 4-Nitrophenol-d4	14.740	143	128738	41.258	ng/u1	0.00
60) Fluorene-d10	15.527	176	662787	40.043	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.651	200	145553	41.480	ng/u1	0.00
73) Anthracene-d10	17.378	188	1166089	40.504	ng/u1	0.00
81) Pyrene-d10	19.669	212	1445692	40.539	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.382	264	1470487	40.785	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.400	88	26779	16.160	ng/uL	92
5) Pyridine	3.788	79	197605	40.029	ng/u1	98
6) Benzaldehyde	7.043	77	140691	44.060	ng/u1	95
8) Phenol	7.102	94	240171	39.421	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.331	93	182093	39.538	ng/u1	98
12) 2-Chlorophenol	7.472	128	181165	41.505	ng/u1	97
13) 2-Methylphenol	8.347	108	182951	40.046	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.424	45	328201	39.592	ng/u1	96
16) Acetophenone	8.723	105	316762	39.905	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.712	70	171480	41.267	ng/u1	94
18) 4-Methylphenol	8.677	108	205120	39.908	ng/u1	99
19) Hexachloroethane	8.988	117	72611	41.323	ng/u1	97
22) Nitrobenzene	9.105	77	238581	41.393	ng/u1#	94
23) Isophorone	9.628	82	473870	40.213	ng/u1	99
25) 2-Nitrophenol	9.816	139	108469	42.236	ng/u1	92
26) 2,4-Dimethylphenol	9.875	107	222049	40.463	ng/u1	95
27) Bis(2-Chloroethoxy)met...	10.110	93	263948	40.079	ng/u1	97
29) 2,4-Dichlorophenol	10.345	162	206976	41.412	ng/u1	98
30) Naphthalene	10.751	128	637057	39.856	ng/u1	99
32) 4-Chloroaniline	10.856	127	280139	40.383	ng/u1	100
33) Hexachlorobutadiene	11.038	225	167160	40.148	ng/u1	98
34) Caprolactam	11.614	113	39302	21.511	ng/u1	89
35) 4-Chloro-3-methylphenol	11.984	107	224526	41.182	ng/u1	99
36) 2-Methylnaphthalene	12.360	142	476017	40.352	ng/u1	97


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090924\
 Data File : BG062691.D
 Acq On : 9 Sep 2024 16:24
 Operator : JU/RC
 Sample : SST04055
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040441

Quant Time: Sep 10 00:52:13 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 00:41:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.578	142	481670	39.861	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.725	216	333369	40.654	ng/ul	97
40) Hexachlorocyclopentadiene	12.707	237	167657	39.958	ng/ul	94
41) 2,4,6-Trichlorophenol	12.966	196	206348	41.544	ng/ul	98
42) 2,4,5-Trichlorophenol	13.042	196	221213	41.481	ng/ul	93
43) 1,1'-Biphenyl	13.365	154	666791	39.753	ng/ul	98
44) 2-Chloronaphthalene	13.412	162	552339	40.496	ng/ul	97
45) 2-Nitroaniline	13.612	65	162243	43.349	ng/ul	97
47) Dimethylphthalate	13.988	163	757147	40.026	ng/ul	98
48) 2,6-Dinitrotoluene	14.111	165	143184	42.964	ng/ul	93
50) Acenaphthylene	14.258	152	867130	40.262	ng/ul	99
51) 3-Nitroaniline	14.434	138	139488	41.806	ng/ul	96
52) Acenaphthene	14.599	153	606438	39.752	ng/ul	98
53) 2,4-Dinitrophenol	14.646	184	91737	41.687	ng/ul	92
55) 4-Nitrophenol	14.752	109	113611	40.506	ng/ul	92
56) Dibenzofuran	14.934	168	901357	39.772	ng/ul	99
57) 2,4-Dinitrotoluene	14.899	165	220415	42.302	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.163	232	225616	41.747	ng/ul	99
59) Diethylphthalate	15.357	149	743647	39.850	ng/ul	100
61) Fluorene	15.586	166	707085	39.717	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.580	204	411409	39.395	ng/ul	99
63) 4-Nitroaniline	15.604	138	152083	42.357	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.662	198	149892	41.209	ng/ul	95
67) N-Nitrosodiphenylamine	15.792	169	638025	40.781	ng/ul	98
68) 4-Bromophenyl-phenylether	16.473	248	285808	40.715	ng/ul	95
69) Hexachlorobenzene	16.591	284	325691	40.636	ng/ul	99
70) Atrazine	16.744	200	283155	40.637	ng/ul	96
71) Pentachlorophenol	16.932	266	200812	41.889	ng/ul	99
72) Phenanthrene	17.325	178	1277358	40.266	ng/ul	100
74) Anthracene	17.413	178	1304190	40.327	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.336	216	341828	41.032	ng/ul	99
76) Pentachlorobenzene	14.857	250	366851	39.739	ng/ul	99
77) Carbazole	17.684	167	1115649	41.599	ng/ul	98
78) Di-n-butylphthalate	18.248	149	1302178	41.056	ng/ul	99
80) Fluoranthene	19.335	202	1596268	40.415	ng/ul	99
82) Pyrene	19.699	202	1712139	40.270	ng/ul	98
83) Butylbenzylphthalate	20.592	149	535443	42.424	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.426	252	564379	42.300	ng/ul	99
85) Benzo(a)anthracene	21.509	228	1764888	40.041	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.426	149	775854	41.748	ng/ul	99
87) Chrysene	21.573	228	1673327	40.017	ng/ul	98
89) Di-n-octyl phthalate	22.584	149	1337746	42.119	ng/ul	100
90) Benzo(b)fluoranthene	23.624	252	1700176	40.329	ng/ul	99
91) Benzo(k)fluoranthene	23.688	252	1782839	41.301	ng/ul	100
93) Benzo(a)pyrene	24.452	252	1628571	40.957	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.054	276	2007274	40.972	ng/ul	98
95) Dibenzo(a,h)anthracene	28.107	278	1615460	41.052	ng/ul	99
96) Benzo(g,h,i)perylene	29.135	276	1525969	40.329	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 16.160 ng/uL
RT: 3.400 min Scan# 51
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion: 88 Resp: 26779
Ion Ratio Lower Upper
88 100
43 31.1 23.1 34.7
58 79.3 57.2 85.8



#5
Pyridine
Concen: 40.029 ng/uL
RT: 3.788 min Scan# 119
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion: 79 Resp: 197605
Ion Ratio Lower Upper
79 100
52 89.3 69.8 104.8
51 41.1 34.0 51.0









#6
Benzaldehyde
Concen: 44.060 ng/ul
RT: 7.043 min Scan# 61
Delta R.T. -0.004 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion: 77 Resp: 140691
Ion Ratio Lower Upper
77 100
105 94.8 79.0 118.4
106 91.9 69.0 103.6



#8
Phenol
Concen: 39.421 ng/ul
RT: 7.102 min Scan# 683
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion: 94 Resp: 240171
Ion Ratio Lower Upper
94 100
65 31.4 25.1 37.7
66 40.1 30.8 46.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 39.538 ng/ul
 RT: 7.331 min Scan# 71
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Instrument :
 BNA_G
 ClientSampleId :
 SSTD040441

Tgt Ion: 93 Resp: 182093




Ion	Ratio	Lower	Upper
93	100		
63	80.2	62.0	93.0
95	32.9	25.9	38.9




#12
 2-Chlorophenol
 Concen: 41.505 ng/ul
 RT: 7.472 min Scan# 746
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Tgt Ion:128 Resp: 181165

Ion	Ratio	Lower	Upper
128	100		
64	49.2	42.0	63.0
130	30.8	24.0	36.0





#13

2-Methylphenol

Concen: 40.046 ng/ul

RT: 8.347 min Scan# 895

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :





SSTD040441

Tgt Ion:108 Resp: 182951

Ion Ratio Lower Upper

108 100

107 88.3 72.1 108.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 39.592 ng/ul

RT: 8.424 min Scan# 908

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion: 45 Resp: 328201


Ion Ratio Lower Upper

45 100

77 12.8 9.0 13.6

79 9.5 6.6 9.8









#16
Acetophenone
Concen: 39.905 ng/ul
RT: 8.723 min Scan# 91
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:105 Resp: 316762
Ion Ratio Lower Upper
105 100
77 80.2 63.3 94.9
51 30.2 26.5 39.7



#17
N-Nitroso-di-n-propylamine
Concen: 41.267 ng/ul
RT: 8.712 min Scan# 957
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion: 70 Resp: 171480
Ion Ratio Lower Upper
70 100
42 68.0 50.2 75.2
101 8.9 6.2 9.4
130 17.2 15.1 22.7





#18

4-Methylphenol

Concen: 39.908 ng/ul

RT: 8.677 min Scan# 91

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :





SSTD040441

Tgt Ion:108 Resp: 205120

Ion Ratio Lower Upper

108 100

107 106.6 86.4 129.6



#19

Hexachloroethane

Concen: 41.323 ng/ul

RT: 8.988 min Scan# 1004

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:117 Resp: 72611


Ion Ratio Lower Upper

117 100

201 131.7 106.6 160.0

199 81.4 62.0 93.0









#22
Nitrobenzene
Concen: 41.393 ng/ul
RT: 9.105 min Scan# 1102
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion: 77 Resp: 238581
Ion Ratio Lower Upper
77 100
123 42.4 30.9 46.3
65 14.8 9.7 14.5#



#23
Isophorone
Concen: 40.213 ng/ul
RT: 9.628 min Scan# 1113
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion: 82 Resp: 473870
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.8
138 15.7 13.0 19.6









#25
 2-Nitrophenol
 Concen: 42.236 ng/ul
 RT: 9.816 min Scan# 1145
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040441


Tgt Ion:139 Resp: 108469
 Ion Ratio Lower Upper
 139 100
 65 50.1 45.4 68.0
 109 33.5 23.9 35.9



#26
 2,4-Dimethylphenol
 Concen: 40.463 ng/ul
 RT: 9.875 min Scan# 1155
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Tgt Ion:107 Resp: 222049
 Ion Ratio Lower Upper
 107 100
 121 50.7 38.8 58.2
 122 82.4 70.5 105.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.079 ng/ul
 RT: 10.110 min Scan# 1195
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040441


Tgt Ion: 93 Resp: 263948
 Ion Ratio Lower Upper
 93 100
 95 32.5 27.4 41.2
 123 10.8 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 41.412 ng/ul
 RT: 10.345 min Scan# 1235
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Tgt Ion:162 Resp: 206976
 Ion Ratio Lower Upper
 162 100
 164 65.6 51.9 77.9
 98 32.8 23.8 35.6









#30
Naphthalene
Concen: 39.856 ng/ul
RT: 10.751 min Scan# 1304
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:128 Resp: 637057
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.6
127 12.9 10.7 16.1



#32
4-Chloroaniline
Concen: 40.383 ng/ul
RT: 10.856 min Scan# 1322
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:127 Resp: 280139
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 40.148 ng/ul

RT: 11.038 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:225 Resp: 167160

Ion Ratio Lower Upper

225 100

223 64.2 52.3 78.5

227 69.0 56.4 84.6



#34

Caprolactam

Concen: 21.511 ng/ul

RT: 11.614 min Scan# 1451

Delta R.T. 0.013 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:113 Resp: 39302


Ion Ratio Lower Upper

113 100

55 207.8 150.8 226.2

56 144.6 125.0 187.6









#35
4-Chloro-3-methylphenol
Concen: 41.182 ng/ul
RT: 11.984 min Scan# 1514
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:107 Resp: 224526
Ion Ratio Lower Upper
107 100
144 27.0 21.8 32.6
142 81.6 64.8 97.2



#36
2-Methylnaphthalene
Concen: 40.352 ng/ul
RT: 12.360 min Scan# 1578
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:142 Resp: 476017
Ion Ratio Lower Upper
142 100
141 88.8 73.4 110.2





#37

1-Methylnaphthalene

Concen: 39.861 ng/ul

RT: 12.578 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:142 Resp: 481670

Ion Ratio Lower Upper

142 100

141 90.2 71.1 106.7

116 3.7 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 40.654 ng/ul

RT: 12.725 min Scan# 1640

Delta R.T. -0.004 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Tgt Ion:216 Resp: 333369




Ion Ratio Lower Upper


216 100

214 74.9 61.8 92.8

179 19.2 13.6 20.4

108 12.7 10.7 16.1









#40
Hexachlorocyclopentadiene
Concen: 39.958 ng/ul
RT: 12.707 min Scan# 1637
Delta R.T. -0.004 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:237 Resp: 167657
Ion Ratio Lower Upper
237 100
235 63.5 47.0 70.6
272 13.7 9.4 14.0



#41
2,4,6-Trichlorophenol
Concen: 41.544 ng/ul
RT: 12.966 min Scan# 1681
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:196 Resp: 206348
Ion Ratio Lower Upper
196 100
198 96.6 76.2 114.4
200 31.2 26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 41.481 ng/ul
RT: 13.042 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:196 Resp: 221213
Ion Ratio Lower Upper
196 100
198 96.3 71.2 106.8
200 33.3 24.6 37.0



#43
1,1'-Biphenyl
Concen: 39.753 ng/ul
RT: 13.365 min Scan# 1749
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:154 Resp: 666791
Ion Ratio Lower Upper
154 100
153 40.2 33.4 50.0
76 12.5 10.1 15.1





#44

2-Chloronaphthalene

Concen: 40.496 ng/ul

RT: 13.412 min Scan# 1757

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:162 Resp: 552339

Ion Ratio Lower Upper

162 100

164 32.7 27.1 40.7

127 33.7 28.5 42.7



#45

2-Nitroaniline

Concen: 43.349 ng/ul

RT: 13.612 min Scan# 1791

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion: 65 Resp: 162243


Ion Ratio Lower Upper

65 100

92 65.5 55.0 82.4

138 97.4 80.0 120.0









#47
Dimethylphthalate
Concen: 40.026 ng/ul
RT: 13.988 min Scan# 1855
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:163 Resp: 757147
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.6
164 9.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 42.964 ng/ul
RT: 14.111 min Scan# 1876
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:165 Resp: 143184
Ion Ratio Lower Upper
165 100
89 54.0 47.4 71.0
121 17.5 17.0 25.6





#50

Acenaphthylene

Concen: 40.262 ng/ul

RT: 14.258 min Scan# 1901

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:152 Resp: 867130

Ion Ratio Lower Upper

152 100

151 20.3 15.9 23.9

153 13.1 10.2 15.2



#51

3-Nitroaniline

Concen: 41.806 ng/ul

RT: 14.434 min Scan# 1931

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:138 Resp: 139488


Ion Ratio Lower Upper

138 100

108 10.3 7.7 11.5

92 119.7 92.6 138.8





#52

Acenaphthene

Concen: 39.752 ng/ul

RT: 14.599 min Scan# 1959

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:153 Resp: 606438

Ion Ratio Lower Upper

153 100

152 48.1 37.2 55.8

154 85.7 67.8 101.6



#53

2,4-Dinitrophenol

Concen: 41.687 ng/ul

RT: 14.646 min Scan# 1967

Delta R.T. 0.008 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:184 Resp: 91737


Ion Ratio Lower Upper

184 100

63 51.9 49.0 73.4

154 59.2 50.4 75.6









#55
4-Nitrophenol
Concen: 40.506 ng/ul
RT: 14.752 min Scan# 1984
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:109 Resp: 113611
Ion Ratio Lower Upper
109 100
139 123.4 92.5 138.7
65 133.1 97.8 146.8



#56
Dibenzofuran
Concen: 39.772 ng/ul
RT: 14.934 min Scan# 2016
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:168 Resp: 901357
Ion Ratio Lower Upper
168 100
139 34.0 27.6 41.4





#57

2,4-Dinitrotoluene

Concen: 42.302 ng/ul

RT: 14.899 min Scan# 2009

Delta R.T. 0.008 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:165 Resp: 220415

Ion Ratio Lower Upper

165 100

63 40.6 33.8 50.8

182 2.3 2.4 3.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 41.747 ng/ul

RT: 15.163 min Scan# 2055

Delta R.T. 0.008 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Tgt Ion:232 Resp: 225616




Ion Ratio Lower Upper


232 100

131 33.2 26.6 40.0

130 1.9 1.6 2.4

166 25.0 20.6 31.0









#59
Diethylphthalate
Concen: 39.850 ng/ul
RT: 15.357 min Scan# 2088
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:149 Resp: 743647
Ion Ratio Lower Upper
149 100
177 22.8 18.4 27.6
150 12.2 9.7 14.5



#61
Fluorene
Concen: 39.717 ng/ul
RT: 15.586 min Scan# 2127
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:166 Resp: 707085
Ion Ratio Lower Upper
166 100
165 103.0 80.9 121.3
167 13.9 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 39.395 ng/ul
RT: 15.580 min Scan# 2110
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:204 Resp: 411409
Ion Ratio Lower Upper
204 100
206 31.8 25.3 37.9
141 45.8 37.4 56.0



#63
4-Nitroaniline
Concen: 42.357 ng/ul
RT: 15.604 min Scan# 2130
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:138 Resp: 152083
Ion Ratio Lower Upper
138 100
92 51.9 42.8 64.2
108 88.5 71.0 106.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 41.209 ng/ul
 RT: 15.662 min Scan# 2140
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040441


Tgt Ion:198 Resp: 149892
 Ion Ratio Lower Upper
 198 100
 182 3.3 3.3 4.9
 77 21.5 15.0 22.6



#67
 N-Nitrosodiphenylamine
 Concen: 40.781 ng/ul
 RT: 15.792 min Scan# 2162
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Tgt Ion:169 Resp: 638025
 Ion Ratio Lower Upper
 169 100
 168 70.5 55.7 83.5
 167 37.8 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 40.715 ng/ul
RT: 16.473 min Scan# 21
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:248 Resp: 285808
Ion Ratio Lower Upper
248 100
250 94.0 77.9 116.9
141 54.3 47.9 71.9



#69
Hexachlorobenzene
Concen: 40.636 ng/ul
RT: 16.591 min Scan# 2298
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:284 Resp: 325691
Ion Ratio Lower Upper
284 100
142 23.7 18.6 27.8
249 29.6 23.7 35.5









#70
Atrazine
Concen: 40.637 ng/ul
RT: 16.744 min Scan# 21
Delta R.T. 0.008 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:200 Resp: 283155
Ion Ratio Lower Upper
200 100
173 24.9 18.6 27.8
215 49.5 37.6 56.4



#71
Pentachlorophenol
Concen: 41.889 ng/ul
RT: 16.932 min Scan# 2356
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:266 Resp: 200812
Ion Ratio Lower Upper
266 100
264 62.1 48.6 72.8
268 62.8 49.5 74.3









#72
Phenanthrene
Concen: 40.266 ng/ul
RT: 17.325 min Scan# 2423
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:178 Resp: 1277358
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 18.9 14.9 22.3



#74
Anthracene
Concen: 40.327 ng/ul
RT: 17.413 min Scan# 2438
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:178 Resp: 1304190
Ion Ratio Lower Upper
178 100
179 16.0 12.4 18.6
176 18.7 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 41.032 ng/ul
RT: 13.336 min Scan# 1
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:216 Resp: 341828
Ion Ratio Lower Upper
216 100
214 75.2 61.5 92.3
218 47.2 38.0 57.0



#76
Pentachlorobenzene
Concen: 39.739 ng/ul
RT: 14.857 min Scan# 2003
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:250 Resp: 366851
Ion Ratio Lower Upper
250 100
248 64.2 51.5 77.3
252 64.4 50.8 76.2





#77

Carbazole

Concen: 41.599 ng/ul

RT: 17.684 min Scan# 2484

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:167 Resp: 1115649

Ion Ratio Lower Upper

167 100

166 22.6 17.2 25.8

139 11.7 9.0 13.6



#78

Di-n-butylphthalate

Concen: 41.056 ng/ul

RT: 18.248 min Scan# 2580

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:149 Resp: 1302178


Ion Ratio Lower Upper

149 100

150 9.3 7.0 10.6

104 5.1 4.1 6.1









#80
Fluoranthene
Concen: 40.415 ng/ul
RT: 19.335 min Scan# 2765
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:202 Resp: 1596268
Ion Ratio Lower Upper
202 100
101 8.9 6.6 10.0
100 6.9 5.3 7.9



#82
Pyrene
Concen: 40.270 ng/ul
RT: 19.699 min Scan# 2827
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:202 Resp: 1712139
Ion Ratio Lower Upper
202 100
101 10.3 7.7 11.5
100 8.6 6.5 9.7









#83
 Butylbenzylphthalate
 Concen: 42.424 ng/ul
 RT: 20.592 min Scan# 2979
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040441


Tgt Ion:149 Resp: 535443
 Ion Ratio Lower Upper
 149 100
 91 70.4 57.3 85.9
 206 23.5 20.4 30.6



#84
 3,3'-Dichlorobenzidine
 Concen: 42.300 ng/ul
 RT: 21.426 min Scan# 3121
 Delta R.T. 0.002 min
 Lab File: BG062691.D
 Acq: 9 Sep 2024 16:24

Tgt Ion:252 Resp: 564379
 Ion Ratio Lower Upper
 252 100
 254 65.4 51.8 77.6
 126 7.7 6.0 9.0









#85
Benzo(a)anthracene
Concen: 40.041 ng/ul
RT: 21.509 min Scan# 3135
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24




Instrument :
BNA_G
ClientSampleId :
SSTD040441


Tgt Ion:228 Resp: 1764888
Ion Ratio Lower Upper
228 100
229 20.2 15.8 23.8
226 28.0 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 41.748 ng/ul
RT: 21.426 min Scan# 3121
Delta R.T. 0.002 min
Lab File: BG062691.D
Acq: 9 Sep 2024 16:24

Tgt Ion:149 Resp: 775854
Ion Ratio Lower Upper
149 100
167 27.2 21.3 31.9
279 4.3 3.8 5.6





#87

Chrysene

Concen: 40.017 ng/ul

RT: 21.573 min Scan# 3146

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:228 Resp: 1673327

Ion Ratio Lower Upper

228 100

226 30.8 24.2 36.2

229 20.5 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 42.119 ng/ul




RT: 22.584 min Scan# 3318


Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Tgt Ion:149 Resp: 1337746





#90

Benzo(b)fluoranthene

Concen: 40.329 ng/ul

RT: 23.624 min Scan# 3495

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:252 Resp: 1700176

Ion Ratio Lower Upper

252 100

253 21.7 17.0 25.6

125 6.9 5.4 8.0



#91

Benzo(k)fluoranthene

Concen: 41.301 ng/ul

RT: 23.688 min Scan# 3506

Delta R.T. 0.002 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:252 Resp: 1782839


Ion Ratio Lower Upper

252 100

253 21.2 16.9 25.3

125 7.0 5.7 8.5





#93

Benzo(a)pyrene

Concen: 40.957 ng/ul

RT: 24.452 min Scan# 3

Delta R.T. 0.007 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:252 Resp: 1628571

Ion Ratio Lower Upper

252 100

253 21.6 17.4 26.0

125 8.0 6.1 9.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 40.972 ng/ul

RT: 28.054 min Scan# 4249

Delta R.T. 0.007 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24




Tgt Ion:276 Resp: 2007274


Ion Ratio Lower Upper

276 100

138 14.0 11.4 17.2

277 24.6 18.9 28.3





#95

Dibenzo(a,h)anthracene

Concen: 41.052 ng/ul

RT: 28.107 min Scan# 41

Delta R.T. 0.008 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Instrument :

BNA_G

ClientSampleId :

SSTD040441





Tgt Ion:278 Resp: 1615460

Ion Ratio Lower Upper

278 100

139 10.3 7.8 11.8

279 23.1 18.6 28.0



#96

Benzo(g,h,i)perylene

Concen: 40.329 ng/ul

RT: 29.135 min Scan# 4433

Delta R.T. 0.008 min

Lab File: BG062691.D

Acq: 9 Sep 2024 16:24

Tgt Ion:276 Resp: 1525969

Ion Ratio Lower Upper

276 100

138 15.3 11.8 17.6

277 23.8 18.4 27.6

