




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090925\
Data File : BG064316.D
Acq On : 9 Sep 2025 8:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 09 11:23:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 28 17:41:58 2025
Response via : Initial Calibration



TIC: BG064316.D\data.ms

(70) Pentachlorophenol (C)




16.873min (-0.005) 401326.33 ng

response 198040

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	64.73
264.00	55.40	64.12
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG090925\
Data File : BG064316.D
Acq On : 9 Sep 2025 8:49
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 09 11:23:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 28 17:41:58 2025
Response via : Initial Calibration



TIC: BG064316.D\data.ms

(77) Benzidine

19.458min (+ 0.001) 754140.66 ng

response 904275

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	14.12
183.00	12.10	12.94
0.00	0.00	0.00