

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091017\
 Data File : BG028630.D
 Acq On : 10 Sep 2017 00:52
 Operator : SJ/JU
 Sample : I5167-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CPS043061

Quant Time: Sep 11 07:17:32 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 18:39:50 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.34	152	31552	20.00	ng	0.00
21) Naphthalene-d8	11.15	136	131508	20.00	ng	-0.01
38) Acenaphthene-d10	14.95	164	95420	20.00	ng	0.00
63) Phenanthrene-d10	17.70	188	259825	20.00	ng	0.00
75) Chrysene-d12	22.03	240	311569	20.00	ng	-0.01
86) Perylene-d12	25.53	264	300882	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.90	112	75111	40.08	ng	-0.01
7) Phenol-d6	7.49	99	61606	22.67	ng	0.00
23) Nitrobenzene-d5	9.51	82	242346	84.62	ng	0.00
41) 2,4,6-Tribromophenol	16.44	330	148978	102.72	ng	0.00
44) 2-Fluorobiphenyl	13.57	172	632089	86.66	ng	0.00
78) Terphenyl-d14	20.28	244	1156202	79.95	ng	0.00
Target Compounds						
35) Caprolactam	12.07	113	45415	53.927	ng	Qvalue 91
83) Bis(2-ethylhexyl)phthalate	21.86	149	29726	2.886	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091017\
Data File : BG028630.D
Acq On : 10 Sep 2017 00:52
Operator : SJ/JU
Sample : I5167-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CPS043061

Quant Time: Sep 11 07:17:32 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 07 18:39:50 2017
Response via : Initial Calibration

