


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG091017\
 Data File : BG028615.D
 Acq On : 9 Sep 2017 15:01
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG090717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 07 18:39:50 2017



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.5	18170	PASS
68	69	0.00	2	1.2	238	PASS
69	198	0.00	100	37.7	19853	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	40.0	21061	PASS
197	198	0.00	2	0.4	217	PASS
198	198	100	100	100.0	52613	PASS
199	198	5	9	7.2	3786	PASS
275	198	10	60	28.0	14713	PASS
365	198	1	100	4.0	2112	PASS
441	442	0.01	24	15.8	7302	PASS
442	198	50	100	88.1	46344	PASS
443	442	15	24	19.2	8904	PASS