




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091025\
Data File : BG064328.D
Acq On : 10 Sep 2025 10:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 12:18:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 28 17:41:58 2025
Response via : Initial Calibration



TIC: BG064328.D\data.ms

(70) Pentachlorophenol (C)




16.866min (-0.012) 579514.13 ng

response 223802

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	62.79
264.00	55.40	65.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091025\
Data File : BG064328.D
Acq On : 10 Sep 2025 10:50
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 10 12:18:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 28 17:41:58 2025
Response via : Initial Calibration



TIC: BG064328.D\data.ms

(77) Benzidine**19.451min (-0.006) 234970.81 ng****response 1028136**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.60	14.53
183.00	12.10	12.70
0.00	0.00	0.00