

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG091115\  
 Data File : BG018615.D  
 Acq On : 10 Sep 2015 23:03  
 Operator : UM/IZ  
 Sample : SSTDICV040  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

Quant Time: Sep 11 10:57:22 2015  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG091115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 11 10:44:16 2015  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 8.01  | 152  | 62572    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 10.81 | 136  | 276595   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10      | 14.64 | 164  | 177309   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10      | 17.37 | 188  | 454242   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 21.64 | 240  | 494581   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 24.82 | 264  | 510507   | 20.00 | ng    | 0.00     |

## System Monitoring Compounds

|                          |       |     |         |       |    |      |
|--------------------------|-------|-----|---------|-------|----|------|
| 5) 2-Fluorophenol        | 5.58  | 112 | 288660  | 79.69 | ng | 0.00 |
| 7) Phenol-d6             | 7.17  | 99  | 417848  | 80.45 | ng | 0.00 |
| 23) Nitrobenzene-d5      | 9.17  | 82  | 390326  | 78.97 | ng | 0.00 |
| 41) 2,4,6-Tribromophenol | 16.12 | 330 | 184102  | 77.11 | ng | 0.00 |
| 44) 2-Fluorobiphenyl     | 13.26 | 172 | 1006390 | 78.79 | ng | 0.00 |
| 78) Terphenyl-d14        | 19.98 | 244 | 1474490 | 78.29 | ng | 0.00 |

## Target Compounds

|                                |       |     |        |       |    | Qvalue |
|--------------------------------|-------|-----|--------|-------|----|--------|
| 2) 1,4-Dioxane                 | 3.48  | 88  | 59095  | 39.03 | ng | # 100  |
| 3) Pyridine                    | 3.87  | 79  | 189761 | 40.43 | ng | 94     |
| 4) n-Nitrosodimethylamine      | 3.79  | 42  | 64987  | 39.76 | ng | # 76   |
| 6) Aniline                     | 7.33  | 93  | 284969 | 39.81 | ng | 95     |
| 8) 2-Chlorophenol              | 7.57  | 128 | 168539 | 40.30 | ng | 98     |
| 9) Benzaldehyde                | 7.14  | 77  | 117348 | 41.97 | ng | 97     |
| 10) Phenol                     | 7.20  | 94  | 223417 | 40.72 | ng | 98     |
| 11) bis(2-Chloroethyl)ether    | 7.42  | 93  | 171427 | 40.32 | ng | 100    |
| 12) 1,3-Dichlorobenzene        | 7.90  | 146 | 190737 | 39.20 | ng | 97     |
| 13) 1,4-Dichlorobenzene        | 8.04  | 146 | 191700 | 39.28 | ng | 99     |
| 14) 1,2-Dichlorobenzene        | 8.36  | 146 | 186788 | 40.09 | ng | 97     |
| 15) Benzyl Alcohol             | 8.24  | 79  | 152936 | 40.85 | ng | 97     |
| 16) 2,2'-oxybis(1-Chloropropan | 8.54  | 45  | 189939 | 39.34 | ng | 98     |
| 17) 2-Methylphenol             | 8.44  | 107 | 155507 | 40.79 | ng | 97     |
| 18) Hexachloroethane           | 9.09  | 117 | 65769  | 37.69 | ng | 93     |
| 19) n-Nitroso-di-n-propylamine | 8.81  | 70  | 149224 | 40.79 | ng | # 99   |
| 20) 3+4-Methylphenols          | 8.77  | 107 | 220103 | 41.12 | ng | 97     |
| 22) Acetophenone               | 8.83  | 105 | 274678 | 39.20 | ng | # 98   |
| 24) Nitrobenzene               | 9.21  | 77  | 208898 | 39.80 | ng | 98     |
| 25) Isophorone                 | 9.73  | 82  | 414252 | 38.97 | ng | 99     |
| 26) 2-Nitrophenol              | 9.92  | 139 | 102038 | 41.67 | ng | 99     |
| 27) 2,4-Dimethylphenol         | 9.98  | 122 | 176936 | 39.79 | ng | 97     |
| 28) bis(2-Chloroethoxy)methane | 10.21 | 93  | 237362 | 38.89 | ng | 95     |
| 29) 2,4-Dichlorophenol         | 10.46 | 162 | 175950 | 39.16 | ng | 96     |
| 30) 1,2,4-Trichlorobenzene     | 10.68 | 180 | 197854 | 39.73 | ng | 95     |
| 31) Naphthalene                | 10.86 | 128 | 582384 | 39.32 | ng | 98     |
| 32) Benzoic acid               | 10.11 | 122 | 131741 | 37.72 | ng | 97     |
| 33) 4-Chloroaniline            | 10.96 | 127 | 262000 | 39.09 | ng | 97     |
| 34) Hexachlorobutadiene        | 11.15 | 225 | 112810 | 38.42 | ng | 91     |
| 35) Caprolactam                | 11.73 | 113 | 75929  | 39.30 | ng | 97     |
| 36) 4-Chloro-3-methylphenol    | 12.09 | 107 | 196627 | 39.22 | ng | 99     |
| 37) 2-Methylnaphthalene        | 12.46 | 142 | 417426 | 38.50 | ng | 99     |
| 39) 1,2,4,5-Tetrachlorobenzene | 12.83 | 216 | 221289 | 39.35 | ng | 99     |
| 40) Hexachlorocyclopentadiene  | 12.81 | 237 | 122965 | 43.54 | ng | 97     |

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG091115\  
 Data File : BG018615.D  
 Acq On : 10 Sep 2015 23:03  
 Operator : UM/IZ  
 Sample : SSTDICV040  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

Quant Time: Sep 11 10:57:22 2015  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG091115.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 11 10:44:16 2015  
 Response via : Initial Calibration


| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 13.07 | 196  | 160351   | 41.17 | ng    | 98       |
| 43) 2,4,5-Trichlorophenol      | 13.14 | 196  | 167508   | 39.21 | ng    | 99       |
| 45) 1,1'-Biphenyl              | 13.47 | 154  | 553373   | 39.25 | ng    | 97       |
| 46) 2-Chloronaphthalene        | 13.51 | 162  | 424792   | 39.10 | ng    | 98       |
| 47) 2-Nitroaniline             | 13.71 | 65   | 133012   | 40.69 | ng    | 92       |
| 48) Acenaphthylene             | 14.35 | 152  | 761706   | 39.67 | ng    | 97       |
| 49) Dimethylphthalate          | 14.08 | 163  | 572503   | 38.99 | ng    | 99       |
| 50) 2,6-Dinitrotoluene         | 14.21 | 165  | 126096   | 39.53 | ng    | 97       |
| 51) Acenaphthene               | 14.70 | 154  | 435732   | 39.00 | ng    | 99       |
| 52) 3-Nitroaniline             | 14.53 | 138  | 151038   | 39.62 | ng    | 94       |
| 53) 2,4-Dinitrophenol          | 14.74 | 184  | 58647    | 40.50 | ng    | 93       |
| 54) Dibenzofuran               | 15.03 | 168  | 668822   | 38.52 | ng    | 99       |
| 55) 4-Nitrophenol              | 14.84 | 139  | 123795   | 42.06 | ng    | 95       |
| 56) 2,4-Dinitrotoluene         | 14.99 | 165  | 182287   | 40.29 | ng    | # 98     |
| 57) Fluorene                   | 15.68 | 166  | 566612   | 37.90 | ng    | 99       |
| 58) 2,3,4,6-Tetrachlorophenol  | 15.26 | 232  | 154759   | 38.63 | ng    | 97       |
| 59) Diethylphthalate           | 15.45 | 149  | 590046   | 38.53 | ng    | 99       |
| 60) 4-Chlorophenyl-phenylether | 15.67 | 204  | 274438   | 38.21 | ng    | 99       |
| 61) 4-Nitroaniline             | 15.69 | 138  | 169030   | 41.13 | ng    | 98       |
| 62) Azobenzene                 | 15.96 | 77   | 516932   | 38.20 | ng    | 99       |
| 64) 4,6-Dinitro-2-methylphenol | 15.76 | 198  | 95778    | 40.13 | ng    | 97       |
| 65) n-Nitrosodiphenylamine     | 15.89 | 169  | 515102   | 39.15 | ng    | 99       |
| 66) 4-Bromophenyl-phenylether  | 16.56 | 248  | 183957   | 39.82 | ng    | 98       |
| 67) Hexachlorobenzene          | 16.69 | 284  | 198611   | 39.57 | ng    | 98       |
| 68) Atrazine                   | 16.83 | 200  | 207844   | 39.73 | ng    | 98       |
| 69) Pentachlorophenol          | 17.03 | 266  | 133545   | 43.14 | ng    | 99       |
| 70) Phenanthrene               | 17.42 | 178  | 922178   | 38.60 | ng    | 99       |
| 71) Anthracene                 | 17.51 | 178  | 940349   | 38.87 | ng    | 98       |
| 72) Carbazole                  | 17.78 | 167  | 928353   | 39.65 | ng    | 99       |
| 73) Di-n-butylphthalate        | 18.34 | 149  | 1130801  | 40.94 | ng    | 100      |
| 74) Fluoranthene               | 19.42 | 202  | 1182700  | 39.97 | ng    | 100      |
| 76) Benzidine                  | 19.60 | 184  | 638389   | 42.98 | ng    | 98       |
| 77) Pyrene                     | 19.79 | 202  | 1197885  | 39.41 | ng    | 100      |
| 79) Butylbenzylphthalate       | 20.67 | 149  | 504397   | 40.82 | ng    | 99       |
| 80) Benzo(a)anthracene         | 21.62 | 228  | 1129445  | 39.67 | ng    | 99       |
| 81) 3,3'-Dichlorobenzidine     | 21.53 | 252  | 440336   | 40.70 | ng    | 99       |
| 82) Chrysene                   | 21.69 | 228  | 1060356  | 39.55 | ng    | 99       |
| 83) Bis(2-ethylhexyl)phthalate | 21.52 | 149  | 730204   | 40.96 | ng    | 98       |
| 84) Di-n-octyl phthalate       | 22.73 | 149  | 1233064  | 40.90 | ng    | 98       |
| 85) Indeno(1,2,3-cd)pyrene     | 28.47 | 276  | 1359567  | 39.93 | ng    | # 100    |
| 87) Benzo(b)fluoranthene       | 23.82 | 252  | 1171417  | 39.57 | ng    | 96       |
| 88) Benzo(k)fluoranthene       | 23.88 | 252  | 1127559  | 38.02 | ng    | 99       |
| 89) Benzo(a)pyrene             | 24.68 | 252  | 1102268  | 39.13 | ng    | 98       |
| 90) Dibenzo(a,h)anthracene     | 28.53 | 278  | 1079332  | 38.91 | ng    | 98       |
| 91) Benzo(g,h,i)perylene       | 29.60 | 276  | 1094311  | 38.74 | ng    | 96       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG091115\  
Data File : BG018615.D  
Acq On : 10 Sep 2015 23:03  
Operator : UM/IZ  
Sample : SSTDICV040  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

Quant Time: Sep 11 10:57:22 2015  
Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG091115.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 11 10:44:16 2015  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.01 min Scan# 880

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:152 Resp: 62572

Ion Ratio Lower Upper

152 100

150 156.4 115.0 172.4




115 57.5 43.9 65.9



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 39.03 ng

RT: 3.48 min Scan# 109

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion: 88 Resp: 59095

Ion Ratio Lower Upper

88 100

58 66.7 0.0 0.0#



43 23.5 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 40.43 ng

RT: 3.87 min Scan# 176

Delta R.T. -0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion: 79 Resp: 189761

Ion Ratio Lower Upper




79 100

52 56.8 50.3 75.5

51 28.9 24.8 37.2



Abundance Ion 79.00 (78.70 to 79.70): BGC  
Ion 52.00 (51.70 to 52.70): BGC  
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 39.76 ng

RT: 3.79 min Scan# 162

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion: 42 Resp: 64987

Ion Ratio Lower Upper



42 100


74 173.3 113.9 170.9#

44 7.4 5.7 8.5



Abundance Ion 42.00 (41.70 to 42.70): BGC  
Ion 74.00 (73.70 to 74.70): BGC  
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 79.69 ng

RT: 5.58 min Scan# 466

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion: 112 Resp: 288660

Ion Ratio Lower Upper

112 100

64 59.7 43.7 65.5




63 30.1 24.0 36.0



Abundance Ion 112.00 (111.70 to 112.70): BGC

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 39.81 ng

RT: 7.33 min Scan# 764

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion: 93 Resp: 284969

Ion Ratio Lower Upper

93 100

66 35.9 26.2 39.2



65 19.6 14.0 21.0




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 80.45 ng

RT: 7.17 min Scan# 737

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion: 99 Resp: 417848

Ion Ratio Lower Upper

99 100

42 15.1 11.5 17.3




71 28.1 22.6 34.0



Abundance Ion 99.00 (98.70 to 99.70): BGC

Ion 42.00 (41.70 to 42.70): BGC

Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 40.30 ng

RT: 7.57 min Scan# 805

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion: 128 Resp: 168539

Ion Ratio Lower Upper

128 100

130 30.9 13.0 53.0



64 47.9 28.2 68.2




Abundance Ion 128.00 (127.70 to 128.70): BGC

Ion 130.00 (129.70 to 130.70): BGC

Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 41.97 ng

RT: 7.14 min Scan# 732

Delta R.T. 0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115





Tgt Ion: 77 Resp: 117348

Ion Ratio Lower Upper

77 100

105 102.9 77.7 117.7

106 94.9 75.8 115.8



#10

Phenol

Concen: 40.72 ng

RT: 7.20 min Scan# 742

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03




Tgt Ion: 94 Resp: 223417

Ion Ratio Lower Upper


94 100

65 25.9 5.4 45.4

66 29.9 11.8 51.8











#11  
 bis(2-Chloroethyl)ether  
 Concen: 40.32 ng  
 RT: 7.42 min Scan# 780  
 Delta R.T. -0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


Tgt Ion: 93 Resp: 171427  
 Ion Ratio Lower Upper  
 93 100  
 63 70.7 50.5 90.5  
 95 32.6 12.2 52.2



#12  
 1,3-Dichlorobenzene  
 Concen: 39.20 ng  
 RT: 7.90 min Scan# 861  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion: 146 Resp: 190737  
 Ion Ratio Lower Upper  
 146 100  
 148 63.0 53.4 80.0  
 75 26.7 21.1 31.7









#13  
 1,4-Dichlorobenzene  
 Concen: 39.28 ng  
 RT: 8.04 min Scan# 886  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 66.6  | 52.5  | 78.7  |
| 111     | 35.3  | 28.3  | 42.5  |



#14  
 1,2-Dichlorobenzene  
 Concen: 40.09 ng  
 RT: 8.36 min Scan# 940  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 63.9  | 52.5  | 78.7  |
| 111     | 39.4  | 29.9  | 44.9  |









#15  
Benzyl Alcohol  
Concen: 40.85 ng  
RT: 8.24 min Scan# 920  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion: 79 Resp: 152936  
Ion Ratio Lower Upper  
79 100  
108 84.5 65.5 98.3  
77 66.9 51.3 76.9



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 39.34 ng  
RT: 8.54 min Scan# 971  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion: 45 Resp: 189939  
Ion Ratio Lower Upper  
45 100  
77 18.3 0.0 37.0  
79 13.6 0.0 33.8









#17  
2-Methylphenol  
Concen: 40.79 ng  
RT: 8.44 min Scan# 954  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 155507 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 105.3 | 86.5  | 129.7  |
| 77       | 39.8  | 34.6  | 51.8   |
| 79       | 41.6  | 32.3  | 48.5   |




Abundance Ion 107.00 (106.70 to 107.70): E  
Ion 108.00 (107.70 to 108.70): E  
Ion 77.00 (76.70 to 77.70): BG  
Ion 79.00 (78.70 to 79.70): BG






#18  
Hexachloroethane  
Concen: 37.69 ng  
RT: 9.09 min Scan# 1064  
Delta R.T. -0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 117   | Resp: | 65769 |
| Ion      | Ratio | Lower | Upper |
| 117      | 100   |       |       |
| 119      | 96.3  | 71.7  | 107.5 |
| 201      | 107.1 | 90.9  | 136.3 |



Abundance Ion 117.00 (116.70 to 117.70): E  
Ion 119.00 (118.70 to 119.70): E  
Ion 201.00 (200.70 to 201.70): E









#19  
n-Nitroso-di-n-propylamine  
Concen: 40.79 ng  
RT: 8.81 min Scan# 1016  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 70    | Resp: | 149224 |
| Ion      | Ratio | Lower | Upper  |
| 70       | 100   |       |        |
| 42       | 50.3  | 40.6  | 60.8   |
| 101      | 8.5   | 8.6   | 13.0#  |
| 130      | 21.3  | 17.0  | 25.6   |




Abundance Ion 70.00 (69.70 to 70.70): BGC  
Ion 42.00 (41.70 to 42.70): BGC  
Ion 101.00 (100.70 to 101.70): BGC  
Ion 130.00 (129.70 to 130.70): BGC






#20  
3+4-Methylphenols  
Concen: 41.12 ng  
RT: 8.77 min Scan# 1010  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 220103 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 89.6  | 72.2  | 112.2  |
| 77       | 30.1  | 13.4  | 53.4   |
| 79       | 25.8  | 7.2   | 47.2   |



Abundance Ion 107.00 (106.70 to 107.70): BGC  
Ion 108.00 (107.70 to 108.70): BGC  
Ion 77.00 (76.70 to 77.70): BGC  
Ion 79.00 (78.70 to 79.70): BGC









#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 10.81 min Scan# 1357  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 276595 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 12.0  | 8.6   | 12.8   |
| 54       | 7.5   | 5.4   | 8.2    |
| 68       | 6.8   | 5.3   | 7.9    |




Abundance Ion 136.00 (135.70 to 136.70): E  
Ion 137.00 (136.70 to 137.70): E  
Ion 54.00 (53.70 to 54.70): BGC  
Ion 68.00 (67.70 to 68.70): BGC






#22  
Acetophenone  
Concen: 39.20 ng  
RT: 8.83 min Scan# 1019  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 105   | Resp: | 274678 |
| Ion      | Ratio | Lower | Upper  |
| 105      | 100   |       |        |
| 71       | 3.2   | 2.1   | 3.1#   |
| 51       | 21.3  | 17.8  | 26.8   |
| 120      | 22.6  | 17.7  | 26.5   |



Abundance Ion 105.00 (104.70 to 105.70): E  
Ion 71.00 (70.70 to 71.70): BGC  
Ion 51.00 (50.70 to 51.70): BGC  
Ion 120.00 (119.70 to 120.70): E









#23  
 Nitrobenzene-d5  
 Concen: 78.97 ng  
 RT: 9.17 min Scan# 1077  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 82      | 100   |       |       |
| 128     | 44.0  | 36.7  | 55.1  |
| 54      | 42.3  | 33.8  | 50.8  |



#24  
 Nitrobenzene  
 Concen: 39.80 ng  
 RT: 9.21 min Scan# 1084  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 77      | 100   |       |       |
| 123     | 46.5  | 38.2  | 57.2  |
| 65      | 12.7  | 10.6  | 16.0  |





#25

Isophorone

Concen: 38.97 ng

RT: 9.73 min Scan# 1173

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115





Tgt Ion: 82 Resp: 414252

Ion Ratio Lower Upper

82 100

95 6.5 5.2 7.8

138 17.5 14.4 21.6



#26

2-Nitrophenol

Concen: 41.67 ng

RT: 9.92 min Scan# 1205

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03




Tgt Ion: 139 Resp: 102038

Ion Ratio Lower Upper


139 100

109 23.5 20.1 30.1

65 42.5 34.1 51.1











#27  
 2,4-Dimethylphenol  
 Concen: 39.79 ng  
 RT: 9.98 min Scan# 1215  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 122     | 100   |       |       |
| 107     | 110.7 | 91.6  | 137.4 |
| 121     | 57.3  | 47.3  | 70.9  |



#28  
 bis(2-Chloroethoxy)methane  
 Concen: 38.89 ng  
 RT: 10.21 min Scan# 1255  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 95      | 34.8  | 25.7  | 38.5  |
| 123     | 12.3  | 8.6   | 12.8  |









#29  
2,4-Dichlorophenol  
Concen: 39.16 ng  
RT: 10.46 min Scan# 1297  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


| Tgt Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 162     | 100   | 175950 |       |       |
| 164     | 65.9  | 42.3   | 82.3  |       |
| 98      | 33.1  | 11.5   | 51.5  |       |



#30  
1,2,4-Trichlorobenzene  
Concen: 39.73 ng  
RT: 10.68 min Scan# 1334  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Resp   | Lower | Upper |
|---------|-------|--------|-------|-------|
| 180     | 100   | 197854 |       |       |
| 182     | 93.7  | 79.7   | 119.5 |       |
| 145     | 28.1  | 22.0   | 33.0  |       |









#31  
Naphthalene  
Concen: 39.32 ng  
RT: 10.86 min Scan# 1366  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion:128 Resp: 582384  
Ion Ratio Lower Upper  
128 100  
129 11.6 8.5 12.7  
127 13.0 10.6 16.0



#32  
Benzoic acid  
Concen: 37.72 ng  
RT: 10.11 min Scan# 1237  
Delta R.T. 0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:122 Resp: 131741  
Ion Ratio Lower Upper  
122 100  
105 121.1 103.8 143.8  
77 91.8 68.8 108.8









#33  
4-Chloroaniline  
Concen: 39.09 ng  
RT: 10.96 min Scan# 1383  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 127   | Resp: | 262000 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 34.5  | 24.6  | 37.0   |
| 65       | 27.6  | 22.6  | 33.8   |
| 92       | 19.2  | 15.1  | 22.7   |




Abundance Ion 127.00 (126.70 to 127.70): E  
Ion 129.00 (128.70 to 129.70): E  
Ion 65.00 (64.70 to 65.70): BGC  
Ion 92.00 (91.70 to 92.70): BGC






#34  
Hexachlorobutadiene  
Concen: 38.42 ng  
RT: 11.15 min Scan# 1415  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 112810 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 57.3  | 51.0  | 76.4   |
| 227      | 57.1  | 51.8  | 77.8   |



Abundance Ion 225.00 (224.70 to 225.70): E  
Ion 223.00 (222.70 to 223.70): E  
Ion 227.00 (226.70 to 227.70): E









#35  
 Caprolactam  
 Concen: 39.30 ng  
 RT: 11.73 min Scan# 1514  
 Delta R.T. 0.04 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


Tgt Ion:113 Resp: 75929  
 Ion Ratio Lower Upper  
 113 100  
 55 150.3 135.0 175.0  
 56 101.3 83.4 123.4



#36  
 4-Chloro-3-methylphenol  
 Concen: 39.22 ng  
 RT: 12.09 min Scan# 1574  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:107 Resp: 196627  
 Ion Ratio Lower Upper  
 107 100  
 144 26.4 20.6 31.0  
 142 77.5 62.6 93.8





#37

2-Methylnaphthalene

Concen: 38.50 ng

RT: 12.46 min Scan# 1638

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:142 Resp: 417426

Ion Ratio Lower Upper

142 100

141 94.2 74.6 112.0




115 33.3 27.4 41.0



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.64 min Scan# 2008

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:164 Resp: 177309

Ion Ratio Lower Upper

164 100

162 103.2 78.8 118.2



160 46.7 36.4 54.6




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39  
 1,2,4,5-Tetrachlorobenzene  
 Concen: 39.35 ng  
 RT: 12.83 min Scan# 1701  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 216   | Resp: | 221289 |
| Ion      | Ratio | Lower | Upper  |
| 216      | 100   |       |        |
| 214      | 78.7  | 62.5  | 93.7   |
| 179      | 19.9  | 15.6  | 23.4   |
| 108      | 17.1  | 13.3  | 19.9   |




Abundance Ion 216.00 (215.70 to 216.70): E  
 Ion 214.00 (213.70 to 214.70): E  
 Ion 179.00 (178.70 to 179.70): E  
 Ion 108.00 (107.70 to 108.70): E






#40  
 Hexachlorocyclopentadiene  
 Concen: 43.54 ng  
 RT: 12.81 min Scan# 1698  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 237   | Resp: | 122965 |
| Ion      | Ratio | Lower | Upper  |
| 237      | 100   |       |        |
| 235      | 62.4  | 44.6  | 84.6   |
| 272      | 11.0  | 0.0   | 32.8   |



Abundance Ion 236.80 (236.50 to 237.50): E  
 Ion 234.90 (234.60 to 235.60): E  
 Ion 271.75 (271.45 to 272.45): E









#41  
 2,4,6-Tribromophenol  
 Concen: 77.11 ng  
 RT: 16.12 min Scan# 2261  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

Tgt Ion:330 Resp: 184102  
 Ion Ratio Lower Upper  
 330 100  
 332 95.4 78.1 117.1  
 141 36.3 28.2 42.2




#42  
 2,4,6-Trichlorophenol  
 Concen: 41.17 ng  
 RT: 13.07 min Scan# 1741  
 Delta R.T. 0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:196 Resp: 160351  
 Ion Ratio Lower Upper  
 196 100  
 198 91.5 74.7 112.1  
 200 32.3 24.1 36.1











#43  
 2,4,5-Trichlorophenol  
 Concen: 39.21 ng  
 RT: 13.14 min Scan# 1754  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


| Tgt Ion | Resp   | Lower | Upper |
|---------|--------|-------|-------|
| 196     | 167508 |       |       |
| 198     | 94.1   | 75.2  | 112.8 |
| 97      | 42.7   | 34.2  | 51.2  |
| 132     | 25.0   | 21.8  | 32.6  |



#44  
 2-Fluorobiphenyl  
 Concen: 78.79 ng  
 RT: 13.26 min Scan# 1774  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

| Tgt Ion | Resp    | Lower | Upper |
|---------|---------|-------|-------|
| 172     | 1006390 |       |       |
| 171     | 37.6    | 29.0  | 43.6  |
| 170     | 25.5    | 19.9  | 29.9  |









#45  
 1,1'-Biphenyl  
 Concen: 39.25 ng  
 RT: 13.47 min Scan# 1809  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

Tgt Ion:154 Resp: 553373  
 Ion Ratio Lower Upper  
 154 100  
 153 43.7 21.9 61.9  
 76 13.9 0.0 33.2




Abundance Ion 154.00 (153.70 to 154.70): E  
 Ion 153.00 (152.70 to 153.70): E  
 Ion 76.00 (75.70 to 76.70): BGC






#46  
 2-Chloronaphthalene  
 Concen: 39.10 ng  
 RT: 13.51 min Scan# 1817  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:162 Resp: 424792  
 Ion Ratio Lower Upper  
 162 100  
 127 37.5 28.3 42.5  
 164 34.1 27.4 41.0



Abundance Ion 162.00 (161.70 to 162.70): E  
 Ion 127.00 (126.70 to 127.70): E  
 Ion 164.00 (163.70 to 164.70): E









#47  
2-Nitroaniline  
Concen: 40.69 ng  
RT: 13.71 min Scan# 1851  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion: 65 Resp: 133012  
Ion Ratio Lower Upper  
65 100  
92 75.5 54.8 82.2  
138 119.8 89.6 134.4



#48  
Acenaphthylene  
Concen: 39.67 ng  
RT: 14.35 min Scan# 1960  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion: 152 Resp: 761706  
Ion Ratio Lower Upper  
152 100  
151 19.9 17.0 25.4  
153 13.9 10.4 15.6





#49

Dimethylphthalate

Concen: 38.99 ng

RT: 14.08 min Scan# 1914

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:163 Resp: 572503

Ion Ratio Lower Upper

163 100

194 4.5 3.1 4.7




164 10.5 8.8 13.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 39.53 ng

RT: 14.21 min Scan# 1935

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:165 Resp: 126096

Ion Ratio Lower Upper

165 100

63 46.0 35.7 53.5



89 52.5 39.7 59.5




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG0

Ion 89.00 (88.70 to 89.70): BG0





#51

Acenaphthene

Concen: 39.00 ng

RT: 14.70 min Scan# 2019

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:154 Resp: 435732

Ion Ratio Lower Upper

154 100

153 114.0 92.9 139.3




152 57.4 45.8 68.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 39.62 ng

RT: 14.53 min Scan# 1990

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:138 Resp: 151038

Ion Ratio Lower Upper

138 100

108 8.2 7.9 11.9



92 111.5 94.2 141.4




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC









#53  
2,4-Dinitrophenol  
Concen: 40.50 ng  
RT: 14.74 min Scan# 2025  
Delta R.T. 0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion:184 Resp: 58647  
Ion Ratio Lower Upper  
184 100  
63 60.1 42.7 64.1  
154 58.7 44.3 66.5



#54  
Dibenzofuran  
Concen: 38.52 ng  
RT: 15.03 min Scan# 2075  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:168 Resp: 668822  
Ion Ratio Lower Upper  
168 100  
139 35.1 27.4 41.0  
169 14.2 11.0 16.4









#55  
4-Nitrophenol  
Concen: 42.06 ng  
RT: 14.84 min Scan# 2043  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 139     | 100   |       |       |
| 109     | 64.7  | 43.2  | 83.2  |
| 65      | 80.8  | 67.8  | 107.8 |



#56  
2,4-Dinitrotoluene  
Concen: 40.29 ng  
RT: 14.99 min Scan# 2068  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 165     | 100   |       |       |
| 63      | 36.8  | 31.0  | 46.4  |
| 89      | 69.2  | 54.4  | 81.6  |
| 182     | 2.5   | 1.6   | 2.4#  |





#57

Fluorene

Concen: 37.90 ng

RT: 15.68 min Scan# 2186

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:166 Resp: 566612

Ion Ratio Lower Upper

166 100

165 99.2 80.0 120.0




167 13.8 11.1 16.7



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#58

2,3,4,6-Tetrachlorophenol

Concen: 38.63 ng

RT: 15.26 min Scan# 2114

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Tgt Ion:232 Resp: 154759


Ion Ratio Lower Upper

232 100

131 39.9 33.0 49.6

130 2.1 2.0 3.0

166 28.5 24.1 36.1





Abundance Ion 232.00 (231.70 to 232.70): E


Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E











#59  
Diethylphthalate  
Concen: 38.53 ng  
RT: 15.45 min Scan# 2146  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion:149 Resp: 590046  
Ion Ratio Lower Upper  
149 100  
177 23.7 19.6 29.4  
150 12.4 10.2 15.2



#60  
4-Chlorophenyl-phenylether  
Concen: 38.21 ng  
RT: 15.67 min Scan# 2184  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:204 Resp: 274438  
Ion Ratio Lower Upper  
204 100  
206 34.1 28.3 42.5  
141 54.6 44.0 66.0









#61  
4-Nitroaniline  
Concen: 41.13 ng  
RT: 15.69 min Scan# 2188  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion: 138 Resp: 169030  
Ion Ratio Lower Upper  
138 100  
92 50.1 29.1 69.1  
108 67.3 49.8 89.8



#62  
Azobenzene  
Concen: 38.20 ng  
RT: 15.96 min Scan# 2234  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion: 77 Resp: 516932  
Ion Ratio Lower Upper  
77 100  
182 27.3 7.0 47.0  
105 21.6 1.4 41.4  
51 24.1 4.9 44.9









#63  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 17.37 min Scan# 2474  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion:188 Resp: 454242  
Ion Ratio Lower Upper  
188 100  
94 9.0 7.7 11.5  
80 9.7 7.8 11.8



#64  
4,6-Dinitro-2-methylphenol  
Concen: 40.13 ng  
RT: 15.76 min Scan# 2199  
Delta R.T. 0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:198 Resp: 95778  
Ion Ratio Lower Upper  
198 100  
51 35.0 14.2 54.2  
105 44.2 21.1 61.1









#65  
n-Nitrosodiphenylamine  
Concen: 39.15 ng  
RT: 15.89 min Scan# 2221  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

Tgt Ion:169 Resp: 515102  
Ion Ratio Lower Upper  
169 100  
168 71.8 58.2 87.2  
167 41.2 32.7 49.1




Abundance Ion 169.00 (168.70 to 169.70): E  
Ion 168.00 (167.70 to 168.70): E  
Ion 167.00 (166.70 to 167.70): E






#66  
4-Bromophenyl-phenylether  
Concen: 39.82 ng  
RT: 16.56 min Scan# 2336  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:248 Resp: 183957  
Ion Ratio Lower Upper  
248 100  
250 98.0 77.8 116.8  
141 60.9 50.4 75.6



Abundance Ion 248.00 (247.70 to 248.70): E  
Ion 250.00 (249.70 to 250.70): E  
Ion 141.00 (140.70 to 141.70): E









#67  
Hexachlorobenzene  
Concen: 39.57 ng  
RT: 16.69 min Scan# 2357  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 34.8  | 26.6  | 39.8  |
| 249     | 32.0  | 26.2  | 39.2  |



#68  
Atrazine  
Concen: 39.73 ng  
RT: 16.83 min Scan# 2381  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 200     | 100   |       |       |
| 173     | 27.7  | 5.2   | 45.2  |
| 215     | 47.1  | 27.0  | 67.0  |









#69  
 Pentachlorophenol  
 Concen: 43.14 ng  
 RT: 17.03 min Scan# 2415  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 266     | 100   |       |       |
| 268     | 59.3  | 47.6  | 71.4  |
| 264     | 60.7  | 48.1  | 72.1  |



#70  
 Phenanthrene  
 Concen: 38.60 ng  
 RT: 17.42 min Scan# 2482  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 178     | 100   |       |       |
| 176     | 21.6  | 16.7  | 25.1  |
| 179     | 17.5  | 13.9  | 20.9  |









#71  
 Anthracene  
 Concen: 38.87 ng  
 RT: 17.51 min Scan# 2497  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


Tgt Ion:178 Resp: 940349  
 Ion Ratio Lower Upper  
 178 100  
 176 20.0 16.8 25.2  
 179 16.9 13.9 20.9



#72  
 Carbazole  
 Concen: 39.65 ng  
 RT: 17.78 min Scan# 2543  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:167 Resp: 928353  
 Ion Ratio Lower Upper  
 167 100  
 166 23.8 19.0 28.6  
 139 11.7 10.2 15.4









#73  
Di-n-butylphthalate  
Concen: 40.94 ng  
RT: 18.34 min Scan# 2638  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115

Tgt Ion:149 Resp: 1130801  
Ion Ratio Lower Upper  
149 100  
150 10.2 8.2 12.4  
104 5.3 4.4 6.6




#74  
Fluoranthene  
Concen: 39.97 ng  
RT: 19.42 min Scan# 2823  
Delta R.T. 0.01 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:202 Resp: 1182700  
Ion Ratio Lower Upper  
202 100  
101 12.0 0.0 32.4  
203 19.4 0.0 39.5







#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.64 min Scan# 3200

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:240 Resp: 494581

Ion Ratio Lower Upper

240 100

120 11.6 7.4 11.0#




236 28.0 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 42.98 ng

RT: 19.60 min Scan# 2853

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:184 Resp: 638389

Ion Ratio Lower Upper

184 100

185 18.0 13.4 20.2



183 12.5 10.1 15.1




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 39.41 ng

RT: 19.79 min Scan# 2885

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:202 Resp: 1197885

Ion Ratio Lower Upper

202 100

200 24.2 19.4 29.0




203 19.3 15.5 23.3



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 78.29 ng

RT: 19.98 min Scan# 2918

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:244 Resp: 1474490

Ion Ratio Lower Upper

244 100

212 8.5 6.9 10.3



122 12.3 10.2 15.2




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79  
Butylbenzylphthalate  
Concen: 40.82 ng  
RT: 20.67 min Scan# 3035  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03




Instrument :  
BNA\_G  
ClientSampleId :  
ICVBG091115


Tgt Ion:149 Resp: 504397  
Ion Ratio Lower Upper  
149 100  
91 71.5 58.1 87.1  
206 22.7 17.8 26.8



#80  
Benzo(a)anthracene  
Concen: 39.67 ng  
RT: 21.62 min Scan# 3196  
Delta R.T. -0.00 min  
Lab File: BG018615.D  
Acq: 10 Sep 2015 23:03

Tgt Ion:228 Resp: 1129445  
Ion Ratio Lower Upper  
228 100  
226 29.7 23.4 35.0  
229 21.0 16.6 25.0









#81  
 3,3'-Dichlorobenzidine  
 Concen: 40.70 ng  
 RT: 21.53 min Scan# 3182  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115

Tgt Ion:252 Resp: 440336  
 Ion Ratio Lower Upper  
 252 100  
 254 64.7 52.5 78.7  
 126 11.1 8.5 12.7




Abundance Ion 252.00 (251.70 to 252.70): E  
 Ion 254.00 (253.70 to 254.70): E  
 Ion 126.00 (125.70 to 126.70): E






#82  
 Chrysene  
 Concen: 39.55 ng  
 RT: 21.69 min Scan# 3208  
 Delta R.T. 0.01 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:228 Resp: 1060356  
 Ion Ratio Lower Upper  
 228 100  
 226 33.2 26.7 40.1  
 229 20.7 17.0 25.6



Abundance Ion 228.00 (227.70 to 228.70): E  
 Ion 226.00 (225.70 to 226.70): E  
 Ion 229.00 (228.70 to 229.70): E









#83  
 Bis(2-ethylhexyl)phthalate  
 Concen: 40.96 ng  
 RT: 21.52 min Scan# 3180  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03




Instrument :  
 BNA\_G  
 ClientSampleId :  
 ICVBG091115


Tgt Ion:149 Resp: 730204  
 Ion Ratio Lower Upper  
 149 100  
 167 28.8 23.8 35.8  
 279 4.0 3.4 5.2



#84  
 Di-n-octyl phthalate  
 Concen: 40.90 ng  
 RT: 22.73 min Scan# 3385  
 Delta R.T. -0.00 min  
 Lab File: BG018615.D  
 Acq: 10 Sep 2015 23:03

Tgt Ion:149 Resp: 1233064  
 Ion Ratio Lower Upper  
 149 100  
 167 1.6 1.2 1.8  
 43 8.2 5.9 8.9





#85

Indeno(1,2,3-cd)pyrene

Concen: 39.93 ng

RT: 28.47 min Scan# 4363

Delta R.T. 0.03 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:276 Resp: 1359567

Ion Ratio Lower Upper

276 100

138 25.5 0.0 0.0#




277 25.5 0.0 0.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 24.82 min Scan# 3742

Delta R.T. -0.00 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:264 Resp: 510507

Ion Ratio Lower Upper

264 100

260 22.6 18.5 27.7



265 21.9 17.2 25.8




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 39.57 ng

RT: 23.82 min Scan# 3571

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:252 Resp: 1171417

Ion Ratio Lower Upper

252 100

253 21.8 19.1 28.7




125 10.3 8.9 13.3



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 38.02 ng

RT: 23.88 min Scan# 3582

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:252 Resp: 1127559

Ion Ratio Lower Upper

252 100

253 24.1 19.0 28.4



125 10.6 8.9 13.3




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 39.13 ng

RT: 24.68 min Scan# 3717

Delta R.T. 0.01 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115


Tgt Ion:252 Resp: 1102268

Ion Ratio Lower Upper

252 100

253 22.1 18.7 28.1




125 11.9 9.1 13.7



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 38.91 ng

RT: 28.53 min Scan# 4373

Delta R.T. 0.02 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03


Tgt Ion:278 Resp: 1079332

Ion Ratio Lower Upper

278 100

139 15.7 13.8 20.6



279 24.1 19.4 29.0




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E







#91

Benzo(g,h,i)perylene

Concen: 38.74 ng

RT: 29.60 min Scan# 4555

Delta R.T. 0.02 min

Lab File: BG018615.D

Acq: 10 Sep 2015 23:03

Instrument :

BNA\_G

ClientSampleId :

ICVBG091115

Tgt Ion: 276 Resp: 1094311

Ion Ratio Lower Upper

276 100

277 25.5 18.4 27.6

138 23.1 17.4 26.0

