

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091117\
 Data File : BG028649.D
 Acq On : 11 Sep 2017 19:50
 Operator : SJ/JU
 Sample : I5167-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C19013061

Manual Integrations
 APPROVED

Sohil
 9/12/2017 6:03:15 PM

Quant Time: Sep 12 05:41:09 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 07 18:39:50 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.34	152	35773	20.00	ng	0.00
21) Naphthalene-d8	11.16	136	142319	20.00	ng	0.00
38) Acenaphthene-d10	14.94	164	104996	20.00	ng	0.00
63) Phenanthrene-d10	17.69	188	275191	20.00	ng	0.00
75) Chrysene-d12	22.02	240	318334	20.00	ng	-0.01
86) Perylene-d12	25.53	264	305902	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.91	112	136576	64.28	ng	0.00
7) Phenol-d6	7.49	99	110042	35.71	ng	0.00
23) Nitrobenzene-d5	9.50	82	294523	95.02	ng	0.00
41) 2,4,6-Tribromophenol	16.44	330	258236	161.82	ng	0.00
44) 2-Fluorobiphenyl	13.56	172	748305	93.24	ng	0.00
78) Terphenyl-d14	20.28	244	1305890	88.38	ng	0.00
Target Compounds						
35) Caprolactam	12.07	113	36255	39.780	ng	98
54) Dibenzofuran	15.34	168	22010	2.264	ng	# 90
57) Fluorene	15.99	166	30912	3.479	ng	96
70) Phenanthrene	17.74	178	35589	2.432	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091117\
Data File : BG028649.D
Acq On : 11 Sep 2017 19:50
Operator : SJ/JU
Sample : I5167-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C19013061

Manual Integrations
APPROVED

Sohil
9/12/2017 6:03:15 PM

Quant Time: Sep 12 05:41:09 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 07 18:39:50 2017
Response via : Initial Calibration

