

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091117\  
 Data File : BG028660.D  
 Acq On : 12 Sep 2017 3:42  
 Operator : SJ/JU  
 Sample : I5180-04  
 Misc :  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 090717-TIDO-F-U-S

Quant Time: Sep 12 06:42:21 2017  
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Sep 07 18:39:50 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.33  | 152  | 39066    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 11.15 | 136  | 152623   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 14.94 | 164  | 108288   | 20.00  | ng    | 0.00     |
| 63) Phenanthrene-d10        | 17.69 | 188  | 296609   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 22.03 | 240  | 351444   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 25.53 | 264  | 336393   | 20.00  | ng    | -0.01    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.91  | 112  | 98684    | 42.53  | ng    | 0.00     |
| 7) Phenol-d6                | 7.49  | 99   | 79204    | 23.53  | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.50  | 82   | 295526   | 88.91  | ng    | 0.00     |
| 41) 2,4,6-Tribromophenol    | 16.43 | 330  | 166715   | 101.29 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 13.56 | 172  | 750774   | 90.70  | ng    | 0.00     |
| 78) Terphenyl-d14           | 20.28 | 244  | 1349322  | 82.72  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA G\DATA\BG091117\  
Data File : BG028660.D  
Acq On : 12 Sep 2017 3:42  
Operator : SJ/JU  
Sample : I5180-04  
Misc :  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
090717-TIDO-F-U-S

Quant Time: Sep 12 06:42:21 2017  
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG090717.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Sep 07 18:39:50 2017  
Response via : Initial Calibration

