

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG091217\
 Data File : BG028672.D
 Acq On : 12 Sep 2017 16:33
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG091217

Quant Time: Sep 12 17:31:07 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 16:57:11 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	38.265	4.3	94	0.00
3	Pyridine	40.000	41.882	-4.7	97	0.00
4	n-Nitrosodimethylamine	40.000	40.298	-0.7	94	0.00
5 S	2-Fluorophenol	80.000	81.167	-1.5	93	0.00
6	Aniline	40.000	41.842	-4.6	96	0.00
7 S	Phenol-d6	80.000	81.896	-2.4	96	0.00
8	2-Chlorophenol	40.000	41.505	-3.8	97	0.00
9	Benzaldehyde	40.000	40.997	-2.5	97	0.00
10 C	Phenol	40.000	40.995	-2.5	95	0.00
11	bis(2-Chloroethyl)ether	40.000	39.265	1.8	94	0.00
12	1,3-Dichlorobenzene	40.000	41.854	-4.6	100	0.00
13 C	1,4-Dichlorobenzene	40.000	41.133	-2.8	95	0.00
14	1,2-Dichlorobenzene	40.000	40.671	-1.7	98	0.00
15	Benzyl Alcohol	40.000	41.153	-2.9	96	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	40.429	-1.1	95	0.00
17	2-Methylphenol	40.000	41.169	-2.9	97	0.00
18	Hexachloroethane	40.000	41.406	-3.5	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.654	-6.6	100	0.00
20	3+4-Methylphenols	40.000	41.860	-4.6	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	38.719	3.2	97	0.00
23 S	Nitrobenzene-d5	80.000	77.750	2.8	98	0.00
24	Nitrobenzene	40.000	39.090	2.3	100	0.00
25	Isophorone	40.000	39.022	2.4	99	0.00
26 C	2-Nitrophenol	40.000	37.891	5.3	95	0.00
27	2,4-Dimethylphenol	40.000	38.327	4.2	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.661	3.3	97	0.00
29 C	2,4-Dichlorophenol	40.000	39.481	1.3	99	0.00
30	1,2,4-Trichlorobenzene	40.000	38.540	3.7	98	0.00
31	Naphthalene	40.000	38.555	3.6	97	0.00
32	Benzoic acid	40.000	40.876	-2.2	106	0.00
33	4-Chloroaniline	40.000	39.001	2.5	99	0.00
34 C	Hexachlorobutadiene	40.000	37.786	5.5	98	0.00
35	Caprolactam	40.000	38.902	2.7	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.359	4.1	99	0.00
37	2-Methylnaphthalene	40.000	39.304	1.7	100	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	106	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	39.187	2.0	100	0.00
40 P	Hexachlorocyclopentadiene	40.000	37.482	6.3	97	0.00
41 S	2,4,6-Tribromophenol	80.000	80.033	-0.0	105	0.00
42 C	2,4,6-Trichlorophenol	40.000	38.429	3.9	100	0.00
43	2,4,5-Trichlorophenol	40.000	38.719	3.2	101	0.00
44 S	2-Fluorobiphenyl	80.000	78.699	1.6	100	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG091217\
 Data File : BG028672.D
 Acq On : 12 Sep 2017 16:33
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG091217

Quant Time: Sep 12 17:31:07 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 12 16:57:11 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	39.723	0.7	102	0.00
46	2-Chloronaphthalene	40.000	39.167	2.1	101	0.00
47	2-Nitroaniline	40.000	39.894	0.3	101	0.00
48	Acenaphthylene	40.000	39.722	0.7	103	0.00
49	Dimethylphthalate	40.000	39.273	1.8	102	0.00
50	2,6-Dinitrotoluene	40.000	39.839	0.4	102	0.00
51 C	Acenaphthene	40.000	38.785	3.0	101	0.00
52	3-Nitroaniline	40.000	41.370	-3.4	106	0.00
53 P	2,4-Dinitrophenol	40.000	39.766	0.6	107	0.00
54	Dibenzofuran	40.000	39.028	2.4	101	0.00
55 P	4-Nitrophenol	40.000	39.086	2.3	95	0.00
56	2,4-Dinitrotoluene	40.000	40.890	-2.2	102	0.00
57	Fluorene	40.000	39.496	1.3	102	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.003	-0.0	104	0.00
59	Diethylphthalate	40.000	39.348	1.6	105	0.00
60	4-Chlorophenyl-phenylether	40.000	39.124	2.2	104	0.00
61	4-Nitroaniline	40.000	39.733	0.7	98	0.00
62	Azobenzene	40.000	39.429	1.4	104	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
64	4,6-Dinitro-2-methylphenol	40.000	41.482	-3.7	104	0.00
65 c	n-Nitrosodiphenylamine	40.000	39.757	0.6	104	0.00
66	4-Bromophenyl-phenylether	40.000	39.285	1.8	102	0.00
67	Hexachlorobenzene	40.000	38.832	2.9	102	0.00
68	Atrazine	40.000	39.329	1.7	100	0.00
69 C	Pentachlorophenol	40.000	39.635	0.9	99	0.00
70	Phenanthrene	40.000	39.925	0.2	104	0.00
71	Anthracene	40.000	39.649	0.9	102	0.00
72	Carbazole	40.000	40.322	-0.8	101	0.00
73	Di-n-butylphthalate	40.000	40.181	-0.5	103	0.00
74 C	Fluoranthene	40.000	39.797	0.5	100	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
76	Benzidine	40.000	43.024	-7.6	102	0.00
77	Pyrene	40.000	39.764	0.6	101	0.00
78 S	Terphenyl-d14	80.000	81.233	-1.5	101	0.00
79	Butylbenzylphthalate	40.000	39.941	0.1	100	0.00
80	Benzo(a)anthracene	40.000	40.423	-1.1	101	0.00
81	3,3'-Dichlorobenzidine	40.000	40.875	-2.2	100	0.00
82	Chrysene	40.000	40.263	-0.7	100	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.225	-0.6	99	0.00
84 c	Di-n-octyl phthalate	40.000	40.074	-0.2	98	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	40.240	-0.6	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Benzo(b)fluoranthene	40.000	39.779	0.6	100	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG091217\
Data File : BG028672.D
Acq On : 12 Sep 2017 16:33
Operator : SJ/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG091217

Quant Time: Sep 12 17:31:07 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 16:57:11 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.491	-1.2	101	0.00
89 C	Benzo(a)pyrene	40.000	39.914	0.2	99	0.00
90	Dibenzo(a,h)anthracene	40.000	40.472	-1.2	100	0.00
91	Benzo(a,h,i)perylene	40.000	40.411	-1.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0