

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG091219\
 Data File : BG042783.D
 Acq On : 12 Sep 2019 15:44
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 06:35:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:06:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.525	0.541	-3.0	107	0.00
3	Pyridine	1.581	1.636	-3.5	106	0.00
4	n-Nitrosodimethylamine	0.769	0.812	-5.6	109	0.00
5 S	2-Fluorophenol	1.150	1.176	-2.3	107	0.00
6	Aniline	2.228	2.227	0.0	104	0.00
7 S	Phenol-d6	1.651	1.693	-2.5	107	0.00
8	2-Chlorophenol	1.244	1.249	-0.4	107	0.00
9	Benzaldehyde	1.075	1.031	4.1	107	0.00
10 C	Phenol	1.693	1.691	0.1	105	0.00
11	bis(2-Chloroethyl)ether	1.300	1.307	-0.5	105	0.00
12	1,3-Dichlorobenzene	1.517	1.528	-0.7	104	0.00
13 C	1,4-Dichlorobenzene	1.518	1.542	-1.6	106	0.00
14	1,2-Dichlorobenzene	1.445	1.447	-0.1	106	0.00
15	Benzyl Alcohol	1.417	1.444	-1.9	106	0.00
16	2,2'-oxybis(1-Chloropropane	2.854	2.887	-1.2	106	0.00
17	2-Methylphenol	1.133	1.132	0.1	106	0.00
18	Hexachloroethane	0.561	0.563	-0.4	105	0.00
19 P	n-Nitroso-di-n-propylamine	1.217	1.222	-0.4	106	0.00
20	3+4-Methylphenols	1.562	1.560	0.1	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.505	0.500	1.0	104	0.00
23 S	Nitrobenzene-d5	0.384	0.397	-3.4	109	0.00
24	Nitrobenzene	0.403	0.415	-3.0	109	0.00
25	Isophorone	0.801	0.796	0.6	103	0.00
26 C	2-Nitrophenol	0.159	0.163	-2.5	110	0.00
27	2,4-Dimethylphenol	0.280	0.278	0.7	103	0.00
28	bis(2-Chloroethoxy)methane	0.447	0.447	0.0	105	0.00
29 C	2,4-Dichlorophenol	0.326	0.323	0.9	104	0.00
30	1,2,4-Trichlorobenzene	0.398	0.400	-0.5	105	0.00
31	Naphthalene	0.987	0.997	-1.0	105	0.00
32	Benzoic acid	0.210	0.216	-2.9	101	0.01
33	4-Chloroaniline	0.462	0.453	1.9	102	0.00
34 C	Hexachlorobutadiene	0.279	0.276	1.1	105	0.00
35	Caprolactam	0.121	0.119	1.7	101	0.00
36 C	4-Chloro-3-methylphenol	0.354	0.354	0.0	103	0.00
37	2-Methylnaphthalene	0.740	0.744	-0.5	104	0.00
38	1-Methylnaphthalene	0.694	0.700	-0.9	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.661	0.665	-0.6	105	0.00
41 P	Hexachlorocyclopentadiene	0.379	0.365	3.7	102	0.00
42 S	2,4,6-Tribromophenol	0.266	0.264	0.8	101	0.00
43 C	2,4,6-Trichlorophenol	0.440	0.441	-0.2	103	0.00
44	2,4,5-Trichlorophenol	0.451	0.453	-0.4	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG091219\
 Data File : BG042783.D
 Acq On : 12 Sep 2019 15:44
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 13 06:35:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG091219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 12 14:06:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.263	1.286	-1.8	103	0.00
46	1,1'-Biphenyl	1.400	1.402	-0.1	104	0.00
47	2-Chloronaphthalene	1.158	1.161	-0.3	104	0.00
48	2-Nitroaniline	0.402	0.421	-4.7	109	0.00
49	Acenaphthylene	1.770	1.805	-2.0	105	0.00
50	Dimethylphthalate	1.498	1.508	-0.7	103	0.00
51	2,6-Dinitrotoluene	0.307	0.313	-2.0	105	0.00
52 C	Acenaphthene	1.156	1.167	-1.0	105	0.00
53	3-Nitroaniline	0.345	0.344	0.3	102	0.00
54 P	2,4-Dinitrophenol	0.151	0.148	2.0	103	0.00
55	Dibenzofuran	1.719	1.744	-1.5	103	0.00
56 P	4-Nitrophenol	0.267	0.277	-3.7	105	0.00
57	2,4-Dinitrotoluene	0.410	0.430	-4.9	108	0.00
58	Fluorene	1.424	1.442	-1.3	104	0.00
59	2,3,4,6-Tetrachlorophenol	0.437	0.448	-2.5	104	0.00
60	Diethylphthalate	1.521	1.538	-1.1	103	0.00
61	4-Chlorophenyl-phenylether	0.842	0.847	-0.6	103	0.00
62	4-Nitroaniline	0.372	0.380	-2.2	103	0.00
63	Azobenzene	1.465	1.484	-1.3	103	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	0.086	0.086	0.0	106	0.00
66 c	n-Nitrosodiphenylamine	0.538	0.541	-0.6	104	0.00
67	4-Bromophenyl-phenylether	0.242	0.239	1.2	102	0.00
68	Hexachlorobenzene	0.259	0.253	2.3	101	0.00
69	Atrazine	0.187	0.168	10.2	101	0.00
70 C	Pentachlorophenol	0.153	0.157	-2.6	105	0.00
71	Phenanthrene	0.956	0.964	-0.8	102	0.00
72	Anthracene	0.956	0.965	-0.9	102	0.00
73	Carbazole	0.908	0.916	-0.9	101	0.00
74	Di-n-butylphthalate	1.088	1.113	-2.3	101	0.00
75 C	Fluoranthene	1.240	1.263	-1.9	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzidine	0.559	0.537	3.9	100	0.00
78	Pyrene	1.252	1.283	-2.5	102	0.00
79 S	Terphenyl-d14	0.902	0.941	-4.3	101	0.00
80	Butylbenzylphthalate	0.507	0.517	-2.0	102	0.00
81	Benzo(a)anthracene	1.272	1.283	-0.9	102	0.00
82	3,3'-Dichlorobenzidine	0.494	0.497	-0.6	104	0.00
83	Chrysene	1.205	1.223	-1.5	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.692	0.703	-1.6	104	0.00
85 c	Di-n-octyl phthalate	1.180	1.195	-1.3	103	0.00
86	Indeno(1,2,3-cd)pyrene	1.492	1.497	-0.3	105	0.00
87 I	Perylene-d12	1.000	1.000	0.0	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG091219\
Data File : BG042783.D
Acq On : 12 Sep 2019 15:44
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 13 06:35:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG091219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 12 14:06:09 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.174	1.176	-0.2	103	0.00
89	Benzo(k)fluoranthene	1.109	1.114	-0.5	104	0.00
90 C	Benzo(a)pyrene	1.100	1.092	0.7	103	0.00
91	Dibenzo(a,h)anthracene	1.082	1.081	0.1	104	0.00
92	Benzo(g,h,i)perylene	1.055	1.057	-0.2	105	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0