

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
 Data File : BG062828.D
 Acq On : 15 Sep 2024 1:54
 Operator : JU/RC
 Sample : P3993-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0UF2

Quant Time: Sep 15 02:35:16 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 01:28:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.900	152	36259	20.000	ng/u1	0.00
20) Naphthalene-d8	10.696	136	150142	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.527	164	132402	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.271	188	352367	20.000	ng/u1	0.00
79) Chrysene-d12	21.519	240	386605	20.000	ng/u1	0.00
88) Perylene-d12	24.580	264	427700	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	2338	2.915	ng/uL	0.00
4) Pyridine-d5	3.769	84	34305	13.436	ng/u1	0.00
7) Phenol-d5	7.065	99	54084	17.010	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.230	67	33088	17.358	ng/u1	0.00
11) 2-Chlorophenol-d4	7.435	132	40915	17.667	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	47096	17.775	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	20868	19.447	ng/u1	0.00
24) 2-Nitrophenol-d4	9.774	143	27749	23.645	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.314	165	51494	20.148	ng/u1	0.00
31) 4-Chloroaniline-d4	10.826	131	67539	19.225	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	227473	22.310	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	239414	21.143	ng/u1	0.00
54) 4-Nitrophenol-d4	14.727	143	31886	18.409	ng/u1	0.00
60) Fluorene-d10	15.520	176	209933	22.849	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.638	200	34308	16.926	ng/u1	0.00
73) Anthracene-d10	17.371	188	398118	23.940	ng/u1	0.00
81) Pyrene-d10	19.662	212	528006	24.270	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.369	264	563550	24.954	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
Data File : BG062828.D
Acq On : 15 Sep 2024 1:54
Operator : JU/RC
Sample : P3993-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0UF2

Quant Time: Sep 15 02:35:16 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 10 01:28:17 2024
Response via : Initial Calibration

