

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
 Data File : BG062847.D
 Acq On : 15 Sep 2024 15:37
 Operator : JU/RC
 Sample : P3952-17
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B24

Quant Time: Sep 15 23:04:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 01:28:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.902	152	40250	20.000	ng/u1	0.00
20) Naphthalene-d8	10.693	136	168934	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.530	164	146065	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.274	188	414330	20.000	ng/u1	0.00
79) Chrysene-d12	21.522	240	482070	20.000	ng/u1	0.00
88) Perylene-d12	24.583	264	539076	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.355	96	3912	4.394	ng/uL	0.00
4) Pyridine-d5	3.766	84	11825	4.172	ng/u1	0.00
7) Phenol-d5	7.068	99	23971	6.792	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.227	67	56137	26.529	ng/u1	0.00
11) 2-Chlorophenol-d4	7.432	132	66828	25.994	ng/u1	0.00
15) 4-Methylphenol-d8	8.602	113	52371	17.806	ng/u1	0.00
21) Nitrobenzene-d5	9.054	128	40496	33.540	ng/u1	0.00
24) 2-Nitrophenol-d4	9.777	143	46823	35.459	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.311	165	90783	31.569	ng/u1	0.00
31) 4-Chloroaniline-d4	10.823	131	66829	16.907	ng/u1	0.00
46) Dimethylphthalate-d6	13.931	166	360401	32.041	ng/u1	0.00
49) Acenaphthylene-d8	14.219	160	379053	30.343	ng/u1	0.00
54) 4-Nitrophenol-d4	14.730	143	13181	6.898	ng/u1	0.00
60) Fluorene-d10	15.523	176	336893	33.238	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.640	200	81904	34.366	ng/u1	0.00
73) Anthracene-d10	17.374	188	708578	36.237	ng/u1	0.00
81) Pyrene-d10	19.665	212	975601	35.964	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.377	264	1069419	37.570	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
Data File : BG062847.D
Acq On : 15 Sep 2024 15:37
Operator : JU/RC
Sample : P3952-17
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0B24

Quant Time: Sep 15 23:04:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 10 01:28:17 2024
Response via : Initial Calibration

