

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
 Data File : BG062852.D
 Acq On : 15 Sep 2024 18:52
 Operator : JU/RC
 Sample : P3997-02
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 COAN8

Quant Time: Sep 15 23:45:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 10 01:28:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.900	152	41576	20.000	ng/u1	0.00
20) Naphthalene-d8	10.691	136	167780	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.527	164	142534	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.271	188	392170	20.000	ng/u1	0.00
79) Chrysene-d12	21.519	240	421931	20.000	ng/u1	0.00
88) Perylene-d12	24.580	264	476589	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	3630	3.947	ng/uL	0.00
4) Pyridine-d5	3.769	84	18015	6.153	ng/u1	0.00
7) Phenol-d5	7.071	99	25485	6.990	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.230	67	54260	24.824	ng/u1	0.00
11) 2-Chlorophenol-d4	7.436	132	67251	25.325	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	52019	17.122	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	39552	32.983	ng/u1	0.00
24) 2-Nitrophenol-d4	9.774	143	45878	34.983	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.315	165	100656	35.243	ng/u1	0.00
31) 4-Chloroaniline-d4	10.826	131	67200	17.118	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	371428	33.840	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	381861	31.325	ng/u1	0.00
54) 4-Nitrophenol-d4	14.733	143	14320	7.680	ng/u1	0.00
60) Fluorene-d10	15.520	176	334828	33.852	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.638	200	74156	32.873	ng/u1	0.00
73) Anthracene-d10	17.371	188	628240	33.944	ng/u1	0.00
81) Pyrene-d10	19.662	212	839725	35.367	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.380	264	918198	36.487	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091324\
Data File : BG062852.D
Acq On : 15 Sep 2024 18:52
Operator : JU/RC
Sample : P3997-02
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
COAN8

Quant Time: Sep 15 23:45:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG090924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 10 01:28:17 2024
Response via : Initial Calibration

