


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG091625\
Data File : BG064381.D
Acq On : 16 Sep 2025 11:55
Operator : RC/JU
Sample : DFTPP459
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP459

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG091125.MA.M
Title : SVOA CALIBRATION
Last Update : Fri Sep 12 11:06:20 2025



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	36819	PASS
68	69	0.00	2	1.6	469	PASS
69	198	0.00	100	34.6	28480	PASS
70	69	0.00	2	0.3	78	PASS
127	198	10	80	41.9	34528	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	82365	PASS
199	198	5	9	7.6	6260	PASS
275	198	10	60	29.7	24498	PASS
365	198	1	100	6.3	5213	PASS
441	198	0.01	100	7.8	6449	PASS
442	198	50	100	93.2	76747	PASS
443	442	15	24	19.3	14789	PASS

m/z	Abundance
43.90	654.0
206.70	196.0

Instrument :

BNA_G

ClientSampleId :

DFTPP459

m/z	Abundance
37.90	344.0
39.00	2432.0
39.80	159.0
43.90	369.0
50.00	7548.0
51.00	28816.0
52.00	1662.0
55.00	297.0
55.90	488.0
56.90	2166.0
60.90	305.0
61.90	256.0
63.00	1424.0
65.20	245.0
67.90	543.0
69.00	21744.0
72.90	182.0
73.10	183.0
73.90	2805.0
75.00	4433.0
76.00	1456.0
77.00	22888.0
78.00	1459.0
78.90	2168.0
80.00	1765.0
80.90	2284.0
82.10	511.0
82.90	397.0
85.10	462.0
86.00	689.0
87.00	351.0
91.10	489.0
91.90	598.0
93.00	3231.0
94.00	348.0
95.90	267.0
98.00	2862.0
99.00	1650.0
100.10	248.0
100.90	1097.0
102.90	197.0
104.00	793.0
105.00	753.0
107.00	9587.0
108.10	1169.0
108.90	274.0
110.00	13816.0
111.00	1967.0
112.10	391.0
116.10	801.0
117.00	9335.0
118.00	836.0
122.00	702.0
123.00	1167.0
123.90	364.0
124.90	390.0
127.00	26440.0
128.10	2014.0
129.00	11956.0
130.00	910.0
134.10	444.0
135.00	954.0
136.10	417.0
137.10	522.0
138.10	162.0
141.00	1683.0
142.00	565.0
143.10	334.0
144.10	353.0
146.00	394.0
147.00	885.0
148.00	2067.0
148.90	447.0
149.10	435.0
153.00	624.0
154.00	466.0
154.90	830.0
156.10	1732.0
157.90	226.0
158.90	196.0
160.00	631.0
161.00	1041.0
162.00	219.0
164.10	186.0
164.90	1093.0
165.90	872.0
167.10	3897.0
168.00	1973.0
170.10	256.0
170.90	197.0
172.00	556.0
173.20	485.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

174.10	680.0
175.00	1920.0
176.00	300.0
176.90	684.0
178.10	151.0
179.00	2640.0
180.10	1949.0
181.00	1096.0
184.30	226.0
185.10	1394.0
186.10	9875.0
187.10	2915.0
188.00	407.0
188.90	547.0
189.80	237.0
191.00	711.0
192.10	639.0
193.10	987.0
194.20	244.0
195.00	270.0
195.90	1280.0
198.00	58464.0
199.00	4130.0
200.10	488.0
201.30	275.0
203.00	587.0
204.00	2801.0
205.00	4346.0
206.10	17304.0
207.10	2540.0
208.10	829.0
208.80	214.0
210.00	384.0
211.20	734.0
212.10	222.0
215.00	260.0
215.80	437.0
217.00	4988.0
217.90	615.0
218.90	159.0
221.00	1852.0
222.00	633.0
223.20	1219.0
224.10	10881.0
225.00	2863.0
226.00	475.0
227.00	5707.0
228.00	1138.0
228.90	861.0
230.10	225.0
231.00	449.0
232.00	196.0
234.00	303.0
234.90	422.0
236.10	367.0
237.10	550.0
239.10	208.0
240.10	188.0
241.00	427.0
242.00	530.0
244.10	7503.0
245.10	836.0
246.00	2337.0
246.80	344.0
247.80	286.0
248.40	217.0
248.80	306.0
249.80	156.0
251.10	221.0
252.90	399.0
255.00	39752.0
256.00	6619.0
256.90	365.0
257.20	353.0
258.00	2746.0
259.10	343.0
265.00	1132.0
266.20	239.0
273.00	1449.0
274.00	3563.0
275.00	15965.0
276.00	2524.0
277.00	1949.0
278.10	470.0
283.00	273.0
285.10	218.0
286.10	253.0
291.70	164.0
293.10	364.0
294.00	227.0
296.00	5912.0
296.90	680.0
302.90	675.0
314.00	423.0
315.10	697.0
316.00	455.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

322.90	1291.0
324.00	470.0
326.80	340.0
327.80	250.0
334.00	1286.0
335.00	355.0
341.10	194.0
345.80	323.0
352.00	592.0
353.00	432.0
353.90	531.0
357.20	165.0
365.00	3273.0
366.10	459.0
372.10	941.0
373.20	231.0
383.00	186.0
389.90	211.0
391.10	229.0
402.30	412.0
403.20	374.0
403.90	245.0
421.00	214.0
422.00	323.0
423.00	2304.0
423.80	457.0
424.20	497.0
441.00	6429.0
442.00	42816.0
443.00	8110.0
444.10	956.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

m/z	Abundance
37.90	483.0
39.00	2983.0
44.00	780.0
50.00	12770.0
51.00	44712.0
52.00	2132.0
54.90	238.0
55.90	1277.0
57.00	2182.0
57.80	264.0
61.00	482.0
62.00	409.0
63.00	1749.0
63.90	390.0
65.00	819.0
68.00	864.0
69.00	35752.0
69.90	235.0
73.10	557.0
74.00	4600.0
74.90	6109.0
76.10	2197.0
77.00	37144.0
78.00	2415.0
79.00	3242.0
79.90	2275.0
81.00	2810.0
81.90	722.0
82.90	690.0
84.80	418.0
86.00	988.0
87.10	487.0
88.00	264.0
88.90	257.0
91.00	790.0
92.00	798.0
93.00	5584.0
94.10	529.0
96.20	308.0
97.10	326.0
98.00	4208.0
99.00	3462.0
100.10	331.0
101.00	1958.0
102.80	568.0
104.00	1248.0
104.90	1731.0
107.00	14856.0
108.00	1962.0
110.00	20880.0
111.00	3896.0
112.00	340.0
116.10	1101.0
117.00	16033.0
118.00	1061.0
119.90	412.0
121.30	213.0
122.00	1167.0
123.00	1951.0
124.00	653.0
125.00	801.0
127.00	40040.0
128.10	3501.0
129.00	18712.0
130.00	1774.0
130.70	237.0
131.20	240.0
132.00	197.0
133.00	190.0
134.00	1130.0
135.00	1551.0
136.00	666.0
137.10	940.0
138.00	447.0
141.00	2465.0
142.00	862.0
142.80	629.0
143.10	567.0
143.90	168.0
146.00	419.0
147.00	1631.0
148.00	4134.0
149.00	701.0
150.00	294.0
150.70	269.0
151.30	385.0
152.00	268.0
153.10	961.0
154.00	902.0
155.00	1379.0
156.00	2614.0
157.00	406.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

158.10	605.0
159.00	504.0
160.00	837.0
161.10	1560.0
162.10	411.0
162.90	180.0
164.00	188.0
164.70	944.0
165.10	995.0
166.20	771.0
167.10	5825.0
168.10	2135.0
169.00	1111.0
169.80	236.0
171.10	210.0
172.20	506.0
173.00	557.0
174.10	1197.0
175.00	2140.0
175.80	598.0
176.90	1208.0
178.00	568.0
179.00	4416.0
180.10	2847.0
181.00	1001.0
181.80	315.0
182.20	220.0
182.80	486.0
183.90	383.0
185.00	2941.0
186.00	16004.0
187.10	4885.0
188.10	749.0
188.90	1006.0
190.00	307.0
190.80	515.0
192.00	2091.0
193.10	2167.0
194.10	465.0
194.80	283.0
196.00	2378.0
198.00	95808.0
199.00	7886.0
200.00	664.0
201.40	536.0
201.90	244.0
203.00	862.0
204.00	5004.0
205.00	6698.0
206.10	30976.0
207.10	4417.0
208.10	1266.0
209.00	430.0
210.00	439.0
211.00	1371.0
212.90	169.0
214.20	308.0
215.00	387.0
217.00	8808.0
217.90	1071.0
218.70	222.0
221.10	2884.0
222.10	1161.0
223.10	2093.0
224.00	18024.0
225.10	4172.0
226.10	559.0
227.00	9127.0
228.10	1631.0
229.00	1577.0
230.10	215.0
231.20	501.0
233.10	165.0
234.00	577.0
235.00	673.0
236.00	596.0
236.70	439.0
237.20	353.0
239.10	390.0
239.60	289.0
240.20	239.0
241.00	629.0
242.00	993.0
242.90	950.0
244.10	13242.0
245.10	1950.0
246.10	4159.0
246.80	403.0
247.00	413.0
249.10	479.0
251.50	278.0
252.20	207.0
253.00	567.0
253.30	546.0
255.00	71560.0
256.10	9936.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

257.00	1210.0
258.00	5231.0
259.00	792.0
260.90	224.0
265.10	2201.0
265.90	485.0
270.20	319.0
271.10	297.0
272.00	416.0
273.10	2625.0
274.10	5233.0
275.00	27912.0
276.10	3763.0
277.00	3726.0
278.00	509.0
282.00	159.0
283.00	262.0
284.10	395.0
285.20	600.0
286.20	251.0
293.20	921.0
294.00	254.0
294.60	264.0
294.80	258.0
296.00	11018.0
297.00	1956.0
301.90	486.0
303.10	1566.0
304.20	388.0
307.90	250.0
309.30	157.0
314.10	803.0
315.10	1721.0
316.10	662.0
321.00	351.0
323.00	3974.0
324.20	465.0
326.00	279.0
327.10	666.0
327.90	262.0
332.10	371.0
333.00	603.0
334.10	2094.0
335.20	511.0
340.80	331.0
341.20	360.0
341.90	153.0
342.10	154.0
346.00	538.0
346.90	176.0
352.00	1423.0
353.10	594.0
354.00	1093.0
354.80	163.0
365.00	6133.0
365.90	667.0
371.20	281.0
372.10	1543.0
372.90	295.0
382.90	615.0
390.00	223.0
391.20	374.0
400.80	161.0
402.00	697.0
403.00	726.0
404.00	573.0
421.10	852.0
422.00	903.0
423.10	5498.0
424.20	1047.0
441.00	12918.0
442.00	84264.0
443.00	16736.0
444.10	1390.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

m/z	Abundance
36.90	243.0
38.00	270.0
39.00	2976.0
44.00	459.0
48.80	429.0
50.00	10211.0
51.00	36928.0
52.00	1995.0
56.00	1221.0
57.00	2150.0
61.00	308.0
61.90	486.0
63.00	1281.0
64.10	221.0
65.00	547.0
69.00	27944.0
72.90	187.0
74.00	3990.0
75.00	4892.0
76.00	1671.0
77.00	28272.0
78.10	1702.0
79.00	2364.0
80.00	1970.0
81.00	2968.0
81.90	977.0
83.00	395.0
85.00	476.0
86.00	824.0
87.00	405.0
87.90	355.0
88.90	200.0
91.00	426.0
92.00	894.0
93.00	5061.0
94.00	236.0
98.00	4417.0
99.00	3282.0
100.00	150.0
101.10	1093.0
102.90	649.0
103.90	994.0
105.00	1206.0
105.80	597.0
107.00	12933.0
108.00	1910.0
110.00	19248.0
111.00	3548.0
112.00	469.0
112.90	206.0
116.00	756.0
117.00	12403.0
118.00	853.0
119.60	174.0
122.10	1126.0
123.00	1184.0
124.10	725.0
124.90	706.0
127.00	37104.0
128.00	3158.0
129.00	17736.0
130.00	2238.0
130.90	271.0
131.90	192.0
133.90	647.0
135.00	1236.0
136.10	794.0
137.10	997.0
137.90	266.0
138.50	200.0
140.20	435.0
141.00	2563.0
141.90	712.0
143.10	450.0
146.10	302.0
147.10	1722.0
148.00	3366.0
149.20	874.0
149.90	350.0
152.10	325.0
152.90	853.0
154.10	743.0
155.10	1389.0
156.00	2060.0
157.00	499.0
158.00	467.0
159.10	532.0
160.00	946.0
161.00	1538.0
162.00	346.0
164.20	183.0
165.00	863.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

166.10	1031.0
167.00	6600.0
168.00	2808.0
169.00	393.0
170.60	288.0
171.10	289.0
172.00	481.0
173.00	798.0
174.00	1177.0
175.10	2541.0
176.10	628.0
177.00	905.0
177.80	325.0
179.00	4310.0
180.00	2653.0
181.10	1289.0
182.00	576.0
184.00	372.0
185.20	1835.0
186.10	14744.0
187.00	4672.0
188.00	446.0
189.00	964.0
190.00	210.0
191.00	676.0
191.90	1503.0
193.10	1615.0
194.00	395.0
195.00	348.0
195.20	358.0
196.00	2241.0
198.00	92824.0
199.00	6765.0
200.00	959.0
200.80	208.0
201.30	501.0
203.00	973.0
204.10	4305.0
205.10	7868.0
206.10	30088.0
207.00	3934.0
207.80	1098.0
209.00	269.0
210.00	451.0
211.00	1416.0
215.00	602.0
215.90	1061.0
217.00	8105.0
217.90	1129.0
218.80	217.0
220.10	152.0
221.10	3370.0
221.90	1032.0
223.10	2321.0
224.10	18144.0
225.10	4576.0
226.10	802.0
227.00	10283.0
228.10	1319.0
228.90	1739.0
230.00	307.0
231.20	518.0
232.90	169.0
234.10	616.0
235.10	685.0
235.90	520.0
236.90	550.0
239.10	371.0
240.00	297.0
241.10	437.0
242.00	1122.0
243.10	1295.0
244.10	13376.0
245.10	1380.0
246.00	3967.0
246.90	727.0
248.10	207.0
248.90	748.0
250.00	211.0
252.80	498.0
255.00	73080.0
256.10	11535.0
256.90	842.0
258.10	6275.0
259.10	846.0
259.80	218.0
261.00	181.0
264.10	197.0
265.00	2161.0
266.00	419.0
267.00	262.0
271.00	401.0
272.10	280.0
273.00	2209.0
274.00	5555.0
275.00	29616.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459

276.00	3808.0
277.10	3729.0
278.20	636.0
279.00	192.0
282.80	487.0
284.00	209.0
285.10	601.0
286.00	157.0
289.10	239.0
293.10	821.0
294.30	270.0
294.90	297.0
296.10	11680.0
297.10	1632.0
302.10	216.0
303.10	1233.0
304.00	587.0
308.00	181.0
308.90	176.0
310.20	288.0
313.00	213.0
314.00	633.0
315.10	1206.0
316.00	910.0
316.90	164.0
321.00	567.0
321.80	209.0
323.10	3235.0
324.10	791.0
324.70	199.0
327.00	656.0
328.00	400.0
328.90	153.0
331.90	233.0
333.10	552.0
334.10	2568.0
335.20	478.0
336.10	153.0
341.20	426.0
346.10	788.0
346.90	236.0
352.00	1568.0
353.00	805.0
354.20	1235.0
365.00	6234.0
366.10	1101.0
366.90	158.0
370.90	296.0
372.10	2101.0
373.10	433.0
383.00	764.0
384.00	474.0
390.10	467.0
391.10	220.0
400.80	207.0
401.90	771.0
403.10	1191.0
404.00	555.0
421.10	976.0
421.90	809.0
423.10	5791.0
423.90	1234.0
426.30	217.0
427.00	203.0
435.40	202.0
436.40	195.0
436.70	165.0
437.30	202.0
442.00	103160.0
443.00	19520.0
444.00	1917.0

Instrument :
BNA_G
ClientSampleId :
DFTPP459