



# 2,4-Dinitrophenol



$R = 9.139e-003 A^2 + 8.864e-002 A - 4.112e-002$   
 Coef of Det ( $r^2$ ) = 0.992777    Curve Fit: Quadratic  
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG091922.M  
 Calibration Table Last Updated: Mon Sep 19 17:41:40 2022


# 4-Nitrophenol



$\text{Response} = 2.298\text{e-}001 * \text{Amt} - 4.770\text{e-}002$   
 Coef of Det ( $r^2$ ) = 0.994640    Curve Fit: Linear  
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG091922.M  
 Calibration Table Last Updated: Mon Sep 19 17:41:40 2022

# 2,3,4,6-Tetrachlorophenol

Response Ratio




$$\text{Response} = 3.821\text{e-}001 * \text{Amt} - 5.415\text{e-}002$$

Coef of Det ( $r^2$ ) = 0.998311 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG091922.M


Calibration Table Last Updated: Mon Sep 19 17:41:40 2022

# Pentachlorophenol



$R = 6.705e-003 A^2 + 4.671e-002 A - 1.568e-002$   
 Coef of Det ( $r^2$ ) = 0.996110    Curve Fit: Quadratic  
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG091922.M  
 Calibration Table Last Updated: Mon Sep 19 17:41:40 2022

# 4,6-Dinitro-2-methylphenol



Response =  $1.137e-001 * \text{Amt} - 2.835e-002$   
 Coef of Det ( $r^2$ ) = 0.996975    Curve Fit: Linear  
 Method Name: Z:\svoasrv\HPCHEM1\BNA G\Methods\8270-BG091922.M  
 Calibration Table Last Updated: Mon Sep 19 17:41:40 2022