

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092017\
 Data File : BG028843.D
 Acq On : 21 Sep 2017 5:09
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 21 11:57:08 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 21 11:41:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.491	0.473	3.7	100	0.00
3	Pyridine	1.400	1.425	-1.8	99	0.00
4	n-Nitrosodimethylamine	0.637	0.621	2.5	96	0.00
5 S	2-Fluorophenol	1.212	1.238	-2.1	98	0.00
6	Aniline	2.124	2.108	0.8	95	0.00
7 S	Phenol-d6	1.825	1.805	1.1	97	0.00
8	2-Chlorophenol	1.351	1.323	2.1	96	0.00
9	Benzaldehyde	1.132	1.040	8.1	91	0.00
10 C	Phenol	1.926	1.923	0.2	97	0.00
11	bis(2-Chloroethyl)ether	1.383	1.338	3.3	97	0.00
12	1,3-Dichlorobenzene	1.455	1.464	-0.6	101	0.00
13 C	1,4-Dichlorobenzene	1.496	1.515	-1.3	98	0.00
14	1,2-Dichlorobenzene	1.478	1.469	0.6	100	0.00
15	Benzyl Alcohol	1.425	1.316	7.6	90	0.00
16	2,2'-oxybis(1-Chloropropane	2.513	2.292	8.8	90	0.00
17	2-Methylphenol	1.259	1.199	4.8	95	0.00
18	Hexachloroethane	0.548	0.544	0.7	96	0.00
19 P	n-Nitroso-di-n-propylamine	1.294	1.210	6.5	92	0.00
20	3+4-Methylphenols	1.760	1.721	2.2	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
22	Acetophenone	0.585	0.573	2.1	93	0.00
23 S	Nitrobenzene-d5	0.418	0.417	0.2	95	0.00
24	Nitrobenzene	0.463	0.457	1.3	95	0.00
25	Isophorone	0.846	0.781	7.7	88	0.00
26 C	2-Nitrophenol	0.197	0.192	2.5	92	0.00
27	2,4-Dimethylphenol	0.360	0.344	4.4	90	0.00
28	bis(2-Chloroethoxy)methane	0.459	0.444	3.3	91	0.00
29 C	2,4-Dichlorophenol	0.358	0.358	0.0	94	0.00
30	1,2,4-Trichlorobenzene	0.409	0.418	-2.2	98	0.00
31	Naphthalene	1.045	1.053	-0.8	96	0.00
32	Benzoic acid	0.234	0.189	19.2	74	0.00
33	4-Chloroaniline	0.438	0.431	1.6	94	0.00
34 C	Hexachlorobutadiene	0.301	0.308	-2.3	100	0.00
35	Caprolactam	0.129	0.123	4.7	90	0.00
36 C	4-Chloro-3-methylphenol	0.466	0.429	7.9	89	0.00
37	2-Methylnaphthalene	0.780	0.763	2.2	94	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
39	1,2,4,5-Tetrachlorobenzene	0.822	0.845	-2.8	95	0.00
40 P	Hexachlorocyclopentadiene	0.264	0.230	12.9	78	0.00
41 S	2,4,6-Tribromophenol	0.331	0.340	-2.7	98	0.00
42 C	2,4,6-Trichlorophenol	0.522	0.510	2.3	93	0.00
43	2,4,5-Trichlorophenol	0.524	0.514	1.9	93	0.00
44 S	2-Fluorobiphenyl	1.500	1.504	-0.3	93	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092017\
 Data File : BG028843.D
 Acq On : 21 Sep 2017 5:09
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 21 11:57:08 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 21 11:41:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.656	1.680	-1.4	95	0.00
46	2-Chloronaphthalene	1.208	1.202	0.5	93	0.00
47	2-Nitroaniline	0.449	0.449	0.0	93	0.00
48	Acenaphthylene	1.930	1.926	0.2	94	0.00
49	Dimethylphthalate	1.677	1.636	2.4	93	0.00
50	2,6-Dinitrotoluene	0.373	0.371	0.5	93	0.00
51 C	Acenaphthene	1.172	1.166	0.5	94	0.00
52	3-Nitroaniline	0.330	0.339	-2.7	96	0.00
53 P	2,4-Dinitrophenol	0.159	0.092	42.1#	52	0.00
54	Dibenzofuran	1.869	1.863	0.3	94	0.00
55 P	4-Nitrophenol	0.242	0.211	12.8	78	0.00
56	2,4-Dinitrotoluene	0.525	0.548	-4.4	96	0.00
57	Fluorene	1.669	1.637	1.9	92	0.00
58	2,3,4,6-Tetrachlorophenol	0.462	0.442	4.3	91	0.00
59	Diethylphthalate	1.723	1.669	3.1	94	0.00
60	4-Chlorophenyl-phenylether	0.950	0.956	-0.6	97	0.00
61	4-Nitroaniline	0.350	0.360	-2.9	93	0.00
62	Azobenzene	1.450	1.429	1.4	95	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
64	4,6-Dinitro-2-methylphenol	0.129	0.118	8.5	84	0.00
65 c	n-Nitrosodiphenylamine	0.573	0.584	-1.9	97	0.00
66	4-Bromophenyl-phenylether	0.257	0.260	-1.2	96	0.00
67	Hexachlorobenzene	0.293	0.294	-0.3	96	0.00
68	Atrazine	0.246	0.250	-1.6	94	0.00
69 C	Pentachlorophenol	0.132	0.118	10.6	82	0.00
70	Phenanthrene	1.023	1.049	-2.5	98	0.00
71	Anthracene	1.033	1.075	-4.1	98	0.00
72	Carbazole	0.930	0.997	-7.2	98	0.00
73	Di-n-butylphthalate	1.141	1.171	-2.6	96	0.00
74 C	Fluoranthene	1.249	1.344	-7.6	99	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76	Benzidine	0.519	0.485	6.6	86	0.00
77	Pyrene	1.154	1.172	-1.6	100	0.00
78 S	Terphenyl-d14	0.938	0.975	-3.9	101	0.00
79	Butylbenzylphthalate	0.466	0.462	0.9	97	0.00
80	Benzo(a)anthracene	1.204	1.224	-1.7	98	0.00
81	3,3'-Dichlorobenzidine	0.488	0.483	1.0	94	0.00
82	Chrysene	1.142	1.172	-2.6	99	0.00
83	Bis(2-ethylhexyl)phthalate	0.662	0.653	1.4	94	0.00
84 c	Di-n-octyl phthalate	1.157	1.153	0.3	95	0.00
85	Indeno(1,2,3-cd)pyrene	1.468	1.440	1.9	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Benzo(b)fluoranthene	1.198	1.219	-1.8	99	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092017\
Data File : BG028843.D
Acq On : 21 Sep 2017 5:09
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 21 11:57:08 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 21 11:41:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.197	1.243	-3.8	100	0.00
89 C	Benzo(a)pyrene	1.137	1.166	-2.6	99	0.00
90	Dibenzo(a,h)anthracene	1.186	1.162	2.0	94	0.00
91	Benzo(a,h,i)perylene	1.157	1.118	3.4	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0