

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092115\
 Data File : BG018801.D
 Acq On : 22 Sep 2015 2:52
 Operator : TP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02099

Quant Time: Sep 22 03:38:20 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 22 02:06:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.450	0.450	0.0	96	0.00
3 S	1,4-Dioxane-d8	0.416	0.397	4.6	97	0.00
4	Benzaldehyde	0.950	1.110	-16.8	105	0.00
5 S	Phenol-d5	1.642	1.693	-3.1	105	0.00
6	Phenol	1.699	1.732	-1.9	105	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.952	0.976	-2.5	101	0.00
8	Bis(2-Chloroethyl)ether	1.269	1.372	-8.1	108	0.00
9 S	2-Chlorophenol-d4	1.244	1.294	-4.0	106	0.00
10	2-Chlorophenol	1.288	1.350	-4.8	103	0.00
11	2-Methylphenol	1.306	1.363	-4.4	107	0.00
12	2,2'-oxybis(1-Chloropropane	1.506	1.529	-1.5	101	0.00
13 S	4-Methylphenol-d8	1.333	1.419	-6.5	110	0.00
14	Acetophenone	2.030	2.203	-8.5	108	0.00
15 P	N-Nitroso-di-n-propylamine	1.047	1.054	-0.7	102	0.00
16	4-Methylphenol	1.436	1.492	-3.9	105	0.00
17	Hexachloroethane	0.513	0.509	0.8	106	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
19 S	Nitrobenzene-d5	0.152	0.154	-1.3	108	0.00
20	Nitrobenzene	0.351	0.341	2.8	104	0.00
21	Isophorone	0.716	0.723	-1.0	108	0.00
22 S	2-Nitrophenol-d4	0.152	0.166	-9.2	118	0.00
23 C	2-Nitrophenol	0.172	0.183	-6.4	117	0.00
24	2,4-Dimethylphenol	0.356	0.354	0.6	109	0.00
25	Bis(2-Chloroethoxy)methane	0.422	0.435	-3.1	110	0.00
26 S	2,4-Dichlorophenol-d3	0.321	0.330	-2.8	111	0.00
27 C	2,4-Dichlorophenol	0.321	0.331	-3.1	111	0.00
28	Naphthalene	1.018	1.018	0.0	107	0.00
29 S	4-Chloroaniline-d4	0.373	0.453	-21.4#	123	0.00
30	4-Chloroaniline	0.386	0.462	-19.7	119	0.00
31 C	Hexachlorobutadiene	0.198	0.202	-2.0	110	0.00
32	Caprolactam	0.131	0.145	-10.7	115	0.00
33 C	4-Chloro-3-methylphenol	0.342	0.378	-10.5	118	0.00
34	2-Methylnaphthalene	0.752	0.789	-4.9	113	0.00
35	1-Methylnaphthalene	0.721	0.746	-3.5	113	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
37	1,2,4,5-Tetrachlorobenzene	0.620	0.602	2.9	116	0.00
38	Hexachlorocyclopentadiene	0.345	0.297	13.9	105	0.00
39 C	2,4,6-Trichlorophenol	0.403	0.402	0.2	113	0.00
40	2,4,5-Trichlorophenol	0.434	0.439	-1.2	116	0.00
41	1,1'-Biphenyl	1.514	1.483	2.0	115	0.00
42	2-Chloronaphthalene	1.154	1.157	-0.3	116	0.00
43	2-Nitroaniline	0.342	0.339	0.9	117	0.00
44 S	Dimethylphthalate-d6	1.610	1.663	-3.3	122	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092115\
 Data File : BG018801.D
 Acq On : 22 Sep 2015 2:52
 Operator : TP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02099

Quant Time: Sep 22 03:38:20 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Sep 22 02:06:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.560	1.637	-4.9	122	0.00
46	2,6-Dinitrotoluene	0.314	0.352	-12.1	132	0.00
47 S	Acenaphthylene-d8	1.907	1.965	-3.0	119	0.00
48	Acenaphthylene	2.029	2.044	-0.7	116	0.00
49	3-Nitroaniline	0.333	0.379	-13.8	124	0.00
50 C	Acenaphthene	1.367	1.381	-1.0	118	0.00
51	2,4-Dinitrophenol	0.141	0.126	10.6	130	0.00
52 S	4-Nitrophenol-d4	0.309	0.309	0.0	118	0.00
53	4-Nitrophenol	0.212	0.202	4.7	112	0.00
54	Dibenzofuran	1.918	1.927	-0.5	116	0.00
55	2,4-Dinitrotoluene	0.467	0.526	-12.6	130	0.00
56	2,3,4,6-Tetrachlorophenol	0.419	0.430	-2.6	123	0.00
57	Diethylphthalate	1.622	1.669	-2.9	120	0.00
58 S	Fluorene-d10	1.407	1.475	-4.8	121	0.00
59	Fluorene	1.620	1.665	-2.8	119	0.00
60	4-Chlorophenyl-phenylether	0.772	0.798	-3.4	122	0.00
61	4-Nitroaniline	0.383	0.431	-12.5	120	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.095	0.099	-4.2	138	0.00
64	4,6-Dinitro-2-methylphenol	0.097	0.103	-6.2	138	0.00
65	N-Nitrosodiphenylamine	0.524	0.528	-0.8	119	0.00
66	4-Bromophenyl-phenylether	0.196	0.200	-2.0	124	0.00
67	Hexachlorobenzene	0.217	0.224	-3.2	127	0.00
68	Atrazine	0.231	0.241	-4.3	119	0.00
69 C	Pentachlorophenol	0.127	0.113	11.0	111	0.00
70	Phenanthrene	1.017	1.044	-2.7	119	0.00
71 S	Anthracene-d10	0.881	0.914	-3.7	123	0.00
72	Anthracene	1.020	1.050	-2.9	119	0.00
73	1,2,3,4-Tetrachlorobenzene	0.238	0.225	5.5	113	0.00
74	Pentachlorobenzene	0.232	0.241	-3.9	122	0.00
75	Carbazole	0.906	1.002	-10.6	118	0.00
76	Di-n-butylphthalate	1.144	1.193	-4.3	118	0.00
77 C	Fluoranthene	1.127	1.307	-16.0	119	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
79 S	Pyrene-d10	0.920	0.968	-5.2	120	0.00
80	Pyrene	1.176	1.236	-5.1	117	0.00
81	Butylbenzylphthalate	0.451	0.475	-5.3	121	0.00
82	3,3'-Dichlorobenzidine	0.355	0.427	-20.3#	132	0.00
83	Benzo(a)anthracene	1.109	1.133	-2.2	116	0.00
84	Bis(2-ethylhexyl)phthalate	0.669	0.698	-4.3	118	0.00
85	Chrysene	1.023	1.049	-2.5	118	0.00
86 I	Perylene-d12	1.000	1.000	0.0	115	0.00
87	Di-n-octyl phthalate	1.058	1.170	-10.6	123	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092115\
Data File : BG018801.D
Acq On : 22 Sep 2015 2:52
Operator : TP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02099

Quant Time: Sep 22 03:38:20 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Sep 22 02:06:40 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.131	1.158	-2.4	122	0.00
89	Benzo(k)fluoranthene	1.083	1.128	-4.2	120	0.00
90 S	Benzo(a)pyrene-d12	0.973	1.020	-4.8	124	0.00
91 C	Benzo(a)pyrene	1.075	1.105	-2.8	121	0.00
92	Indeno(1,2,3-cd)pyrene	1.246	1.309	-5.1	125	0.00
93	Dibenzo(a,h)anthracene	1.000	1.070	-7.0	129	0.00
94	Benzo(a,h,i)perylene	1.040	1.087	-4.5	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0