


Data Path : Z:\HPCHEM1\BNA G\DATA\BG092115\
Data File : BG018775.D
Acq On : 21 Sep 2015 10:23
Operator : TP
Sample : DFTPP48
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP48

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
Title : SVOA CALIBRATION
Last Update : Tue Sep 22 03:49:40 2015



AutoFind: Scans 2547, 2548, 2549; Background Corrected with Scan 2541

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.0	17880	PASS
68	69	0.00	2	1.2	251	PASS
69	198	0.00	100	30.4	20900	PASS
70	69	0.00	2	1.2	251	PASS
127	198	10	80	41.2	28290	PASS
197	198	0.00	2	0.2	123	PASS
198	198	100	100	100.0	68685	PASS
199	198	5	9	6.5	4472	PASS
275	198	10	60	27.9	19162	PASS
365	198	1	100	3.6	2496	PASS
441	442	0.01	24	17.9	9374	PASS
442	198	50	100	76.2	52333	PASS
443	442	15	24	18.1	9460	PASS

m/z	Abundance
37.90	243.0
39.00	1523.0
39.90	788.0
41.00	431.0
42.90	543.0
44.00	870.0
49.10	373.0
50.00	5056.0
51.00	21424.0
52.00	1168.0
55.00	296.0
56.00	611.0
57.00	1973.0
58.10	209.0
60.00	324.0
60.80	224.0
61.90	380.0
63.00	905.0
64.00	182.0
64.90	458.0
67.90	333.0
68.90	22936.0
70.00	199.0
71.00	199.0
73.00	766.0
74.00	2085.0
75.00	4573.0
76.00	1476.0
77.00	29264.0
78.00	1904.0
79.00	1414.0
79.90	1113.0
80.90	2065.0
81.80	329.0
83.00	626.0
83.90	278.0
84.90	425.0
85.90	639.0
86.80	241.0
88.00	170.0
90.90	571.0
91.90	533.0
93.00	3107.0
94.00	317.0
95.90	297.0
97.10	208.0
97.90	2882.0
98.90	1972.0
100.00	495.0
100.90	1020.0
102.00	176.0
103.00	533.0
103.90	526.0
105.00	1015.0
105.90	479.0
106.90	8294.0
107.90	1296.0
109.00	189.0
109.90	15340.0
111.00	2973.0
111.90	286.0
115.10	166.0
116.00	473.0
116.90	6428.0
118.00	569.0
119.00	291.0
121.90	540.0
122.90	1406.0
124.00	403.0
125.00	455.0
127.00	30696.0
128.00	2086.0
129.00	12466.0
129.90	1001.0
130.90	349.0
133.90	388.0
134.90	840.0
135.90	435.0
137.00	477.0
137.80	238.0
139.90	265.0
140.90	1945.0
141.80	541.0
143.00	321.0
146.00	256.0
147.00	759.0
147.90	1822.0
148.90	351.0
151.00	214.0
151.80	192.0
152.90	498.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

154.00	924.0
155.00	1418.0
156.00	1395.0
157.00	356.0
157.90	529.0
159.90	425.0
160.90	643.0
161.90	266.0
164.90	406.0
166.00	535.0
167.00	3292.0
168.00	1400.0
168.90	357.0
171.80	503.0
172.90	492.0
174.00	653.0
175.00	1585.0
175.90	413.0
177.00	619.0
177.90	214.0
178.90	2711.0
180.00	2091.0
181.00	725.0
181.90	157.0
183.90	278.0
185.00	1065.0
186.00	9618.0
187.00	2257.0
187.90	246.0
189.00	589.0
191.00	474.0
192.00	1028.0
193.00	1025.0
194.00	359.0
195.90	2692.0
196.90	369.0
197.90	69296.0
198.90	4799.0
199.80	349.0
201.40	507.0
202.90	445.0
203.90	2905.0
205.00	4185.0
206.00	19304.0
207.00	2286.0
207.90	865.0
208.90	344.0
210.00	394.0
210.30	289.0
211.10	492.0
215.90	368.0
216.90	5048.0
218.00	667.0
218.90	184.0
220.00	168.0
220.90	2293.0
221.60	664.0
222.90	1338.0
224.00	10720.0
225.00	2653.0
225.90	373.0
226.90	3627.0
228.00	764.0
229.00	816.0
230.00	182.0
231.00	246.0
233.90	319.0
234.90	367.0
236.00	226.0
236.90	378.0
239.00	166.0
239.80	205.0
241.00	276.0
242.00	547.0
243.00	431.0
244.00	8814.0
245.00	1211.0
246.00	1468.0
246.90	219.0
247.90	151.0
249.00	311.0
253.00	256.0
253.90	168.0
255.00	41440.0
255.90	7019.0
256.90	727.0
258.00	2300.0
259.00	564.0
265.00	799.0
265.80	224.0
271.90	173.0
273.00	1411.0
274.00	3173.0
275.00	18496.0
276.00	1840.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

276.90	1763.0
277.90	219.0
278.90	151.0
280.90	334.0
282.90	345.0
283.80	153.0
285.00	220.0
293.00	303.0
294.10	214.0
296.00	5342.0
296.90	837.0
303.00	591.0
303.90	464.0
314.00	280.0
315.00	569.0
316.00	371.0
323.00	1814.0
324.00	365.0
327.00	267.0
332.00	194.0
334.00	902.0
340.90	317.0
345.80	210.0
352.00	461.0
353.00	292.0
354.00	289.0
357.00	208.0
365.00	2358.0
366.00	410.0
371.00	232.0
372.00	827.0
373.10	192.0
382.90	251.0
389.90	157.0
401.00	210.0
402.00	485.0
403.00	412.0
421.10	307.0
421.90	376.0
423.10	2386.0
424.10	517.0
441.10	7770.0
442.00	42160.0
443.00	8127.0
444.10	824.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	179.0
39.00	1623.0
40.00	703.0
41.00	307.0
43.00	491.0
44.00	1048.0
45.00	191.0
49.10	431.0
50.00	4831.0
51.00	19864.0
52.00	1039.0
55.00	755.0
55.90	780.0
56.90	2034.0
58.00	156.0
60.00	321.0
60.90	607.0
61.90	378.0
63.00	1168.0
65.00	495.0
66.90	205.0
68.10	480.0
68.90	23840.0
70.00	298.0
71.00	172.0
73.00	654.0
74.00	2271.0
74.90	3986.0
76.00	1480.0
77.00	29856.0
78.00	1600.0
79.00	1870.0
79.90	1498.0
81.00	1882.0
82.00	530.0
83.00	553.0
83.90	282.0
85.00	354.0
85.90	545.0
87.00	270.0
91.00	507.0
91.90	660.0
93.00	2919.0
93.90	285.0
95.00	226.0
96.10	195.0
96.90	163.0
97.90	2386.0
99.00	1785.0
100.90	1196.0
102.90	388.0
103.90	704.0
105.00	874.0
106.00	282.0
107.00	9454.0
108.00	1266.0
109.90	17568.0
110.90	2905.0
111.90	378.0
113.00	185.0
116.00	725.0
117.00	6669.0
117.90	365.0
122.00	438.0
122.90	1184.0
123.90	463.0
125.00	600.0
127.00	32744.0
128.00	2608.0
128.90	11605.0
129.90	1163.0
130.80	301.0
131.90	170.0
132.80	150.0
133.90	501.0
135.00	950.0
135.90	561.0
136.80	399.0
137.80	182.0
138.80	254.0
140.90	1352.0
141.90	891.0
142.90	593.0
145.00	275.0
146.00	214.0
147.00	912.0
147.90	1735.0
149.00	386.0
150.90	374.0
151.90	202.0
152.90	481.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

153.90	405.0
155.00	1176.0
155.90	1333.0
157.00	243.0
158.00	369.0
158.90	447.0
160.00	497.0
160.90	582.0
162.00	203.0
162.90	229.0
164.90	808.0
165.90	648.0
167.00	3252.0
168.00	1869.0
169.00	412.0
169.90	161.0
171.10	234.0
171.90	313.0
173.00	539.0
174.00	1029.0
175.00	1555.0
176.00	486.0
176.90	676.0
177.90	291.0
178.90	3631.0
180.00	2251.0
180.90	944.0
181.90	299.0
184.00	300.0
184.90	1548.0
186.00	11354.0
187.00	3491.0
189.00	969.0
192.00	865.0
193.00	1523.0
193.90	325.0
195.10	213.0
195.90	2555.0
196.60	503.0
197.90	80560.0
198.90	4960.0
200.00	486.0
201.50	502.0
203.00	740.0
204.00	3102.0
205.00	4979.0
206.00	21392.0
207.00	3106.0
208.00	878.0
208.90	358.0
211.00	678.0
211.60	336.0
214.80	205.0
215.90	412.0
216.90	6024.0
217.90	744.0
218.80	196.0
220.90	3605.0
221.70	719.0
222.90	1252.0
224.00	13154.0
224.90	2713.0
226.00	256.0
227.00	4952.0
227.80	961.0
228.90	996.0
231.00	461.0
233.90	359.0
234.90	365.0
236.00	301.0
237.00	419.0
238.90	283.0
239.90	193.0
241.00	413.0
242.00	795.0
243.00	836.0
244.00	9823.0
245.00	1493.0
245.90	1511.0
246.90	394.0
249.00	273.0
250.00	330.0
252.10	250.0
252.90	306.0
255.00	51216.0
256.00	7656.0
257.00	505.0
257.90	3044.0
258.90	358.0
259.90	166.0
265.00	1170.0
266.10	167.0
273.00	1629.0
274.00	3714.0
275.00	21984.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

276.00	2653.0
276.90	1816.0
277.80	339.0
282.90	310.0
283.90	235.0
285.00	552.0
290.90	157.0
292.90	335.0
296.00	6405.0
297.10	830.0
303.00	807.0
304.00	156.0
309.90	150.0
314.00	354.0
314.90	933.0
316.10	201.0
320.90	357.0
322.10	190.0
323.00	2584.0
324.00	484.0
326.90	518.0
328.00	243.0
333.10	212.0
334.00	1115.0
334.90	294.0
340.80	307.0
341.90	196.0
345.90	425.0
351.10	155.0
351.90	289.0
353.00	286.0
353.90	508.0
365.00	3038.0
365.90	263.0
370.90	229.0
372.00	912.0
372.80	240.0
373.10	287.0
382.90	274.0
389.90	163.0
402.00	402.0
403.00	575.0
404.10	287.0
421.10	479.0
421.90	367.0
423.00	3151.0
424.00	517.0
438.50	182.0
440.00	250.0
441.00	11057.0
442.00	61696.0
443.10	10627.0
444.00	1246.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
37.90	159.0
39.00	1427.0
40.00	495.0
41.00	360.0
43.00	699.0
44.00	1139.0
45.00	199.0
49.00	202.0
50.00	3023.0
51.00	12352.0
52.00	542.0
55.00	477.0
56.00	563.0
57.00	1331.0
59.90	234.0
61.00	374.0
61.90	410.0
63.00	412.0
65.00	255.0
68.00	274.0
69.00	15924.0
70.00	258.0
72.90	580.0
73.90	1353.0
75.00	2847.0
76.00	758.0
77.00	17680.0
78.00	1205.0
78.90	1243.0
79.90	866.0
81.00	1488.0
81.90	318.0
83.00	467.0
83.90	207.0
84.90	400.0
85.90	216.0
90.90	379.0
91.90	356.0
92.90	2222.0
94.10	246.0
95.00	173.0
96.00	197.0
97.00	267.0
97.90	1711.0
98.90	1456.0
99.90	200.0
100.90	934.0
101.90	157.0
102.90	309.0
104.00	564.0
105.00	713.0
106.00	208.0
107.00	6550.0
108.00	805.0
108.80	365.0
109.90	10449.0
110.90	1848.0
112.00	283.0
113.00	201.0
115.90	522.0
117.00	4494.0
117.80	186.0
120.00	238.0
121.90	282.0
122.90	672.0
124.00	330.0
124.80	287.0
126.10	200.0
127.00	21432.0
128.00	1495.0
128.90	7884.0
129.90	589.0
130.90	331.0
133.90	553.0
135.00	846.0
135.90	305.0
137.00	338.0
137.80	207.0
140.90	1283.0
141.90	313.0
142.90	294.0
145.90	266.0
147.00	540.0
147.90	1694.0
148.90	185.0
150.00	206.0
151.00	297.0
152.90	337.0
154.90	489.0
156.00	1055.0
157.00	276.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

157.90	344.0
159.00	337.0
159.90	609.0
160.90	657.0
165.00	521.0
165.90	507.0
167.00	2428.0
168.00	1423.0
169.00	241.0
171.00	236.0
172.00	199.0
173.00	295.0
174.00	569.0
174.90	941.0
176.00	237.0
176.60	189.0
177.00	288.0
178.90	2534.0
180.00	941.0
181.00	504.0
183.80	233.0
185.00	758.0
186.00	7322.0
187.00	2347.0
189.00	518.0
190.30	164.0
191.00	422.0
191.90	840.0
192.90	1070.0
194.10	251.0
195.90	1347.0
197.90	56200.0
198.90	3659.0
199.90	420.0
201.40	347.0
203.00	676.0
204.00	2451.0
204.90	3608.0
206.00	15200.0
207.00	2642.0
207.90	652.0
209.00	295.0
209.90	332.0
211.00	659.0
215.00	348.0
216.00	297.0
216.90	3495.0
217.80	497.0
220.90	2653.0
222.00	534.0
222.90	1313.0
224.00	8712.0
225.00	2193.0
227.00	3731.0
227.90	469.0
228.90	669.0
231.00	356.0
233.80	317.0
235.00	410.0
235.80	292.0
239.80	239.0
240.80	188.0
241.10	215.0
241.90	543.0
243.00	400.0
244.00	7164.0
244.90	964.0
245.90	1534.0
246.90	324.0
249.00	377.0
250.00	193.0
253.00	244.0
253.80	193.0
255.00	37976.0
256.00	5555.0
256.90	284.0
258.00	1928.0
258.80	350.0
264.90	585.0
265.90	204.0
273.00	1308.0
274.00	2738.0
275.00	17008.0
276.00	2667.0
277.00	969.0
277.90	231.0
281.90	170.0
282.90	407.0
285.00	389.0
290.80	156.0
292.90	391.0
293.90	167.0
296.00	4845.0
296.90	724.0
303.00	797.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

304.00	274.0
309.90	172.0
313.90	369.0
314.90	474.0
315.90	521.0
321.00	264.0
322.00	171.0
323.00	1633.0
324.00	418.0
326.90	437.0
333.10	152.0
334.00	977.0
341.10	214.0
345.80	508.0
351.90	496.0
353.00	493.0
354.00	449.0
364.90	2093.0
366.00	204.0
372.00	808.0
373.00	534.0
382.90	294.0
391.20	154.0
402.00	539.0
402.90	500.0
421.00	385.0
421.90	504.0
423.00	2840.0
424.00	549.0
439.20	188.0
441.10	9296.0
442.00	53144.0
443.00	9627.0
444.00	943.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48

m/z	Abundance
39.90	583.0
41.10	409.0
43.00	543.0
44.00	989.0
49.00	192.0
55.00	415.0
57.00	431.0
57.80	178.0
59.90	167.0
67.00	233.0
71.20	295.0
73.00	395.0
83.90	169.0
85.90	242.0
91.00	251.0
92.90	288.0
104.80	172.0
129.00	206.0
133.90	169.0
207.10	269.0
271.50	285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP48