



Data Path : Z:\HPCHEM1\BNA G\DATA\BG092117\
Data File : BG028860.D
Acq On : 21 Sep 2017 16:21
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 22 07:13:12 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 16:57:11 2017
Response via : Initial Calibration



TIC: BG028860.D

(69) Pentachlorophenol (C)



17.527min (+0.178) 0.00ng

response 558

Ion	Exp%	Act%
265.70	100	100
268.00	60.70	57.38
264.00	62.60	70.49
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG092117\
Data File : BG028860.D
Acq On : 21 Sep 2017 16:21
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Sep 22 07:19:54 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG091217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 12 16:57:11 2017
Response via : Initial Calibration



TIC: BG028860.D

(76) Benzidine

19.877min (-0.033) 0.00ng

response 228081

Ion	Exp%	Act%
184.00	100	100
185.00	16.30	14.84
183.00	13.80	10.90#
0.00	0.00	0.00