


Data Path : Z:\HPCHEM1\BNA G\Data\BG092117\
 Data File : BG028860.D
 Acq On : 21 Sep 2017 16:21
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG091217.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Sep 12 16:57:11 2017



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.4	12816	PASS
68	69	0.00	2	1.2	161	PASS
69	198	0.00	100	29.8	13456	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	34.9	15767	PASS
197	198	0.00	2	0.7	329	PASS
198	198	100	100	100.0	45205	PASS
199	198	5	9	6.6	2963	PASS
275	198	10	60	27.8	12575	PASS
365	198	1	100	3.7	1671	PASS
441	442	0.01	24	15.8	7038	PASS
442	198	50	100	98.4	44485	PASS
443	442	15	24	19.1	8515	PASS