


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092215\
Data File : BG018817.D
Acq On : 22 Sep 2015 15:29
Operator : TP
Sample : DFTPP49
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP49

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
Title : SVOA CALIBRATION
Last Update : Wed Sep 23 07:13:57 2015



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2541

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.0	12953	PASS
68	69	0.00	2	0.6	85	PASS
69	198	0.00	100	29.9	14928	PASS
70	69	0.00	2	1.0	143	PASS
127	198	10	80	41.7	20831	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	49904	PASS
199	198	5	9	6.4	3192	PASS
275	198	10	60	27.0	13497	PASS
365	198	1	100	3.8	1881	PASS
441	442	0.01	24	17.6	6969	PASS
442	198	50	100	79.4	39640	PASS
443	442	15	24	19.7	7804	PASS

Scan 2541 (17.767 min): BG018817.D
DFTPP49

m/z	Abundance
40.00	445.0
44.00	397.0
81.10	195.0
160.20	158.0
179.90	185.0
207.00	164.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

m/z	Abundance
36.90	198.0
38.00	239.0
39.00	1419.0
39.90	503.0
44.00	207.0
49.00	343.0
50.00	3644.0
51.00	15004.0
52.00	853.0
55.90	540.0
56.90	917.0
57.90	190.0
61.00	316.0
62.00	164.0
63.00	955.0
63.80	156.0
65.00	402.0
67.80	179.0
68.90	16992.0
69.90	223.0
72.90	242.0
74.00	1126.0
75.00	2949.0
76.00	1157.0
77.00	22120.0
78.00	1435.0
78.90	1085.0
80.00	1064.0
80.90	1627.0
82.00	690.0
83.00	263.0
84.90	237.0
85.80	287.0
87.10	308.0
91.00	458.0
91.90	495.0
92.90	2733.0
93.90	264.0
97.00	178.0
97.90	1663.0
98.90	1480.0
101.00	1077.0
102.90	259.0
103.90	260.0
105.00	448.0
107.00	5634.0
107.90	920.0
108.80	215.0
110.00	11036.0
111.00	1860.0
112.00	338.0
116.00	338.0
116.90	4441.0
117.80	300.0
122.00	401.0
122.90	689.0
124.00	283.0
124.90	393.0
126.00	167.0
127.00	24616.0
128.00	1501.0
129.00	8042.0
129.90	876.0
133.90	664.0
135.00	858.0
135.90	284.0
137.00	401.0
139.90	211.0
140.90	857.0
141.90	428.0
142.90	370.0
145.90	294.0
146.90	725.0
147.90	1010.0
149.00	385.0
151.00	272.0
152.90	604.0
154.00	317.0
154.90	907.0
156.00	1067.0
157.00	181.0
158.00	297.0
160.00	448.0
160.90	652.0
161.90	186.0
163.80	159.0
165.00	628.0
165.90	557.0
166.90	2986.0
168.00	1373.0
168.90	252.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

171.90	227.0
172.90	572.0
174.00	617.0
175.00	1014.0
175.90	317.0
177.00	584.0
178.90	2038.0
179.90	1586.0
181.00	660.0
183.90	229.0
185.00	1018.0
186.00	7470.0
187.00	1845.0
188.80	484.0
190.90	181.0
191.20	173.0
191.90	615.0
193.00	805.0
194.00	303.0
194.60	211.0
195.90	782.0
197.90	52864.0
198.90	3476.0
199.80	333.0
202.90	511.0
204.00	2258.0
204.90	3463.0
206.00	14810.0
207.00	2104.0
208.00	456.0
208.90	181.0
209.90	318.0
210.90	548.0
216.90	3681.0
217.90	656.0
220.90	2395.0
222.90	797.0
224.00	8062.0
225.00	1728.0
226.00	347.0
226.90	3277.0
227.90	391.0
228.90	692.0
231.00	225.0
234.10	169.0
235.00	256.0
236.90	151.0
240.90	273.0
242.00	526.0
242.80	501.0
244.00	6886.0
245.00	1012.0
245.90	1413.0
246.90	186.0
252.00	246.0
255.00	30096.0
255.90	4902.0
257.00	423.0
257.90	1510.0
258.80	424.0
264.90	983.0
273.00	941.0
274.00	2232.0
275.00	12230.0
276.00	1562.0
277.00	1131.0
282.90	198.0
284.90	183.0
288.90	197.0
293.00	255.0
296.00	4209.0
296.80	444.0
298.10	152.0
299.50	154.0
303.00	679.0
314.90	394.0
316.10	159.0
320.90	208.0
323.00	1515.0
323.90	337.0
327.00	349.0
328.00	223.0
334.00	872.0
335.00	158.0
341.00	174.0
345.90	381.0
351.90	541.0
352.90	416.0
354.00	433.0
364.90	1542.0
366.00	302.0
372.00	489.0
383.00	236.0
402.10	378.0
403.00	372.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

420.80	185.0
421.90	224.0
423.00	1717.0
423.90	234.0
441.00	5543.0
442.00	31392.0
443.00	6103.0
444.00	650.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

m/z	Abundance
37.90	174.0
39.00	884.0
39.90	460.0
44.00	383.0
48.90	212.0
50.00	3495.0
51.00	15410.0
52.00	707.0
56.00	458.0
56.90	1006.0
61.00	166.0
63.00	720.0
64.00	200.0
64.90	286.0
69.00	17416.0
69.90	206.0
72.90	180.0
74.00	1992.0
74.90	2599.0
76.00	1284.0
77.00	22048.0
78.00	2104.0
79.00	1191.0
79.90	961.0
80.90	1314.0
81.90	350.0
83.00	387.0
85.00	272.0
86.00	626.0
87.00	278.0
91.00	448.0
92.00	272.0
92.90	2317.0
93.70	153.0
93.90	160.0
95.00	157.0
96.00	176.0
97.90	1759.0
98.90	1839.0
99.90	208.0
101.00	1279.0
103.00	287.0
104.00	518.0
104.90	406.0
106.10	188.0
107.00	6255.0
107.90	1119.0
109.90	12738.0
111.00	1880.0
111.90	260.0
116.00	495.0
116.90	6090.0
117.90	442.0
121.90	722.0
122.90	707.0
123.90	363.0
124.90	483.0
127.00	24488.0
128.00	2632.0
128.90	9881.0
129.90	728.0
133.90	361.0
135.00	780.0
135.90	279.0
137.00	299.0
138.00	184.0
140.00	272.0
140.90	1104.0
141.90	446.0
142.70	174.0
143.00	174.0
146.90	606.0
148.00	1359.0
149.00	547.0
151.60	237.0
152.90	446.0
154.00	389.0
155.00	645.0
156.00	1002.0
158.00	258.0
159.90	519.0
161.00	610.0
161.90	229.0
162.80	180.0
164.00	181.0
165.00	578.0
165.90	495.0
166.90	2492.0
168.00	1389.0
169.00	328.0
171.10	168.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

172.00	322.0
172.80	254.0
174.00	644.0
175.00	1370.0
175.90	242.0
177.00	387.0
178.00	199.0
178.90	2472.0
180.00	1435.0
180.90	611.0
185.00	794.0
186.00	7898.0
187.00	1932.0
188.00	344.0
189.00	386.0
190.80	218.0
191.90	530.0
193.00	950.0
196.00	1651.0
197.90	59832.0
198.90	3904.0
199.90	416.0
201.40	253.0
202.90	630.0
204.00	2696.0
205.00	3552.0
206.00	15517.0
207.00	1927.0
207.90	734.0
209.00	170.0
209.90	188.0
210.20	224.0
210.90	761.0
216.00	223.0
216.90	4915.0
217.90	635.0
221.00	2347.0
222.90	1392.0
224.00	8966.0
225.00	1917.0
225.80	256.0
226.90	3924.0
228.00	542.0
229.00	727.0
231.00	317.0
234.00	258.0
234.80	205.0
237.00	505.0
239.00	165.0
240.90	496.0
241.90	461.0
243.00	565.0
244.00	7509.0
245.00	844.0
245.90	1382.0
246.90	344.0
248.90	280.0
252.00	392.0
253.00	243.0
255.00	38016.0
256.00	5487.0
257.00	592.0
257.90	2224.0
258.90	319.0
260.70	152.0
265.00	1038.0
272.90	1374.0
273.90	2773.0
275.00	15691.0
276.00	1755.0
277.00	1508.0
282.00	160.0
282.90	261.0
285.00	332.0
289.70	169.0
293.00	278.0
296.00	5084.0
296.90	777.0
303.00	522.0
303.90	242.0
307.80	166.0
314.00	442.0
315.00	497.0
316.00	271.0
323.00	2006.0
324.00	466.0
326.90	181.0
327.90	169.0
332.00	212.0
332.90	274.0
333.90	1180.0
334.90	194.0
346.00	410.0
351.90	465.0
353.00	547.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

353.90	586.0
355.10	197.0
365.00	2521.0
366.10	199.0
371.00	207.0
372.00	1319.0
373.10	311.0
382.90	151.0
389.90	151.0
401.90	518.0
402.90	461.0
403.80	207.0
420.90	461.0
422.10	411.0
423.00	2143.0
424.00	457.0
438.70	154.0
441.00	8636.0
442.00	48504.0
443.00	9717.0
444.10	1011.0
449.70	151.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

m/z	Abundance
36.90	177.0
39.00	673.0
40.00	575.0
43.90	210.0
49.00	211.0
50.00	2241.0
51.00	8445.0
52.00	395.0
55.00	217.0
56.00	203.0
57.00	593.0
61.90	229.0
63.00	541.0
65.00	288.0
68.10	255.0
68.90	10376.0
74.00	1061.0
75.00	1809.0
76.10	766.0
77.00	12736.0
78.00	904.0
78.90	978.0
79.90	689.0
81.00	1076.0
81.90	224.0
82.90	258.0
84.80	204.0
85.90	253.0
91.00	175.0
92.00	328.0
92.90	1408.0
98.00	865.0
98.90	879.0
99.90	152.0
100.90	489.0
104.00	387.0
104.90	240.0
107.00	3915.0
108.00	696.0
110.00	7206.0
110.90	914.0
115.90	246.0
117.00	3279.0
118.00	450.0
121.80	232.0
123.00	464.0
123.90	232.0
125.00	154.0
127.00	13391.0
128.00	1137.0
129.00	5773.0
129.90	325.0
134.00	226.0
135.00	494.0
136.00	330.0
136.90	314.0
137.60	161.0
140.90	699.0
141.90	168.0
146.00	163.0
147.00	479.0
147.90	916.0
149.00	238.0
152.80	255.0
153.90	197.0
154.90	593.0
155.90	817.0
157.90	268.0
158.80	231.0
159.90	215.0
161.00	473.0
166.10	162.0
166.90	1868.0
167.90	1017.0
169.00	226.0
172.90	201.0
174.10	216.0
175.00	632.0
175.90	235.0
176.90	194.0
178.00	158.0
179.00	1342.0
179.90	644.0
180.90	705.0
185.00	601.0
186.00	4715.0
187.00	1320.0
189.00	422.0
191.90	363.0
193.00	407.0
195.90	888.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49

196.60	287.0
197.90	37016.0
198.90	2196.0
199.90	194.0
201.50	185.0
203.00	298.0
204.00	1411.0
205.00	2224.0
206.00	10839.0
207.00	1454.0
208.00	276.0
208.90	186.0
209.90	232.0
211.10	344.0
216.90	3447.0
217.90	269.0
218.60	152.0
219.90	161.0
220.90	1255.0
221.90	353.0
222.90	709.0
224.00	5946.0
225.00	1711.0
225.80	156.0
226.90	2532.0
228.00	253.0
228.90	379.0
231.00	262.0
235.00	170.0
237.00	177.0
242.00	372.0
243.00	379.0
244.00	4987.0
245.00	888.0
245.90	1171.0
246.90	412.0
253.20	240.0
254.90	25560.0
255.90	3502.0
256.90	543.0
258.00	1738.0
258.90	183.0
265.00	412.0
270.80	185.0
273.00	822.0
274.00	2135.0
275.00	12570.0
276.00	1503.0
277.00	868.0
278.00	225.0
284.10	153.0
285.00	210.0
295.90	3426.0
297.00	562.0
303.00	755.0
314.00	300.0
314.90	462.0
316.00	270.0
321.00	254.0
323.00	1011.0
323.90	385.0
327.00	323.0
332.80	186.0
334.10	1145.0
335.10	187.0
341.10	160.0
346.00	269.0
347.00	170.0
352.00	638.0
353.00	178.0
354.00	453.0
365.00	1581.0
365.70	198.0
372.00	774.0
373.00	197.0
389.90	239.0
403.00	211.0
404.00	161.0
421.00	316.0
422.10	463.0
423.00	2157.0
424.00	275.0
439.50	154.0
441.00	6730.0
442.00	39024.0
443.00	7593.0
444.10	494.0

Instrument :
BNA_G
ClientSampleId :
DFTPP49