


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092222\
 Data File : BG055006.D
 Acq On : 22 Sep 2022 11:19
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Sep 19 17:41:40 2022






AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.1	18755	PASS
68	69	0.00	2	1.3	263	PASS
69	198	0.00	100	27.8	19960	PASS
70	69	0.00	2	0.3	58	PASS
127	198	10	80	40.7	29224	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	71848	PASS
199	198	5	9	7.1	5093	PASS
275	198	10	60	28.8	20726	PASS
365	198	1	100	3.7	2688	PASS
441	198	0.01	100	13.8	9922	PASS
442	442	50	100	100.0	60227	PASS
443	442	15	24	20.2	12186	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092222\
Data File : BG055006.D
Acq On : 22 Sep 2022 11:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Sep 23 04:12:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration



TIC: BG055006.D\data.ms

(70) Pentachlorophenol (C)

17.110min (+ 0.001) 3109.79 ng




response 59443

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.00	67.12
264.00	61.70	65.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092222\
Data File : BG055006.D
Acq On : 22 Sep 2022 11:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Sep 23 04:12:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration



TIC: BG055006.D\data.ms

(77) Benzidine

19.677min (-0.005) 0.00 ng

response 532613

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.68
183.00	12.30	12.18
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/22/2022	BNA_G	BG055006.D
Compound Name	Response	Retention Time
DDT	259679	20.899
DDD	10769	20.459
DDE	57	19.959
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10826	270505	4.00

Instrument :
BNA_G
ClientSampleId :
DFTPP