

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG092316\  
 Data File : BG024105.D  
 Acq On : 24 Sep 2016 2:26  
 Operator : UM/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

Quant Time: Sep 24 05:11:14 2016  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG092316.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 23 15:35:28 2016  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 8.04  | 152  | 59086    | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8        | 10.87 | 136  | 261574   | 20.00 | ng    | 0.00     |
| 38) Acenaphthene-d10      | 14.71 | 164  | 209962   | 20.00 | ng    | 0.00     |
| 63) Phenanthrene-d10      | 17.48 | 188  | 440951   | 20.00 | ng    | 0.00     |
| 75) Chrysene-d12          | 21.79 | 240  | 433607   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12          | 25.12 | 264  | 432499   | 20.00 | ng    | -0.02    |

## System Monitoring Compounds

|                          |       |     |         |       |    |      |
|--------------------------|-------|-----|---------|-------|----|------|
| 5) 2-Fluorophenol        | 5.59  | 112 | 290705  | 85.24 | ng | 0.00 |
| 7) Phenol-d6             | 7.21  | 99  | 435959  | 75.97 | ng | 0.00 |
| 23) Nitrobenzene-d5      | 9.22  | 82  | 541686  | 81.62 | ng | 0.00 |
| 41) 2,4,6-Tribromophenol | 16.22 | 330 | 223123  | 73.93 | ng | 0.00 |
| 44) 2-Fluorobiphenyl     | 13.33 | 172 | 1039011 | 83.18 | ng | 0.00 |
| 78) Terphenyl-d14        | 20.09 | 244 | 1400016 | 66.11 | ng | 0.00 |

## Target Compounds

|                                |       |     |        |       |    | Qvalue |
|--------------------------------|-------|-----|--------|-------|----|--------|
| 2) 1,4-Dioxane                 | 3.44  | 88  | 55866  | 42.80 | ng | 94     |
| 3) Pyridine                    | 3.85  | 79  | 194652 | 42.19 | ng | 99     |
| 4) n-Nitrosodimethylamine      | 3.77  | 42  | 89845  | 38.42 | ng | # 94   |
| 6) Aniline                     | 7.37  | 93  | 301639 | 37.42 | ng | 97     |
| 8) 2-Chlorophenol              | 7.61  | 128 | 158938 | 38.93 | ng | 94     |
| 9) Benzaldehyde                | 7.18  | 77  | 127932 | 40.94 | ng | 95     |
| 10) Phenol                     | 7.24  | 94  | 230839 | 38.26 | ng | 97     |
| 11) bis(2-Chloroethyl)ether    | 7.45  | 93  | 183585 | 38.59 | ng | 85     |
| 12) 1,3-Dichlorobenzene        | 7.92  | 146 | 178951 | 39.31 | ng | 95     |
| 13) 1,4-Dichlorobenzene        | 8.08  | 146 | 186902 | 40.22 | ng | 98     |
| 14) 1,2-Dichlorobenzene        | 8.40  | 146 | 176361 | 39.00 | ng | 91     |
| 15) Benzyl Alcohol             | 8.29  | 79  | 197742 | 35.60 | ng | 99     |
| 16) 2,2'-oxybis(1-Chloropropan | 8.57  | 45  | 251126 | 38.23 | ng | 92     |
| 17) 2-Methylphenol             | 8.50  | 107 | 150284 | 37.47 | ng | 97     |
| 18) Hexachloroethane           | 9.12  | 117 | 77454  | 39.57 | ng | 97     |
| 19) n-Nitroso-di-n-propylamine | 8.85  | 70  | 195665 | 34.71 | ng | 96     |
| 20) 3+4-Methylphenols          | 8.83  | 107 | 213022 | 36.57 | ng | 99     |
| 22) Acetophenone               | 8.88  | 105 | 310872 | 41.20 | ng | # 98   |
| 24) Nitrobenzene               | 9.26  | 77  | 295055 | 41.47 | ng | # 97   |
| 25) Isophorone                 | 9.79  | 82  | 529584 | 39.12 | ng | 97     |
| 26) 2-Nitrophenol              | 9.99  | 139 | 107109 | 42.45 | ng | # 93   |
| 27) 2,4-Dimethylphenol         | 10.04 | 122 | 170418 | 40.39 | ng | 99     |
| 28) bis(2-Chloroethoxy)methane | 10.27 | 93  | 267173 | 40.59 | ng | 97     |
| 29) 2,4-Dichlorophenol         | 10.53 | 162 | 182095 | 41.46 | ng | 99     |
| 30) 1,2,4-Trichlorobenzene     | 10.73 | 180 | 193793 | 41.23 | ng | 95     |
| 31) Naphthalene                | 10.92 | 128 | 539775 | 40.74 | ng | 99     |
| 32) Benzoic acid               | 10.23 | 122 | 125637 | 39.13 | ng | 95     |
| 33) 4-Chloroaniline            | 11.04 | 127 | 229582 | 38.82 | ng | 97     |
| 34) Hexachlorobutadiene        | 11.19 | 225 | 147561 | 39.88 | ng | 95     |
| 35) Caprolactam                | 11.84 | 113 | 68271  | 38.22 | ng | # 80   |
| 36) 4-Chloro-3-methylphenol    | 12.18 | 107 | 242073 | 39.17 | ng | 98     |
| 37) 2-Methylnaphthalene        | 12.53 | 142 | 402359 | 39.66 | ng | 94     |
| 39) 1,2,4,5-Tetrachlorobenzene | 12.90 | 216 | 252809 | 41.24 | ng | 98     |
| 40) Hexachlorocyclopentadiene  | 12.87 | 237 | 98649  | 41.17 | ng | 92     |

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG092316\  
 Data File : BG024105.D  
 Acq On : 24 Sep 2016 2:26  
 Operator : UM/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1


Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

Quant Time: Sep 24 05:11:14 2016  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG092316.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 23 15:35:28 2016  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 42) 2,4,6-Trichlorophenol      | 13.15 | 196  | 166142   | 41.91 | ng    | 93       |
| 43) 2,4,5-Trichlorophenol      | 13.24 | 196  | 167921   | 41.17 | ng    | 93       |
| 45) 1,1'-Biphenyl              | 13.54 | 154  | 607814   | 42.33 | ng    | 98       |
| 46) 2-Chloronaphthalene        | 13.59 | 162  | 442646   | 41.29 | ng    | 97       |
| 47) 2-Nitroaniline             | 13.81 | 65   | 200909   | 40.28 | ng    | 98       |
| 48) Acenaphthylene             | 14.44 | 152  | 750865   | 41.26 | ng    | 99       |
| 49) Dimethylphthalate          | 14.16 | 163  | 635162   | 39.43 | ng    | 100      |
| 50) 2,6-Dinitrotoluene         | 14.30 | 165  | 137321   | 40.58 | ng    | 98       |
| 51) Acenaphthene               | 14.78 | 154  | 445987   | 40.41 | ng    | 96       |
| 52) 3-Nitroaniline             | 14.64 | 138  | 135793   | 40.18 | ng    | 92       |
| 53) 2,4-Dinitrophenol          | 14.87 | 184  | 75404    | 37.52 | ng    | 94       |
| 54) Dibenzofuran               | 15.11 | 168  | 687421   | 40.00 | ng    | 97       |
| 55) 4-Nitrophenol              | 14.98 | 139  | 88768    | 36.22 | ng    | 90       |
| 56) 2,4-Dinitrotoluene         | 15.10 | 165  | 205963   | 40.00 | ng    | 90       |
| 57) Fluorene                   | 15.77 | 166  | 597768   | 39.96 | ng    | 96       |
| 58) 2,3,4,6-Tetrachlorophenol  | 15.36 | 232  | 155257   | 39.41 | ng    | 90       |
| 59) Diethylphthalate           | 15.53 | 149  | 669959   | 38.62 | ng    | 98       |
| 60) 4-Chlorophenyl-phenylether | 15.75 | 204  | 311003   | 38.81 | ng    | 98       |
| 61) 4-Nitroaniline             | 15.81 | 138  | 125541   | 35.87 | ng    | 89       |
| 62) Azobenzene                 | 16.05 | 77   | 691176   | 39.22 | ng    | 93       |
| 64) 4,6-Dinitro-2-methylphenol | 15.87 | 198  | 125911   | 41.12 | ng    | 76       |
| 65) n-Nitrosodiphenylamine     | 15.98 | 169  | 540647   | 44.82 | ng    | 95       |
| 66) 4-Bromophenyl-phenylether  | 16.65 | 248  | 218064   | 42.63 | ng    | 98       |
| 67) Hexachlorobenzene          | 16.78 | 284  | 243716   | 42.29 | ng    | 96       |
| 68) Atrazine                   | 16.93 | 200  | 206354   | 38.75 | ng    | 97       |
| 69) Pentachlorophenol          | 17.14 | 266  | 101749   | 36.11 | ng    | 98       |
| 70) Phenanthrene               | 17.52 | 178  | 933955   | 41.58 | ng    | 98       |
| 71) Anthracene                 | 17.62 | 178  | 928453   | 40.24 | ng    | 98       |
| 72) Carbazole                  | 17.89 | 167  | 818081   | 38.09 | ng    | 96       |
| 73) Di-n-butylphthalate        | 18.43 | 149  | 956224   | 34.01 | ng    | 97       |
| 74) Fluoranthene               | 19.54 | 202  | 996596   | 33.99 | ng    | 94       |
| 76) Benzidine                  | 19.72 | 184  | 448549   | 36.54 | ng    | 96       |
| 77) Pyrene                     | 19.90 | 202  | 1006740  | 33.25 | ng    | 94       |
| 79) Butylbenzylphthalate       | 20.77 | 149  | 446377   | 40.04 | ng    | 97       |
| 80) Benzo(a)anthracene         | 21.76 | 228  | 999207   | 40.48 | ng    | 95       |
| 81) 3,3'-Dichlorobenzidine     | 21.68 | 252  | 393413   | 44.47 | ng    | # 95     |
| 82) Chrysene                   | 21.84 | 228  | 959745   | 41.36 | ng    | 99       |
| 83) Bis(2-ethylhexyl)phthalate | 21.64 | 149  | 638097   | 47.90 | ng    | 94       |
| 84) Di-n-octyl phthalate       | 22.88 | 149  | 1115925  | 48.55 | ng    | 99       |
| 85) Indeno(1,2,3-cd)pyrene     | 28.94 | 276  | 1194193  | 46.15 | ng    | # 100    |
| 87) Benzo(b)fluoranthene       | 24.06 | 252  | 1019649  | 40.52 | ng    | # 99     |
| 88) Benzo(k)fluoranthene       | 24.13 | 252  | 1025059  | 41.02 | ng    | # 97     |
| 89) Benzo(a)pyrene             | 24.97 | 252  | 993388   | 40.84 | ng    | # 95     |
| 90) Dibenzo(a,h)anthracene     | 29.00 | 278  | 965882   | 39.77 | ng    | # 95     |
| 91) Benzo(g,h,i)perylene       | 30.15 | 276  | 978990   | 40.01 | ng    | # 93     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.04 min Scan# 886

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:152 Resp: 59086

Ion Ratio Lower Upper

152 100

150 157.2 144.7 217.1




115 58.5 64.0 96.0#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 42.80 ng

RT: 3.44 min Scan# 103

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion: 88 Resp: 55866

Ion Ratio Lower Upper

88 100

58 80.0 60.4 90.6



43 34.0 31.1 46.7



Abundance Ion 88.00 (87.70 to 88.70): BGC

Ion 58.00 (57.70 to 58.70): BGC

Ion 43.00 (42.70 to 43.70): BGC





#3

Pyridine

Concen: 42.19 ng

RT: 3.85 min Scan# 172

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 79 Resp: 194652

Ion Ratio Lower Upper

79 100

52 64.4 51.8 77.8

51 40.1 32.7 49.1



#4

n-Nitrosodimethylamine

Concen: 38.42 ng

RT: 3.77 min Scan# 159

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26




Tgt Ion: 42 Resp: 89845


Ion Ratio Lower Upper

42 100

74 130.6 100.6 150.8

44 21.7 4.5 6.7#





#5

2-Fluorophenol

Concen: 85.24 ng

RT: 5.59 min Scan# 468

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 112 Resp: 290705

Ion Ratio Lower Upper

112 100

64 78.4 59.0 88.4




63 41.1 37.8 56.8



Abundance Ion 112.00 (111.70 to 112.70): BGC

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 37.42 ng

RT: 7.37 min Scan# 771

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion: 93 Resp: 301639

Ion Ratio Lower Upper

93 100

66 48.0 37.5 56.3



65 24.7 17.9 26.9




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 75.97 ng

RT: 7.21 min Scan# 744

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 99 Resp: 435959

Ion Ratio Lower Upper

99 100

42 22.0 17.8 26.6

71 49.6 41.5 62.3



#8

2-Chlorophenol

Concen: 38.93 ng

RT: 7.61 min Scan# 812

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26




Tgt Ion: 128 Resp: 158938


Ion Ratio Lower Upper

128 100

130 31.6 12.9 52.9

64 68.1 42.0 82.0





#9

Benzaldehyde

Concen: 40.94 ng

RT: 7.18 min Scan# 739

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 77 Resp: 127932

Ion Ratio Lower Upper

77 100

105 77.0 58.7 98.7

106 80.9 67.5 107.5



#10

Phenol

Concen: 38.26 ng

RT: 7.24 min Scan# 749

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26




Tgt Ion: 94 Resp: 230839


Ion Ratio Lower Upper

94 100

65 37.9 18.1 58.1

66 58.0 42.1 82.1









#11  
bis(2-Chloroethyl)ether  
Concen: 38.59 ng  
RT: 7.45 min Scan# 786  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 63      | 93.1  | 55.0  | 95.0  |
| 95      | 30.9  | 11.0  | 51.0  |



#12  
1,3-Dichlorobenzene  
Concen: 39.31 ng  
RT: 7.92 min Scan# 866  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 68.9  | 50.3  | 75.5  |
| 75      | 46.6  | 37.0  | 55.6  |









#13  
 1,4-Dichlorobenzene  
 Concen: 40.22 ng  
 RT: 8.08 min Scan# 892  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 65.6  | 54.5  | 81.7  |
| 111     | 48.2  | 37.8  | 56.8  |



#14  
 1,2-Dichlorobenzene  
 Concen: 39.00 ng  
 RT: 8.40 min Scan# 947  
 Delta R.T. 0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 62.1  | 52.9  | 79.3  |
| 111     | 44.6  | 43.5  | 65.3  |





#15

Benzyl Alcohol

Concen: 35.60 ng

RT: 8.29 min Scan# 928

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 79 Resp: 197742

Ion Ratio Lower Upper




79 100

108 58.2 46.5 69.7

77 66.3 52.3 78.5



Abundance Ion 79.00 (78.70 to 79.70): BGC  
Ion 108.00 (107.70 to 108.70): BGC  
Ion 77.00 (76.70 to 77.70): BGC



#16

2,2'-oxybis(1-Chloropropane)

Concen: 38.23 ng

RT: 8.57 min Scan# 976

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion: 45 Resp: 251126

Ion Ratio Lower Upper



45 100


77 16.6 0.6 40.6

79 13.3 0.0 36.2



Abundance Ion 45.00 (44.70 to 45.70): BGC  
Ion 77.00 (76.70 to 77.70): BGC  
Ion 79.00 (78.70 to 79.70): BGC









#17  
2-Methylphenol  
Concen: 37.47 ng  
RT: 8.50 min Scan# 964  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 107     | 100   |       |       |
| 108     | 110.8 | 92.0  | 138.0 |
| 77      | 71.0  | 55.8  | 83.6  |
| 79      | 67.2  | 55.0  | 82.6  |




Abundance Ion 107.00 (106.70 to 107.70): E  
Ion 108.00 (107.70 to 108.70): E  
Ion 77.00 (76.70 to 77.70): BG  
Ion 79.00 (78.70 to 79.70): BG






#18  
Hexachloroethane  
Concen: 39.57 ng  
RT: 9.12 min Scan# 1070  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 119     | 87.3  | 73.7  | 110.5 |
| 201     | 110.3 | 88.7  | 133.1 |



Abundance Ion 117.00 (116.70 to 117.70): E  
Ion 119.00 (118.70 to 119.70): E  
Ion 201.00 (200.70 to 201.70): E









#19  
n-Nitroso-di-n-propylamine  
Concen: 34.71 ng  
RT: 8.85 min Scan# 1024  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 70    | Resp: | 195665 |
| Ion      | Ratio | Lower | Upper  |
| 70       | 100   |       |        |
| 42       | 55.8  | 42.1  | 63.1   |
| 101      | 9.5   | 7.2   | 10.8   |
| 130      | 16.2  | 14.2  | 21.2   |




Abundance Ion 70.00 (69.70 to 70.70): BGC  
Ion 42.00 (41.70 to 42.70): BGC  
Ion 101.00 (100.70 to 101.70): BGC  
Ion 130.00 (129.70 to 130.70): BGC






#20  
3+4-Methylphenols  
Concen: 36.57 ng  
RT: 8.83 min Scan# 1021  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 213022 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 93.2  | 72.5  | 112.5  |
| 77       | 47.9  | 30.1  | 70.1   |
| 79       | 43.7  | 24.2  | 64.2   |



Abundance Ion 107.00 (106.70 to 107.70): BGC  
Ion 108.00 (107.70 to 108.70): BGC  
Ion 77.00 (76.70 to 77.70): BGC  
Ion 79.00 (78.70 to 79.70): BGC









#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 10.87 min Scan# 1368  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 136   | Resp: | 261574 |
| Ion      | Ratio | Lower | Upper  |
| 136      | 100   |       |        |
| 137      | 12.4  | 9.6   | 14.4   |
| 54       | 9.6   | 7.3   | 10.9   |
| 68       | 6.2   | 5.3   | 7.9    |



#22  
Acetophenone  
Concen: 41.20 ng  
RT: 8.88 min Scan# 1028  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 105   | Resp: | 310872 |
| Ion      | Ratio | Lower | Upper  |
| 105      | 100   |       |        |
| 71       | 0.9   | 2.6   | 3.8#   |
| 51       | 32.4  | 27.1  | 40.7   |
| 120      | 19.1  | 15.1  | 22.7   |









#23  
 Nitrobenzene-d5  
 Concen: 81.62 ng  
 RT: 9.22 min Scan# 1087  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 82      | 100   |       |       |
| 128     | 30.5  | 24.4  | 36.6  |
| 54      | 51.2  | 41.4  | 62.0  |



#24  
 Nitrobenzene  
 Concen: 41.47 ng  
 RT: 9.26 min Scan# 1094  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 77      | 100   |       |       |
| 123     | 31.7  | 25.0  | 37.6  |
| 65      | 13.3  | 13.4  | 20.0  |





#25

Isophorone

Concen: 39.12 ng

RT: 9.79 min Scan# 1183

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion: 82 Resp: 529584

Ion Ratio Lower Upper

82 100

95 9.0 8.2 12.2

138 14.5 10.7 16.1



#26

2-Nitrophenol

Concen: 42.45 ng

RT: 9.99 min Scan# 1217

Delta R.T. 0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26




Tgt Ion: 139 Resp: 107109

Ion Ratio Lower Upper

139 100

109 42.5 43.5 65.3#

65 79.6 64.3 96.5









#27  
2,4-Dimethylphenol  
Concen: 40.39 ng  
RT: 10.04 min Scan# 1227  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 122     | 100   |       |       |
| 107     | 146.1 | 117.0 | 175.4 |
| 121     | 60.8  | 50.1  | 75.1  |



#28  
bis(2-Chloroethoxy)methane  
Concen: 40.59 ng  
RT: 10.27 min Scan# 1265  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 95      | 34.0  | 25.4  | 38.2  |
| 123     | 9.9   | 8.2   | 12.2  |









#29  
 2,4-Dichlorophenol  
 Concen: 41.46 ng  
 RT: 10.53 min Scan# 1310  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 162     | 100   |       |       |
| 164     | 64.7  | 44.3  | 84.3  |
| 98      | 40.6  | 19.5  | 59.5  |



#30  
 1,2,4-Trichlorobenzene  
 Concen: 41.23 ng  
 RT: 10.73 min Scan# 1344  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 100   |       |       |
| 182     | 87.6  | 73.8  | 110.8 |
| 145     | 28.1  | 24.4  | 36.6  |









#31  
Naphthalene  
Concen: 40.74 ng  
RT: 10.92 min Scan# 1376  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 128     | 100   |       |       |
| 129     | 11.8  | 9.4   | 14.0  |
| 127     | 14.8  | 11.3  | 16.9  |



#32  
Benzoic acid  
Concen: 39.13 ng  
RT: 10.23 min Scan# 1259  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 122     | 100   |       |       |
| 105     | 144.7 | 117.2 | 157.2 |
| 77      | 125.4 | 109.2 | 149.2 |









#33  
4-Chloroaniline  
Concen: 38.82 ng  
RT: 11.04 min Scan# 1397  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 127   | Resp: | 229582 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 35.3  | 27.9  | 41.9   |
| 65       | 49.0  | 36.8  | 55.2   |
| 92       | 25.2  | 21.0  | 31.6   |




Abundance Ion 127.00 (126.70 to 127.70): E  
Ion 129.00 (128.70 to 129.70): E  
Ion 65.00 (64.70 to 65.70): BGC  
Ion 92.00 (91.70 to 92.70): BGC






#34  
Hexachlorobutadiene  
Concen: 39.88 ng  
RT: 11.19 min Scan# 1422  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 147561 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 65.6  | 49.6  | 74.4   |
| 227      | 63.2  | 54.5  | 81.7   |



Abundance Ion 225.00 (224.70 to 225.70): E  
Ion 223.00 (222.70 to 223.70): E  
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 38.22 ng

RT: 11.84 min Scan# 1532

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:113 Resp: 68271

Ion Ratio Lower Upper

113 100

55 201.6 154.5 194.5#




56 170.7 125.1 165.1#



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#36

4-Chloro-3-methylphenol

Concen: 39.17 ng

RT: 12.18 min Scan# 1590

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:107 Resp: 242073

Ion Ratio Lower Upper

107 100

144 22.6 17.0 25.6



142 68.2 53.0 79.6




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37  
 2-Methylnaphthalene  
 Concen: 39.66 ng  
 RT: 12.53 min Scan# 1650  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 142     | 100   |       |       |
| 141     | 91.4  | 70.2  | 105.2 |
| 115     | 45.2  | 41.7  | 62.5  |



#38  
 Acenaphthene-d10  
 Concen: 20.00 ng  
 RT: 14.71 min Scan# 2022  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 164     | 100   |       |       |
| 162     | 101.6 | 87.7  | 131.5 |
| 160     | 42.8  | 37.2  | 55.8  |





#39

1,2,4,5-Tetrachlorobenzene

Concen: 41.24 ng

RT: 12.90 min Scan# 1713

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC

Tgt Ion:216 Resp: 252809


Ion Ratio Lower Upper

216 100

214 81.4 62.9 94.3

179 22.0 17.2 25.8

108 20.2 16.3 24.5




Abundance Ion 216.00 (215.70 to 216.70): E


Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



Time--> 12.80 12.90 13.00



#40

Hexachlorocyclopentadiene

Concen: 41.17 ng

RT: 12.87 min Scan# 1708

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:237 Resp: 98649

Ion Ratio Lower Upper

237 100

235 56.0 43.1 83.1


272 9.7 0.0 30.9




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E




Time--> 12.80 12.85 12.90






#41  
 2,4,6-Tribromophenol  
 Concen: 73.93 ng  
 RT: 16.22 min Scan# 2278  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 330     | 100   |       |       |
| 332     | 97.5  | 77.4  | 116.2 |
| 141     | 42.8  | 31.7  | 47.5  |




Abundance Ion 329.80 (329.50 to 330.50): E  
 Ion 331.80 (331.50 to 332.50): E  
 Ion 141.00 (140.70 to 141.70): E






#42  
 2,4,6-Trichlorophenol  
 Concen: 41.91 ng  
 RT: 13.15 min Scan# 1756  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 107.1 | 78.2  | 117.2 |
| 200     | 34.3  | 27.0  | 40.4  |



Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 200.00 (199.70 to 200.70): E









#43  
 2,4,5-Trichlorophenol  
 Concen: 41.17 ng  
 RT: 13.24 min Scan# 1771  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

Tgt Ion:196 Resp: 167921  
 Ion Ratio Lower Upper  
 196 100  
 198 100.0 85.7 128.5  
 97 63.2 45.5 68.3  
 132 26.0 18.9 28.3




Abundance Ion 196.00 (195.70 to 196.70): E  
 Ion 198.00 (197.70 to 198.70): E  
 Ion 97.00 (96.70 to 97.70): BG  
 Ion 132.00 (131.70 to 132.70): E






#44  
 2-Fluorobiphenyl  
 Concen: 83.18 ng  
 RT: 13.33 min Scan# 1786  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

Tgt Ion:172 Resp: 1039011  
 Ion Ratio Lower Upper  
 172 100  
 171 39.0 31.6 47.4  
 170 26.9 21.3 31.9



Abundance Ion 172.00 (171.70 to 172.70): E  
 Ion 171.00 (170.70 to 171.70): E  
 Ion 170.00 (169.70 to 170.70): E









#45  
 1,1'-Biphenyl  
 Concen: 42.33 ng  
 RT: 13.54 min Scan# 1822  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 154     | 100   |       |       |
| 153     | 41.6  | 20.2  | 60.2  |
| 76      | 13.9  | 0.0   | 32.9  |



#46  
 2-Chloronaphthalene  
 Concen: 41.29 ng  
 RT: 13.59 min Scan# 1831  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 162     | 100   |       |       |
| 127     | 41.6  | 32.4  | 48.6  |
| 164     | 34.0  | 25.2  | 37.8  |





#47

2-Nitroaniline

Concen: 40.28 ng

RT: 13.81 min Scan# 1868

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion: 65 Resp: 200909

Ion Ratio Lower Upper




65 100

92 59.5 49.4 74.0

138 73.7 58.0 87.0



Abundance Ion 65.00 (64.70 to 65.70): BGC  
Ion 92.00 (91.70 to 92.70): BGC  
Ion 138.00 (137.70 to 138.70): BGC



#48

Acenaphthylene

Concen: 41.26 ng

RT: 14.44 min Scan# 1975

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion: 152 Resp: 750865

Ion Ratio Lower Upper



152 100


151 21.0 17.6 26.4

153 14.2 11.4 17.2



Abundance Ion 152.00 (151.70 to 152.70): BGC  
Ion 151.00 (150.70 to 151.70): BGC  
Ion 153.00 (152.70 to 153.70): BGC





#49

Dimethylphthalate

Concen: 39.43 ng

RT: 14.16 min Scan# 1928

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:163 Resp: 635162

Ion Ratio Lower Upper

163 100

194 4.5 3.6 5.4




164 10.1 8.1 12.1



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 40.58 ng

RT: 14.30 min Scan# 1951

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:165 Resp: 137321

Ion Ratio Lower Upper

165 100

63 84.1 64.3 96.5



89 80.7 64.3 96.5




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC





#51

Acenaphthene

Concen: 40.41 ng

RT: 14.78 min Scan# 2033

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:154 Resp: 445987

Ion Ratio Lower Upper

154 100

153 114.6 87.5 131.3


152 54.2 42.7 64.1





Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



Time-->



#52

3-Nitroaniline

Concen: 40.18 ng

RT: 14.64 min Scan# 2010

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:138 Resp: 135793

Ion Ratio Lower Upper

138 100

108 12.9 11.7 17.5


92 151.4 129.7 194.5




Abundance Ion 138.00 (137.70 to 138.70): E


Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BG



Time-->









#53  
2,4-Dinitrophenol  
Concen: 37.52 ng  
RT: 14.87 min Scan# 2048  
Delta R.T. 0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:184 Resp: 75404  
Ion Ratio Lower Upper  
184 100  
63 81.3 59.6 89.4  
154 80.0 67.4 101.0



#54  
Dibenzofuran  
Concen: 40.00 ng  
RT: 15.11 min Scan# 2090  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion:168 Resp: 687421  
Ion Ratio Lower Upper  
168 100  
139 43.6 36.4 54.6  
169 13.7 11.9 17.9









#55  
4-Nitrophenol  
Concen: 36.22 ng  
RT: 14.98 min Scan# 2067  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:139 Resp: 88768  
Ion Ratio Lower Upper  
139 100  
109 115.9 103.0 143.0  
65 148.8 112.7 152.7



#56  
2,4-Dinitrotoluene  
Concen: 40.00 ng  
RT: 15.10 min Scan# 2087  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion:165 Resp: 205963  
Ion Ratio Lower Upper  
165 100  
63 65.8 45.8 68.6  
89 93.9 68.6 102.8  
182 2.5 2.0 3.0





#57

Fluorene

Concen: 39.96 ng

RT: 15.77 min Scan# 2202

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion:166 Resp: 597768

Ion Ratio Lower Upper

166 100

165 96.7 81.0 121.6

167 13.4 12.0 18.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 39.41 ng

RT: 15.36 min Scan# 2131

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Tgt Ion:232 Resp: 155257




Ion Ratio Lower Upper


232 100

131 47.3 32.7 49.1

130 3.0 2.6 3.8

166 34.1 23.0 34.6





#59

Diethylphthalate

Concen: 38.62 ng

RT: 15.53 min Scan# 2160

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 669959

Ion Ratio Lower Upper

149 100

177 23.3 19.2 28.8




150 14.0 10.6 16.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#60

4-Chlorophenyl-phenylether

Concen: 38.81 ng

RT: 15.75 min Scan# 2199

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:204 Resp: 311003

Ion Ratio Lower Upper

204 100

206 36.6 28.5 42.7



141 66.4 51.4 77.0




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61  
4-Nitroaniline  
Concen: 35.87 ng  
RT: 15.81 min Scan# 2209  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion: 138 Resp: 125541  
Ion Ratio Lower Upper  
138 100  
92 71.2 54.1 94.1  
108 135.8 133.9 173.9



#62  
Azobenzene  
Concen: 39.22 ng  
RT: 16.05 min Scan# 2249  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion: 77 Resp: 691176  
Ion Ratio Lower Upper  
77 100  
182 21.1 3.5 43.5  
105 16.5 0.0 38.5  
51 31.1 17.1 57.1





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.48 min Scan# 2493

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:188 Resp: 440951

Ion Ratio Lower Upper

188 100

94 7.2 5.2 7.8




80 9.2 7.2 10.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#64

4,6-Dinitro-2-methylphenol

Concen: 41.12 ng

RT: 15.87 min Scan# 2219

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:198 Resp: 125911

Ion Ratio Lower Upper

198 100

51 60.3 26.0 66.0



105 57.2 20.1 60.1




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BGC

Ion 105.00 (104.70 to 105.70): E









#65  
 n-Nitrosodiphenylamine  
 Concen: 44.82 ng  
 RT: 15.98 min Scan# 2237  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 169     | 100  |       |       |
| 168     | 72.2 | 55.0  | 82.4  |
| 167     | 37.5 | 27.4  | 41.0  |



#66  
 4-Bromophenyl-phenylether  
 Concen: 42.63 ng  
 RT: 16.65 min Scan# 2352  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Resp | Lower | Upper |
|---------|------|-------|-------|
| 248     | 100  |       |       |
| 250     | 95.2 | 77.8  | 116.8 |
| 141     | 81.1 | 65.7  | 98.5  |









#67  
Hexachlorobenzene  
Concen: 42.29 ng  
RT: 16.78 min Scan# 2374  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 36.4  | 26.8  | 40.2  |
| 249     | 38.2  | 28.9  | 43.3  |




Abundance Ion 283.80 (283.50 to 284.50): E  
Ion 142.00 (141.70 to 142.70): E  
Ion 249.00 (248.70 to 249.70): E






#68  
Atrazine  
Concen: 38.75 ng  
RT: 16.93 min Scan# 2399  
Delta R.T. -0.00 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 200     | 100   |       |       |
| 173     | 25.4  | 1.6   | 41.6  |
| 215     | 49.1  | 28.7  | 68.7  |



Abundance Ion 200.00 (199.70 to 200.70): E  
Ion 173.00 (172.70 to 173.70): E  
Ion 215.00 (214.70 to 215.70): E





#69

Pentachlorophenol

Concen: 36.11 ng

RT: 17.14 min Scan# 2435

Delta R.T. -0.00 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:266 Resp: 101749

Ion Ratio Lower Upper

266 100

268 64.9 51.9 77.9




264 59.9 50.6 75.8



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70

Phenanthrene

Concen: 41.58 ng

RT: 17.52 min Scan# 2500

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:178 Resp: 933955

Ion Ratio Lower Upper

178 100

176 21.4 16.8 25.2



179 16.3 14.3 21.5




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#71  
 Anthracene  
 Concen: 40.24 ng  
 RT: 17.62 min Scan# 2516  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC


Tgt Ion:178 Resp: 928453  
 Ion Ratio Lower Upper  
 178 100  
 176 20.8 17.3 25.9  
 179 16.6 14.0 21.0



#72  
 Carbazole  
 Concen: 38.09 ng  
 RT: 17.89 min Scan# 2563  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

Tgt Ion:167 Resp: 818081  
 Ion Ratio Lower Upper  
 167 100  
 166 23.5 20.7 31.1  
 139 15.6 11.9 17.9









#73  
Di-n-butylphthalate  
Concen: 34.01 ng  
RT: 18.43 min Scan# 2654  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:149 Resp: 956224  
Ion Ratio Lower Upper  
149 100  
150 10.0 9.1 13.7  
104 7.6 6.9 10.3



#74  
Fluoranthene  
Concen: 33.99 ng  
RT: 19.54 min Scan# 2843  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion:202 Resp: 996596  
Ion Ratio Lower Upper  
202 100  
101 8.7 0.0 27.4  
203 18.2 1.7 41.7





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.79 min Scan# 3226

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:240 Resp: 433607

Ion Ratio Lower Upper

240 100

120 6.3 4.1 6.1#




236 27.9 21.1 31.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 36.54 ng

RT: 19.72 min Scan# 2874

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:184 Resp: 448549

Ion Ratio Lower Upper

184 100

185 13.6 12.6 19.0



183 12.3 10.7 16.1




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 33.25 ng

RT: 19.90 min Scan# 2905

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :


BNA\_G

ClientSampleId :

SSTDCCC040EC

Tgt Ion:202 Resp: 1006740

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 202 | 100   |       |       |
| 200 | 22.8  | 21.2  | 31.8  |
| 203 | 19.0  | 16.8  | 25.2  |




Abundance



Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



Time-->



#78

Terphenyl-d14

Concen: 66.11 ng

RT: 20.09 min Scan# 2937


Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Tgt Ion:244 Resp: 1400016

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 244 | 100   |       |       |
| 212 | 9.6   | 10.8  | 16.2# |
| 122 | 8.1   | 8.0   | 12.0  |




Abundance


Ion 244.00 (243.70 to 244.70): E


Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



Time-->









#79  
Butylbenzylphthalate  
Concen: 40.04 ng  
RT: 20.77 min Scan# 3053  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC


Tgt Ion:149 Resp: 446377  
Ion Ratio Lower Upper  
149 100  
91 82.5 68.1 102.1  
206 22.1 19.8 29.6



#80  
Benzo(a)anthracene  
Concen: 40.48 ng  
RT: 21.76 min Scan# 3222  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion:228 Resp: 999207  
Ion Ratio Lower Upper  
228 100  
226 29.9 25.3 37.9  
229 21.2 19.9 29.9









#81  
 3,3'-Dichlorobenzidine  
 Concen: 44.47 ng  
 RT: 21.68 min Scan# 3207  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26




Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 254     | 60.4  | 51.2  | 76.8  |
| 126     | 9.2   | 5.3   | 7.9#  |



#82  
 Chrysene  
 Concen: 41.36 ng  
 RT: 21.84 min Scan# 3235  
 Delta R.T. -0.00 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 228     | 100   |       |       |
| 226     | 34.0  | 26.7  | 40.1  |
| 229     | 21.3  | 17.4  | 26.2  |





#83

Bis(2-ethylhexyl)phthalate

Concen: 47.90 ng

RT: 21.64 min Scan# 3201

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:149 Resp: 638097

Ion Ratio Lower Upper

149 100

167 26.4 24.0 36.0

279 5.5 4.8 7.2






Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E

21.64



#84

Di-n-octyl phthalate

Concen: 48.55 ng

RT: 22.88 min Scan# 3412

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:149 Resp: 1115925

Ion Ratio Lower Upper

149 100

167 1.6 1.5 2.3

43 10.6 8.8 13.2






Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BG

22.88





#85

Indeno(1,2,3-cd)pyrene

Concen: 46.15 ng

RT: 28.94 min Scan# 4444

Delta R.T. -0.03 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC





Tgt Ion:276 Resp: 1194193

Ion Ratio Lower Upper

276 100

138 15.1 0.0 0.0#

277 25.4 0.0 0.0#



#86

Perylene-d12

Concen: 20.00 ng

RT: 25.12 min Scan# 3793

Delta R.T. -0.02 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26




Tgt Ion:264 Resp: 432499


Ion Ratio Lower Upper

264 100

260 24.2 18.4 27.6

265 22.5 18.9 28.3









#87  
Benzo(b)fluoranthene  
Concen: 40.52 ng  
RT: 24.06 min Scan# 3613  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26




Instrument :  
BNA\_G  
ClientSampleId :  
SSTDCCC040EC

Tgt Ion:252 Resp: 1019649  
Ion Ratio Lower Upper  
252 100  
253 23.8 18.9 28.3  
125 6.7 3.9 5.9#



#88  
Benzo(k)fluoranthene  
Concen: 41.02 ng  
RT: 24.13 min Scan# 3625  
Delta R.T. -0.01 min  
Lab File: BG024105.D  
Acq: 24 Sep 2016 2:26

Tgt Ion:252 Resp: 1025059  
Ion Ratio Lower Upper  
252 100  
253 22.8 18.8 28.2  
125 7.3 3.2 4.8#





#89

Benzo(a)pyrene

Concen: 40.84 ng

RT: 24.97 min Scan# 3767

Delta R.T. -0.01 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26

Instrument :

BNA\_G

ClientSampleId :

SSTDCCC040EC


Tgt Ion:252 Resp: 993388

Ion Ratio Lower Upper

252 100

253 20.8 18.7 28.1




125 6.8 3.9 5.9#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 39.77 ng

RT: 29.00 min Scan# 4453

Delta R.T. -0.03 min

Lab File: BG024105.D

Acq: 24 Sep 2016 2:26


Tgt Ion:278 Resp: 965882

Ion Ratio Lower Upper

278 100

139 9.0 4.7 7.1#



279 24.7 18.3 27.5




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91  
 Benzo(g,h,i)perylene  
 Concen: 40.01 ng  
 RT: 30.15 min Scan# 4650  
 Delta R.T. -0.01 min  
 Lab File: BG024105.D  
 Acq: 24 Sep 2016 2:26

Instrument :  
 BNA\_G  
 ClientSampleId :  
 SSTDCCC040EC

Tgt Ion: 276 Resp: 978990  
 Ion Ratio Lower Upper  
 276 100  
 277 25.3 18.6 27.8  
 138 13.5 6.8 10.2#

