

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092316\
 Data File : BG024107.D
 Acq On : 24 Sep 2016 3:43
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 05:16:11 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG092316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 23 15:35:28 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.442	0.519	-17.4	100	0.00
3	Pyridine	1.562	1.693	-8.4	95	0.00
4	n-Nitrosodimethylamine	0.792	0.853	-7.7	100	0.00
5 S	2-Fluorophenol	1.154	1.241	-7.5	93	0.00
6	Aniline	2.728	2.597	4.8	90	0.00
7 S	Phenol-d6	1.942	1.907	1.8	91	0.00
8	2-Chlorophenol	1.382	1.345	2.7	92	0.00
9	Benzaldehyde	1.058	1.122	-6.0	99	0.00
10 C	Phenol	2.042	2.031	0.5	92	0.00
11	bis(2-Chloroethyl)ether	1.610	1.564	2.9	90	0.00
12	1,3-Dichlorobenzene	1.541	1.583	-2.7	96	0.00
13 C	1,4-Dichlorobenzene	1.573	1.617	-2.8	96	0.00
14	1,2-Dichlorobenzene	1.531	1.543	-0.8	93	0.00
15	Benzyl Alcohol	1.880	1.681	10.6	86	0.00
16	2,2'-oxybis(1-Chloropropane	2.223	2.223	0.0	93	0.00
17	2-Methylphenol	1.358	1.307	3.8	90	0.00
18	Hexachloroethane	0.663	0.688	-3.8	96	0.00
19 P	n-Nitroso-di-n-propylamine	1.908	1.722	9.7	86	0.00
20	3+4-Methylphenols	1.972	1.906	3.3	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.577	0.589	-2.1	91	0.00
23 S	Nitrobenzene-d5	0.507	0.511	-0.8	89	0.00
24	Nitrobenzene	0.544	0.549	-0.9	89	0.00
25	Isophorone	1.035	0.998	3.6	88	0.00
26 C	2-Nitrophenol	0.193	0.198	-2.6	90	0.00
27	2,4-Dimethylphenol	0.323	0.322	0.3	88	0.00
28	bis(2-Chloroethoxy)methane	0.503	0.511	-1.6	91	0.00
29 C	2,4-Dichlorophenol	0.336	0.353	-5.1	90	0.00
30	1,2,4-Trichlorobenzene	0.359	0.370	-3.1	90	0.00
31	Naphthalene	1.013	1.032	-1.9	91	0.00
32	Benzoic acid	0.241	0.222	7.9	80	-0.01
33	4-Chloroaniline	0.452	0.448	0.9	88	0.00
34 C	Hexachlorobutadiene	0.283	0.277	2.1	87	0.00
35	Caprolactam	0.137	0.126	8.0	87	0.00
36 C	4-Chloro-3-methylphenol	0.472	0.468	0.8	86	0.00
37	2-Methylnaphthalene	0.776	0.764	1.5	88	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
39	1,2,4,5-Tetrachlorobenzene	0.584	0.593	-1.5	85	0.00
40 P	Hexachlorocyclopentadiene	0.220	0.233	-5.9	82	0.00
41 S	2,4,6-Tribromophenol	0.287	0.258	10.1	78	0.00
42 C	2,4,6-Trichlorophenol	0.378	0.393	-4.0	86	0.00
43	2,4,5-Trichlorophenol	0.389	0.375	3.6	80	0.00
44 S	2-Fluorobiphenyl	1.190	1.219	-2.4	87	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092316\
 Data File : BG024107.D
 Acq On : 24 Sep 2016 3:43
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 24 05:16:11 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG092316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 23 15:35:28 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.368	1.444	-5.6	89	0.00
46	2-Chloronaphthalene	1.021	1.061	-3.9	88	0.00
47	2-Nitroaniline	0.475	0.465	2.1	82	0.00
48	Acenaphthylene	1.734	1.738	-0.2	87	0.00
49	Dimethylphthalate	1.534	1.486	3.1	83	0.00
50	2,6-Dinitrotoluene	0.322	0.320	0.6	84	0.00
51 C	Acenaphthene	1.051	1.058	-0.7	86	0.00
52	3-Nitroaniline	0.322	0.319	0.9	87	0.00
53 P	2,4-Dinitrophenol	0.190	0.177	6.8	78	0.00
54	Dibenzofuran	1.637	1.651	-0.9	87	0.00
55 P	4-Nitrophenol	0.225	0.201	10.7	77	0.00
56	2,4-Dinitrotoluene	0.490	0.473	3.5	84	0.00
57	Fluorene	1.425	1.397	2.0	86	0.00
58	2,3,4,6-Tetrachlorophenol	0.375	0.357	4.8	79	0.00
59	Diethylphthalate	1.652	1.572	4.8	85	0.00
60	4-Chlorophenyl-phenylether	0.763	0.717	6.0	83	0.00
61	4-Nitroaniline	0.333	0.312	6.3	83	0.00
62	Azobenzene	1.679	1.644	2.1	86	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	81	0.00
64	4,6-Dinitro-2-methylphenol	0.139	0.142	-2.2	78	0.00
65 c	n-Nitrosodiphenylamine	0.547	0.586	-7.1	82	0.00
66	4-Bromophenyl-phenylether	0.232	0.240	-3.4	81	0.00
67	Hexachlorobenzene	0.261	0.263	-0.8	79	0.00
68	Atrazine	0.242	0.230	5.0	75	0.00
69 C	Pentachlorophenol	0.124	0.119	4.0	75	0.00
70	Phenanthrene	1.019	1.044	-2.5	81	0.00
71	Anthracene	1.046	1.053	-0.7	80	0.00
72	Carbazole	0.974	0.968	0.6	80	0.00
73	Di-n-butylphthalate	1.275	1.186	7.0	76	0.00
74 C	Fluoranthene	1.330	1.235	7.1	78	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
76	Benzidine	0.566	0.550	2.8	92	0.00
77	Pyrene	1.397	1.165	16.6	79	0.00
78 S	Terphenyl-d14	0.977	0.808	17.3	81	0.00
79	Butylbenzylphthalate	0.514	0.515	-0.2	101	0.00
80	Benzo(a)anthracene	1.139	1.152	-1.1	102	-0.01
81	3,3'-Dichlorobenzidine	0.408	0.457	-12.0	111	0.00
82	Chrysene	1.070	1.115	-4.2	106	0.00
83	Bis(2-ethylhexyl)phthalate	0.614	0.745	-21.3	118	0.00
84 c	Di-n-octyl phthalate	1.060	1.276	-20.4#	113	0.00
85	Indeno(1,2,3-cd)pyrene	1.193	1.374	-15.2	109	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	112	-0.01
87	Benzo(b)fluoranthene	1.164	1.182	-1.5	109	-0.01

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092316\
Data File : BG024107.D
Acq On : 24 Sep 2016 3:43
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 24 05:16:11 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG092316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 23 15:35:28 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.156	1.181	-2.2	110	-0.01
89 C	Benzo(a)pyrene	1.125	1.159	-3.0	112	-0.02
90	Dibenzo(a,h)anthracene	1.123	1.127	-0.4	110	-0.02
91	Benzo(g,h,i)perylene	1.132	1.125	0.6	109	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1