

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092415\
 Data File : BG018883.D
 Acq On : 25 Sep 2015 00:33
 Operator : TP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: Sep 25 01:09:44 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 24 04:11:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.450	0.431	4.2	136	0.00
3 S	1,4-Dioxane-d8	0.416	0.372	10.6	134	0.00
4	Benzaldehyde	0.950	1.071	-12.7	149	0.00
5 S	Phenol-d5	1.642	1.629	0.8	148	0.00
6	Phenol	1.699	1.726	-1.6	153#	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.952	0.960	-0.8	146	0.00
8	Bis(2-Chloroethyl)ether	1.269	1.329	-4.7	153#	0.00
9 S	2-Chlorophenol-d4	1.244	1.330	-6.9	160#	0.00
10	2-Chlorophenol	1.288	1.338	-3.9	150	0.00
11	2-Methylphenol	1.306	1.368	-4.7	158#	0.00
12	2,2'-oxybis(1-Chloropropane	1.506	1.477	1.9	143	0.00
13 S	4-Methylphenol-d8	1.333	1.356	-1.7	155#	0.00
14	Acetophenone	2.030	2.150	-5.9	155#	0.00
15 P	N-Nitroso-di-n-propylamine	1.047	1.031	1.5	146	0.00
16	4-Methylphenol	1.436	1.475	-2.7	152#	0.00
17	Hexachloroethane	0.513	0.536	-4.5	164#	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	150	0.00
19 S	Nitrobenzene-d5	0.152	0.152	0.0	153#	0.00
20	Nitrobenzene	0.351	0.334	4.8	145	0.00
21	Isophorone	0.716	0.706	1.4	152#	0.00
22 S	2-Nitrophenol-d4	0.152	0.176	-15.8	179#	0.00
23 C	2-Nitrophenol	0.172	0.186	-8.1	170#	0.00
24	2,4-Dimethylphenol	0.356	0.353	0.8	156#	0.00
25	Bis(2-Chloroethoxy)methane	0.422	0.428	-1.4	154#	0.00
26 S	2,4-Dichlorophenol-d3	0.321	0.333	-3.7	161#	0.00
27 C	2,4-Dichlorophenol	0.321	0.331	-3.1	160#	0.00
28	Naphthalene	1.018	1.019	-0.1	153#	0.00
29 S	4-Chloroaniline-d4	0.373	0.433	-16.1	168#	0.00
30	4-Chloroaniline	0.386	0.443	-14.8	164#	0.00
31 C	Hexachlorobutadiene	0.198	0.209	-5.6	164#	0.00
32	Caprolactam	0.131	0.127	3.1	145	0.00
33 C	4-Chloro-3-methylphenol	0.342	0.342	0.0	153#	0.00
34	2-Methylnaphthalene	0.752	0.767	-2.0	158#	0.00
35	1-Methylnaphthalene	0.721	0.730	-1.2	158#	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	149	0.00
37	1,2,4,5-Tetrachlorobenzene	0.620	0.653	-5.3	165#	0.00
38	Hexachlorocyclopentadiene	0.345	0.318	7.8	148	0.00
39 C	2,4,6-Trichlorophenol	0.403	0.435	-7.9	160#	0.00
40	2,4,5-Trichlorophenol	0.434	0.456	-5.1	158#	0.00
41	1,1'-Biphenyl	1.514	1.553	-2.6	158#	0.00
42	2-Chloronaphthalene	1.154	1.181	-2.3	156#	0.00
43	2-Nitroaniline	0.342	0.328	4.1	149	0.00
44 S	Dimethylphthalate-d6	1.610	1.634	-1.5	157#	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092415\
 Data File : BG018883.D
 Acq On : 25 Sep 2015 00:33
 Operator : TP
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02008

Quant Time: Sep 25 01:09:44 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 24 04:11:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.560	1.573	-0.8	154#	0.00
46	2,6-Dinitrotoluene	0.314	0.355	-13.1	174#	0.00
47 S	Acenaphthylene-d8	1.907	1.970	-3.3	157#	0.00
48	Acenaphthylene	2.029	2.021	0.4	151#	0.00
49	3-Nitroaniline	0.333	0.383	-15.0	165#	0.00
50 C	Acenaphthene	1.367	1.374	-0.5	154#	0.00
51	2,4-Dinitrophenol	0.141	0.130	7.8	176#	0.00
52 S	4-Nitrophenol-d4	0.309	0.304	1.6	152#	0.00
53	4-Nitrophenol	0.212	0.188	11.3	136	0.00
54	Dibenzofuran	1.918	1.913	0.3	151#	0.00
55	2,4-Dinitrotoluene	0.467	0.504	-7.9	163#	0.00
56	2,3,4,6-Tetrachlorophenol	0.419	0.414	1.2	155#	0.00
57	Diethylphthalate	1.622	1.644	-1.4	155#	0.00
58 S	Fluorene-d10	1.407	1.428	-1.5	153#	0.00
59	Fluorene	1.620	1.582	2.3	148	0.00
60	4-Chlorophenyl-phenylether	0.772	0.799	-3.5	161#	0.00
61	4-Nitroaniline	0.383	0.418	-9.1	153#	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.095	0.106	-11.6	184#	0.00
64	4,6-Dinitro-2-methylphenol	0.097	0.110	-13.4	184#	0.00
65	N-Nitrosodiphenylamine	0.524	0.539	-2.9	152#	0.00
66	4-Bromophenyl-phenylether	0.196	0.204	-4.1	158#	0.00
67	Hexachlorobenzene	0.217	0.235	-8.3	167#	0.00
68	Atrazine	0.231	0.250	-8.2	155#	0.00
69 C	Pentachlorophenol	0.127	0.105	17.3	129	0.00
70	Phenanthrene	1.017	1.023	-0.6	146	0.00
71 S	Anthracene-d10	0.881	0.902	-2.4	152#	0.00
72	Anthracene	1.020	1.033	-1.3	146	0.00
73	1,2,3,4-Tetrachlorobenzene	0.238	0.255	-7.1	159#	0.00
74	Pentachlorobenzene	0.232	0.257	-10.8	163#	0.00
75	Carbazole	0.906	0.988	-9.1	145	0.00
76	Di-n-butylphthalate	1.144	1.180	-3.1	146	0.00
77 C	Fluoranthene	1.127	1.295	-14.9	148	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	143	0.00
79 S	Pyrene-d10	0.920	0.943	-2.5	150	0.00
80	Pyrene	1.176	1.171	0.4	141	0.00
81	Butylbenzylphthalate	0.451	0.482	-6.9	157#	0.00
82	3,3'-Dichlorobenzidine	0.355	0.463	-30.4#	182#	0.00
83	Benzo(a)anthracene	1.109	1.150	-3.7	151#	0.00
84	Bis(2-ethylhexyl)phthalate	0.669	0.723	-8.1	156#	0.00
85	Chrysene	1.023	1.057	-3.3	152#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
87	Di-n-octyl phthalate	1.058	1.162	-9.8	163#	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG092415\
Data File : BG018883.D
Acq On : 25 Sep 2015 00:33
Operator : TP
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02008

Quant Time: Sep 25 01:09:44 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 24 04:11:41 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.131	1.134	-0.3	159#	0.00
89	Benzo(k)fluoranthene	1.083	1.089	-0.6	155#	0.00
90 S	Benzo(a)pyrene-d12	0.973	1.005	-3.3	164#	0.00
91 C	Benzo(a)pyrene	1.075	1.090	-1.4	159#	0.00
92	Indeno(1,2,3-cd)pyrene	1.246	1.304	-4.7	167#	0.00
93	Dibenzo(a,h)anthracene	1.000	1.079	-7.9	174#	0.00
94	Benzo(g,h,i)perylene	1.040	1.088	-4.6	167#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0