

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG092518\
 Data File : BG036967.D
 Acq On : 26 Sep 2018 3:22
 Operator : JU/SJ
 Sample : J4779-08
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SU-02-092418-B

Quant Time: Sep 26 07:50:33 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 10:41:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	47564	20.00	ng	0.00
21) Naphthalene-d8	10.85	136	193604	20.00	ng	0.00
38) Acenaphthene-d10	14.67	164	122022	20.00	ng	0.00
63) Phenanthrene-d10	17.41	188	312458	20.00	ng	0.00
75) Chrysene-d12	21.68	240	323246	20.00	ng	-0.01
86) Perylene-d12	24.89	264	319629	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	308945	124.57	ng	0.00
7) Phenol-d6	7.21	99	376893	98.22	ng	0.00
23) Nitrobenzene-d5	9.20	82	321460	88.92	ng	0.00
41) 2,4,6-Tribromophenol	16.15	330	192056	131.55	ng	-0.01
44) 2-Fluorobiphenyl	13.29	172	688856	84.08	ng	0.00
78) Terphenyl-d14	20.02	244	1156355	94.07	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG092518\
Data File : BG036967.D
Acq On : 26 Sep 2018 3:22
Operator : JU/SJ
Sample : J4779-08
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SU-02-092418-B

Quant Time: Sep 26 07:50:33 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG092018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 10:41:23 2018
Response via : Initial Calibration






#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.05 min Scan# 810
Delta R.T. -0.00 min
Lab File: BG036967.D
Acq: 26 Sep 2018 3:22

Instrument :
BNA_G
ClientSampleId :
SU-02-092418-B

Tot Ion	Ratio	Lower	Upper
152	100		
150	155.8	119.5	179.3
115	63.1	44.5	66.7






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 124.567 ng
RT: 5.62 min Scan# 397
Delta R.T. -0.00 min
Lab File: BG036967.D
Acq: 26 Sep 2018 3:22

Tgt Ion	Ratio	Lower	Upper
112	100		
64	70.1	57.2	85.8
63	38.9	30.8	46.2





Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BG

Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 98.220 ng

RT: 7.21 min Scan# 667

Delta R.T. -0.01 min

Lab File: BG036967.D

Acq: 26 Sep 2018 3:22

Instrument :

BNA_G

ClientSampleId :

SU-02-092418-B


Tot Ion: 99 Resp: 376893

Ion Ratio Lower Upper



99 100

42 21.2 16.5 24.7

71 37.8 29.4 44.2



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.85 min Scan# 1287

Delta R.T. -0.01 min

Lab File: BG036967.D

Acq: 26 Sep 2018 3:22

Tgt Ion: 136 Resp: 193604


Ion Ratio Lower Upper

136 100



137 10.7 8.8 13.2

54 11.9 9.2 13.8

68 5.7 5.8 8.6#



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC








#23
 Nitrobenzene-d5
 Concen: 88.916 ng
 RT: 9.20 min Scan# 1007
 Delta R.T. -0.01 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22




Instrument :
 BNA_G
 ClientSampleId :
 SU-02-092418-B

Tot Ion	82	Resp	321460
Ion Ratio	Lower	Upper	
82	100		
128	39.2	33.3	49.9
54	56.2	45.1	67.7



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.67 min Scan# 1937
 Delta R.T. -0.01 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22

Tgt Ion	164	Resp	122022
Ion Ratio	Lower	Upper	
164	100		
162	107.5	80.7	121.1
160	46.4	35.3	52.9








#41
 2,4,6-Tribromophenol
 Concen: 131.553 ng
 RT: 16.15 min Scan# 2190
 Delta R.T. -0.01 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22

Instrument :
 BNA_G
 ClientSampleId :
 SU-02-092418-B

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.7	75.4	113.2
141	37.8	30.7	46.1




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#44
 2-Fluorobiphenyl
 Concen: 84.081 ng
 RT: 13.29 min Scan# 1703
 Delta R.T. -0.01 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.9	31.3	46.9
170	25.7	21.0	31.6



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.41 min Scan# 2404
Delta R.T. -0.01 min
Lab File: BG036967.D
Acq: 26 Sep 2018 3:22




Instrument :
BNA_G
ClientSampleId :
SU-02-092418-B

Tot Ion	Ratio	Lower	Upper
188	100		
94	7.5	6.7	10.1
80	9.5	8.1	12.1



#75
Chrysene-d12
Concen: 20.000 ng
RT: 21.68 min Scan# 3130
Delta R.T. -0.01 min
Lab File: BG036967.D
Acq: 26 Sep 2018 3:22

Tgt Ion	Ratio	Lower	Upper
240	100		
120	8.1	6.9	10.3
236	26.9	21.2	31.8









#78
 Terphenyl-d14
 Concen: 94.065 ng
 RT: 20.02 min Scan# 2848
 Delta R.T. -0.01 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22

Instrument :
 BNA_G
 ClientSampleId :
 SU-02-092418-B

Tot Ion:244 Resp: 1156355

Ion	Ratio	Lower	Upper
244	100		
212	8.4	7.0	10.6
122	8.8	7.1	10.7



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.89 min Scan# 3676
 Delta R.T. -0.02 min
 Lab File: BG036967.D
 Acq: 26 Sep 2018 3:22

Tgt Ion:264 Resp: 319629

Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.6	29.4
265	21.8	17.4	26.0

