

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092822\
 Data File : BG055029.D
 Acq On : 28 Sep 2022 10:39
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/29/2022
 Supervised By : Jagrut Upadhyay 09/29/2022

Quant Time: Sep 28 14:24:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 14:23:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							09/29/2022
1) 1,4-Dichlorobenzene-d4	8.352	152	32109	20.000	ng	0.00	Supervised By : Jagrut Upadhyay
21) Naphthalene-d8	11.178	136	127715	20.000	ng	0.00	
39) Acenaphthene-d10	14.956	164	87173	20.000	ng	0.00	
64) Phenanthrene-d10	17.688	188	212706	20.000	ng	0.00	
76) Chrysene-d12	21.989	240	244921	20.000	ng	# 0.00	
86) Perylene-d12	25.455	264	255522	20.000	ng	0.00	09/29/2022
System Monitoring Compounds							
5) 2-Fluorophenol	5.860	112	17062	10.397	ng	0.00	
7) Phenol-d6	7.470	99	21589	9.815	ng	0.00	
23) Nitrobenzene-d5	9.521	82	16528	7.501	ng	0.00	
42) 2,4,6-Tribromophenol	16.430	330	8174	8.004	ng	0.00	
45) 2-Fluorobiphenyl	13.587	172	58515	10.579	ng	0.00	
79) Terphenyl-d14	20.267	244	122782	10.487	ng	0.00	
Target Compounds							Qvalue
2) 1,4-Dioxane	3.698	88	4956	6.217	ng	# 88	
3) Pyridine	4.121	79	11507m	5.312	ng		
4) n-Nitrosodimethylamine	4.027	42	5689	7.197	ng	# 83	
6) Aniline	7.658	93	15011	5.456	ng	95	
8) 2-Chlorophenol	7.905	128	9452	4.890	ng	95	
10) Phenol	7.494	94	12861	5.396	ng	89	
11) bis(2-Chloroethyl)ether	7.752	93	10402	5.518	ng	98	
12) 1,3-Dichlorobenzene	8.234	146	12087	5.361	ng	97	
13) 1,4-Dichlorobenzene	8.381	146	12054	5.246	ng	96	
14) 1,2-Dichlorobenzene	8.710	146	11942	5.516	ng	96	
15) Benzyl Alcohol	8.575	79	8761	5.446	ng	97	
16) 2,2'-oxybis(1-Chloropr...	8.863	45	16891	7.311	ng	98	
17) 2-Methylphenol	8.769	107	8975	5.341	ng	96	
18) Hexachloroethane	9.450	117	3609	4.570	ng	91	
19) n-Nitroso-di-n-propyla...	9.145	70	8906	5.883	ng	93	
20) 3+4-Methylphenols	9.098	107	11650	4.957	ng	95	
22) Acetophenone	9.168	105	15918	5.274	ng	# 97	
24) Nitrobenzene	9.556	77	8862	4.014	ng	98	
25) Isophorone	10.079	82	22217	5.311	ng	# 94	
26) 2-Nitrophenol	10.279	139	3214	2.765	ng	# 89	
27) 2,4-Dimethylphenol	10.320	122	8244	5.024	ng	93	
28) bis(2-Chloroethoxy)met...	10.567	93	11386	4.832	ng	99	
29) 2,4-Dichlorophenol	10.813	162	8411	4.294	ng	98	
30) 1,2,4-Trichlorobenzene	11.037	180	11175	4.950	ng	94	
31) Naphthalene	11.236	128	34482	5.163	ng	96	
33) 4-Chloroaniline	11.325	127	12944	4.757	ng	97	
34) Hexachlorobutadiene	11.513	225	7322	4.884	ng	99	
35) Caprolactam	12.053	113	2760	3.867	ng	# 80	
36) 4-Chloro-3-methylphenol	12.406	107	9454	4.519	ng	89	
37) 2-Methylnaphthalene	12.811	142	24118	5.069	ng	97	
38) 1-Methylnaphthalene	13.028	142	23210m	4.988	ng		
40) 1,2,4,5-Tetrachloroben...	13.164	216	13196	5.177	ng	98	
41) Hexachlorocyclopentadiene	13.152	237	5888	23.749	ng	86	
43) 2,4,6-Trichlorophenol	13.393	196	6780	4.629	ng	97	
44) 2,4,5-Trichlorophenol	13.463	196	8115	4.826	ng	91	
46) 1,1'-Biphenyl	13.798	154	31223	5.309	ng	99	

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092822\
 Data File : BG055029.D
 Acq On : 28 Sep 2022 10:39
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005

Quant Time: Sep 28 14:24:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 14:23:01 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Christian Giraldo 09/29/2022
 Supervised By :Jagrut Upadhyay 09/29/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
47) 2-Chloronaphthalene	13.845	162	25205	5.482	ng	95	09/29/2022
48) 2-Nitroaniline	14.027	65	4129	3.060	ng	#	Supervised By :Jagrut Upadhyay
49) Acenaphthylene	14.680	152	40038	5.153	ng	98	
50) Dimethylphthalate	14.397	163	32641	4.945	ng	98	
51) 2,6-Dinitrotoluene	14.515	165	3863	2.752	ng	#	86
52) Acenaphthene	15.020	154	24151	5.045	ng	93	
53) 3-Nitroaniline	14.838	138	4647	3.052	ng	#	81
54) 2,4-Dinitrophenol	15.044	184	2351	6.881	ng	#	60
55) Dibenzofuran	15.349	168	40582	5.396	ng	96	
57) 2,4-Dinitrotoluene	15.291	165	5070	2.540	ng	#	89
58) Fluorene	15.996	166	32963	5.164	ng	95	
59) 2,3,4,6-Tetrachlorophenol	15.561	232	6633	4.680	ng	#	98
60) Diethylphthalate	15.749	149	33436	5.073	ng	98	
61) 4-Chlorophenyl-phenyle...	15.984	204	16735	5.188	ng	99	
62) 4-Nitroaniline	15.990	138	5599	3.531	ng	93	
63) Azobenzene	16.272	77	33740	6.259	ng	97	
66) n-Nitrosodiphenylamine	16.189	169	29378	5.166	ng	99	
67) 4-Bromophenyl-phenylether	16.877	248	11666	4.973	ng	99	
68) Hexachlorobenzene	16.994	284	12931	4.799	ng	99	
69) Atrazine	17.124	200	10931	5.064	ng	97	
70) Pentachlorophenol	17.329	266	5740	10.842	ng	89	
71) Phenanthrene	17.735	178	57001	5.262	ng	97	
72) Anthracene	17.823	178	56596	5.356	ng	96	
73) Carbazole	18.081	167	53492	5.528	ng	95	
74) Di-n-butylphthalate	18.628	149	59917	4.993	ng	98	
75) Fluoranthene	19.721	202	73751	5.375	ng	94	
77) Benzidine	19.885	184	36763	6.058	ng	99	
78) Pyrene	20.079	202	76643	4.760	ng	99	
80) Butylbenzylphthalate	20.954	149	21975	3.512	ng	95	
81) Benzo(a)anthracene	21.971	228	80828	5.131	ng	97	
82) 3,3'-Dichlorobenzidine	21.871	252	23980	4.389	ng	#	95
83) Chrysene	22.036	228	75296	4.994	ng	99	
84) Bis(2-ethylhexyl)phtha...	21.865	149	35239	3.927	ng	#	99
85) Di-n-octyl phthalate	23.170	149	58331	3.760	ng	97	
87) Indeno(1,2,3-cd)pyrene	29.433	276	89951	4.735	ng	#	93
88) Benzo(b)fluoranthene	24.345	252	75427	5.005	ng	#	97
89) Benzo(k)fluoranthene	24.415	252	77280	5.079	ng	#	97
90) Benzo(a)pyrene	25.285	252	62247	4.797	ng	#	97
91) Dibenzo(a,h)anthracene	29.509	278	76798	4.911	ng	97	
92) Benzo(g,h,i)perylene	30.678	276	73316	4.651	ng	#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations


Quant Time: Sep 28 14:24:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 14:23:01 2022
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.352 min Scan# 80
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 152 Resp: 32109
Ion Ratio Lower Upper
152 100
150 154.0 125.2 187.8
115 56.1 47.9 71.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#2
1,4-Dioxane
Concen: 6.217 ng
RT: 3.698 min Scan# 70
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion: 88 Resp: 4956
Ion Ratio Lower Upper
88 100
58 73.2 65.9 98.9
43 36.6 22.7 34.1#





#3
Pyridine
Concen: 5.312 ng m
RT: 4.121 min Scan# 141
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 79 Resp: 11507
Ion Ratio Lower Upper
79 100
52 69.9 61.4 92.0
51 39.1 31.4 47.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#4
n-Nitrosodimethylamine
Concen: 7.197 ng
RT: 4.027 min Scan# 126
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion: 42 Resp: 5689
Ion Ratio Lower Upper
42 100
74 117.9 109.6 164.4
44 22.1 5.5 8.3#





#5
2-Fluorophenol
Concen: 10.397 ng
RT: 5.860 min Scan# 439
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 112 Resp: 17062
Ion Ratio Lower Upper
112 100
64 64.0 49.3 73.9
63 38.4 26.3 39.5


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#6
Aniline
Concen: 5.456 ng
RT: 7.658 min Scan# 744
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion: 93 Resp: 15011
Ion Ratio Lower Upper
93 100
66 41.4 31.8 47.8
65 24.8 16.6 25.0





#7
Phenol-d6
Concen: 9.815 ng
RT: 7.470 min Scan# 712
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 99 Resp: 21589
Ion Ratio Lower Upper
99 100
42 19.6 14.6 22.0
71 35.6 24.6 37.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#8
2-Chlorophenol
Concen: 4.890 ng
RT: 7.905 min Scan# 786
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:128 Resp: 9452
Ion Ratio Lower Upper
128 100
130 29.4 12.5 52.5
64 45.2 28.6 68.6





#10
Phenol
Concen: 5.396 ng
RT: 7.494 min Scan# 716
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 94 Resp: 1286
Ion Ratio Lower Upper
94 100
65 35.8 9.9 49.9
66 43.2 16.4 56.4


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#11
bis(2-Chloroethyl)ether
Concen: 5.518 ng
RT: 7.752 min Scan# 760
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Tgt Ion: 93 Resp: 10402
Ion Ratio Lower Upper
93 100
63 79.6 62.1 102.1
95 31.5 12.1 52.1





#12

1,3-Dichlorobenzene

Concen: 5.361 ng

RT: 8.234 min Scan# 84

Delta R.T. -0.007 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :

SSTDICC005

Tgt Ion:146 Resp: 1208

Ion Ratio Lower Upper

146 100

148 62.2 50.6 76.0



75 34.2 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 09/29/2022

Supervised By :Jagrut Upadhyay 09/29/2022





09/29/2022

Supervised By :Jagrut

Upadhyay

09/29/2022



#13

1,4-Dichlorobenzene

Concen: 5.246 ng

RT: 8.381 min Scan# 867

Delta R.T. -0.007 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39




Tgt Ion:146 Resp: 12054


Ion Ratio Lower Upper

146 100

148 66.2 50.4 75.6


111 38.7 33.0 49.4





#14
1,2-Dichlorobenzene
Concen: 5.516 ng
RT: 8.710 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:146 Resp: 11942
Ion Ratio Lower Upper
146 100
148 59.1 51.2 76.8
111 42.0 34.3 51.5


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#15
Benzyl Alcohol
Concen: 5.446 ng
RT: 8.575 min Scan# 900
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion: 79 Resp: 8761
Ion Ratio Lower Upper
79 100
108 82.0 63.4 95.0
77 70.4 55.0 82.4







#16
2,2'-oxybis(1-Chloropropane)
Concen: 7.311 ng
RT: 8.863 min Scan# 949
Delta R.T. -0.013 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Tgt Ion: 45 Resp: 1689
Ion Ratio Lower Upper
45 100
77 12.4 0.0 32.8
79 9.3 0.0 30.4



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#17
2-Methylphenol
Concen: 5.341 ng
RT: 8.769 min Scan# 933
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:107 Resp: 8975
Ion Ratio Lower Upper
107 100
108 107.3 85.6 128.4
77 47.7 35.8 53.8
79 50.1 33.8 50.6





#18
Hexachloroethane
Concen: 4.570 ng
RT: 9.450 min Scan# 1049
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 117 Resp: 3609
Ion Ratio Lower Upper
117 100
119 98.1 77.0 115.6
201 126.0 87.2 130.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#19
n-Nitroso-di-n-propylamine
Concen: 5.883 ng
RT: 9.145 min Scan# 997
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion: 70 Resp: 8906
Ion Ratio Lower Upper
70 100
42 64.6 46.6 69.8
101 7.7 7.7 11.5
130 19.5 13.9 20.9





#20
3+4-Methylphenols
Concen: 4.957 ng
RT: 9.098 min Scan# 989
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion	Ratio	Lower	Upper
107	100		
108	87.6	71.6	111.6
77	34.1	13.1	53.1
79	33.2	7.9	47.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.178 min Scan# 1343
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.2	13.8
54	9.4	7.7	11.5
68	5.7	5.6	8.4





#22
Acetophenone
Concen: 5.274 ng
RT: 9.168 min Scan# 1002
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 105 Resp: 15918
Ion Ratio Lower Upper
105 100
71 0.0 1.2 1.8
51 29.9 22.9 34.3
120 19.8 17.4 26.2


Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 09/29/2022
Supervised By : Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By : Jagrut
Upadhyay


09/29/2022



#23
Nitrobenzene-d5
Concen: 7.501 ng
RT: 9.521 min Scan# 1061
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion: 82 Resp: 16528
Ion Ratio Lower Upper
82 100
128 38.2 32.2 48.2
54 49.3 41.4 62.0





#24

Nitrobenzene

Concen: 4.014 ng

RT: 9.556 min Scan# 1067

Delta R.T. -0.007 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :



SSTDICC005

Tgt Ion: 77	Resp: 8862
Ion Ratio	Lower Upper
77 100	
123 42.7	35.0 52.6
65 14.2	10.2 15.2

Manual Integrations



APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022



09/29/2022
Supervised By :Jagrut
Upadhyay

09/29/2022



#25

Isophorone

Concen: 5.311 ng




RT: 10.079 min Scan# 1156


Delta R.T. -0.007 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Tgt Ion: 82	Resp: 22217
Ion Ratio	Lower Upper
82 100	
95 8.4	5.6 8.4#
138 14.4	13.9 20.9







#26
2-Nitrophenol
Concen: 2.765 ng
RT: 10.279 min Scan# 1190
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Tgt Ion:139 Resp: 3214
Ion Ratio Lower Upper
139 100
109 26.2 21.8 32.6
65 35.9 37.4 56.0



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#27
2,4-Dimethylphenol
Concen: 5.024 ng
RT: 10.320 min Scan# 1197
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:122 Resp: 8244
Ion Ratio Lower Upper
122 100
107 111.4 94.6 142.0
121 56.3 49.8 74.6





#28
bis(2-Chloroethoxy)methane
Concen: 4.832 ng
RT: 10.567 min Scan# 1239
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion: 93 Resp: 11386
Ion Ratio Lower Upper
93 100
95 32.7 26.2 39.2
123 9.5 8.4 12.6


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#29
2,4-Dichlorophenol
Concen: 4.294 ng
RT: 10.813 min Scan# 1281
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:162 Resp: 8411
Ion Ratio Lower Upper
162 100
164 66.4 47.2 87.2
98 34.4 16.6 56.6





#30
1,2,4-Trichlorobenzene
Concen: 4.950 ng
RT: 11.037 min Scan# 1319
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G



ClientSampleId :

SSTDICC005

Tgt Ion:180 Resp: 11175
Ion Ratio Lower Upper
180 100
182 104.2 77.0 115.6
145 30.6 24.0 36.0



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#31
Naphthalene
Concen: 5.163 ng
RT: 11.236 min Scan# 1353
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:128 Resp: 34482
Ion Ratio Lower Upper
128 100
129 10.5 9.0 13.4
127 11.3 11.1 16.7






#33
4-Chloroaniline
Concen: 4.757 ng
RT: 11.325 min Scan# 1368
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :

SSTDICC005





Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.0	39.0
65	36.2	25.8	38.8
92	19.5	15.1	22.7

Manual Integrations


APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#34
Hexachlorobutadiene
Concen: 4.884 ng
RT: 11.513 min Scan# 1400
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.9	51.4	77.0
227	64.3	52.6	79.0





#35
Caprolactam
Concen: 3.867 ng
RT: 12.053 min Scan# 1498
Delta R.T. -0.036 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:113 Resp: 2760
Ion Ratio Lower Upper
113 100
55 174.0 179.1 219.1
56 110.8 120.8 160.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#36
4-Chloro-3-methylphenol
Concen: 4.519 ng
RT: 12.406 min Scan# 1552
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:107 Resp: 9454
Ion Ratio Lower Upper
107 100
144 24.9 22.4 33.6
142 71.2 65.8 98.8





#37
2-Methylnaphthalene
Concen: 5.069 ng
RT: 12.811 min Scan# 1621
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G



ClientSampleId :

SSTDICC005

Tgt Ion:142 Resp: 24118
Ion Ratio Lower Upper
142 100
141 88.2 73.0 109.4
115 35.7 27.3 40.9



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#38
1-Methylnaphthalene
Concen: 4.988 ng m
RT: 13.028 min Scan# 1658
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:142 Resp: 23210
Ion Ratio Lower Upper
142 100
141 97.0 77.4 116.0
116 4.5 3.0 4.4#





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.956 min Scan# 1986
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G



ClientSampleId :

SSTDICC005

Tgt Ion:164 Resp: 87173
Ion Ratio Lower Upper
164 100
162 109.6 83.0 124.4
160 43.5 38.0 57.0



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#40
1,2,4,5-Tetrachlorobenzene
Concen: 5.177 ng
RT: 13.164 min Scan# 1681
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:216 Resp: 13196
Ion Ratio Lower Upper
216 100
214 76.6 63.2 94.8
179 20.4 16.5 24.7
108 15.8 13.9 20.9





#41
Hexachlorocyclopentadiene
Concen: 23.749 ng
RT: 13.152 min Scan# 1679
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:237 Resp: 588
Ion Ratio Lower Upper
237 100
235 50.8 42.2 82.2
272 9.4 0.0 34.0


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#42
2,4,6-Tribromophenol
Concen: 8.004 ng
RT: 16.430 min Scan# 2237
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion:330 Resp: 8174
Ion Ratio Lower Upper
330 100
332 97.1 79.8 119.6
141 29.8 26.8 40.2





#43
2,4,6-Trichlorophenol
Concen: 4.629 ng
RT: 13.393 min Scan# 1720
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:196 Resp: 6780
Ion Ratio Lower Upper
196 100
198 98.2 77.1 115.7
200 34.9 24.2 36.4


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#44
2,4,5-Trichlorophenol
Concen: 4.826 ng
RT: 13.463 min Scan# 1732
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion:196 Resp: 8115
Ion Ratio Lower Upper
196 100
198 82.2 76.7 115.1
97 45.0 36.6 55.0
132 21.2 19.5 29.3





#45
2-Fluorobiphenyl
Concen: 10.579 ng
RT: 13.587 min Scan# 1753
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:172 Resp: 58515
Ion Ratio Lower Upper
172 100
171 36.3 31.2 46.8
170 25.7 20.6 30.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#46
1,1'-Biphenyl
Concen: 5.309 ng
RT: 13.798 min Scan# 1789
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Tgt Ion:154 Resp: 31223
Ion Ratio Lower Upper
154 100
153 41.8 22.2 62.2
76 12.6 0.0 34.0





#47
2-Chloronaphthalene
Concen: 5.482 ng
RT: 13.845 min Scan# 1797
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G



ClientSampleId :

SSTDICC005

Tgt Ion:162 Resp: 25205
Ion Ratio Lower Upper
162 100
127 35.5 31.5 47.3
164 32.2 27.8 41.6



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#48
2-Nitroaniline
Concen: 3.060 ng
RT: 14.027 min Scan# 1828
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion: 65 Resp: 4129
Ion Ratio Lower Upper
65 100
92 75.6 54.7 82.1
138 126.2 84.1 126.1#





#49
Acenaphthylene
Concen: 5.153 ng
RT: 14.680 min Scan# 1940
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:152 Resp: 40038
Ion Ratio Lower Upper
152 100
151 20.1 17.3 25.9
153 13.6 10.8 16.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#50
Dimethylphthalate
Concen: 4.945 ng
RT: 14.397 min Scan# 1891
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:163 Resp: 32641
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.4
164 9.0 8.0 12.0





#51
2,6-Dinitrotoluene
Concen: 2.752 ng
RT: 14.515 min Scan# 1911
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:165 Resp: 3863
Ion Ratio Lower Upper
165 100
63 45.8 46.0 69.0
89 52.8 50.3 75.5


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#52
Acenaphthene
Concen: 5.045 ng
RT: 15.020 min Scan# 1997
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:154 Resp: 24151
Ion Ratio Lower Upper
154 100
153 125.6 92.8 139.2
152 58.2 45.4 68.2







#53
3-Nitroaniline
Concen: 3.052 ng
RT: 14.838 min Scan# 1966
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Tgt Ion:138 Resp: 464
Ion Ratio Lower Upper
138 100
108 24.9 7.6 11.4#
92 100.8 96.0 144.0



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#54
2,4-Dinitrophenol
Concen: 6.881 ng
RT: 15.044 min Scan# 2001
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:184 Resp: 2351
Ion Ratio Lower Upper
184 100
63 38.2 41.8 62.8#
154 99.1 45.2 67.8#





#55
Dibenzenofuran
Concen: 5.396 ng
RT: 15.349 min Scan# 2053
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:168 Resp: 40582
Ion Ratio Lower Upper
168 100
139 40.6 30.2 45.2
169 12.8 11.3 16.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#57
2,4-Dinitrotoluene
Concen: 2.540 ng
RT: 15.291 min Scan# 2043
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Tgt Ion:165 Resp: 5070
Ion Ratio Lower Upper
165 100
63 40.6 36.8 55.2
89 68.7 64.3 96.5
182 0.0 1.8 2.6#





#58

Fluorene

Concen: 5.164 ng

RT: 15.996 min Scan# 2163

Delta R.T. -0.001 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :

SSTDICC005



Tgt Ion	Ratio	Lower	Upper
166	100		
165	106.6	80.8	121.2
167	13.4	11.4	17.0

Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/29/2022

Supervised By :Jagrut Upadhyay 09/29/2022



09/29/2022
Supervised By :Jagrut
Upadhyay

09/29/2022



#59

2,3,4,6-Tetrachlorophenol

Concen: 4.680 ng




RT: 15.561 min Scan# 2089


Delta R.T. -0.001 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Tgt Ion	Ratio	Lower	Upper
232	100		
131	41.3	31.3	46.9
130	0.0	1.8	2.6
166	28.7	23.1	34.7






#60
Diethylphthalate
Concen: 5.073 ng
RT: 15.749 min Scan# 2121
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :



SSTDICC005



Tgt Ion:149 Resp: 33436
Ion Ratio Lower Upper
149 100
177 21.8 17.8 26.6
150 13.9 10.1 15.1


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#61
4-Chlorophenyl-phenylether
Concen: 5.188 ng
RT: 15.984 min Scan# 2161
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion:204 Resp: 16735
Ion Ratio Lower Upper
204 100
206 31.2 26.6 39.8
141 56.3 44.9 67.3





#62
4-Nitroaniline
Concen: 3.531 ng
RT: 15.990 min Scan# 2162
Delta R.T. -0.013 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:138 Resp: 5599
Ion Ratio Lower Upper
138 100
92 47.3 34.7 74.7
108 80.5 56.2 96.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#63
Azobenzene
Concen: 6.259 ng
RT: 16.272 min Scan# 2210
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion: 77 Resp: 33740
Ion Ratio Lower Upper
77 100
182 23.5 5.1 45.1
105 19.4 0.7 40.7
51 31.6 9.9 49.9





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.688 min Scan# 2451
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :

BNA_G



ClientSampleId :

SSTDICC005

Tgt Ion:188 Resp: 212700
Ion Ratio Lower Upper
188 100
94 7.7 7.0 10.4
80 9.7 7.9 11.9



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#66
n-Nitrosodiphenylamine
Concen: 5.166 ng
RT: 16.189 min Scan# 2196
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:169 Resp: 29378
Ion Ratio Lower Upper
169 100
168 69.5 55.6 83.4
167 37.2 30.9 46.3





#67
4-Bromophenyl-phenylether
Concen: 4.973 ng
RT: 16.877 min Scan# 2313
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:248 Resp: 11660
Ion Ratio Lower Upper
248 100
250 96.7 78.8 118.2
141 64.7 51.9 77.9


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#68
Hexachlorobenzene
Concen: 4.799 ng
RT: 16.994 min Scan# 2333
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:284 Resp: 12931
Ion Ratio Lower Upper
284 100
142 31.6 25.0 37.6
249 30.0 24.5 36.7





#69
 Atrazine
 Concen: 5.064 ng
 RT: 17.124 min Scan# 2355
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39



Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005



Tgt Ion:200 Resp: 10931
 Ion Ratio Lower Upper
 200 100
 173 21.9 5.6 45.6
 215 48.7 28.2 68.2


Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/29/2022
 Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
 Supervised By :Jagrut
 Upadhyay



09/29/2022




#70
 Pentachlorophenol
 Concen: 10.842 ng
 RT: 17.329 min Scan# 2390
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39



Tgt Ion:266 Resp: 5740
 Ion Ratio Lower Upper
 266 100
 268 72.8 51.2 76.8
 264 69.3 49.4 74.0







#71
Phenanthrene
Concen: 5.262 ng
RT: 17.735 min Scan# 2459
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Instrument :
BNA_G
ClientSampleId :
SSTDICC005

Tgt Ion:178 Resp: 57001
Ion Ratio Lower Upper
178 100
176 18.8 16.6 25.0
179 15.6 12.8 19.2



Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#72
Anthracene
Concen: 5.356 ng
RT: 17.823 min Scan# 2474
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:178 Resp: 56596
Ion Ratio Lower Upper
178 100
176 18.2 16.2 24.4
179 17.3 12.9 19.3





#73
 Carbazole
 Concen: 5.528 ng
 RT: 18.081 min Scan# 2518
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39



Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005



Tgt Ion:167 Resp: 53492
 Ion Ratio Lower Upper
 167 100
 166 21.0 18.8 28.2
 139 14.4 10.2 15.4


Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/29/2022
 Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
 Supervised By :Jagrut
 Upadhyay



09/29/2022




#74
 Di-n-butylphthalate
 Concen: 4.993 ng
 RT: 18.628 min Scan# 2611
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39




Tgt Ion:149 Resp: 59917
 Ion Ratio Lower Upper
 149 100
 150 9.5 8.0 12.0
 104 4.5 4.4 6.6





#75
Fluoranthene
Concen: 5.375 ng
RT: 19.721 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:202 Resp: 7375
Ion Ratio Lower Upper
202 100
101 7.8 0.0 31.4
203 16.6 0.0 38.6


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.989 min Scan# 3183
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion:240 Resp: 244921
Ion Ratio Lower Upper
240 100
120 7.9 8.2 12.4#
236 26.4 21.2 31.8





#77
Benzidine
Concen: 6.058 ng
RT: 19.885 min Scan# 2825
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:184 Resp: 36763
Ion Ratio Lower Upper
184 100
185 14.0 11.4 17.0
183 12.1 9.8 14.8


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022




09/29/2022
Supervised By :Jagrut
Upadhyay



09/29/2022




#78
Pyrene
Concen: 4.760 ng
RT: 20.079 min Scan# 2858
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39




Tgt Ion:202 Resp: 76643
Ion Ratio Lower Upper
202 100
200 22.0 17.8 26.6
203 17.7 14.6 21.8





#79
 Terphenyl-d14
 Concen: 10.487 ng
 RT: 20.267 min Scan# 2890
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39



Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005



Tgt Ion:244 Resp: 122782
 Ion Ratio Lower Upper
 244 100
 212 6.9 6.2 9.2
 122 7.3 9.0 13.6


Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 09/29/2022
 Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
 Supervised By :Jagrut
 Upadhyay


09/29/2022



#80
 Butylbenzylphthalate
 Concen: 3.512 ng
 RT: 20.954 min Scan# 3007
 Delta R.T. -0.001 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39


Tgt Ion:149 Resp: 21975
 Ion Ratio Lower Upper
 149 100
 91 69.6 59.5 89.3
 206 22.8 16.6 25.0





#81
Benzo(a)anthracene
Concen: 5.131 ng
RT: 21.971 min Scan# 3180
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:228 Resp: 80828
Ion Ratio Lower Upper
228 100
226 27.4 23.4 35.2
229 19.0 16.5 24.7


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#82
3,3'-Dichlorobenzidine
Concen: 4.389 ng
RT: 21.871 min Scan# 3163
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:252 Resp: 23980
Ion Ratio Lower Upper
252 100
254 60.8 51.4 77.2
126 8.1 8.2 12.2#





#83
Chrysene
Concen: 4.994 ng
RT: 22.036 min Scan# 3192
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:228 Resp: 75296
Ion Ratio Lower Upper
228 100
226 31.8 25.5 38.3
229 19.4 16.8 25.2


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#84
Bis(2-ethylhexyl)phthalate
Concen: 3.927 ng
RT: 21.865 min Scan# 3162
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:149 Resp: 35239
Ion Ratio Lower Upper
149 100
167 26.0 21.4 32.0
279 4.8 3.1 4.7#





#85
Di-n-octyl phthalate
Concen: 3.760 ng
RT: 23.170 min Scan# 3383
Delta R.T. 0.005 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:149 Resp: 5833
Ion Ratio Lower Upper
149 100
167 1.3 1.2 1.8
43 10.2 9.0 13.6


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 25.455 min Scan# 3773
Delta R.T. -0.001 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39


Tgt Ion:264 Resp: 255522
Ion Ratio Lower Upper
264 100
260 23.3 19.9 29.9
265 21.1 16.9 25.3





#87
 Indeno(1,2,3-cd)pyrene
 Concen: 4.735 ng
 RT: 29.433 min Scan# 4453
 Delta R.T. -0.019 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39



Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC005



Tgt Ion: 276 Resp: 8995
 Ion Ratio Lower Upper
 276 100
 138 8.4 12.0 18.0
 277 26.2 20.2 30.4


Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 09/29/2022
 Supervised By : Jagrut Upadhyay 09/29/2022






09/29/2022
 Supervised By : Jagrut
 Upadhyay


09/29/2022



#88
 Benzo(b)fluoranthene
 Concen: 5.005 ng
 RT: 24.345 min Scan# 3584
 Delta R.T. -0.007 min
 Lab File: BG055029.D
 Acq: 28 Sep 2022 10:39


Tgt Ion: 252 Resp: 75427
 Ion Ratio Lower Upper
 252 100
 253 22.1 18.0 27.0
 125 7.4 8.1 12.1#





#89
Benzo(k)fluoranthene
Concen: 5.079 ng
RT: 24.415 min Scan# 3596
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39



Instrument :
BNA_G
ClientSampleId :
SSTDICC005



Tgt Ion:252 Resp: 77280
Ion Ratio Lower Upper
252 100
253 23.3 17.8 26.8
125 8.0 8.2 12.4


Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 09/29/2022
Supervised By :Jagrut Upadhyay 09/29/2022






09/29/2022
Supervised By :Jagrut
Upadhyay


09/29/2022



#90
Benzo(a)pyrene
Concen: 4.797 ng
RT: 25.285 min Scan# 3744
Delta R.T. -0.007 min
Lab File: BG055029.D
Acq: 28 Sep 2022 10:39

Tgt Ion:252 Resp: 62247
Ion Ratio Lower Upper
252 100
253 23.3 18.3 27.5
125 8.4 8.8 13.2





#91

Dibenzo(a,h)anthracene

Concen: 4.911 ng

RT: 29.509 min Scan# 4463

Delta R.T. -0.025 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Instrument :

BNA_G

ClientSampleId :

SSTDICC005



Tgt Ion:	278	Resp:	76798
Ion Ratio	Lower	Upper	
278	100		
139	11.8	11.4	17.2
279	24.9	19.3	28.9

Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 09/29/2022

Supervised By :Jagrut Upadhyay 09/29/2022



09/29/2022
Supervised By :Jagrut
Upadhyay

09/29/2022



#92

Benzo(g,h,i)perylene

Concen: 4.651 ng

RT: 30.678 min Scan# 4662

Delta R.T. -0.019 min

Lab File: BG055029.D

Acq: 28 Sep 2022 10:39

Tgt Ion:	276	Resp:	73316
Ion Ratio	Lower	Upper	
276	100		
277	25.0	19.4	29.0
138	15.8	16.6	25.0#

