

Data Path : Z:\HPCHEM1\BNA G\DATA\BG093015\
 Data File : BG018967.D
 Acq On : 30 Sep 2015 10:43
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02019

Quant Time: Sep 30 17:14:04 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 30 04:08:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.450	0.405	10.0	72	0.00
3 S	1,4-Dioxane-d8	0.416	0.398	4.3	81	0.00
4	Benzaldehyde	0.950	1.040	-9.5	81	0.00
5 S	Phenol-d5	1.642	1.684	-2.6	86	0.00
6	Phenol	1.699	1.730	-1.8	86	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.952	0.939	1.4	81	0.00
8	Bis(2-Chloroethyl)ether	1.269	1.315	-3.6	85	0.00
9 S	2-Chlorophenol-d4	1.244	1.330	-6.9	90	0.00
10	2-Chlorophenol	1.288	1.365	-6.0	86	0.00
11	2-Methylphenol	1.306	1.352	-3.5	88	0.00
12	2,2'-oxybis(1-Chloropropane	1.506	1.391	7.6	76	0.00
13 S	4-Methylphenol-d8	1.333	1.393	-4.5	90	0.00
14	Acetophenone	2.030	2.127	-4.8	87	0.00
15 P	N-Nitroso-di-n-propylamine	1.047	0.987	5.7	79	0.00
16	4-Methylphenol	1.436	1.548	-7.8	90	0.00
17	Hexachloroethane	0.513	0.520	-1.4	90	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
19 S	Nitrobenzene-d5	0.152	0.157	-3.3	91	0.00
20	Nitrobenzene	0.351	0.331	5.7	83	0.00
21	Isophorone	0.716	0.680	5.0	84	0.00
22 S	2-Nitrophenol-d4	0.152	0.170	-11.8	99	0.00
23 C	2-Nitrophenol	0.172	0.187	-8.7	98	0.00
24	2,4-Dimethylphenol	0.356	0.360	-1.1	91	0.00
25	Bis(2-Chloroethoxy)methane	0.422	0.419	0.7	87	0.00
26 S	2,4-Dichlorophenol-d3	0.321	0.351	-9.3	97	0.00
27 C	2,4-Dichlorophenol	0.321	0.345	-7.5	96	0.00
28	Naphthalene	1.018	1.032	-1.4	89	0.00
29 S	4-Chloroaniline-d4	0.373	0.448	-20.1#	100	0.00
30	4-Chloroaniline	0.386	0.455	-17.9	97	0.00
31 C	Hexachlorobutadiene	0.198	0.212	-7.1	96	0.00
32	Caprolactam	0.131	0.133	-1.5	87	0.00
33 C	4-Chloro-3-methylphenol	0.342	0.359	-5.0	92	0.00
34	2-Methylnaphthalene	0.752	0.787	-4.7	93	0.00
35	1-Methylnaphthalene	0.721	0.750	-4.0	93	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
37	1,2,4,5-Tetrachlorobenzene	0.620	0.622	-0.3	100	0.00
38	Hexachlorocyclopentadiene	0.345	0.167	51.6#	50	0.00
39 C	2,4,6-Trichlorophenol	0.403	0.400	0.7	94	0.00
40	2,4,5-Trichlorophenol	0.434	0.435	-0.2	96	0.00
41	1,1'-Biphenyl	1.514	1.476	2.5	95	0.00
42	2-Chloronaphthalene	1.154	1.146	0.7	96	0.00
43	2-Nitroaniline	0.342	0.309	9.6	89	0.00
44 S	Dimethylphthalate-d6	1.610	1.594	1.0	98	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG093015\
 Data File : BG018967.D
 Acq On : 30 Sep 2015 10:43
 Operator : UM/NP
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02019

Quant Time: Sep 30 17:14:04 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Sep 30 04:08:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.560	1.554	0.4	97	0.00
46	2,6-Dinitrotoluene	0.314	0.352	-12.1	110	0.00
47 S	Acenaphthylene-d8	1.907	1.915	-0.4	97	0.00
48	Acenaphthylene	2.029	2.025	0.2	96	0.00
49	3-Nitroaniline	0.333	0.392	-17.7	107	0.00
50 C	Acenaphthene	1.367	1.345	1.6	96	0.00
51	2,4-Dinitrophenol	0.141	0.095	32.6#	82	0.00
52 S	4-Nitrophenol-d4	0.309	0.299	3.2	95	0.00
53	4-Nitrophenol	0.212	0.188	11.3	87	0.00
54	Dibenzofuran	1.918	1.931	-0.7	97	0.00
55	2,4-Dinitrotoluene	0.467	0.525	-12.4	108	0.00
56	2,3,4,6-Tetrachlorophenol	0.419	0.404	3.6	96	0.00
57	Diethylphthalate	1.622	1.597	1.5	96	0.00
58 S	Fluorene-d10	1.407	1.477	-5.0	101	0.00
59	Fluorene	1.620	1.655	-2.2	99	0.00
60	4-Chlorophenyl-phenylether	0.772	0.810	-4.9	104	0.00
61	4-Nitroaniline	0.383	0.439	-14.6	102	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.095	0.099	-4.2	118	0.00
64	4,6-Dinitro-2-methylphenol	0.097	0.101	-4.1	115	0.00
65	N-Nitrosodiphenylamine	0.524	0.521	0.6	100	0.00
66	4-Bromophenyl-phenylether	0.196	0.201	-2.6	106	0.00
67	Hexachlorobenzene	0.217	0.234	-7.8	113	0.00
68	Atrazine	0.231	0.243	-5.2	102	0.00
69 C	Pentachlorophenol	0.127	0.075	40.9#	62	0.00
70	Phenanthrene	1.017	1.046	-2.9	102	0.00
71 S	Anthracene-d10	0.881	0.916	-4.0	105	0.00
72	Anthracene	1.020	1.055	-3.4	102	0.00
73	1,2,3,4-Tetrachlorobenzene	0.238	0.223	6.3	95	0.00
74	Pentachlorobenzene	0.232	0.246	-6.0	106	0.00
75	Carbazole	0.906	1.019	-12.5	102	0.00
76	Di-n-butylphthalate	1.144	1.170	-2.3	99	0.00
77 C	Fluoranthene	1.127	1.335	-18.5	104	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
79 S	Pyrene-d10	0.920	0.979	-6.4	107	0.00
80	Pyrene	1.176	1.218	-3.6	101	0.00
81	Butylbenzylphthalate	0.451	0.488	-8.2	109	0.00
82	3,3'-Dichlorobenzidine	0.355	0.450	-26.8#	121	0.00
83	Benzo(a)anthracene	1.109	1.157	-4.3	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.669	0.696	-4.0	103	0.00
85	Chrysene	1.023	1.071	-4.7	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Di-n-octyl phthalate	1.058	1.159	-9.5	107	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG093015\
Data File : BG018967.D
Acq On : 30 Sep 2015 10:43
Operator : UM/NP
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02019

Quant Time: Sep 30 17:14:04 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG091015.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Sep 30 04:08:14 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.131	1.169	-3.4	108	0.00
89	Benzo(k)fluoranthene	1.083	1.084	-0.1	101	0.00
90 S	Benzo(a)pyrene-d12	0.973	1.009	-3.7	108	0.00
91 C	Benzo(a)pyrene	1.075	1.100	-2.3	105	0.00
92	Indeno(1,2,3-cd)pyrene	1.246	1.312	-5.3	110	0.00
93	Dibenzo(a,h)anthracene	1.000	1.064	-6.4	112	0.00
94	Benzo(a,h,i)perylene	1.040	1.082	-4.0	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1