

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050333.D
 Acq On : 30 Sep 2021 21:17
 Operator : CG/JU
 Sample : PB139240BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS240

Manual Integrations
 APPROVED

mohammad
 10/1/2021 11:30:11 AM

Quant Time: Oct 01 02:47:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.279	152	41207	20.000	ng/u1	0.00
20) Naphthalene-d8	11.099	136	166251	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.889	164	109707	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.638	188	259754	20.000	ng/u1	0.00
79) Chrysene-d12	21.939	240	253290	20.000	ng/u1	0.00
88) Perylene-d12	25.365	264	254813	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.643	96	7377m	6.577	ng/uL	0.00
4) Pyridine-d5	4.066	84	100248	29.995	ng/u1	0.00
7) Phenol-d5	7.403	99	122909	34.139	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.585	67	78616	34.696	ng/u1	0.00
11) 2-Chlorophenol-d4	7.797	132	87283	34.600	ng/u1	0.00
15) 4-Methylphenol-d8	8.960	113	91190	33.464	ng/u1	0.00
21) Nitrobenzene-d5	9.442	128	43243	37.181	ng/u1	0.00
24) 2-Nitrophenol-d4	10.171	143	45360	36.340	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.705	165	83700	33.958	ng/u1	0.00
31) 4-Chloroaniline-d4	11.222	131	117300	33.313	ng/u1	0.00
46) Dimethylphthalate-d6	14.289	166	263153	35.049	ng/u1	0.00
49) Acenaphthylene-d8	14.589	160	324354	34.400	ng/u1	0.00
54) 4-Nitrophenol-d4	15.053	143	52559	36.346	ng/u1	0.00
60) Fluorene-d10	15.876	176	234990	34.739	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.981	200	44748	34.539	ng/u1	0.00
73) Anthracene-d10	17.732	188	386383	35.210	ng/u1	0.00
81) Pyrene-d10	20.006	212	487851	36.222	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.130	264	435692	34.004	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.672	88	17184	12.664	ng/uL	94
5) Pyridine	4.084	79	111785	33.020	ng/u1	92
6) Benzaldehyde	7.403	77	90471	38.103	ng/u1	98
8) Phenol	7.433	94	131610	35.665	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.685	93	99073	35.793	ng/u1	98
12) 2-Chlorophenol	7.832	128	90965	35.246	ng/u1	93
13) 2-Methylphenol	8.702	108	96125	35.347	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.796	45	152854	36.807	ng/u1	98
16) Acetophenone	9.101	105	148768	34.360	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.078	70	88632	34.063	ng/u1	94
18) 4-Methylphenol	9.025	108	99390	34.510	ng/u1	99
19) Hexachloroethane	9.371	117	38685	34.876	ng/u1	95
22) Nitrobenzene	9.489	77	127413	36.554	ng/u1	97
23) Isophorone	10.012	82	229911	34.322	ng/u1	99
25) 2-Nitrophenol	10.200	139	49970	38.951	ng/u1	98
26) 2,4-Dimethylphenol	10.247	107	110309	34.639	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.488	93	126042	34.851	ng/u1	97
29) 2,4-Dichlorophenol	10.735	162	87957	36.044	ng/u1	99
30) Naphthalene	11.152	128	297402	34.843	ng/u1	98
32) 4-Chloroaniline	11.252	127	118297	33.370	ng/u1	99
33) Hexachlorobutadiene	11.428	225	61097	34.452	ng/u1	97
34) Caprolactam	12.015	113	32073m	34.019	ng/u1	
35) 4-Chloro-3-methylphenol	12.344	107	102503	35.628	ng/u1	96
36) 2-Methylnaphthalene	12.738	142	205593	35.429	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050333.D
 Acq On : 30 Sep 2021 21:17
 Operator : CG/JU
 Sample : PB139240BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS240


Manual Integrations
 APPROVED

mohammad
 10/1/2021 11:30:11 AM

Quant Time: Oct 01 02:47:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.950	142	197238	33.762	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	13.091	216	117141	37.000	ng/ul	96
40) Hexachlorocyclopentadiene	13.073	237	65822	31.350	ng/ul	90
41) 2,4,6-Trichlorophenol	13.326	196	75344	37.757	ng/ul	99
42) 2,4,5-Trichlorophenol	13.396	196	81315	37.252	ng/ul	98
43) 1,1'-Biphenyl	13.731	154	263266	35.659	ng/ul	98
44) 2-Chloronaphthalene	13.772	162	205586	35.544	ng/ul	99
45) 2-Nitroaniline	13.972	65	75187	39.705	ng/ul	95
47) Dimethylphthalate	14.336	163	272736	36.104	ng/ul	100
48) 2,6-Dinitrotoluene	14.454	165	55891	39.332	ng/ul	92
50) Acenaphthylene	14.618	152	313902	32.820	ng/ul	98
51) 3-Nitroaniline	14.783	138	59541	38.520	ng/ul#	89
52) Acenaphthene	14.953	153	229945	36.376	ng/ul	98
53) 2,4-Dinitrophenol	14.988	184	32543	36.742	ng/ul#	84
55) 4-Nitrophenol	15.071	109	54420	38.057	ng/ul	97
56) Dibenzofuran	15.288	168	326777	35.544	ng/ul	98
57) 2,4-Dinitrotoluene	15.235	165	80429	38.979	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.505	232	71042	37.846	ng/ul	97
59) Diethylphthalate	15.688	149	286951	36.404	ng/ul	99
61) Fluorene	15.934	166	260382	35.599	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.923	204	140829	35.553	ng/ul	96
63) 4-Nitroaniline	15.946	138	60830	38.970	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.999	198	46268	36.128	ng/ul	92
67) N-Nitrosodiphenylamine	16.134	169	237155	36.110	ng/ul	98
68) 4-Bromophenyl-phenylether	16.816	248	88275	36.502	ng/ul	95
69) Hexachlorobenzene	16.933	284	92345	36.262	ng/ul	98
70) Atrazine	17.074	200	98970	35.393	ng/ul	96
71) Pentachlorophenol	17.274	266	61749	37.025	ng/ul	98
72) Phenanthrene	17.679	178	463444	35.743	ng/ul	99
74) Anthracene	17.768	178	463024	36.058	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.696	216	122777	35.852	ng/uL	99
76) Pentachlorobenzene	15.200	250	106382	34.003	ng/uL	99
77) Carbazole	18.032	167	439077	38.201	ng/ul	99
78) Di-n-butylphthalate	18.578	149	514305	37.761	ng/ul	99
80) Fluoranthene	19.671	202	609023	36.008	ng/ul	100
82) Pyrene	20.035	202	597550	35.827	ng/ul	98
83) Butylbenzylphthalate	20.911	149	237731	38.670	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.822	252	198608	38.043	ng/ul	99
85) Benzo(a)anthracene	21.916	228	555027	36.348	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.804	149	333936	37.801	ng/ul	100
87) Chrysene	21.986	228	536462	36.660	ng/ul	100
89) Di-n-octyl phthalate	23.091	149	570882	38.990	ng/ul	100
90) Benzo(b)fluoranthene	24.272	252	566603	36.506	ng/ul	98
91) Benzo(k)fluoranthene	24.348	252	547334	37.898	ng/ul	99
93) Benzo(a)pyrene	25.206	252	504348	34.057	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	29.295	276	612680	36.850	ng/ul	99
95) Dibenzo(a,h)anthracene	29.383	278	515040	36.401	ng/ul	98
96) Benzo(g,h,i)perylene	30.535	276	506831	36.161	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.664 ng/uL
RT: 3.672 min Scan# 31
Delta R.T. -0.004 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion: 88 Resp: 17184
Ion Ratio Lower Upper
88 100
43 29.6 23.8 35.8
58 73.9 53.7 80.5

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#5
Pyridine
Concen: 33.020 ng/uL
RT: 4.084 min Scan# 101
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion: 79 Resp: 111785
Ion Ratio Lower Upper
79 100
52 86.2 61.8 92.8
51 40.7 30.6 45.8









#6
Benzaldehyde
Concen: 38.103 ng/ul
RT: 7.403 min Scan# 666
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion: 77 Resp: 90471
Ion Ratio Lower Upper
77 100
105 80.8 64.2 96.4
106 81.1 62.1 93.1

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#8
Phenol
Concen: 35.665 ng/ul
RT: 7.433 min Scan# 671
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17


Tgt Ion: 94 Resp: 131610
Ion Ratio Lower Upper
94 100
65 32.9 25.3 37.9
66 46.8 34.9 52.3





#10
Bis(2-Chloroethyl)ether
Concen: 35.793 ng/ul
RT: 7.685 min Scan# 714
Delta R.T. 0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:	93	Resp:	99073
Ion	Ratio	Lower	Upper
93	100		
63	82.3	66.9	100.3
95	30.8	26.1	39.1

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#12
2-Chlorophenol
Concen: 35.246 ng/ul
RT: 7.832 min Scan# 739
Delta R.T. 0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion:	128	Resp:	90965
Ion	Ratio	Lower	Upper
128	100		
64	59.0	43.2	64.8
130	32.5	22.4	33.6









#13
2-Methylphenol
Concen: 35.347 ng/ul
RT: 8.702 min Scan# 887
Delta R.T. 0.002 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion: 108 Resp: 96125
Ion Ratio Lower Upper
108 100
107 90.4 75.5 113.3

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM




#14
2,2'-oxybis(1-Chloropropane)
Concen: 36.807 ng/ul
RT: 8.796 min Scan# 903
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion: 45 Resp: 152854
Ion Ratio Lower Upper
45 100
77 15.1 11.4 17.2
79 11.3 8.6 13.0







#16
Acetophenone
Concen: 34.360 ng/ul
RT: 9.101 min Scan# 955
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:105 Resp: 148768
Ion Ratio Lower Upper
105 100
77 91.1 72.1 108.1
51 30.7 24.6 36.8




Instrument :
BNA_G
ClientSampleId :
SLCS240

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#17
N-Nitroso-di-n-propylamine
Concen: 34.063 ng/ul
RT: 9.078 min Scan# 951
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion: 70 Resp: 88632
Ion Ratio Lower Upper
70 100
42 58.8 50.4 75.6
101 8.0 7.8 11.6
130 16.5 15.9 23.9









#18
4-Methylphenol
Concen: 34.510 ng/ul
RT: 9.025 min Scan# 942
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:108 Resp: 99390
Ion Ratio Lower Upper
108 100
107 116.2 93.8 140.8

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#19
Hexachloroethane
Concen: 34.876 ng/ul
RT: 9.371 min Scan# 1001
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17


Tgt Ion:117 Resp: 38685
Ion Ratio Lower Upper
117 100
201 109.8 83.0 124.4
199 69.2 53.8 80.6





#22
Nitrobenzene
Concen: 36.554 ng/ul
RT: 9.489 min Scan# 1021
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion: 77 Resp: 127413
Ion Ratio Lower Upper
77 100
123 35.2 29.9 44.9
65 14.4 11.0 16.4

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#23
Isophorone
Concen: 34.322 ng/ul
RT: 10.012 min Scan# 1110
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion: 82 Resp: 229911
Ion Ratio Lower Upper
82 100
95 7.5 5.5 8.3
138 14.1 11.6 17.4









#25
2-Nitrophenol
Concen: 38.951 ng/ul
RT: 10.200 min Scan# 1142
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	139	Resp:	49970
Ion Ratio	Lower	Upper	
139	100		
65	62.7	49.1	73.7
109	37.0	28.6	42.8

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#26
2,4-Dimethylphenol
Concen: 34.639 ng/ul
RT: 10.247 min Scan# 1150
Delta R.T. 0.002 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	107	Resp:	110309
Ion Ratio	Lower	Upper	
107	100		
121	46.9	39.0	58.6
122	76.8	64.6	97.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.851 ng/ul
 RT: 10.488 min Scan# 1191
 Delta R.T. -0.003 min
 Lab File: BG050333.D
 Acq: 30 Sep 2021 21:17

Tgt Ion:	93	Resp:	126042
Ion	Ratio	Lower	Upper
93	100		
95	32.5	24.6	37.0
123	9.3	8.3	12.5

Instrument :
 BNA_G
 ClientSampled :
 SLCS240




Manual Integrations
 APPROVED


mohammad
 10/1/2021 11:30:11 AM



#29
 2,4-Dichlorophenol
 Concen: 36.044 ng/ul
 RT: 10.735 min Scan# 1233
 Delta R.T. -0.003 min
 Lab File: BG050333.D
 Acq: 30 Sep 2021 21:17


Tgt Ion:	162	Resp:	87957
Ion	Ratio	Lower	Upper
162	100		
164	62.4	49.4	74.0
98	36.3	29.0	43.4





#30
Naphthalene
Concen: 34.843 ng/uI
RT: 11.152 min Scan# 1304
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:128 Resp: 297402
Ion Ratio Lower Upper
128 100
129 11.5 8.2 12.2
127 14.1 11.0 16.4

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#32
4-Chloroaniline
Concen: 33.370 ng/uI
RT: 11.252 min Scan# 1321
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:127 Resp: 118297
Ion Ratio Lower Upper
127 100
129 30.7 24.3 36.5





#33
Hexachlorobutadiene
Concen: 34.452 ng/ul
RT: 11.428 min Scan# 1351
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:225 Resp: 61097
Ion Ratio Lower Upper
225 100
223 65.4 51.1 76.7
227 62.7 52.7 79.1

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#34
Caprolactam
Concen: 34.019 ng/ul m
RT: 12.015 min Scan# 1451
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:113 Resp: 32073
Ion Ratio Lower Upper
113 100
55 183.3 144.1 216.1
56 136.8 105.8 158.8





#35
4-Chloro-3-methylphenol
Concen: 35.628 ng/ul
RT: 12.344 min Scan# 1507
Delta R.T. 0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:107 Resp: 102503
Ion Ratio Lower Upper
107 100
144 23.0 16.8 25.2
142 72.1 55.4 83.2

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#36
2-Methylnaphthalene
Concen: 35.429 ng/ul
RT: 12.738 min Scan# 1574
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:142 Resp: 205593
Ion Ratio Lower Upper
142 100
141 94.2 72.0 108.0





#37
1-Methylnaphthalene
Concen: 33.762 ng/ul
RT: 12.950 min Scan# 1610
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:142 Resp: 197238
Ion Ratio Lower Upper
142 100
141 93.6 78.9 118.3
116 4.3 3.8 5.6

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.000 ng/ul
RT: 13.091 min Scan# 1634
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion:216 Resp: 117141
Ion Ratio Lower Upper
216 100
214 76.9 65.0 97.4
179 20.1 16.5 24.7
108 18.1 12.2 18.4









#40
Hexachlorocyclopentadiene
Concen: 31.350 ng/ul
RT: 13.073 min Scan# 1631
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	237	Resp:	65822
Ion Ratio	Lower	Upper	
237	100		
235	65.2	45.6	68.4
272	12.8	9.5	14.3

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM




#41
2,4,6-Trichlorophenol
Concen: 37.757 ng/ul
RT: 13.326 min Scan# 1674
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	196	Resp:	75344
Ion Ratio	Lower	Upper	
196	100		
198	94.6	75.1	112.7
200	30.3	23.7	35.5







#42
2,4,5-Trichlorophenol
Concen: 37.252 ng/ul
RT: 13.396 min Scan# 1686
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:196 Resp: 81315
Ion Ratio Lower Upper
196 100
198 95.0 78.3 117.5
200 30.6 24.6 37.0




Instrument :
BNA_G
ClientSampled :
SLCS240

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#43
1,1'-Biphenyl
Concen: 35.659 ng/ul
RT: 13.731 min Scan# 1743
Delta R.T. 0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:154 Resp: 263266
Ion Ratio Lower Upper
154 100
153 40.7 33.4 50.2
76 14.3 12.4 18.6





#44
2-Chloronaphthalene
Concen: 35.544 ng/ul
RT: 13.772 min Scan# 1750
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240




Tgt Ion:162 Resp: 205586
Ion Ratio Lower Upper
162 100
164 32.0 24.5 36.7
127 40.7 32.6 48.8

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#45
2-Nitroaniline
Concen: 39.705 ng/ul
RT: 13.972 min Scan# 1784
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion: 65 Resp: 75187
Ion Ratio Lower Upper
65 100
92 63.0 54.8 82.2
138 91.3 76.1 114.1





#47
Dimethylphthalate
Concen: 36.104 ng/ul
RT: 14.336 min Scan# 1846
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17






Tgt Ion:163 Resp: 272736
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.6
164 10.1 8.0 12.0


Instrument :
BNA_G
ClientSampleId :
SLCS240

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#48
2,6-Dinitrotoluene
Concen: 39.332 ng/ul
RT: 14.454 min Scan# 1866
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:165 Resp: 55891
Ion Ratio Lower Upper
165 100
89 70.9 62.3 93.5
121 22.6 20.8 31.2





#50
Acenaphthylene
Concen: 32.820 ng/ul
RT: 14.618 min Scan# 1894
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:152 Resp: 313902
Ion Ratio Lower Upper
152 100
151 21.4 17.4 26.2
153 14.1 10.3 15.5

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#51
3-Nitroaniline
Concen: 38.520 ng/ul
RT: 14.783 min Scan# 1922
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:138 Resp: 59541
Ion Ratio Lower Upper
138 100
108 13.3 8.6 12.8#
92 133.4 97.0 145.4





#52
Acenaphthene
Concen: 36.376 ng/uI
RT: 14.953 min Scan# 1951
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240




Tgt Ion:153 Resp: 229945
Ion Ratio Lower Upper
153 100
152 46.4 40.2 60.2
154 85.7 68.2 102.2

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#53
2,4-Dinitrophenol
Concen: 36.742 ng/uI
RT: 14.988 min Scan# 1957
Delta R.T. 0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:184 Resp: 32543
Ion Ratio Lower Upper
184 100
63 70.2 62.1 93.1
154 65.2 68.3 102.5#





#55
4-Nitrophenol
Concen: 38.057 ng/ul
RT: 15.071 min Scan# 1971
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240



Tgt Ion:109 Resp: 54420
Ion Ratio Lower Upper
109 100
139 104.4 81.4 122.0
65 125.6 97.8 146.6




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM




#56
Dibenzofuran
Concen: 35.544 ng/ul
RT: 15.288 min Scan# 2008
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:168 Resp: 326777
Ion Ratio Lower Upper
168 100
139 38.9 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 38.979 ng/ul
RT: 15.235 min Scan# 1999
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17





Tgt Ion:165 Resp: 80429
Ion Ratio Lower Upper
165 100
63 53.7 43.4 65.2
182 2.3 1.7 2.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.846 ng/ul
RT: 15.505 min Scan# 2045
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:232 Resp: 71042
Ion Ratio Lower Upper
232 100
131 41.3 34.3 51.5
130 2.9 2.8 4.2
166 26.3 22.4 33.6



Instrument :
BNA_G
ClientSampleId :
SLCS240

Manual Integrations
APPROVED

mohammad
10/1/2021 11:30:11 AM







#59
Diethylphthalate
Concen: 36.404 ng/ul
RT: 15.688 min Scan# 2076
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	149	Resp:	286951
Ion Ratio	Lower	Upper	
149	100		
177	22.5	18.2	27.4
150	13.2	10.2	15.2

Instrument :
BNA_G
ClientSampled :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#61
Fluorene
Concen: 35.599 ng/ul
RT: 15.934 min Scan# 2118
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	166	Resp:	260382
Ion Ratio	Lower	Upper	
166	100		
165	98.4	78.6	118.0
167	13.6	10.9	16.3









#62
4-Chlorophenyl-phenylether
Concen: 35.553 ng/ul
RT: 15.923 min Scan# 2116
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Instrument :
BNA_G
ClientSampled :
SLCS240

Tgt Ion:204 Resp: 140829
Ion Ratio Lower Upper
204 100
206 34.9 24.8 37.2
141 56.5 44.1 66.1




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#63
4-Nitroaniline
Concen: 38.970 ng/ul
RT: 15.946 min Scan# 2120
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17


Tgt Ion:138 Resp: 60830
Ion Ratio Lower Upper
138 100
92 57.7 47.0 70.4
108 96.6 77.2 115.8





#66
4,6-Dinitro-2-methylphenol
Concen: 36.128 ng/ul
RT: 15.999 min Scan# 2129
Delta R.T. 0.002 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240




Tgt Ion:198 Resp: 46268
Ion Ratio Lower Upper
198 100
182 3.1 3.0 4.6
77 27.3 25.8 38.8

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#67
N-Nitrosodiphenylamine
Concen: 36.110 ng/ul
RT: 16.134 min Scan# 2152
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:169 Resp: 237155
Ion Ratio Lower Upper
169 100
168 70.5 58.3 87.5
167 39.3 31.4 47.2





#68
4-Bromophenyl-phenylether
Concen: 36.502 ng/ul
RT: 16.816 min Scan# 2268
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampled :
SLCS240




Tgt Ion:248 Resp: 88275
Ion Ratio Lower Upper
248 100
250 98.2 73.3 109.9
141 77.2 60.2 90.2

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#69
Hexachlorobenzene
Concen: 36.262 ng/ul
RT: 16.933 min Scan# 2288
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion:284 Resp: 92345
Ion Ratio Lower Upper
284 100
142 41.4 33.3 49.9
249 37.7 28.3 42.5









#70
Atrazine
Concen: 35.393 ng/ul
RT: 17.074 min Scan# 2312
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:200 Resp: 98970
Ion Ratio Lower Upper
200 100
173 23.7 19.9 29.9
215 47.5 40.4 60.6

Instrument :
BNA_G
ClientSampled :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#71
Pentachlorophenol
Concen: 37.025 ng/ul
RT: 17.274 min Scan# 2346
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17


Tgt Ion:266 Resp: 61749
Ion Ratio Lower Upper
266 100
264 62.3 51.0 76.4
268 64.9 51.1 76.7





#72
Phenanthrene
Concen: 35.743 ng/u1
RT: 17.679 min Scan# 2415
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240




Tgt Ion:178 Resp: 463444
Ion Ratio Lower Upper
178 100
179 16.1 12.7 19.1
176 19.3 16.2 24.2

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#74
Anthracene
Concen: 36.058 ng/u1
RT: 17.768 min Scan# 2430
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion:178 Resp: 463024
Ion Ratio Lower Upper
178 100
179 16.0 12.7 19.1
176 19.0 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.852 ng/uL
RT: 13.696 min Scan# 1737
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:216 Resp: 122777
Ion Ratio Lower Upper
216 100
214 76.8 61.0 91.6
218 47.1 37.1 55.7

Instrument :
BNA_G
ClientSampled :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#76
Pentachlorobenzene
Concen: 34.003 ng/uL
RT: 15.200 min Scan# 1993
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:250 Resp: 106382
Ion Ratio Lower Upper
250 100
248 62.4 50.2 75.2
252 63.6 51.9 77.9





#77
 Carbazole
 Concen: 38.201 ng/ul
 RT: 18.032 min Scan# 2475
 Delta R.T. -0.003 min
 Lab File: BG050333.D
 Acq: 30 Sep 2021 21:17





Instrument :
 BNA_G
 ClientSampleId :
 SLCS240

Tgt Ion:167 Resp: 439077

Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.6	26.4
139	13.3	10.5	15.7

Manual Integrations
 APPROVED




mohammad
 10/1/2021 11:30:11 AM




#78
 Di-n-butylphthalate
 Concen: 37.761 ng/ul
 RT: 18.578 min Scan# 2568
 Delta R.T. -0.003 min
 Lab File: BG050333.D
 Acq: 30 Sep 2021 21:17

Tgt Ion:149 Resp: 514305

Ion	Ratio	Lower	Upper
149	100		
150	9.6	7.5	11.3
104	6.4	4.9	7.3









#80
Fluoranthene
Concen: 36.008 ng/u1
RT: 19.671 min Scan# 2754
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:202 Resp: 609023
Ion Ratio Lower Upper
202 100
101 9.6 7.7 11.5
100 8.2 6.6 10.0

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM




#82
Pyrene
Concen: 35.827 ng/u1
RT: 20.035 min Scan# 2816
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:202 Resp: 597550
Ion Ratio Lower Upper
202 100
101 11.2 9.7 14.5
100 10.3 7.6 11.4









#83
Butylbenzylphthalate
Concen: 38.670 ng/ul
RT: 20.911 min Scan# 2965
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17





Tgt Ion:149 Resp: 237731
Ion Ratio Lower Upper
149 100
91 78.7 65.8 98.6
206 22.8 17.2 25.8



#84
3,3'-Dichlorobenzidine
Concen: 38.043 ng/ul
RT: 21.822 min Scan# 3120
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Tgt Ion:252 Resp: 198608
Ion Ratio Lower Upper
252 100
254 64.6 52.2 78.2
126 10.5 8.3 12.5



Instrument :
BNA_G
ClientSampleId :
SLCS240

Manual Integrations
APPROVED

mohammad
10/1/2021 11:30:11 AM







#85
Benzo(a)anthracene
Concen: 36.348 ng/ul
RT: 21.916 min Scan# 3136
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	228	Resp:	555027
Ion Ratio	Lower	Upper	
228	100		
229	20.8	15.9	23.9
226	29.5	22.2	33.2

Instrument :
BNA_G
ClientSampled :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.801 ng/ul
RT: 21.804 min Scan# 3117
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:	149	Resp:	333936
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.3	31.9
279	3.9	2.8	4.2









#87
Chrysene
Concen: 36.660 ng/ul
RT: 21.986 min Scan# 3148
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:228 Resp: 536462
Ion Ratio Lower Upper
228 100
226 31.5 25.4 38.0
229 20.5 16.2 24.2

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#89
Di-n-octyl phthalate
Concen: 38.990 ng/ul
RT: 23.091 min Scan# 3336
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17


Tgt Ion:149 Resp: 570882





#90
Benzo(b)fluoranthene
Concen: 36.506 ng/ul
RT: 24.272 min Scan# 3537
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17




Instrument :
BNA_G
ClientSampleId :
SLCS240




Tgt Ion:252 Resp: 566603
Ion Ratio Lower Upper
252 100
253 23.1 17.8 26.8
125 9.1 6.7 10.1

Manual Integrations
APPROVED



mohammad
10/1/2021 11:30:11 AM




#91
Benzo(k)fluoranthene
Concen: 37.898 ng/ul
RT: 24.348 min Scan# 3550
Delta R.T. -0.003 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17



Tgt Ion:252 Resp: 547334
Ion Ratio Lower Upper
252 100
253 21.6 17.1 25.7
125 8.3 7.4 11.0









#93
Benzo(a)pyrene
Concen: 34.057 ng/ul
RT: 25.206 min Scan# 3696
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion: 252 Resp: 504348
Ion Ratio Lower Upper
252 100
253 23.2 17.8 26.6
125 9.0 8.3 12.5

Instrument :
BNA_G
ClientSampleId :
SLCS240




Manual Integrations
APPROVED


mohammad
10/1/2021 11:30:11 AM



#94
Indeno(1,2,3-cd)pyrene
Concen: 36.850 ng/ul
RT: 29.295 min Scan# 4392
Delta R.T. -0.021 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion: 276 Resp: 612680
Ion Ratio Lower Upper
276 100
138 16.1 12.4 18.6
277 23.6 18.5 27.7









#95
Dibenzo(a,h)anthracene
Concen: 36.401 ng/ul
RT: 29.383 min Scan# 4407
Delta R.T. -0.009 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:278 Resp: 515040
Ion Ratio Lower Upper
278 100
139 12.8 11.3 16.9
279 23.9 18.3 27.5

Instrument :
BNA_G
ClientSampleId :
SLCS240

Manual Integrations
APPROVED

mohammad
10/1/2021 11:30:11 AM



#96
Benzo(g,h,i)perylene
Concen: 36.161 ng/ul
RT: 30.535 min Scan# 4603
Delta R.T. -0.021 min
Lab File: BG050333.D
Acq: 30 Sep 2021 21:17

Tgt Ion:276 Resp: 506831
Ion Ratio Lower Upper
276 100
138 17.4 14.1 21.1
277 22.3 17.9 26.9

