

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050353.D
 Acq On : 1 Oct 2021 11:35
 Operator : CG/JU
 Sample : M3876-02MS
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW7W1MS

Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:50:07 AM

Quant Time: Oct 01 12:59:12 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.274	152	40865	20.00	ng/ul	0.00
20) Naphthalene-d8	11.100	136	175124	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.890	164	123607	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.634	188	282929	20.00	ng/ul	0.00
79) Chrysene-d12	21.934	240	245167	20.00	ng/ul	-0.01
88) Perylene-d12	25.366	264	244620	20.00	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.632	96	2855	2.57	ng/uL	0.00
4) Pyridine-d5	4.061	84	22742	6.86	ng/ul	0.00
7) Phenol-d5	7.398	99	55357	15.50	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.586	67	36388	16.19	ng/ul	0.00
11) 2-Chlorophenol-d4	7.798	132	38143	15.25	ng/ul	0.00
15) 4-Methylphenol-d8	8.961	113	39393	14.58	ng/ul	0.00
21) Nitrobenzene-d5	9.437	128	21023	17.16	ng/ul	0.00
24) 2-Nitrophenol-d4	10.166	143	21496	16.35	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.706	165	41532	16.00	ng/ul	0.00
31) 4-Chloroaniline-d4	11.223	131	55443	14.95	ng/ul	0.00
46) Dimethylphthalate-d6	14.285	166	140723	16.64	ng/ul	0.00
49) Acenaphthylene-d8	14.584	160	174709	16.45	ng/ul	0.00
54) 4-Nitrophenol-d4	15.054	143	27163	16.67	ng/ul	0.00
60) Fluorene-d10	15.877	176	131452	17.25	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.983	200	22498	15.94	ng/ul	0.00
73) Anthracene-d10	17.733	188	209010	17.49	ng/ul	0.00
81) Pyrene-d10	20.007	212	249323	19.13	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.125	264	204827	16.65	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.673	88	9154	6.80	ng/uL#	88
5) Pyridine	4.085	79	54977	16.38	ng/ul	94
6) Benzaldehyde	7.404	77	57193	24.29	ng/ul	96
8) Phenol	7.428	94	84901	23.20	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.680	93	60180	21.92	ng/ul	97
12) 2-Chlorophenol	7.827	128	56588	22.11	ng/ul	93
13) 2-Methylphenol	8.697	108	59467	22.05	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.797	45	95134	23.10	ng/ul#	95
16) Acetophenone	9.096	105	99636	23.21	ng/ul	98
17) N-Nitroso-di-n-propyla...	9.073	70	58742	22.76	ng/ul#	94
18) 4-Methylphenol	9.026	108	64931	22.73	ng/ul	94
19) Hexachloroethane	9.367	117	23918	21.74	ng/ul	93
22) Nitrobenzene	9.484	77	83119	22.64	ng/ul	99
23) Isophorone	10.007	82	155471	22.03	ng/ul	100
25) 2-Nitrophenol	10.201	139	31585	23.37	ng/ul	98
26) 2,4-Dimethylphenol	10.242	107	58693	17.50	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.489	93	84552	22.19	ng/ul	97
29) 2,4-Dichlorophenol	10.730	162	59362	23.09	ng/ul	97
30) Naphthalene	11.153	128	199235	22.16	ng/ul	99
32) 4-Chloroaniline	11.247	127	81097	21.72	ng/ul	96
33) Hexachlorobutadiene	11.429	225	39305	21.04	ng/ul	93
34) Caprolactam	12.011	113	23038	23.20	ng/ul	80
35) 4-Chloro-3-methylphenol	12.340	107	74000	24.42	ng/ul	93
36) 2-Methylnaphthalene	12.733	142	143709	23.51	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050353.D
 Acq On : 1 Oct 2021 11:35
 Operator : CG/JU
 Sample : M3876-02MS
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW7W1MS


Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:50:07 AM

Quant Time: Oct 01 12:59:12 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.951	142	138750	22.55	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.092	216	80059	22.44	ng/ul#	97
40) Hexachlorocyclopentadiene	13.068	237	37874	16.01	ng/ul	91
41) 2,4,6-Trichlorophenol	13.327	196	54404	24.20	ng/ul	99
42) 2,4,5-Trichlorophenol	13.391	196	57834	23.52	ng/ul	98
43) 1,1'-Biphenyl	13.726	154	190035	22.85	ng/ul	98
44) 2-Chloronaphthalene	13.773	162	150114	23.03	ng/ul	98
45) 2-Nitroaniline	13.967	65	52767	24.73	ng/ul	99
47) Dimethylphthalate	14.332	163	197322	23.18	ng/ul	97
48) 2,6-Dinitrotoluene	14.455	165	39006	24.36	ng/ul	98
50) Acenaphthylene	14.614	152	231259	21.46	ng/ul	98
51) 3-Nitroaniline	14.784	138	42400	24.35	ng/ul	99
52) Acenaphthene	14.954	153	165347	23.22	ng/ul	99
53) 2,4-Dinitrophenol	14.984	184	22808	22.85	ng/ul	91
55) 4-Nitrophenol	15.066	109	38176	23.70	ng/ul	95
56) Dibenzofuran	15.283	168	237350	22.91	ng/ul	95
57) 2,4-Dinitrotoluene	15.236	165	55629	23.93	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.501	232	51480	24.34	ng/ul	91
59) Diethylphthalate	15.689	149	213180	24.00	ng/ul	99
61) Fluorene	15.936	166	193519	23.48	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.918	204	102419	22.95	ng/ul	94
63) 4-Nitroaniline	15.941	138	44170	25.11	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.994	198	32327	23.17	ng/ul	95
67) N-Nitrosodiphenylamine	16.129	169	174642	24.41	ng/ul	97
68) 4-Bromophenyl-phenylether	16.817	248	63274	24.02	ng/ul	93
69) Hexachlorobenzene	16.934	284	66076	23.82	ng/ul	98
70) Atrazine	17.069	200	72933	23.95	ng/ul	98
71) Pentachlorophenol	17.269	266	41457	22.82	ng/ul	98
72) Phenanthrene	17.675	178	339289	24.02	ng/ul	99
74) Anthracene	17.769	178	335614	24.00	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.697	216	85302	22.87	ng/uL	94
76) Pentachlorobenzene	15.201	250	76931	22.58	ng/uL	98
77) Carbazole	18.033	167	314965	25.16	ng/ul	100
78) Di-n-butylphthalate	18.574	149	381955	25.75	ng/ul	99
80) Fluoranthene	19.672	202	419782	25.64	ng/ul	99
82) Pyrene	20.037	202	415159	25.72	ng/ul	98
83) Butylbenzylphthalate	20.906	149	155081	26.06	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.823	252	95756	18.95	ng/ul	98
85) Benzo(a)anthracene	21.917	228	369396	24.99	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.805	149	228812	26.76	ng/ul	97
87) Chrysene	21.987	228	354313	25.01	ng/ul	100
89) Di-n-octyl phthalate	23.092	149	379167	26.98	ng/ul	100
90) Benzo(b)fluoranthene	24.267	252	365599	24.54	ng/ul	98
91) Benzo(k)fluoranthene	24.343	252	356099	25.68	ng/ul	98
93) Benzo(a)pyrene	25.201	252	319080	22.44	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	29.296	276	402330m	25.21	ng/ul	
95) Dibenzo(a,h)anthracene	29.379	278	337341	24.84	ng/ul	98
96) Benzo(g,h,i)perylene	30.536	276	283056	21.04	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 6.80 ng/uL
RT: 3.673 min Scan# 31
Delta R.T. -0.003 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 88 Resp: 9154
Ion Ratio Lower Upper
88 100
43 30.0 23.8 35.8
58 80.8 53.7 80.5#

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#5
Pyridine
Concen: 16.38 ng/uL
RT: 4.085 min Scan# 101
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35


Tgt Ion: 79 Resp: 54977
Ion Ratio Lower Upper
79 100
52 70.7 61.8 92.8
51 40.2 30.6 45.8





#6
Benzaldehyde
Concen: 24.29 ng/ul
RT: 7.404 min Scan# 666
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion: 77 Resp: 57193
Ion Ratio Lower Upper
77 100
105 83.7 64.2 96.4
106 80.8 62.1 93.1

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#8
Phenol
Concen: 23.20 ng/ul
RT: 7.428 min Scan# 670
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion: 94 Resp: 84901
Ion Ratio Lower Upper
94 100
65 34.1 25.3 37.9
66 43.4 34.9 52.3





#10
Bis(2-Chloroethyl)ether
Concen: 21.92 ng/ul
RT: 7.680 min Scan# 713
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Instrument :
BNA_G
ClientSampleId :
EW7W1MS






Tgt Ion: 93 Resp: 60180


Ion	Ratio	Lower	Upper
93	100		
63	86.7	66.9	100.3
95	34.3	26.1	39.1

Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM






#12
2-Chlorophenol
Concen: 22.11 ng/ul
RT: 7.827 min Scan# 738
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion: 128 Resp: 56588

Ion	Ratio	Lower	Upper
128	100		
64	58.9	43.2	64.8
130	32.2	22.4	33.6









#13
2-Methylphenol
Concen: 22.05 ng/ul
RT: 8.697 min Scan# 886
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 108 Resp: 59467
Ion Ratio Lower Upper
108 100
107 94.2 75.5 113.3

Instrument :
BNA_G
ClientSampled :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#14
2,2'-oxybis(1-chloropropane)
Concen: 23.10 ng/ul
RT: 8.797 min Scan# 903
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 45 Resp: 95134
Ion Ratio Lower Upper
45 100
77 16.3 11.4 17.2
79 8.6 8.6 13.0#









#16
Acetophenone
Concen: 23.21 ng/ul
RT: 9.096 min Scan# 954
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:105 Resp: 99636
Ion Ratio Lower Upper
105 100
77 87.6 72.1 108.1
51 30.9 24.6 36.8

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#17
N-Nitroso-di-n-propylamine
Concen: 22.76 ng/ul
RT: 9.073 min Scan# 950
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 70 Resp: 58742
Ion Ratio Lower Upper
70 100
42 59.3 50.4 75.6
101 8.5 7.8 11.6
130 14.2 15.9 23.9#









#18
4-Methylphenol
Concen: 22.73 ng/ul
RT: 9.026 min Scan# 942
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Instrument :
BNA_G
ClientSampled :
EW7W1MS

Tgt Ion:108 Resp: 64931
Ion Ratio Lower Upper
108 100
107 111.0 93.8 140.8




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#19
Hexachloroethane
Concen: 21.74 ng/ul
RT: 9.367 min Scan# 1000
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:117 Resp: 23918
Ion Ratio Lower Upper
117 100
201 112.1 83.0 124.4
199 71.7 53.8 80.6









#22
Nitrobenzene
Concen: 22.64 ng/ul
RT: 9.484 min Scan# 1020
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 77	Resp: 83119
Ion Ratio	Lower Upper
77 100	
123 36.9	29.9 44.9
65 14.7	11.0 16.4

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM




#23
Isophorone
Concen: 22.03 ng/ul
RT: 10.007 min Scan# 1109
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion: 82	Resp: 155471
Ion Ratio	Lower Upper
82 100	
95 6.8	5.5 8.3
138 14.7	11.6 17.4









#25
2-Nitrophenol
Concen: 23.37 ng/ul
RT: 10.201 min Scan# 1142
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35





Tgt Ion:139 Resp: 31585
Ion Ratio Lower Upper
139 100
65 62.7 49.1 73.7
109 34.5 28.6 42.8



#26
2,4-Dimethylphenol
Concen: 17.50 ng/ul
RT: 10.242 min Scan# 1149
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:107 Resp: 58693
Ion Ratio Lower Upper
107 100
121 45.4 39.0 58.6
122 78.6 64.6 97.0



Instrument :
BNA_G
ClientSampled :
EW7W1MS


Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM



#27
Bis(2-Chloroethoxy)methane
Concen: 22.19 ng/ul
RT: 10.489 min Scan# 1191
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion: 93 Resp: 84552
Ion Ratio Lower Upper
93 100
95 32.9 24.6 37.0
123 10.2 8.3 12.5

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#29
2,4-Dichlorophenol
Concen: 23.09 ng/ul
RT: 10.730 min Scan# 1232
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion: 162 Resp: 59362
Ion Ratio Lower Upper
162 100
164 63.5 49.4 74.0
98 39.0 29.0 43.4





#30
Naphthalene
Concen: 22.16 ng/ul
RT: 11.153 min Scan# 1304
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:128 Resp: 199235
Ion Ratio Lower Upper
128 100
129 11.3 8.2 12.2
127 13.6 11.0 16.4

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#32
4-Chloroaniline
Concen: 21.72 ng/ul
RT: 11.247 min Scan# 1320
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:127 Resp: 81097
Ion Ratio Lower Upper
127 100
129 32.3 24.3 36.5









#33
Hexachlorobutadiene
Concen: 21.04 ng/ul
RT: 11.429 min Scan# 1351
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35





Tgt Ion:225 Resp: 39305
Ion Ratio Lower Upper
225 100
223 59.0 51.1 76.7
227 60.0 52.7 79.1



#34
Caprolactam
Concen: 23.20 ng/ul
RT: 12.011 min Scan# 1450
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:113 Resp: 23038
Ion Ratio Lower Upper
113 100
55 212.7 144.1 216.1
56 151.9 105.8 158.8



Instrument :
BNA_G
ClientSampleId :
EW7W1MS


Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM



#35
4-Chloro-3-methylphenol
Concen: 24.42 ng/ul
RT: 12.340 min Scan# 1506
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:107 Resp: 74000
Ion Ratio Lower Upper
107 100
144 22.8 16.8 25.2
142 76.0 55.4 83.2

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#36
2-Methylnaphthalene
Concen: 23.51 ng/ul
RT: 12.733 min Scan# 1573
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:142 Resp: 143709
Ion Ratio Lower Upper
142 100
141 92.7 72.0 108.0





#37
1-Methylnaphthalene
Concen: 22.55 ng/ul
RT: 12.951 min Scan# 1610
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:142 Resp: 138750
Ion Ratio Lower Upper
142 100
141 98.3 78.9 118.3
116 4.7 3.8 5.6

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.44 ng/ul
RT: 13.092 min Scan# 1634
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion:216 Resp: 80059
Ion Ratio Lower Upper
216 100
214 79.1 65.0 97.4
179 20.0 16.5 24.7
108 19.1 12.2 18.4#









#40
Hexachlorocyclopentadiene
Concen: 16.01 ng/ul
RT: 13.068 min Scan# 1630
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:	237	Resp:	37874
Ion	Ratio	Lower	Upper
237	100		
235	64.7	45.6	68.4
272	13.6	9.5	14.3

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#41
2,4,6-Trichlorophenol
Concen: 24.20 ng/ul
RT: 13.327 min Scan# 1674
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35


Tgt Ion:	196	Resp:	54404
Ion	Ratio	Lower	Upper
196	100		
198	93.5	75.1	112.7
200	29.0	23.7	35.5





#42
2,4,5-Trichlorophenol
Concen: 23.52 ng/ul
RT: 13.391 min Scan# 1685
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:196 Resp: 57834
Ion Ratio Lower Upper
196 100
198 98.7 78.3 117.5
200 28.0 24.6 37.0

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#43
1,1'-Biphenyl
Concen: 22.85 ng/ul
RT: 13.726 min Scan# 1742
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:154 Resp: 190035
Ion Ratio Lower Upper
154 100
153 43.6 33.4 50.2
76 15.1 12.4 18.6





#44
2-Chloronaphthalene
Concen: 23.03 ng/ul
RT: 13.773 min Scan# 1750
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:162 Resp: 150114
Ion Ratio Lower Upper
162 100
164 30.6 24.5 36.7
127 38.1 32.6 48.8

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#45
2-Nitroaniline
Concen: 24.73 ng/ul
RT: 13.967 min Scan# 1783
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion: 65 Resp: 52767
Ion Ratio Lower Upper
65 100
92 69.2 54.8 82.2
138 96.0 76.1 114.1





#47
Dimethylphthalate
Concen: 23.18 ng/ul
RT: 14.332 min Scan# 1845
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:163 Resp: 197322
Ion Ratio Lower Upper
163 100
194 4.2 3.0 4.6
164 11.1 8.0 12.0

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#48
2,6-Dinitrotoluene
Concen: 24.36 ng/ul
RT: 14.455 min Scan# 1866
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:165 Resp: 39006
Ion Ratio Lower Upper
165 100
89 76.3 62.3 93.5
121 24.3 20.8 31.2





#50
Acenaphthylene
Concen: 21.46 ng/ul
RT: 14.614 min Scan# 1893
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:152 Resp: 231259
Ion Ratio Lower Upper
152 100
151 21.4 17.4 26.2
153 14.1 10.3 15.5

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#51
3-Nitroaniline
Concen: 24.35 ng/ul
RT: 14.784 min Scan# 1922
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:138 Resp: 42400
Ion Ratio Lower Upper
138 100
108 11.7 8.6 12.8
92 122.2 97.0 145.4





#52
Acenaphthene
Concen: 23.22 ng/ul
RT: 14.954 min Scan# 1951
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:153 Resp: 165347
Ion Ratio Lower Upper
153 100
152 49.0 40.2 60.2
154 86.4 68.2 102.2

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#53
2,4-Dinitrophenol
Concen: 22.85 ng/ul
RT: 14.984 min Scan# 1956
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:184 Resp: 22808
Ion Ratio Lower Upper
184 100
63 74.6 62.1 93.1
154 72.9 68.3 102.5





#55
4-Nitrophenol
Concen: 23.70 ng/ul
RT: 15.066 min Scan# 1970
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:109 Resp: 38176
Ion Ratio Lower Upper
109 100
139 105.8 81.4 122.0
65 129.3 97.8 146.6

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#56
Dibenzofuran
Concen: 22.91 ng/ul
RT: 15.283 min Scan# 2007
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:168 Resp: 237350
Ion Ratio Lower Upper
168 100
139 40.8 30.2 45.4







#57
2,4-Dinitrotoluene
Concen: 23.93 ng/ul
RT: 15.236 min Scan# 1999
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:165 Resp: 55629
Ion Ratio Lower Upper
165 100
63 56.1 43.4 65.2
182 1.9 1.7 2.5




Instrument :
BNA_G
ClientSampleId :
EW7W1MS

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#58
2,3,4,6-Tetrachlorophenol
Concen: 24.34 ng/ul
RT: 15.501 min Scan# 2044
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:232 Resp: 51480
Ion Ratio Lower Upper
232 100
131 48.6 34.3 51.5
130 3.6 2.8 4.2
166 33.2 22.4 33.6





#59
Diethylphthalate
Concen: 24.00 ng/ul
RT: 15.689 min Scan# 2076
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:149 Resp: 213180
Ion Ratio Lower Upper
149 100
177 22.5 18.2 27.4
150 12.2 10.2 15.2

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#61
Fluorene
Concen: 23.48 ng/ul
RT: 15.936 min Scan# 2118
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion:166 Resp: 193519
Ion Ratio Lower Upper
166 100
165 100.9 78.6 118.0
167 13.6 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 22.95 ng/ul
RT: 15.918 min Scan# 2115
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Instrument :
BNA_G
ClientSampled :
EW7W1MS

Tgt Ion:204 Resp: 102419
Ion Ratio Lower Upper
204 100
206 33.1 24.8 37.2
141 60.0 44.1 66.1




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#63
4-Nitroaniline
Concen: 25.11 ng/ul
RT: 15.941 min Scan# 2119
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35


Tgt Ion:138 Resp: 44170
Ion Ratio Lower Upper
138 100
92 58.1 47.0 70.4
108 96.4 77.2 115.8





#66
4,6-Dinitro-2-methylphenol
Concen: 23.17 ng/ul
RT: 15.994 min Scan# 2128
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:198 Resp: 32327
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.6
77 29.1 25.8 38.8

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#67
N-Nitrosodiphenylamine
Concen: 24.41 ng/ul
RT: 16.129 min Scan# 2151
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:169 Resp: 174642
Ion Ratio Lower Upper
169 100
168 71.7 58.3 87.5
167 35.9 31.4 47.2





#68
4-Bromophenyl-phenylether
Concen: 24.02 ng/ul
RT: 16.817 min Scan# 2268
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampled :
EW7W1MS




Tgt Ion:248 Resp: 63274
Ion Ratio Lower Upper
248 100
250 100.0 73.3 109.9
141 78.9 60.2 90.2

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#69
Hexachlorobenzene
Concen: 23.82 ng/ul
RT: 16.934 min Scan# 2288
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion:284 Resp: 66076
Ion Ratio Lower Upper
284 100
142 39.6 33.3 49.9
249 35.9 28.3 42.5









#70
Atrazine
Concen: 23.95 ng/ul
RT: 17.069 min Scan# 2311
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:200 Resp: 72933
Ion Ratio Lower Upper
200 100
173 25.1 19.9 29.9
215 48.6 40.4 60.6

Instrument :
BNA_G
ClientSampled :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#71
Pentachlorophenol
Concen: 22.82 ng/ul
RT: 17.269 min Scan# 2345
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35


Tgt Ion:266 Resp: 41457
Ion Ratio Lower Upper
266 100
264 62.9 51.0 76.4
268 65.8 51.1 76.7





#72
Phenanthrene
Concen: 24.02 ng/u1
RT: 17.675 min Scan# 2414
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:178 Resp: 339289
Ion Ratio Lower Upper
178 100
179 16.0 12.7 19.1
176 19.1 16.2 24.2

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#74
Anthracene
Concen: 24.00 ng/u1
RT: 17.769 min Scan# 2430
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:178 Resp: 335614
Ion Ratio Lower Upper
178 100
179 15.7 12.7 19.1
176 19.1 15.0 22.4





#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.87 ng/uL
RT: 13.697 min Scan# 1737
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:216 Resp: 85302
Ion Ratio Lower Upper
216 100
214 79.9 61.0 91.6
218 51.8 37.1 55.7

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#76
Pentachlorobenzene
Concen: 22.58 ng/uL
RT: 15.201 min Scan# 1993
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion:250 Resp: 76931
Ion Ratio Lower Upper
250 100
248 64.4 50.2 75.2
252 63.2 51.9 77.9









#77
Carbazole
Concen: 25.16 ng/ul
RT: 18.033 min Scan# 2475
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:167 Resp: 314965
Ion Ratio Lower Upper
167 100
166 21.9 17.6 26.4
139 12.9 10.5 15.7

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM



#78
Di-n-butylphthalate
Concen: 25.75 ng/ul
RT: 18.574 min Scan# 2567
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:149 Resp: 381955
Ion Ratio Lower Upper
149 100
150 9.1 7.5 11.3
104 6.6 4.9 7.3









#80
Fluoranthene
Concen: 25.64 ng/ul
RT: 19.672 min Scan# 2754
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:202 Resp: 419782
Ion Ratio Lower Upper
202 100
101 9.4 7.7 11.5
100 7.6 6.6 10.0

Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM




#82
Pyrene
Concen: 25.72 ng/ul
RT: 20.037 min Scan# 2816
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:202 Resp: 415159
Ion Ratio Lower Upper
202 100
101 10.7 9.7 14.5
100 9.7 7.6 11.4







#83
Butylbenzylphthalate
Concen: 26.06 ng/ul
RT: 20.906 min Scan# 2964
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:149 Resp: 155081
Ion Ratio Lower Upper
149 100
91 81.1 65.8 98.6
206 20.5 17.2 25.8




Instrument :
BNA_G
ClientSampleId :
EW7W1MS

Manual Integrations
APPROVED



mohammad
10/4/2021 11:50:07 AM




#84
3,3'-Dichlorobenzidine
Concen: 18.95 ng/ul
RT: 21.823 min Scan# 3120
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:252 Resp: 95756
Ion Ratio Lower Upper
252 100
254 63.6 52.2 78.2
126 9.2 8.3 12.5








#85
Benzo(a)anthracene
Concen: 24.99 ng/u1
RT: 21.917 min Scan# 3136
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35


Instrument :
BNA_G
ClientSampled :
EW7W1MS





Tgt Ion:228 Resp: 369396
Ion Ratio Lower Upper
228 100
229 20.4 15.9 23.9
226 28.3 22.2 33.2




#86
Bis(2-ethylhexyl)phthalate
Concen: 26.76 ng/u1
RT: 21.805 min Scan# 3117
Delta R.T. -0.002 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:149 Resp: 228812
Ion Ratio Lower Upper
149 100
167 24.6 21.3 31.9
279 3.5 2.8 4.2









#87
Chrysene
Concen: 25.01 ng/u1
RT: 21.987 min Scan# 3148
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35





Tgt Ion:228 Resp: 354313
Ion Ratio Lower Upper
228 100
226 31.3 25.4 38.0
229 20.2 16.2 24.2



#89
Di-n-octyl phthalate
Concen: 26.98 ng/u1
RT: 23.092 min Scan# 3336
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:149 Resp: 379167



Instrument :
BNA_G
ClientSampleId :
EW7W1MS

Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM







#90
Benzo(b)fluoranthene
Concen: 24.54 ng/ul
RT: 24.267 min Scan# 3536
Delta R.T. -0.014 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:252 Resp: 365599
Ion Ratio Lower Upper
252 100
253 23.2 17.8 26.8
125 9.7 6.7 10.1

Instrument :
BNA_G
ClientSampled :
EW7W1MS




Manual Integrations
APPROVED


mohammad
10/4/2021 11:50:07 AM




#91
Benzo(k)fluoranthene
Concen: 25.68 ng/ul
RT: 24.343 min Scan# 3549
Delta R.T. -0.008 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35

Tgt Ion:252 Resp: 356099
Ion Ratio Lower Upper
252 100
253 22.8 17.1 25.7
125 9.6 7.4 11.0









#93
Benzo(a)pyrene
Concen: 22.44 ng/u1
RT: 25.201 min Scan# 3695
Delta R.T. -0.014 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35





Tgt Ion:252 Resp: 319080
Ion Ratio Lower Upper
252 100
253 22.8 17.8 26.6
125 9.6 8.3 12.5



#94
Indeno(1,2,3-cd)pyrene
Concen: 25.21 ng/u1 m
RT: 29.296 min Scan# 4392
Delta R.T. -0.020 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Tgt Ion:276 Resp: 402330
Ion Ratio Lower Upper
276 100
138 16.4 12.4 18.6
277 24.3 18.5 27.7



Instrument :
BNA_G
ClientSampleId :
EW7W1MS


Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM



#95
Dibenzo(a,h)anthracene
Concen: 24.84 ng/ul
RT: 29.379 min Scan# 4406
Delta R.T. -0.014 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35




Instrument :
BNA_G
ClientSampleId :
EW7W1MS




Tgt Ion:278 Resp: 337341
Ion Ratio Lower Upper
278 100
139 12.3 11.3 16.9
279 23.1 18.3 27.5

Manual Integrations
APPROVED

mohammad
10/4/2021 11:50:07 AM



#96
Benzo(g,h,i)perylene
Concen: 21.04 ng/ul
RT: 30.536 min Scan# 4603
Delta R.T. -0.020 min
Lab File: BG050353.D
Acq: 1 Oct 2021 11:35



Tgt Ion:276 Resp: 283056
Ion Ratio Lower Upper
276 100
138 17.7 14.1 21.1
277 21.9 17.9 26.9

