

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050359.D
 Acq On : 1 Oct 2021 15:40
 Operator : CG/JU
 Sample : M3876-07
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW7W3

Quant Time: Oct 01 17:06:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.273	152	48227	20.00	ng/ul	0.00
20) Naphthalene-d8	11.099	136	198533	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.889	164	135116	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.633	188	312288	20.00	ng/ul	0.00
79) Chrysene-d12	21.934	240	308691	20.00	ng/ul	-0.01
88) Perylene-d12	25.365	264	301531	20.00	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.644	96	6968	5.31	ng/uL	0.00
4) Pyridine-d5	4.067	84	104405	26.69	ng/ul	0.00
7) Phenol-d5	7.404	99	125905	29.88	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.586	67	82309	31.04	ng/ul	0.00
11) 2-Chlorophenol-d4	7.797	132	86719	29.37	ng/ul	0.00
15) 4-Methylphenol-d8	8.961	113	100059	31.37	ng/ul	0.00
21) Nitrobenzene-d5	9.443	128	48875	35.19	ng/ul	0.00
24) 2-Nitrophenol-d4	10.171	143	53797	36.09	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.706	165	94431	32.08	ng/ul	0.00
31) 4-Chloroaniline-d4	11.223	131	139097	33.08	ng/ul	0.00
46) Dimethylphthalate-d6	14.284	166	346490	37.47	ng/ul	0.00
49) Acenaphthylene-d8	14.584	160	420237	36.19	ng/ul	0.00
54) 4-Nitrophenol-d4	15.054	143	63259	35.52	ng/ul	0.00
60) Fluorene-d10	15.876	176	306454	36.78	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.982	200	50877	32.66	ng/ul	0.00
73) Anthracene-d10	17.733	188	498576	37.79	ng/ul	0.00
81) Pyrene-d10	20.007	212	610461	37.19	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.124	264	540992	35.68	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
Data File : BG050359.D
Acq On : 1 Oct 2021 15:40
Operator : CG/JU
Sample : M3876-07
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
EW7W3

Quant Time: Oct 01 17:06:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 30 16:27:24 2021
Response via : Initial Calibration

