

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
 Data File : BG050390.D
 Acq On : 2 Oct 2021 14:52
 Operator : CG/JU
 Sample : M3877-11
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 EW2M7

Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:51:20 AM

Quant Time: Oct 04 01:07:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 30 16:27:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.271	152	53261	20.00	ng/ul	0.00
20) Naphthalene-d8	11.098	136	223254	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.887	164	146424	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.631	188	311422	20.00	ng/ul	0.00
79) Chrysene-d12	21.932	240	284619	20.00	ng/ul	-0.02
88) Perylene-d12	25.357	264	282159	20.00	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.642	96	3274m	2.26	ng/uL	0.00
4) Pyridine-d5	4.065	84	26281	6.08	ng/ul	0.00
7) Phenol-d5	7.402	99	61261	13.16	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.584	67	39901	13.62	ng/ul	0.00
11) 2-Chlorophenol-d4	7.796	132	42603	13.07	ng/ul	0.00
15) 4-Methylphenol-d8	8.959	113	38950	11.06	ng/ul	0.00
21) Nitrobenzene-d5	9.441	128	23005	14.73	ng/ul	0.00
24) 2-Nitrophenol-d4	10.163	143	23727m	14.16	ng/ul	-0.01
28) 2,4-Dichlorophenol-d3	10.698	165	40950	12.37	ng/ul	-0.01
31) 4-Chloroaniline-d4	11.221	131	56055	11.85	ng/ul	0.00
46) Dimethylphthalate-d6	14.282	166	139656	13.94	ng/ul	-0.01
49) Acenaphthylene-d8	14.582	160	176926	14.06	ng/ul	-0.01
54) 4-Nitrophenol-d4	15.046	143	22070m	11.43	ng/ul	-0.01
60) Fluorene-d10	15.874	176	124121	13.75	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.974	200	16250	10.46	ng/ul	-0.01
73) Anthracene-d10	17.725	188	192634	14.64	ng/ul	-0.01
81) Pyrene-d10	19.999	212	231769	15.31	ng/ul	-0.01
92) Benzo(a)pyrene-d12	25.116	264	198726	14.01	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093021\
Data File : BG050390.D
Acq On : 2 Oct 2021 14:52
Operator : CG/JU
Sample : M3877-11
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
EW2M7

Manual Integrations
APPROVED

mohammad
10/4/2021 11:51:20 AM

Quant Time: Oct 04 01:07:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG093021.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 30 16:27:24 2021
Response via : Initial Calibration

