

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093022\
 Data File : BG055039.D
 Acq On : 30 Sep 2022 13:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 02:11:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 15:43:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	42.567	-6.4	110	0.00
3	Pyridine	40.000	43.189	-8.0	107	0.00
4	n-Nitrosodimethylamine	40.000	42.493	-6.2	107	0.00
5 S	2-Fluorophenol	80.000	83.468	-4.3	105	0.00
6	Aniline	40.000	42.206	-5.5	106	0.00
7 S	Phenol-d6	80.000	84.628	-5.8	105	0.00
8	2-Chlorophenol	40.000	40.508	-1.3	102	0.00
9	Benzaldehyde	40.000	45.173	-12.9	110	0.00
10 C	Phenol	40.000	41.363	-3.4	104	0.00
11	bis(2-Chloroethyl)ether	40.000	42.530	-6.3	108	0.00
12	1,3-Dichlorobenzene	40.000	42.596	-6.5	109	0.00
13 C	1,4-Dichlorobenzene	40.000	42.113	-5.3	107	0.00
14	1,2-Dichlorobenzene	40.000	41.979	-4.9	108	0.00
15	Benzyl Alcohol	40.000	41.268	-3.2	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.458	-3.6	105	0.00
17	2-Methylphenol	40.000	41.471	-3.7	103	0.00
18	Hexachloroethane	40.000	41.824	-4.6	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.480	-3.7	102	0.00
20	3+4-Methylphenols	40.000	41.498	-3.7	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	43.550	-8.9	105	0.00
23 S	Nitrobenzene-d5	80.000	88.316	-10.4	103	0.00
24	Nitrobenzene	40.000	44.515	-11.3	103	0.00
25	Isophorone	40.000	42.854	-7.1	102	0.00
26 C	2-Nitrophenol	40.000	34.863	12.8	91	0.00
27	2,4-Dimethylphenol	40.000	42.854	-7.1	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.579	-6.4	103	0.00
29 C	2,4-Dichlorophenol	40.000	42.051	-5.1	98	0.00
30	1,2,4-Trichlorobenzene	40.000	43.837	-9.6	107	0.00
31	Naphthalene	40.000	42.677	-6.7	104	0.00
32	Benzoic acid	40.000	34.408	14.0	88	-0.02
33	4-Chloroaniline	40.000	43.396	-8.5	102	0.00
34 C	Hexachlorobutadiene	40.000	43.682	-9.2	106	0.00
35	Caprolactam	40.000	40.430	-1.1	92	-0.01
36 C	4-Chloro-3-methylphenol	40.000	42.544	-6.4	98	0.00
37	2-Methylnaphthalene	40.000	43.052	-7.6	103	0.00
38	1-Methylnaphthalene	40.000	42.794	-7.0	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.713	-6.8	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.507	3.7	91	0.00
42 S	2,4,6-Tribromophenol	80.000	76.388	4.5	87	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.861	0.3	91	0.00
44	2,4,5-Trichlorophenol	40.000	40.825	-2.1	93	0.00
45 S	2-Fluorobiphenyl	80.000	85.140	-6.4	102	0.00
46	1,1'-Biphenyl	40.000	42.886	-7.2	103	0.00
47	2-Chloronaphthalene	40.000	42.960	-7.4	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093022\
 Data File : BG055039.D
 Acq On : 30 Sep 2022 13:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 01 02:11:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 28 15:43:39 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.973	5.1	95	0.00
49	Acenaphthylene	40.000	43.057	-7.6	102	0.00
50	Dimethylphthalate	40.000	41.985	-5.0	98	0.00
51	2,6-Dinitrotoluene	40.000	40.250	-0.6	100	0.00
52 C	Acenaphthene	40.000	42.700	-6.8	101	0.00
53	3-Nitroaniline	40.000	41.184	-3.0	100	0.00
54 P	2,4-Dinitrophenol	40.000	40.313	-0.8	95	0.00
55	Dibenzofuran	40.000	42.687	-6.7	101	0.00
56 P	4-Nitrophenol	40.000	40.123	-0.3	92	0.00
57	2,4-Dinitrotoluene	40.000	40.829	-2.1	102	0.00
58	Fluorene	40.000	42.838	-7.1	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.394	-1.0	90	0.00
60	Diethylphthalate	40.000	42.549	-6.4	100	0.00
61	4-Chlorophenyl-phenylether	40.000	43.013	-7.5	102	0.00
62	4-Nitroaniline	40.000	46.199	-15.5	102	0.00
63	Azobenzene	40.000	42.239	-5.6	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.144	7.1	90	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.374	-5.9	102	0.00
67	4-Bromophenyl-phenylether	40.000	42.399	-6.0	102	0.00
68	Hexachlorobenzene	40.000	42.535	-6.3	101	0.00
69	Atrazine	40.000	40.828	-2.1	95	0.00
70 C	Pentachlorophenol	40.000	38.026	4.9	86	0.00
71	Phenanthrene	40.000	42.741	-6.9	101	0.00
72	Anthracene	40.000	42.480	-6.2	102	0.00
73	Carbazole	40.000	42.144	-5.4	101	0.00
74	Di-n-butylphthalate	40.000	41.714	-4.3	96	0.00
75 C	Fluoranthene	40.000	42.112	-5.3	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	34.154	14.6	77	0.00
78	Pyrene	40.000	44.298	-10.7	101	0.00
79 S	Terphenyl-d14	80.000	87.758	-9.7	101	0.00
80	Butylbenzylphthalate	40.000	41.040	-2.6	89	0.00
81	Benzo(a)anthracene	40.000	42.969	-7.4	99	0.00
82	3,3'-Dichlorobenzidine	40.000	43.582	-9.0	97	0.00
83	Chrysene	40.000	42.872	-7.2	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.313	-3.3	90	0.00
85 c	Di-n-octyl phthalate	40.000	40.258	-0.6	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	42.340	-5.9	95	-0.03
88	Benzo(b)fluoranthene	40.000	43.133	-7.8	97	-0.01
89	Benzo(k)fluoranthene	40.000	43.786	-9.5	99	0.00
90 C	Benzo(a)pyrene	40.000	43.748	-9.4	97	-0.01
91	Dibenzo(a,h)anthracene	40.000	41.981	-5.0	95	-0.03
92	Benzo(g,h,i)perylene	40.000	42.282	-5.7	95	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG093022\
Data File : BG055039.D
Acq On : 30 Sep 2022 13:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 01 02:11:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG092822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 28 15:43:39 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0