


Data Path : Z:\HPCHEM1\BNA G\Data\BG100115\
 Data File : BG018968.D
 Acq On : 30 Sep 2015 11:33
 Operator : UM/NP
 Sample : DFTPP56
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP56

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG091015.M
 Title : SVOA CALIBRATION
 Last Update : Thu Oct 01 04:12:22 2015



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2538

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.3	18340	PASS
68	69	0.00	2	1.4	306	PASS
69	198	0.00	100	29.4	22242	PASS
70	69	0.00	2	0.6	130	PASS
127	198	10	80	40.6	30656	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	75546	PASS
199	198	5	9	6.8	5139	PASS
275	198	10	60	28.1	21205	PASS
365	198	1	100	4.0	3042	PASS
441	442	0.01	24	16.5	11313	PASS
442	198	50	100	90.6	68434	PASS
443	442	15	24	19.6	13387	PASS

m/z	Abundance
38.00	198.0
39.00	1711.0
39.90	404.0
42.90	172.0
44.00	298.0
49.00	261.0
50.00	5703.0
51.00	21888.0
52.00	1463.0
54.90	225.0
55.90	739.0
57.00	1712.0
61.00	332.0
62.00	493.0
63.00	1049.0
65.00	530.0
68.90	26536.0
69.90	211.0
73.00	163.0
74.00	2985.0
75.00	4051.0
76.00	1307.0
77.00	32272.0
78.00	2222.0
78.90	1789.0
79.90	1413.0
80.90	2343.0
81.90	422.0
83.00	621.0
83.90	154.0
85.00	342.0
85.90	496.0
87.00	235.0
88.00	188.0
90.90	464.0
92.00	525.0
92.90	2714.0
94.00	309.0
95.90	259.0
98.00	2281.0
98.90	1882.0
99.90	204.0
100.90	1276.0
103.10	284.0
103.90	862.0
104.90	884.0
106.00	290.0
107.00	9292.0
108.00	1571.0
110.00	16744.0
111.00	3230.0
111.90	430.0
113.80	168.0
116.00	624.0
117.00	7068.0
117.90	832.0
120.00	204.0
121.90	787.0
122.90	1135.0
123.80	511.0
125.00	635.0
127.00	35768.0
128.00	2709.0
129.00	13338.0
129.90	1268.0
130.90	287.0
131.90	175.0
133.90	526.0
134.90	1030.0
135.90	427.0
137.00	579.0
140.00	157.0
140.90	1868.0
141.90	477.0
142.90	453.0
145.90	309.0
146.90	1230.0
147.90	2090.0
148.90	500.0
151.00	300.0
151.40	238.0
151.70	228.0
151.90	226.0
153.00	546.0
154.00	331.0
155.00	1091.0
155.90	1342.0
156.90	235.0
157.90	290.0
159.00	248.0
160.00	578.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

161.00	1078.0
161.90	220.0
164.90	783.0
166.00	522.0
166.90	4283.0
167.90	1580.0
168.80	382.0
170.90	323.0
172.00	331.0
173.00	446.0
173.90	783.0
175.00	2030.0
175.90	495.0
176.90	739.0
177.90	281.0
178.90	3249.0
179.90	2110.0
181.00	1147.0
182.10	178.0
183.80	243.0
185.00	1346.0
186.00	11676.0
187.00	3293.0
187.90	489.0
189.00	553.0
190.90	474.0
191.90	1132.0
193.00	1591.0
193.90	429.0
195.00	202.0
195.90	2575.0
197.90	89680.0
198.90	6369.0
199.90	681.0
200.90	219.0
201.40	490.0
202.00	169.0
202.90	498.0
203.90	3564.0
205.00	5296.0
206.00	23472.0
207.00	3141.0
207.90	643.0
208.90	268.0
210.30	435.0
211.00	931.0
211.70	158.0
211.90	162.0
215.00	263.0
215.90	681.0
216.90	6454.0
217.90	564.0
221.00	3675.0
221.90	892.0
222.90	1217.0
224.00	13527.0
225.00	3473.0
225.80	388.0
227.00	5323.0
228.00	854.0
229.00	943.0
229.90	242.0
231.00	528.0
234.10	451.0
235.00	451.0
235.90	206.0
236.90	410.0
239.00	254.0
240.00	157.0
240.90	198.0
241.90	720.0
243.00	804.0
244.00	10417.0
245.00	1602.0
245.90	1904.0
246.90	438.0
247.80	194.0
248.80	400.0
252.00	175.0
253.00	172.0
255.00	55312.0
256.00	7983.0
257.00	554.0
257.90	3408.0
258.90	692.0
260.00	205.0
264.90	1422.0
271.00	152.0
272.90	1685.0
274.00	4669.0
275.00	25424.0
275.90	3187.0
276.90	2016.0
277.90	510.0
283.00	194.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

284.10	170.0
285.00	303.0
292.00	195.0
292.90	525.0
293.90	164.0
295.90	6526.0
296.90	1238.0
301.90	184.0
302.90	970.0
304.00	231.0
305.00	207.0
308.00	162.0
310.10	244.0
313.90	321.0
314.90	1007.0
315.90	356.0
321.10	192.0
321.90	166.0
323.00	2158.0
324.10	209.0
326.90	407.0
328.00	222.0
331.90	187.0
333.00	289.0
334.00	1489.0
335.00	325.0
341.10	340.0
345.90	434.0
352.00	716.0
353.00	494.0
354.00	871.0
355.00	223.0
365.00	3313.0
365.90	235.0
370.00	171.0
372.00	1436.0
373.00	213.0
374.00	166.0
383.00	392.0
389.90	158.0
390.20	152.0
402.00	420.0
403.00	796.0
420.90	689.0
421.80	372.0
423.00	4257.0
423.90	869.0
436.70	181.0
441.00	14081.0
442.00	81608.0
443.00	15303.0
444.00	1428.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

m/z	Abundance
38.00	188.0
39.00	1589.0
39.90	368.0
43.10	200.0
44.00	338.0
48.90	188.0
50.00	5072.0
51.00	18464.0
52.00	1142.0
54.90	257.0
55.90	727.0
57.00	1492.0
60.80	210.0
62.00	160.0
62.90	1018.0
64.00	187.0
65.00	434.0
67.90	491.0
68.90	22096.0
70.00	181.0
71.20	167.0
72.90	240.0
74.00	2252.0
74.90	3948.0
76.00	1337.0
77.00	29208.0
78.00	1824.0
79.00	1430.0
79.90	1172.0
81.00	2220.0
81.90	452.0
82.90	436.0
85.00	361.0
86.00	424.0
87.00	223.0
90.90	651.0
91.90	394.0
92.90	2282.0
93.90	184.0
98.00	2540.0
98.90	1942.0
99.80	260.0
101.00	1101.0
102.90	418.0
103.90	675.0
104.90	728.0
106.20	424.0
107.00	7831.0
108.00	1665.0
108.90	266.0
109.90	15854.0
110.90	2868.0
112.00	297.0
116.00	478.0
117.00	5833.0
117.90	533.0
121.90	431.0
122.90	1024.0
123.90	480.0
125.00	530.0
127.00	29304.0
128.00	2680.0
129.00	11070.0
129.90	1309.0
131.00	222.0
134.00	393.0
134.90	995.0
135.90	427.0
136.90	596.0
140.90	1247.0
141.90	334.0
143.00	391.0
144.00	182.0
145.90	299.0
146.90	778.0
147.90	1690.0
148.90	309.0
151.00	473.0
153.00	482.0
153.90	353.0
155.00	971.0
155.90	1017.0
157.10	369.0
157.80	375.0
158.90	323.0
159.90	452.0
161.00	308.0
162.10	242.0
164.90	654.0
165.90	429.0
167.00	3507.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

168.00	1498.0
168.90	405.0
169.50	198.0
171.90	310.0
172.90	279.0
174.00	665.0
174.90	1115.0
175.90	430.0
176.90	320.0
178.00	260.0
178.90	2315.0
180.00	1423.0
180.90	610.0
184.00	339.0
184.90	1063.0
186.00	9159.0
187.00	2651.0
188.00	269.0
189.00	515.0
190.90	266.0
191.90	755.0
193.00	1201.0
194.00	166.0
194.90	301.0
196.00	1826.0
197.90	66680.0
198.90	4878.0
200.00	566.0
201.20	423.0
202.80	386.0
203.90	2801.0
204.90	3698.0
206.00	19760.0
207.00	2806.0
207.90	554.0
208.90	242.0
210.00	336.0
210.90	928.0
211.80	197.0
215.00	275.0
215.90	618.0
216.90	4693.0
218.00	601.0
220.90	2968.0
221.70	779.0
223.00	1100.0
224.00	9568.0
225.00	2288.0
225.90	213.0
226.90	3829.0
227.90	658.0
228.90	927.0
231.00	355.0
234.00	280.0
235.00	379.0
236.00	357.0
236.90	401.0
238.90	211.0
240.90	457.0
242.00	419.0
243.00	634.0
244.00	7358.0
245.00	1060.0
246.00	1358.0
246.90	431.0
248.80	274.0
253.90	290.0
255.00	40392.0
255.90	6052.0
257.00	639.0
257.90	2125.0
259.00	530.0
263.80	164.0
265.00	1117.0
271.00	235.0
272.00	190.0
272.90	1038.0
273.90	2807.0
275.00	17376.0
276.00	2321.0
277.00	1598.0
277.90	278.0
283.00	227.0
284.00	185.0
292.90	403.0
296.00	5051.0
297.00	658.0
301.90	174.0
302.90	881.0
303.90	150.0
314.10	213.0
314.90	662.0
316.00	384.0
323.00	1666.0
324.00	310.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

325.90	183.0
327.00	485.0
328.00	209.0
332.90	197.0
334.00	1069.0
335.00	194.0
340.90	442.0
345.90	402.0
352.00	538.0
352.90	263.0
354.00	644.0
364.90	2281.0
366.00	258.0
371.00	155.0
371.90	692.0
373.00	247.0
382.90	260.0
390.10	218.0
402.00	209.0
402.90	343.0
404.00	236.0
420.90	252.0
421.90	365.0
423.00	2256.0
424.10	491.0
441.00	7898.0
442.00	46392.0
443.00	9479.0
444.00	846.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

Scan 2538 (17.750 min): BG018968.D
DFTPP56

m/z	Abundance
39.90	573.0
43.90	260.0
83.80	211.0
280.90	204.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

m/z	Abundance
38.10	272.0
39.00	1100.0
39.90	584.0
43.90	296.0
49.00	308.0
50.00	3871.0
51.00	14668.0
52.00	619.0
54.90	313.0
56.00	506.0
57.00	1216.0
60.90	298.0
63.00	781.0
65.00	469.0
68.00	428.0
68.90	18096.0
72.90	169.0
74.00	1601.0
75.00	2852.0
76.00	1159.0
77.00	21960.0
78.00	1239.0
78.90	1302.0
80.00	1263.0
80.90	1505.0
82.00	454.0
82.90	307.0
83.90	173.0
85.00	372.0
86.00	442.0
90.90	370.0
92.00	225.0
92.90	2161.0
94.00	310.0
96.00	185.0
97.00	162.0
98.00	2211.0
99.00	1346.0
100.10	176.0
100.90	809.0
102.90	383.0
103.90	536.0
105.00	557.0
106.00	289.0
107.00	6612.0
107.90	915.0
109.90	13466.0
111.00	1738.0
112.00	204.0
115.90	319.0
116.90	5403.0
117.90	311.0
121.90	377.0
122.90	594.0
123.90	489.0
124.90	221.0
127.00	26896.0
128.00	1511.0
128.90	10555.0
129.90	940.0
130.80	155.0
131.70	162.0
132.80	155.0
133.90	399.0
134.90	868.0
135.90	375.0
137.00	435.0
140.90	1280.0
141.90	547.0
142.90	433.0
145.90	186.0
146.90	652.0
148.00	1036.0
149.00	425.0
151.00	216.0
153.00	381.0
154.00	260.0
155.00	823.0
156.00	1034.0
157.00	220.0
157.80	171.0
159.00	199.0
159.90	359.0
160.90	792.0
164.90	578.0
166.10	317.0
167.00	2867.0
168.00	1646.0
168.90	322.0
169.90	265.0
171.90	286.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

173.00	339.0
173.90	474.0
175.00	1160.0
175.90	433.0
177.00	643.0
177.90	295.0
178.90	2171.0
180.00	1388.0
180.90	466.0
183.90	240.0
185.00	1137.0
186.00	8665.0
187.00	2393.0
187.90	195.0
188.90	544.0
190.90	297.0
192.00	677.0
192.90	1063.0
193.90	169.0
195.90	1407.0
197.90	70280.0
198.90	4172.0
200.00	416.0
201.20	278.0
202.90	786.0
204.00	3043.0
205.00	4626.0
206.00	19224.0
207.00	2791.0
207.90	902.0
209.00	217.0
210.20	364.0
210.90	556.0
212.00	150.0
214.00	154.0
214.80	192.0
215.90	476.0
216.90	5542.0
217.90	568.0
221.00	3542.0
222.00	573.0
222.90	1143.0
224.00	11309.0
225.00	2869.0
225.90	253.0
226.90	4505.0
227.90	501.0
229.00	877.0
231.00	556.0
234.80	426.0
235.90	172.0
236.90	372.0
239.00	231.0
239.90	168.0
240.90	339.0
242.00	732.0
243.00	930.0
244.00	8854.0
245.00	1418.0
245.90	1794.0
247.00	296.0
248.90	382.0
251.80	200.0
253.00	316.0
253.80	279.0
255.00	46560.0
256.00	6633.0
256.90	522.0
258.00	2486.0
259.00	632.0
264.90	832.0
265.70	242.0
272.90	1345.0
274.00	3635.0
275.00	20816.0
276.00	3166.0
277.00	1581.0
277.90	344.0
282.90	259.0
285.10	410.0
292.90	454.0
295.90	6257.0
297.00	718.0
301.90	207.0
303.00	627.0
304.00	225.0
313.80	283.0
314.90	492.0
316.00	469.0
320.90	194.0
323.00	2100.0
323.90	406.0
326.90	471.0
327.80	255.0
332.90	275.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56

334.00	1409.0
335.00	573.0
340.10	228.0
341.00	341.0
346.00	534.0
347.00	157.0
352.00	662.0
352.90	417.0
354.00	893.0
365.00	3532.0
366.00	296.0
372.00	1102.0
372.90	296.0
383.00	361.0
389.90	161.0
391.10	208.0
401.80	349.0
402.00	344.0
403.00	912.0
403.90	282.0
421.00	803.0
422.00	760.0
423.00	3738.0
424.00	858.0
438.90	168.0
439.40	188.0
441.00	11962.0
442.00	77304.0
443.00	15380.0
444.00	1294.0

Instrument :
BNA_G
ClientSampleId :
DFTPP56