

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100220\
 Data File : BG046779.D
 Acq On : 3 Oct 2020 19:51
 Operator : CG/JU
 Sample : L4150-06
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CB7J2

Quant Time: Oct 05 01:00:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG092920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 05 00:58:20 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.115	152	85866	20.00	ng/ul	0.00
18) Naphthalene-d8	10.923	136	339036	20.00	ng/ul	# 0.00
36) Acenaphthene-d10	14.731	164	237418	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.474	188	573302	20.00	ng/ul	0.00
78) Chrysene-d12	21.752	240	614227	20.00	ng/ul	0.00
86) Perylene-d12	25.024	264	629299	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.526	96	6918	3.95	ng/uL	0.00
5) Phenol-d5	7.257	99	140879	19.17	ng/ul	0.00
7) Bis-(2-Chloroethyl)eth...	7.427	67	88411	19.72	ng/ul	0.00
9) 2-Chlorophenol-d4	7.639	132	106594	19.87	ng/ul	0.00
13) 4-Methylphenol-d8	8.802	113	106870	18.59	ng/ul	0.00
19) Nitrobenzene-d5	9.266	128	57675	22.73	ng/ul	0.00
22) 2-Nitrophenol-d4	9.995	143	65508	25.40	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.536	165	120068	19.79	ng/ul	0.00
29) 4-Chloroaniline-d4	11.047	131	101581	14.09	ng/ul	0.00
44) Dimethylphthalate-d6	14.131	166	418590	23.07	ng/ul	0.00
47) Acenaphthylene-d8	14.425	160	467123	21.69	ng/ul	0.00
52) 4-Nitrophenol-d4	14.901	143	70517	22.80	ng/ul	0.00
58) Fluorene-d10	15.724	176	367070	21.95	ng/ul	0.00
63) 4,6-Dinitro-2-methylph...	15.823	200	62773	21.91	ng/ul	0.00
71) Anthracene-d10	17.574	188	562583	21.39	ng/ul	0.00
79) Pyrene-d10	19.854	212	702663	22.05	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.795	264	672379	19.82	ng/ul	0.00
Target Compounds						
						Qvalue
14) Acetophenone	8.932	105	11291	1.21	ng/ul#	93
45) Dimethylphthalate	14.178	163	44580	2.43	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100220\
Data File : BG046779.D
Acq On : 3 Oct 2020 19:51
Operator : CG/JU
Sample : L4150-06
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CB7J2

Quant Time: Oct 05 01:00:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG092920MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 05 00:58:20 2020
Response via : Initial Calibration

