



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG100220\
 Data File : BG046729.D
 Acq On : 2 Oct 2020 8:32
 Operator : CG/JU
 Sample : DFTPP36
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP36

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG092920MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Oct 02 10:06:22 2020



AutoFind: Scans 2523, 2524, 2525; Background Corrected with Scan 2517

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.7	66752	PASS
68	69	0.00	2	0.4	306	PASS
69	198	0.00	100	36.0	73563	PASS
70	69	0.00	2	0.5	400	PASS
127	198	10	80	41.5	84763	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	204288	PASS
199	198	5	9	7.0	14371	PASS
275	198	10	60	29.8	60795	PASS
365	198	1	100	3.6	7274	PASS
441	198	0.01	100	12.3	25205	PASS
442	198	50	100	75.8	154944	PASS
443	442	15	24	19.7	30557	PASS

m/z	Abundance
40.10	163.0
44.00	459.0
48.90	160.0
55.00	254.0
57.00	176.0
67.60	169.0
73.00	201.0
80.90	172.0
85.10	271.0
91.10	186.0
95.90	490.0
106.70	240.0
115.20	201.0
191.50	159.0
207.00	947.0
207.80	385.0
209.00	303.0
280.80	223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

m/z	Abundance
36.90	157.0
38.00	704.0
39.10	6971.0
40.00	515.0
41.10	349.0
43.10	379.0
44.10	1034.0
49.00	857.0
50.10	20168.0
51.10	79056.0
52.10	4478.0
55.00	455.0
56.00	3157.0
57.00	6081.0
58.00	264.0
59.20	153.0
61.10	1180.0
62.10	1074.0
63.10	4258.0
64.00	553.0
65.10	1486.0
67.20	320.0
69.00	87216.0
73.00	1164.0
74.00	8252.0
75.00	14493.0
76.10	5048.0
77.10	90160.0
78.00	5839.0
79.00	5961.0
80.00	5268.0
81.00	8126.0
82.00	1744.0
82.90	1318.0
84.30	288.0
85.00	1709.0
86.00	2489.0
86.90	1364.0
87.90	554.0
89.30	167.0
90.90	1577.0
92.00	2456.0
93.00	11703.0
94.00	848.0
94.90	351.0
95.90	378.0
96.70	200.0
98.00	7522.0
99.00	7067.0
100.00	502.0
101.00	4192.0
102.00	360.0
103.00	1489.0
104.00	2421.0
105.00	2462.0
107.00	28512.0
108.00	3945.0
110.00	50408.0
111.00	8063.0
111.90	986.0
113.00	322.0
116.00	1725.0
117.00	24448.0
118.00	1888.0
118.90	271.0
119.90	366.0
121.90	2505.0
123.00	3264.0
124.00	1313.0
124.90	1357.0
127.00	98528.0
128.00	8093.0
129.00	41824.0
129.90	4094.0
131.00	875.0
131.90	484.0
132.90	159.0
133.90	952.0
135.00	2787.0
136.00	1373.0
137.00	1428.0
138.00	224.0
140.00	666.0
141.00	4998.0
141.80	1550.0
142.90	1033.0
143.90	404.0
145.10	423.0
145.90	965.0
147.00	3027.0
148.00	5673.0
148.90	1441.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

150.00	571.0
151.00	624.0
151.90	321.0
152.80	1791.0
154.00	1214.0
154.90	2973.0
156.00	4833.0
157.10	960.0
157.90	1355.0
158.90	808.0
159.90	1995.0
161.00	2485.0
162.00	790.0
162.80	286.0
165.00	1944.0
166.00	1946.0
167.00	9574.0
168.00	3910.0
168.90	1080.0
170.00	454.0
170.80	442.0
171.90	1234.0
172.90	1123.0
173.90	2805.0
175.00	3864.0
176.00	1195.0
176.90	2402.0
178.00	704.0
178.90	9378.0
179.90	6336.0
180.90	2881.0
181.80	323.0
183.20	394.0
183.90	585.0
185.00	4696.0
185.90	30296.0
186.90	9173.0
187.90	1137.0
188.90	2054.0
189.80	682.0
191.00	938.0
191.90	2668.0
192.90	2805.0
194.00	455.0
194.90	469.0
195.90	5560.0
197.90	220608.0
198.90	15970.0
199.80	1247.0
201.30	1236.0
202.90	2085.0
203.90	9255.0
204.90	14819.0
205.90	63000.0
206.90	9437.0
207.90	2821.0
208.80	732.0
209.70	863.0
210.90	2212.0
214.80	1076.0
215.90	1453.0
216.90	16048.0
217.90	2636.0
219.00	457.0
220.90	8118.0
221.80	1761.0
222.90	4815.0
223.90	33912.0
225.00	9606.0
225.80	655.0
226.90	15054.0
227.80	2696.0
228.90	2617.0
229.90	436.0
230.90	1262.0
231.90	439.0
232.90	442.0
233.80	1265.0
234.90	1190.0
235.80	939.0
236.90	1258.0
237.90	236.0
238.90	712.0
239.80	429.0
240.90	715.0
241.90	1694.0
243.00	1961.0
243.90	28608.0
244.90	3259.0
245.90	5522.0
246.80	1016.0
247.70	477.0
248.80	1076.0
249.90	207.0
251.00	396.0
252.00	375.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

252.30	389.0
252.90	726.0
254.90	140800.0
255.90	19760.0
256.90	1880.0
257.90	10103.0
258.80	1541.0
259.90	443.0
261.00	486.0
264.80	2687.0
265.80	261.0
270.00	194.0
270.80	407.0
271.80	553.0
272.80	3070.0
273.90	10260.0
274.90	59296.0
275.90	8252.0
276.80	5948.0
277.80	689.0
278.70	177.0
280.80	351.0
282.00	252.0
282.80	417.0
284.00	258.0
284.80	912.0
285.90	230.0
288.60	183.0
289.80	300.0
292.90	1306.0
293.90	571.0
294.80	915.0
295.90	18960.0
296.90	2853.0
300.70	257.0
301.90	401.0
302.90	2090.0
303.90	525.0
305.00	259.0
307.80	424.0
308.80	163.0
309.70	229.0
312.80	228.0
313.80	846.0
314.80	1755.0
315.90	1331.0
320.70	238.0
322.90	5526.0
323.90	892.0
326.80	907.0
327.80	735.0
328.80	195.0
331.70	324.0
332.80	807.0
333.80	2891.0
334.80	1058.0
335.90	170.0
339.80	151.0
340.80	799.0
345.90	1047.0
346.90	280.0
349.90	161.0
351.90	1583.0
352.80	664.0
353.90	1501.0
354.80	252.0
355.90	165.0
358.70	211.0
364.80	7591.0
365.80	1021.0
370.60	155.0
371.00	259.0
371.90	2131.0
372.90	844.0
382.90	717.0
383.60	298.0
389.90	467.0
390.80	204.0
391.70	245.0
396.70	177.0
400.80	287.0
401.80	1017.0
402.80	1212.0
403.80	411.0
420.80	1185.0
421.80	1164.0
422.80	6598.0
423.80	1373.0
440.80	23040.0
441.80	131264.0
442.80	28272.0
443.80	2994.0
444.70	195.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

m/z	Abundance
37.10	484.0
38.10	834.0
39.10	5635.0
40.00	407.0
44.00	853.0
45.10	221.0
50.10	20336.0
51.10	77776.0
52.00	3736.0
52.90	286.0
55.00	409.0
56.00	2425.0
57.00	6141.0
58.10	205.0
59.10	366.0
61.00	1430.0
62.10	1531.0
63.00	3326.0
64.00	709.0
65.10	1467.0
66.10	295.0
69.00	83848.0
69.90	773.0
73.10	600.0
74.00	8293.0
75.00	14697.0
76.10	4797.0
77.00	90304.0
78.00	6526.0
79.00	6026.0
80.00	4982.0
81.00	7382.0
81.90	1734.0
82.90	1322.0
83.80	197.0
85.00	1123.0
86.00	1815.0
87.00	943.0
88.00	543.0
88.80	192.0
89.10	202.0
91.00	1921.0
92.00	1817.0
93.00	12038.0
94.00	555.0
95.00	165.0
96.00	660.0
97.00	266.0
98.00	9755.0
99.00	6576.0
100.00	579.0
101.00	3402.0
101.90	204.0
102.80	958.0
104.00	2690.0
105.00	2657.0
105.90	1056.0
107.00	32656.0
108.00	4951.0
110.00	45008.0
111.00	7568.0
111.90	1439.0
112.90	511.0
113.80	196.0
116.00	1692.0
117.00	25752.0
118.00	1570.0
120.00	334.0
121.90	1897.0
122.90	3706.0
124.00	1260.0
125.00	1639.0
127.00	97968.0
127.90	8828.0
128.90	40864.0
130.00	3641.0
130.90	603.0
132.10	257.0
132.80	356.0
134.00	1485.0
135.00	3523.0
136.00	1634.0
137.10	1160.0
138.10	523.0
138.90	190.0
139.90	948.0
140.90	5299.0
142.00	1472.0
142.90	1019.0
143.80	343.0
144.80	301.0
146.00	1250.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

147.00	3684.0
148.00	6401.0
148.90	1099.0
150.00	395.0
151.00	547.0
151.50	382.0
152.00	587.0
152.90	1729.0
154.00	1388.0
155.00	3710.0
155.90	4397.0
156.90	950.0
157.90	1093.0
158.90	1144.0
159.90	1841.0
161.00	2124.0
161.90	887.0
162.60	172.0
163.10	157.0
164.00	205.0
164.90	2662.0
166.00	1630.0
166.90	11134.0
167.90	5246.0
168.90	1248.0
169.90	398.0
170.70	493.0
171.90	1434.0
172.90	1455.0
173.90	2227.0
175.00	5136.0
176.00	1261.0
177.00	2478.0
177.80	653.0
178.90	9495.0
180.00	5565.0
181.00	2788.0
181.80	997.0
182.70	203.0
184.00	998.0
185.00	4792.0
185.90	34320.0
186.90	9250.0
187.90	1125.0
188.90	2228.0
189.90	542.0
190.90	1527.0
191.90	2106.0
192.90	3705.0
193.90	500.0
195.90	6876.0
197.90	233600.0
198.90	15706.0
199.80	1340.0
201.60	1208.0
203.00	1607.0
203.90	9266.0
205.00	16936.0
205.90	68544.0
206.90	9215.0
207.90	2882.0
208.80	561.0
209.90	790.0
210.80	2704.0
212.00	509.0
215.00	804.0
215.90	1504.0
216.90	19936.0
217.90	2351.0
218.90	259.0
219.60	494.0
220.90	7618.0
221.70	3138.0
222.90	5088.0
223.90	39136.0
224.90	10201.0
226.10	1342.0
226.90	16131.0
227.90	2408.0
228.90	2588.0
230.00	412.0
230.80	1452.0
232.70	360.0
233.90	908.0
234.80	1352.0
235.90	893.0
236.90	1705.0
238.80	861.0
239.90	462.0
240.80	804.0
241.90	2067.0
242.90	1864.0
243.90	32864.0
244.90	4209.0
245.90	6523.0
246.80	1340.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

247.80	498.0
248.80	1200.0
249.90	407.0
250.70	378.0
251.60	522.0
252.20	271.0
252.90	939.0
254.90	159744.0
255.90	23976.0
256.90	2017.0
257.90	9757.0
258.90	1411.0
259.90	487.0
260.80	331.0
263.10	227.0
264.00	315.0
264.90	3672.0
265.70	322.0
266.10	440.0
271.00	708.0
271.90	585.0
272.80	5563.0
273.90	12222.0
274.90	73608.0
275.90	8788.0
276.90	5338.0
277.90	981.0
278.70	321.0
282.80	608.0
283.70	474.0
284.00	440.0
284.90	1291.0
285.80	178.0
289.00	199.0
289.90	388.0
290.90	161.0
291.80	293.0
292.70	1229.0
293.90	271.0
295.90	20208.0
296.80	2842.0
297.70	282.0
301.80	339.0
302.90	2180.0
303.90	563.0
305.00	190.0
308.00	279.0
308.90	220.0
313.90	1060.0
314.80	3035.0
315.90	1084.0
317.00	372.0
318.60	150.0
320.70	699.0
321.90	195.0
322.90	6250.0
323.80	970.0
326.80	1156.0
327.80	570.0
331.80	653.0
333.00	882.0
333.80	4391.0
334.90	1103.0
339.90	233.0
340.90	1138.0
345.80	1402.0
346.60	230.0
351.80	1820.0
352.80	1074.0
353.90	1869.0
364.80	7712.0
365.80	1242.0
369.90	188.0
370.80	562.0
371.90	2807.0
372.90	998.0
376.90	156.0
382.90	631.0
383.80	273.0
389.80	475.0
390.80	347.0
391.60	243.0
401.80	1437.0
402.80	1843.0
403.90	785.0
420.60	679.0
421.90	1130.0
422.90	11735.0
423.90	1731.0
425.00	243.0
432.20	203.0
440.80	27960.0
441.80	179520.0
442.80	35056.0
443.80	3144.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

m/z	Abundance
38.00	638.0
39.10	2979.0
40.00	405.0
41.20	216.0
43.90	540.0
49.00	440.0
50.10	13276.0
51.10	43424.0
51.90	1956.0
53.10	162.0
55.00	203.0
56.00	1121.0
57.00	3125.0
58.00	248.0
61.00	529.0
62.00	841.0
63.00	2241.0
63.90	291.0
64.40	248.0
65.10	716.0
66.90	325.0
68.00	918.0
69.00	49624.0
70.00	426.0
72.90	488.0
74.00	4928.0
75.00	7884.0
76.10	2921.0
77.00	52112.0
78.00	3524.0
79.00	4145.0
80.00	3474.0
81.00	4067.0
81.90	646.0
83.10	769.0
83.90	154.0
85.00	538.0
86.00	1123.0
87.00	614.0
90.90	1045.0
92.00	822.0
93.00	6637.0
94.10	326.0
95.90	287.0
98.00	6015.0
99.00	3454.0
100.00	413.0
101.00	2495.0
101.90	301.0
102.90	741.0
104.00	1551.0
104.90	1262.0
105.90	424.0
107.00	18144.0
108.00	2729.0
108.80	750.0
110.00	31928.0
111.00	4667.0
111.90	499.0
113.10	221.0
116.00	1593.0
117.00	14940.0
118.00	1198.0
120.00	222.0
122.00	1511.0
122.90	1874.0
124.00	969.0
124.90	995.0
127.00	57792.0
128.00	4250.0
128.90	25272.0
130.00	1631.0
131.00	823.0
131.90	341.0
133.00	242.0
133.80	374.0
134.90	1755.0
136.00	936.0
137.10	1032.0
140.00	550.0
141.00	3162.0
142.10	884.0
143.00	584.0
144.00	246.0
145.90	328.0
146.90	2069.0
147.90	4709.0
148.90	952.0
149.90	188.0
151.00	684.0
151.80	432.0
153.00	856.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

154.00	1203.0
154.90	1631.0
155.90	2293.0
156.90	403.0
158.00	839.0
158.90	551.0
159.90	1039.0
161.00	1667.0
161.90	607.0
162.90	374.0
164.10	390.0
164.90	1910.0
165.90	1112.0
166.90	7158.0
167.90	3706.0
168.90	664.0
170.00	239.0
170.80	518.0
171.80	701.0
172.90	617.0
173.90	1524.0
174.90	2941.0
175.90	306.0
176.90	1474.0
178.00	740.0
178.90	6050.0
179.90	4833.0
180.90	2045.0
181.80	178.0
184.00	516.0
185.00	3759.0
185.90	22816.0
187.00	6685.0
187.80	765.0
188.90	1400.0
189.80	242.0
190.90	1086.0
191.90	2120.0
192.90	1713.0
193.90	677.0
195.00	411.0
195.90	4869.0
197.90	158656.0
198.90	11436.0
199.70	951.0
200.80	385.0
201.40	801.0
202.90	972.0
203.90	6863.0
204.90	10127.0
205.90	43832.0
206.90	6668.0
208.00	2045.0
209.00	330.0
209.90	966.0
210.20	860.0
210.90	2027.0
212.00	347.0
214.80	841.0
216.00	1086.0
216.90	12661.0
217.80	1891.0
218.80	287.0
220.90	6391.0
221.70	1547.0
222.80	2261.0
223.90	28664.0
225.00	6784.0
225.90	827.0
226.90	12644.0
227.90	1187.0
228.90	2059.0
229.80	432.0
230.80	740.0
231.60	184.0
233.00	155.0
233.90	1081.0
234.80	953.0
235.80	457.0
236.90	836.0
237.90	197.0
238.90	599.0
239.80	522.0
240.80	620.0
241.90	1181.0
242.90	1732.0
243.90	20632.0
244.90	2712.0
245.90	5138.0
246.80	676.0
248.60	586.0
250.00	476.0
251.00	227.0
251.80	294.0
252.70	592.0
252.90	627.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36

254.90	110696.0
255.90	15509.0
256.90	1072.0
257.80	6492.0
258.80	1212.0
259.90	209.0
260.90	223.0
264.80	3774.0
265.90	637.0
270.70	199.0
272.90	3340.0
273.90	7045.0
274.90	49480.0
275.90	6087.0
276.90	3447.0
277.90	889.0
278.80	300.0
280.90	214.0
282.90	423.0
283.90	175.0
284.70	417.0
286.70	184.0
288.70	211.0
289.80	266.0
290.90	246.0
292.00	345.0
292.90	856.0
293.70	468.0
294.90	721.0
295.90	14399.0
296.90	2030.0
301.00	278.0
301.90	505.0
302.90	1574.0
304.00	586.0
307.70	265.0
309.70	263.0
310.80	174.0
313.10	277.0
313.90	1171.0
314.90	1721.0
315.90	1104.0
316.90	203.0
320.80	735.0
322.90	4519.0
323.90	781.0
326.00	159.0
326.90	833.0
327.90	544.0
328.80	311.0
331.80	380.0
332.80	331.0
333.90	3319.0
334.80	837.0
340.80	561.0
345.80	1193.0
351.80	1221.0
352.80	1047.0
353.80	1813.0
355.00	345.0
357.00	169.0
360.70	196.0
363.70	204.0
364.80	6520.0
365.80	659.0
369.90	256.0
370.70	244.0
371.80	2084.0
372.80	537.0
382.90	765.0
389.90	333.0
390.60	181.0
391.90	450.0
400.80	316.0
401.80	859.0
402.70	1280.0
403.80	552.0
420.80	968.0
421.80	1149.0
422.80	7363.0
423.90	2237.0
425.00	274.0
433.10	151.0
439.60	289.0
440.80	24616.0
441.80	154048.0
442.80	28344.0
443.80	2524.0
444.70	154.0

Instrument :
BNA_G
ClientSampleId :
DFTPP36